

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP110924\
 Data File : BP022944.D
 Acq On : 08 Nov 2024 22:18
 Operator : RC/JU
 Sample : SST08053
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST080632

Quant Time: Nov 08 23:18:54 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP110924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Nov 08 21:50:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.593	152	66769	20.000	ng/u1	0.00
20) Naphthalene-d8	10.340	136	270960	20.000	ng/u1	-0.01
38) Acenaphthene-d10	14.204	164	221632	20.000	ng/u1	-0.02
64) Phenanthrene-d10	17.022	188	573990	20.000	ng/u1	-0.01
79) Chrysene-d12	21.457	240	634223	20.000	ng/u1	0.00
88) Perylene-d12	24.680	264	742140	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.140	96	45235	26.327	ng/uL	0.00
4) Pyridine-d5	3.546	84	344309	72.995	ng/u1	-0.01
7) Phenol-d5	6.787	99	418385	74.440	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.940	67	237427	71.874	ng/u1	0.00
11) 2-Chlorophenol-d4	7.134	132	314849	78.947	ng/u1	0.00
15) 4-Methylphenol-d8	8.299	113	343152	76.505	ng/u1	0.00
21) Nitrobenzene-d5	8.734	128	160950	77.253	ng/u1	0.00
24) 2-Nitrophenol-d4	9.446	143	197079	81.706	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	9.975	165	403670	78.741	ng/u1	-0.02
31) 4-Chloroaniline-d4	10.487	131	465313	76.814	ng/u1	-0.01
46) Dimethylphthalate-d6	13.616	166	1293926	73.480	ng/u1	0.00
49) Acenaphthylene-d8	13.893	160	1415177	75.666	ng/u1	-0.01
54) 4-Nitrophenol-d4	14.463	143	213745	72.354	ng/u1	0.00
60) Fluorene-d10	15.228	176	1144129	74.909	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.369	200	289246	76.621	ng/u1	0.00
73) Anthracene-d10	17.116	188	1966992	72.846	ng/u1	0.00
81) Pyrene-d10	19.516	212	2486998	74.153	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.474	264	3015432	76.082	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.176	88	48834	26.433	ng/uL	93
5) Pyridine	3.564	79	347522	71.855	ng/u1	98
6) Benzaldehyde	6.752	77	189671	62.448	ng/u1	97
8) Phenol	6.817	94	435703	74.511	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.028	93	322419	73.673	ng/u1	97
12) 2-Chlorophenol	7.169	128	323149	78.359	ng/u1	99
13) 2-Methylphenol	8.034	108	323632	76.262	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.099	45	354396	69.679	ng/u1	96
16) Acetophenone	8.405	105	560319	75.668	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.393	70	282368	75.264	ng/u1	99
18) 4-Methylphenol	8.363	108	355262	75.462	ng/u1	98
19) Hexachloroethane	8.640	117	148872	77.512	ng/u1	96
22) Nitrobenzene	8.775	77	446547	71.805	ng/u1	98
23) Isophorone	9.299	82	801200	71.855	ng/u1	99
25) 2-Nitrophenol	9.481	139	203890	78.433	ng/u1	93
26) 2,4-Dimethylphenol	9.540	107	420908	74.494	ng/u1	99
27) Bis(2-Chloroethoxy)met...	9.769	93	436663	68.971	ng/u1	99
29) 2,4-Dichlorophenol	10.005	162	391261	77.501	ng/u1	98
30) Naphthalene	10.387	128	1049576	72.725	ng/u1	99
32) 4-Chloroaniline	10.510	127	452375	75.654	ng/u1	97
33) Hexachlorobutadiene	10.663	225	328501	77.204	ng/u1	100
34) Caprolactam	11.310	113	63613	39.085	ng/u1	98
35) 4-Chloro-3-methylphenol	11.652	107	422899	76.777	ng/u1	97
36) 2-Methylnaphthalene	12.004	142	809492	76.411	ng/u1	99


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP110924\
 Data File : BP022944.D
 Acq On : 08 Nov 2024 22:18
 Operator : RC/JU
 Sample : SST08053
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST080632

Quant Time: Nov 08 23:18:54 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP110924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Nov 08 21:50:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.228	142	815966	75.462	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.381	216	604935	73.817	ng/ul	96
40) Hexachlorocyclopentadiene	12.357	237	347624	69.623	ng/ul	97
41) 2,4,6-Trichlorophenol	12.634	196	395458	76.716	ng/ul	97
42) 2,4,5-Trichlorophenol	12.716	196	433792	79.033	ng/ul	96
43) 1,1'-Biphenyl	13.034	154	1147100	71.806	ng/ul	98
44) 2-Chloronaphthalene	13.081	162	950330	72.669	ng/ul	98
45) 2-Nitroaniline	13.293	65	291860	74.743	ng/ul	92
47) Dimethylphthalate	13.681	163	1259619	71.787	ng/ul	99
48) 2,6-Dinitrotoluene	13.804	165	269152	81.096	ng/ul	98
50) Acenaphthylene	13.922	152	1391595	71.692	ng/ul	99
51) 3-Nitroaniline	14.146	138	210712	67.596	ng/ul	92
52) Acenaphthene	14.269	153	1000502	72.891	ng/ul	98
53) 2,4-Dinitrophenol	14.357	184	200983	82.306	ng/ul	98
55) 4-Nitrophenol	14.475	109	255810	78.021	ng/ul	99
56) Dibenzofuran	14.616	168	1505778	73.032	ng/ul	99
57) 2,4-Dinitrotoluene	14.598	165	417904	81.031	ng/ul	92
58) 2,3,4,6-Tetrachlorophenol	14.863	232	410637	80.153	ng/ul	98
59) Diethylphthalate	15.057	149	1256779	74.657	ng/ul	99
61) Fluorene	15.275	166	1267454	75.446	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.269	204	737753	76.794	ng/ul	97
63) 4-Nitroaniline	15.316	138	188262	60.019	ng/ul	95
66) 4,6-Dinitro-2-methylph...	15.387	198	309053	76.005	ng/ul#	94
67) N-Nitrosodiphenylamine	15.504	169	1065405	69.313	ng/ul	98
68) 4-Bromophenyl-phenylether	16.192	248	519392	75.036	ng/ul	96
69) Hexachlorobenzene	16.310	284	646352	75.862	ng/ul	96
70) Atrazine	16.481	200	514378	80.506	ng/ul	99
71) Pentachlorophenol	16.669	266	420144	76.692	ng/ul	98
72) Phenanthrene	17.057	178	2149575	69.998	ng/ul	100
74) Anthracene	17.151	178	2170791	70.835	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	12.999	216	613467	70.193	ng/ul	98
76) Pentachlorobenzene	14.546	250	693132	71.742	ng/ul	97
77) Carbazole	17.439	167	1885652	69.539	ng/ul	100
78) Di-n-butylphthalate	18.028	149	2177918	72.588	ng/ul	99
80) Fluoranthene	19.169	202	2870008	74.436	ng/ul	100
82) Pyrene	19.545	202	2994375	72.459	ng/ul	99
83) Butylbenzylphthalate	20.486	149	1085655	75.456	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.363	252	1141249	75.487	ng/ul	100
85) Benzo(a)anthracene	21.439	228	3156500	70.994	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.369	149	1542613	74.695	ng/ul	99
87) Chrysene	21.504	228	2912927	70.658	ng/ul	99
89) Di-n-octyl phthalate	22.592	149	2526099	71.559	ng/ul	100
90) Benzo(b)fluoranthene	23.669	252	3476847	74.427	ng/ul	100
91) Benzo(k)fluoranthene	23.745	252	3364566	73.564	ng/ul	99
93) Benzo(a)pyrene	24.545	252	3136980	72.963	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	28.339	276	1422051	24.786	ng/ul	98
95) Dibenzo(a,h)anthracene	28.433	278	3412638	73.454	ng/ul	99
96) Benzo(g,h,i)perylene	29.498	276	3228735	72.350	ng/ul	99





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 26.433 ng/uL
RT: 3.176 min Scan# 31
Delta R.T. 0.000 min
Lab File: BP022944.D
Acq: 08 Nov 2024 22:18




Instrument :
BNA_P
ClientSampleId :
SSTD080632


Tgt Ion: 88 Resp: 48834
Ion Ratio Lower Upper
88 100
43 28.6 24.2 36.4
58 67.0 59.2 88.8



#5
Pyridine
Concen: 71.855 ng/uL
RT: 3.564 min Scan# 98
Delta R.T. -0.012 min
Lab File: BP022944.D
Acq: 08 Nov 2024 22:18

Tgt Ion: 79 Resp: 347522
Ion Ratio Lower Upper
79 100
52 73.7 59.5 89.3
51 35.3 30.4 45.6









#6
Benzaldehyde
Concen: 62.448 ng/ul
RT: 6.752 min Scan# 641
Delta R.T. -0.006 min
Lab File: BP022944.D
Acq: 08 Nov 2024 22:18




Instrument :
BNA_P
ClientSampleId :
SSTD080632


Tgt Ion: 77 Resp: 189671
Ion Ratio Lower Upper
77 100
105 90.3 72.5 108.7
106 86.8 65.5 98.3



#8
Phenol
Concen: 74.511 ng/ul
RT: 6.817 min Scan# 651
Delta R.T. -0.006 min
Lab File: BP022944.D
Acq: 08 Nov 2024 22:18

Tgt Ion: 94 Resp: 435703
Ion Ratio Lower Upper
94 100
65 33.0 26.2 39.4
66 49.0 40.2 60.4





#10

Bis(2-Chloroethyl)ether

Concen: 73.673 ng/ul

RT: 7.028 min Scan# 61

Delta R.T. -0.006 min

Lab File: BP022944.D

Acq: 08 Nov 2024 22:18

Instrument :

BNA_P

ClientSampleId :

SSTD080632





Tgt Ion: 93 Resp: 322419

Ion Ratio Lower Upper

93 100

63 73.9 60.6 91.0

95 32.9 29.0 43.4



#12

2-Chlorophenol

Concen: 78.359 ng/ul

RT: 7.169 min Scan# 711

Delta R.T. -0.006 min

Lab File: BP022944.D

Acq: 08 Nov 2024 22:18




Tgt Ion:128 Resp: 323149


Ion Ratio Lower Upper

128 100

64 49.7 39.0 58.6

130 33.5 26.4 39.6





#13

2-Methylphenol

Concen: 76.262 ng/ul

RT: 8.034 min Scan# 8

Delta R.T. -0.006 min

Lab File: BP022944.D

Acq: 08 Nov 2024 22:18

Instrument :

BNA_P

ClientSampleId :





SSTD080632

Tgt Ion:108 Resp: 323632

Ion Ratio Lower Upper

108 100

107 90.4 73.3 109.9



#14

2,2'-oxybis(1-Chloropropane)

Concen: 69.679 ng/ul

RT: 8.099 min Scan# 869

Delta R.T. 0.000 min

Lab File: BP022944.D

Acq: 08 Nov 2024 22:18




Tgt Ion: 45 Resp: 354396


Ion Ratio Lower Upper

45 100

77 20.2 18.2 27.2

79 13.5 11.7 17.5









#16
Acetophenone
Concen: 75.668 ng/ul
RT: 8.405 min Scan# 919
Delta R.T. 0.000 min
Lab File: BP022944.D
Acq: 08 Nov 2024 22:18




Instrument :
BNA_P
ClientSampleId :
SSTD080632


Tgt Ion:105 Resp: 560319
Ion Ratio Lower Upper
105 100
77 84.3 69.4 104.0
51 28.2 23.2 34.8



#17
N-Nitroso-di-n-propylamine
Concen: 75.264 ng/ul
RT: 8.393 min Scan# 919
Delta R.T. 0.006 min
Lab File: BP022944.D
Acq: 08 Nov 2024 22:18

Tgt Ion: 70 Resp: 282368
Ion Ratio Lower Upper
70 100
42 52.3 42.3 63.5
101 9.7 7.4 11.2
130 18.2 15.4 23.0









#18
4-Methylphenol
Concen: 75.462 ng/ul
RT: 8.363 min Scan# 914
Delta R.T. 0.000 min
Lab File: BP022944.D
Acq: 08 Nov 2024 22:18




Instrument :
BNA_P
ClientSampleId :
SSTD080632


Tgt Ion:108 Resp: 355262
Ion Ratio Lower Upper
108 100
107 112.2 91.5 137.3



#19
Hexachloroethane
Concen: 77.512 ng/ul
RT: 8.640 min Scan# 961
Delta R.T. -0.006 min
Lab File: BP022944.D
Acq: 08 Nov 2024 22:18

Tgt Ion:117 Resp: 148872
Ion Ratio Lower Upper
117 100
201 137.8 105.0 157.6
199 84.7 65.9 98.9









#22
Nitrobenzene
Concen: 71.805 ng/ul
RT: 8.775 min Scan# 984
Delta R.T. 0.000 min
Lab File: BP022944.D
Acq: 08 Nov 2024 22:18




Instrument :
BNA_P
ClientSampleId :
SSTD080632


Tgt Ion: 77 Resp: 446547
Ion Ratio Lower Upper
77 100
123 39.3 32.2 48.4
65 15.1 10.9 16.3



#23
Isophorone
Concen: 71.855 ng/ul
RT: 9.299 min Scan# 1073
Delta R.T. 0.000 min
Lab File: BP022944.D
Acq: 08 Nov 2024 22:18

Tgt Ion: 82 Resp: 801200
Ion Ratio Lower Upper
82 100
95 9.3 7.6 11.4
138 16.5 13.4 20.2









#25
2-Nitrophenol
Concen: 78.433 ng/ul
RT: 9.481 min Scan# 1104
Delta R.T. 0.006 min
Lab File: BP022944.D
Acq: 08 Nov 2024 22:18




Instrument :
BNA_P
ClientSampleId :
SSTD080632


Tgt Ion:139 Resp: 203890
Ion Ratio Lower Upper
139 100
65 58.4 52.8 79.2
109 45.7 38.9 58.3



#26
2,4-Dimethylphenol
Concen: 74.494 ng/ul
RT: 9.540 min Scan# 1114
Delta R.T. -0.006 min
Lab File: BP022944.D
Acq: 08 Nov 2024 22:18

Tgt Ion:107 Resp: 420908
Ion Ratio Lower Upper
107 100
121 42.8 34.4 51.6
122 71.8 56.7 85.1









#27
Bis(2-Chloroethoxy)methane
Concen: 68.971 ng/ul
RT: 9.769 min Scan# 1153
Delta R.T. 0.000 min
Lab File: BP022944.D
Acq: 08 Nov 2024 22:18




Instrument :
BNA_P
ClientSampleId :
SSTD080632


Tgt Ion: 93 Resp: 436663
Ion Ratio Lower Upper
93 100
95 32.8 27.0 40.6
123 10.5 8.2 12.4



#29
2,4-Dichlorophenol
Concen: 77.501 ng/ul
RT: 10.005 min Scan# 1193
Delta R.T. -0.012 min
Lab File: BP022944.D
Acq: 08 Nov 2024 22:18

Tgt Ion: 162 Resp: 391261
Ion Ratio Lower Upper
162 100
164 66.1 51.4 77.0
98 34.9 27.0 40.6









#30
Naphthalene
Concen: 72.725 ng/ul
RT: 10.387 min Scan# 1258
Delta R.T. -0.012 min
Lab File: BP022944.D
Acq: 08 Nov 2024 22:18




Instrument :
BNA_P
ClientSampleId :
SSTD080632


Tgt Ion:128 Resp: 1049576
Ion Ratio Lower Upper
128 100
129 11.1 8.7 13.1
127 13.2 10.2 15.2



#32
4-Chloroaniline
Concen: 75.654 ng/ul
RT: 10.510 min Scan# 1279
Delta R.T. -0.006 min
Lab File: BP022944.D
Acq: 08 Nov 2024 22:18

Tgt Ion:127 Resp: 452375
Ion Ratio Lower Upper
127 100
129 31.2 26.4 39.6









#33
Hexachlorobutadiene
Concen: 77.204 ng/ul
RT: 10.663 min Scan# 11
Delta R.T. -0.012 min
Lab File: BP022944.D
Acq: 08 Nov 2024 22:18




Instrument :
BNA_P
ClientSampleId :
SSTD080632


Tgt Ion:225 Resp: 328501
Ion Ratio Lower Upper
225 100
223 62.2 49.7 74.5
227 64.8 51.4 77.0



#34
Caprolactam
Concen: 39.085 ng/ul
RT: 11.310 min Scan# 1415
Delta R.T. 0.000 min
Lab File: BP022944.D
Acq: 08 Nov 2024 22:18

Tgt Ion:113 Resp: 63613
Ion Ratio Lower Upper
113 100
55 150.7 123.0 184.6
56 126.9 99.8 149.8









#35
4-Chloro-3-methylphenol
Concen: 76.777 ng/ul
RT: 11.652 min Scan# 1473
Delta R.T. -0.006 min
Lab File: BP022944.D
Acq: 08 Nov 2024 22:18




Instrument :
BNA_P
ClientSampleId :
SSTD080632


Tgt Ion:107 Resp: 422899
Ion Ratio Lower Upper
107 100
144 23.1 18.8 28.2
142 72.0 59.8 89.8



#36
2-Methylnaphthalene
Concen: 76.411 ng/ul
RT: 12.004 min Scan# 1533
Delta R.T. -0.012 min
Lab File: BP022944.D
Acq: 08 Nov 2024 22:18

Tgt Ion:142 Resp: 809492
Ion Ratio Lower Upper
142 100
141 90.5 73.1 109.7





#37

1-Methylnaphthalene

Concen: 75.462 ng/ul

RT: 12.228 min Scan# 1571

Delta R.T. 0.000 min

Lab File: BP022944.D

Acq: 08 Nov 2024 22:18

Instrument :

BNA_P

ClientSampleId :

SSTD080632





Tgt Ion:142 Resp: 815966

Ion Ratio Lower Upper

142 100

141 92.9 73.8 110.8

116 4.9 4.2 6.4



#39

1,2,4,5-Tetrachlorobenzene

Concen: 73.817 ng/ul

RT: 12.381 min Scan# 1597

Delta R.T. -0.006 min

Lab File: BP022944.D

Acq: 08 Nov 2024 22:18

Tgt Ion:216 Resp: 604935




Ion Ratio Lower Upper


216 100

214 78.7 59.4 89.2

179 20.9 17.3 25.9

108 14.1 11.8 17.6





#40

Hexachlorocyclopentadiene

Concen: 69.623 ng/ul

RT: 12.357 min Scan# 1593

Delta R.T. 0.000 min

Lab File: BP022944.D

Acq: 08 Nov 2024 22:18

Instrument :

BNA_P

ClientSampleId :

SSTD080632





Tgt Ion:237 Resp: 347624

Ion Ratio Lower Upper

237 100

235 62.8 48.6 73.0

272 14.7 10.2 15.2



#41

2,4,6-Trichlorophenol

Concen: 76.716 ng/ul

RT: 12.634 min Scan# 1640

Delta R.T. -0.006 min

Lab File: BP022944.D

Acq: 08 Nov 2024 22:18




Tgt Ion:196 Resp: 395458


Ion Ratio Lower Upper

196 100

198 96.4 80.5 120.7

200 31.5 25.2 37.8









#42
2,4,5-Trichlorophenol
Concen: 79.033 ng/ul
RT: 12.716 min Scan# 1
Delta R.T. -0.012 min
Lab File: BP022944.D
Acq: 08 Nov 2024 22:18




Instrument :
BNA_P
ClientSampleId :
SSTD080632


Tgt Ion:196 Resp: 433792
Ion Ratio Lower Upper
196 100
198 94.9 79.6 119.4
200 31.1 25.0 37.6



#43
1,1'-Biphenyl
Concen: 71.806 ng/ul
RT: 13.034 min Scan# 1708
Delta R.T. 0.000 min
Lab File: BP022944.D
Acq: 08 Nov 2024 22:18

Tgt Ion:154 Resp: 1147100
Ion Ratio Lower Upper
154 100
153 41.1 32.0 48.0
76 11.1 8.4 12.6









#44
 2-Chloronaphthalene
 Concen: 72.669 ng/ul
 RT: 13.081 min Scan# 1716
 Delta R.T. 0.000 min
 Lab File: BP022944.D
 Acq: 08 Nov 2024 22:18

Instrument :
 BNA_P
 ClientSampleId :
 SSTD080632

Tgt Ion: 162 Resp: 950330




Ion	Ratio	Lower	Upper
162	100		
164	32.0	26.6	40.0
127	37.0	30.2	45.2




#45
 2-Nitroaniline
 Concen: 74.743 ng/ul
 RT: 13.293 min Scan# 1752
 Delta R.T. 0.006 min
 Lab File: BP022944.D
 Acq: 08 Nov 2024 22:18

Tgt Ion: 65 Resp: 291860

Ion	Ratio	Lower	Upper
65	100		
92	67.5	54.2	81.2
138	94.0	65.1	97.7





#47

Dimethylphthalate

Concen: 71.787 ng/ul

RT: 13.681 min Scan# 1818

Delta R.T. 0.018 min

Lab File: BP022944.D

Acq: 08 Nov 2024 22:18

Instrument :

BNA_P

ClientSampleId :

SSTD080632





Tgt Ion:163 Resp: 1259619

Ion Ratio Lower Upper

163 100

194 4.3 3.8 5.6

164 9.9 8.2 12.4



#48

2,6-Dinitrotoluene

Concen: 81.096 ng/ul

RT: 13.804 min Scan# 1839

Delta R.T. -0.006 min

Lab File: BP022944.D

Acq: 08 Nov 2024 22:18




Tgt Ion:165 Resp: 269152


Ion Ratio Lower Upper

165 100

89 61.2 50.6 75.8

121 20.2 15.8 23.8





#50

Acenaphthylene

Concen: 71.692 ng/ul

RT: 13.922 min Scan# 1859

Delta R.T. -0.012 min

Lab File: BP022944.D

Acq: 08 Nov 2024 22:18

Instrument :

BNA_P

ClientSampleId :

SSTD080632





Tgt Ion:152 Resp: 1391595

Ion Ratio Lower Upper

152 100

151 19.9 15.9 23.9

153 13.2 10.2 15.2



#51

3-Nitroaniline

Concen: 67.596 ng/ul

RT: 14.146 min Scan# 1897

Delta R.T. -0.006 min

Lab File: BP022944.D

Acq: 08 Nov 2024 22:18




Tgt Ion:138 Resp: 210712


Ion Ratio Lower Upper

138 100

108 12.8 12.7 19.1

92 124.6 107.8 161.8









#52
Acenaphthene
Concen: 72.891 ng/ul
RT: 14.269 min Scan# 1918
Delta R.T. -0.006 min
Lab File: BP022944.D
Acq: 08 Nov 2024 22:18




Instrument :
BNA_P
ClientSampleId :
SSTD080632


Tgt Ion:153 Resp: 1000502
Ion Ratio Lower Upper
153 100
152 46.3 39.6 59.4
154 87.8 70.9 106.3



#53
2,4-Dinitrophenol
Concen: 82.306 ng/ul
RT: 14.357 min Scan# 1933
Delta R.T. -0.006 min
Lab File: BP022944.D
Acq: 08 Nov 2024 22:18

Tgt Ion:184 Resp: 200983
Ion Ratio Lower Upper
184 100
63 59.4 50.1 75.1
154 73.5 58.2 87.4









#55
4-Nitrophenol
Concen: 78.021 ng/ul
RT: 14.475 min Scan# 1954
Delta R.T. -0.006 min
Lab File: BP022944.D
Acq: 08 Nov 2024 22:18




Instrument :
BNA_P
ClientSampleId :
SSTD080632


Tgt Ion:109 Resp: 255810
Ion Ratio Lower Upper
109 100
139 82.0 64.6 96.8
65 101.2 80.6 120.8



#56
Dibenzofuran
Concen: 73.032 ng/ul
RT: 14.616 min Scan# 1977
Delta R.T. 0.000 min
Lab File: BP022944.D
Acq: 08 Nov 2024 22:18

Tgt Ion:168 Resp: 1505778
Ion Ratio Lower Upper
168 100
139 37.9 29.7 44.5









#57
2,4-Dinitrotoluene
Concen: 81.031 ng/ul
RT: 14.598 min Scan# 1974
Delta R.T. -0.006 min
Lab File: BP022944.D
Acq: 08 Nov 2024 22:18




Instrument :
BNA_P
ClientSampleId :
SSTD080632


Tgt Ion:165 Resp: 417904
Ion Ratio Lower Upper
165 100
63 45.5 41.3 61.9
182 2.8 2.5 3.7



#58
2,3,4,6-Tetrachlorophenol
Concen: 80.153 ng/ul
RT: 14.863 min Scan# 2019
Delta R.T. -0.012 min
Lab File: BP022944.D
Acq: 08 Nov 2024 22:18

Tgt Ion:232 Resp: 410637
Ion Ratio Lower Upper
232 100
131 36.1 27.0 40.6
130 2.3 1.9 2.9
166 27.1 21.8 32.6









#59
Diethylphthalate
Concen: 74.657 ng/ul
RT: 15.057 min Scan# 2052
Delta R.T. 0.000 min
Lab File: BP022944.D
Acq: 08 Nov 2024 22:18




Instrument :
BNA_P
ClientSampleId :
SSTD080632


Tgt Ion:149 Resp: 1256779
Ion Ratio Lower Upper
149 100
177 24.2 19.7 29.5
150 12.7 9.8 14.6



#61
Fluorene
Concen: 75.446 ng/ul
RT: 15.275 min Scan# 2089
Delta R.T. -0.017 min
Lab File: BP022944.D
Acq: 08 Nov 2024 22:18

Tgt Ion:166 Resp: 1267454
Ion Ratio Lower Upper
166 100
165 97.7 79.5 119.3
167 13.4 10.3 15.5









#62
4-Chlorophenyl-phenylether
Concen: 76.794 ng/ul
RT: 15.269 min Scan# 2088
Delta R.T. -0.012 min
Lab File: BP022944.D
Acq: 08 Nov 2024 22:18




Instrument :
BNA_P
ClientSampleId :
SSTD080632


Tgt Ion:204 Resp: 737753
Ion Ratio Lower Upper
204 100
206 33.0 27.2 40.8
141 52.6 44.6 66.8



#63
4-Nitroaniline
Concen: 60.019 ng/ul
RT: 15.316 min Scan# 2096
Delta R.T. -0.006 min
Lab File: BP022944.D
Acq: 08 Nov 2024 22:18

Tgt Ion:138 Resp: 188262
Ion Ratio Lower Upper
138 100
92 62.1 53.9 80.9
108 128.2 106.6 159.8









#66
4,6-Dinitro-2-methylphenol
Concen: 76.005 ng/ul
RT: 15.387 min Scan# 2108
Delta R.T. 0.000 min
Lab File: BP022944.D
Acq: 08 Nov 2024 22:18




Instrument :
BNA_P
ClientSampleId :
SSTD080632


Tgt Ion:198 Resp: 309053
Ion Ratio Lower Upper
198 100
182 3.8 3.8 5.8#
77 21.9 20.0 30.0



#67
N-Nitrosodiphenylamine
Concen: 69.313 ng/ul
RT: 15.504 min Scan# 2128
Delta R.T. 0.000 min
Lab File: BP022944.D
Acq: 08 Nov 2024 22:18

Tgt Ion:169 Resp: 1065405
Ion Ratio Lower Upper
169 100
168 67.9 53.4 80.0
167 35.9 29.4 44.2









#68
4-Bromophenyl-phenylether
Concen: 75.036 ng/ul
RT: 16.192 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP022944.D
Acq: 08 Nov 2024 22:18




Instrument :
BNA_P
ClientSampleId :
SSTD080632


Tgt Ion:248 Resp: 519392
Ion Ratio Lower Upper
248 100
250 99.2 77.1 115.7
141 58.8 50.7 76.1



#69
Hexachlorobenzene
Concen: 75.862 ng/ul
RT: 16.310 min Scan# 2265
Delta R.T. -0.006 min
Lab File: BP022944.D
Acq: 08 Nov 2024 22:18

Tgt Ion:284 Resp: 646352
Ion Ratio Lower Upper
284 100
142 24.4 21.5 32.3
249 29.8 25.6 38.4





#70

Atrazine

Concen: 80.506 ng/ul

RT: 16.481 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP022944.D

Acq: 08 Nov 2024 22:18

Instrument :

BNA_P

ClientSampleId :

SSTD080632





Tgt Ion:200 Resp: 514378

Ion Ratio Lower Upper

200 100

173 22.9 18.6 27.8

215 48.4 38.3 57.5



#71

Pentachlorophenol

Concen: 76.692 ng/ul

RT: 16.669 min Scan# 2326

Delta R.T. -0.017 min

Lab File: BP022944.D

Acq: 08 Nov 2024 22:18




Tgt Ion:266 Resp: 420144


Ion Ratio Lower Upper

266 100

264 62.3 50.7 76.1

268 62.1 51.9 77.9





#72

Phenanthrene

Concen: 69.998 ng/ul

RT: 17.057 min Scan# 21

Delta R.T. -0.012 min

Lab File: BP022944.D

Acq: 08 Nov 2024 22:18

Instrument :

BNA_P

ClientSampleId :

SSTD080632





Tgt Ion:178 Resp: 2149575

Ion Ratio Lower Upper

178 100

179 15.7 12.7 19.1

176 19.3 15.4 23.2



#74

Anthracene

Concen: 70.835 ng/ul

RT: 17.151 min Scan# 2408

Delta R.T. 0.000 min

Lab File: BP022944.D

Acq: 08 Nov 2024 22:18




Tgt Ion:178 Resp: 2170791


Ion Ratio Lower Upper

178 100

179 15.2 12.4 18.6

176 18.9 14.5 21.7





#75

1,2,3,4-Tetrachlorobenzene

Concen: 70.193 ng/uL

RT: 12.999 min Scan# 1702

Delta R.T. 0.000 min

Lab File: BP022944.D

Acq: 08 Nov 2024 22:18

Instrument :

BNA_P

ClientSampleId :

SSTD080632





Tgt Ion:216 Resp: 613467

Ion Ratio Lower Upper

216 100

214 77.0 60.2 90.2

218 48.0 38.1 57.1



#76

Pentachlorobenzene

Concen: 71.742 ng/uL

RT: 14.546 min Scan# 1965

Delta R.T. 0.000 min

Lab File: BP022944.D

Acq: 08 Nov 2024 22:18




Tgt Ion:250 Resp: 693132


Ion Ratio Lower Upper

250 100

248 63.0 52.2 78.2

252 64.7 50.1 75.1









#77
 Carbazole
 Concen: 69.539 ng/ul
 RT: 17.439 min Scan# 2458
 Delta R.T. -0.006 min
 Lab File: BP022944.D
 Acq: 08 Nov 2024 22:18




Instrument :
 BNA_P
 ClientSampleId :
 SSTD080632


Tgt Ion:167 Resp: 1885652
 Ion Ratio Lower Upper
 167 100
 166 21.9 17.5 26.3
 139 12.1 9.6 14.4



#78
 Di-n-butylphthalate
 Concen: 72.588 ng/ul
 RT: 18.028 min Scan# 2557
 Delta R.T. -0.006 min
 Lab File: BP022944.D
 Acq: 08 Nov 2024 22:18

Tgt Ion:149 Resp: 2177918
 Ion Ratio Lower Upper
 149 100
 150 9.3 7.2 10.8
 104 5.8 5.0 7.4









#80
Fluoranthene
Concen: 74.436 ng/ul
RT: 19.169 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP022944.D
Acq: 08 Nov 2024 22:18




Instrument :
BNA_P
ClientSampleId :
SSTD080632


Tgt Ion:202 Resp: 2870008
Ion Ratio Lower Upper
202 100
101 6.6 5.2 7.8
100 5.2 4.1 6.1



#82
Pyrene
Concen: 72.459 ng/ul
RT: 19.545 min Scan# 2815
Delta R.T. 0.006 min
Lab File: BP022944.D
Acq: 08 Nov 2024 22:18

Tgt Ion:202 Resp: 2994375
Ion Ratio Lower Upper
202 100
101 7.0 6.2 9.2
100 6.0 4.9 7.3





#83

Butylbenzylphthalate

Concen: 75.456 ng/ul

RT: 20.486 min Scan# 2975

Delta R.T. -0.006 min

Lab File: BP022944.D

Acq: 08 Nov 2024 22:18

Instrument :

BNA_P

ClientSampleId :

SSTD080632





Tgt Ion:149 Resp: 1085655

Ion Ratio Lower Upper

149 100

91 68.1 55.6 83.4

206 25.5 20.2 30.2



#84

3,3'-Dichlorobenzidine

Concen: 75.487 ng/ul

RT: 21.363 min Scan# 3124

Delta R.T. 0.000 min

Lab File: BP022944.D

Acq: 08 Nov 2024 22:18




Tgt Ion:252 Resp: 1141249


Ion Ratio Lower Upper

252 100

254 64.6 52.0 78.0

126 6.3 5.0 7.6









#85
Benzo(a)anthracene
Concen: 70.994 ng/ul
RT: 21.439 min Scan# 3137
Delta R.T. 0.000 min
Lab File: BP022944.D
Acq: 08 Nov 2024 22:18




Instrument :
BNA_P
ClientSampleId :
SSTD080632


Tgt Ion:228 Resp: 3156500
Ion Ratio Lower Upper
228 100
229 19.9 15.5 23.3
226 26.9 21.9 32.9



#86
Bis(2-ethylhexyl)phthalate
Concen: 74.695 ng/ul
RT: 21.369 min Scan# 3125
Delta R.T. 0.000 min
Lab File: BP022944.D
Acq: 08 Nov 2024 22:18

Tgt Ion:149 Resp: 1542613
Ion Ratio Lower Upper
149 100
167 30.2 24.5 36.7
279 6.6 5.5 8.3





#87

Chrysene

Concen: 70.658 ng/ul

RT: 21.504 min Scan# 3147

Delta R.T. 0.006 min

Lab File: BP022944.D

Acq: 08 Nov 2024 22:18

Instrument :

BNA_P

ClientSampleId :

SSTD080632





Tgt Ion:228 Resp: 2912927

Ion Ratio Lower Upper

228 100

226 29.9 24.2 36.2

229 19.6 16.1 24.1



#89

Di-n-octyl phthalate

Concen: 71.559 ng/ul




RT: 22.592 min Scan# 3333


Delta R.T. 0.000 min

Lab File: BP022944.D

Acq: 08 Nov 2024 22:18

Tgt Ion:149 Resp: 2526099









#90
Benzo(b)fluoranthene
Concen: 74.427 ng/ul
RT: 23.669 min Scan# 3516
Delta R.T. -0.006 min
Lab File: BP022944.D
Acq: 08 Nov 2024 22:18




Instrument :
BNA_P
ClientSampleId :
SSTD080632


Tgt Ion:252 Resp: 3476847
Ion Ratio Lower Upper
252 100
253 22.0 17.4 26.2
125 4.8 3.7 5.5



#91
Benzo(k)fluoranthene
Concen: 73.564 ng/ul
RT: 23.745 min Scan# 3529
Delta R.T. 0.006 min
Lab File: BP022944.D
Acq: 08 Nov 2024 22:18

Tgt Ion:252 Resp: 3364566
Ion Ratio Lower Upper
252 100
253 21.9 17.8 26.8
125 4.1 3.0 4.4









#93
Benzo(a)pyrene
Concen: 72.963 ng/ul
RT: 24.545 min Scan# 30
Delta R.T. 0.006 min
Lab File: BP022944.D
Acq: 08 Nov 2024 22:18




Instrument :
BNA_P
ClientSampleId :
SSTD080632


Tgt Ion:252 Resp: 3136980
Ion Ratio Lower Upper
252 100
253 21.6 17.6 26.4
125 4.9 4.2 6.2



#94
Indeno(1,2,3-cd)pyrene
Concen: 24.786 ng/ul
RT: 28.339 min Scan# 4310
Delta R.T. -0.006 min
Lab File: BP022944.D
Acq: 08 Nov 2024 22:18

Tgt Ion:276 Resp: 1422051
Ion Ratio Lower Upper
276 100
138 8.1 7.0 10.4
277 23.9 19.8 29.8





#95

Dibenzo(a,h)anthracene

Concen: 73.454 ng/ul

RT: 28.433 min Scan# 41

Delta R.T. 0.030 min

Lab File: BP022944.D

Acq: 08 Nov 2024 22:18

Instrument :

BNA_P

ClientSampleId :

SSTD080632





Tgt Ion:278 Resp: 3412638

Ion Ratio Lower Upper

278 100

139 6.3 5.4 8.0

279 24.0 19.4 29.2



#96

Benzo(g,h,i)perylene

Concen: 72.350 ng/ul

RT: 29.498 min Scan# 4507

Delta R.T. 0.018 min

Lab File: BP022944.D

Acq: 08 Nov 2024 22:18

Tgt Ion:276 Resp: 3228735

Ion Ratio Lower Upper

276 100

138 9.6 7.1 10.7

277 23.7 19.4 29.2

