

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP110924\
 Data File : BP023005.D
 Acq On : 10 Nov 2024 18:16
 Operator : RC/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 92 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020709

Quant Time: Nov 10 22:42:08 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP110924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Nov 08 23:49:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.587	152	95524	20.000	ng/u1	-0.01
20) Naphthalene-d8	10.334	136	382500	20.000	ng/u1	-0.02
38) Acenaphthene-d10	14.204	164	317887	20.000	ng/u1	-0.02
64) Phenanthrene-d10	17.016	188	789679	20.000	ng/u1	-0.02
79) Chrysene-d12	21.451	240	861650	20.000	ng/u1	0.00
88) Perylene-d12	24.674	264	995202	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.140	96	16069	8.086	ng/uL	0.00
4) Pyridine-d5	3.552	84	114416	19.378	ng/u1	0.00
7) Phenol-d5	6.793	99	141586	20.048	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.934	67	83384	20.096	ng/u1	-0.01
11) 2-Chlorophenol-d4	7.134	132	108210	20.921	ng/u1	0.00
15) 4-Methylphenol-d8	8.293	113	117794	20.268	ng/u1	-0.01
21) Nitrobenzene-d5	8.728	128	54667	20.529	ng/u1	0.00
24) 2-Nitrophenol-d4	9.440	143	66113	20.826	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	9.981	165	138791	20.876	ng/u1	-0.01
31) 4-Chloroaniline-d4	10.481	131	158262	19.954	ng/u1	-0.02
46) Dimethylphthalate-d6	13.604	166	467298	20.396	ng/u1	-0.01
49) Acenaphthylene-d8	13.893	160	504116	21.087	ng/u1	-0.01
54) 4-Nitrophenol-d4	14.457	143	67411	19.579	ng/u1	-0.01
60) Fluorene-d10	15.222	176	410161	20.419	ng/u1	-0.01
65) 4,6-Dinitro-2-methylph...	15.357	200	95919	20.276	ng/u1	-0.01
73) Anthracene-d10	17.110	188	700960	20.964	ng/u1	-0.01
81) Pyrene-d10	19.498	212	888963	22.399	ng/u1	-0.01
92) Benzo(a)pyrene-d12	24.451	264	1016306	21.439	ng/u1	-0.02
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.175	88	17032	7.417	ng/uL	95
5) Pyridine	3.575	79	116798	19.697	ng/u1	91
6) Benzaldehyde	6.758	77	102506	25.222	ng/u1	96
8) Phenol	6.816	94	147552	20.007	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.028	93	113552	20.490	ng/u1	92
12) 2-Chlorophenol	7.164	128	112667	20.981	ng/u1	96
13) 2-Methylphenol	8.040	108	111326	20.223	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.099	45	130260	20.763	ng/u1	95
16) Acetophenone	8.399	105	198733	20.350	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.381	70	100744	20.660	ng/u1	98
18) 4-Methylphenol	8.363	108	123247	20.269	ng/u1	98
19) Hexachloroethane	8.634	117	52312	20.142	ng/u1	99
22) Nitrobenzene	8.775	77	158496	20.743	ng/u1	98
23) Isophorone	9.287	82	286809	20.993	ng/u1	98
25) 2-Nitrophenol	9.475	139	69388	20.708	ng/u1	91
26) 2,4-Dimethylphenol	9.540	107	149310	20.635	ng/u1	99
27) Bis(2-Chloroethoxy)met...	9.757	93	160305	21.325	ng/u1	98
29) 2,4-Dichlorophenol	10.005	162	137709	20.912	ng/u1	97
30) Naphthalene	10.387	128	374287	20.142	ng/u1	99
32) 4-Chloroaniline	10.505	127	151885	19.766	ng/u1	98
33) Hexachlorobutadiene	10.663	225	114289	19.181	ng/u1	98
34) Caprolactam	11.293	113	38291	19.549	ng/u1	96
35) 4-Chloro-3-methylphenol	11.652	107	149321	21.078	ng/u1	98
36) 2-Methylnaphthalene	11.999	142	288632	20.574	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP110924\
 Data File : BP023005.D
 Acq On : 10 Nov 2024 18:16
 Operator : RC/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 92 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020709

Quant Time: Nov 10 22:42:08 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP110924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Nov 08 23:49:48 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.216	142	294093	20.506	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.375	216	215729	20.102	ng/ul	97
40) Hexachlorocyclopentadiene	12.346	237	107089	18.998	ng/ul	99
41) 2,4,6-Trichlorophenol	12.622	196	136490	20.886	ng/ul	98
42) 2,4,5-Trichlorophenol	12.710	196	147755	20.973	ng/ul	96
43) 1,1'-Biphenyl	13.016	154	420603	20.649	ng/ul	99
44) 2-Chloronaphthalene	13.069	162	340004	20.419	ng/ul	98
45) 2-Nitroaniline	13.281	65	98089	20.980	ng/ul	90
47) Dimethylphthalate	13.657	163	458459	20.290	ng/ul	98
48) 2,6-Dinitrotoluene	13.798	165	94548	21.512	ng/ul	98
50) Acenaphthylene	13.916	152	513037	21.231	ng/ul	100
51) 3-Nitroaniline	14.140	138	81407	21.504	ng/ul	96
52) Acenaphthene	14.269	153	357956	20.404	ng/ul	99
53) 2,4-Dinitrophenol	14.345	184	56318	18.059	ng/ul	95
55) 4-Nitrophenol	14.475	109	78371	19.455	ng/ul	94
56) Dibenzofuran	14.616	168	547137	20.185	ng/ul	100
57) 2,4-Dinitrotoluene	14.598	165	147630	21.075	ng/ul	93
58) 2,3,4,6-Tetrachlorophenol	14.857	232	140064	20.339	ng/ul	97
59) Diethylphthalate	15.040	149	454307	20.330	ng/ul	99
61) Fluorene	15.275	166	451537	20.353	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.263	204	261262	20.000	ng/ul	98
63) 4-Nitroaniline	15.310	138	79974	22.704	ng/ul	89
66) 4,6-Dinitro-2-methylph...	15.375	198	101945	20.174	ng/ul	97
67) N-Nitrosodiphenylamine	15.481	169	389573	21.349	ng/ul	99
68) 4-Bromophenyl-phenylether	16.169	248	183148	20.481	ng/ul	97
69) Hexachlorobenzene	16.304	284	230202	20.449	ng/ul	98
70) Atrazine	16.463	200	181588	20.587	ng/ul	99
71) Pentachlorophenol	16.669	266	136105	19.474	ng/ul	99
72) Phenanthrene	17.057	178	772994	20.732	ng/ul	100
74) Anthracene	17.145	178	784992	20.941	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	12.987	216	218650	20.971	ng/uL	98
76) Pentachlorobenzene	14.534	250	244608	20.547	ng/uL	97
77) Carbazole	17.439	167	687680	21.027	ng/ul	100
78) Di-n-butylphthalate	18.022	149	769950	21.215	ng/ul	100
80) Fluoranthene	19.151	202	1002481	21.925	ng/ul	100
82) Pyrene	19.533	202	1064838	21.679	ng/ul	100
83) Butylbenzylphthalate	20.480	149	359404	22.224	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.357	252	418758	21.651	ng/ul	99
85) Benzo(a)anthracene	21.433	228	1108751	21.073	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.357	149	508674	22.134	ng/ul	97
87) Chrysene	21.498	228	1037414	20.955	ng/ul	99
89) Di-n-octyl phthalate	22.580	149	828649	20.524	ng/ul	100
90) Benzo(b)fluoranthene	23.657	252	1205322	21.760	ng/ul	99
91) Benzo(k)fluoranthene	23.657	252	1205322	22.104	ng/ul	99
93) Benzo(a)pyrene	24.527	252	1074684	21.462	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	28.315	276	1419424	20.346	ng/ul	99
95) Dibenzo(a,h)anthracene	28.380	278	1139738	20.121	ng/ul	98
96) Benzo(g,h,i)perylene	29.456	276	884276	16.688	ng/ul	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP110924\
 Data File : BP023005.D
 Acq On : 10 Nov 2024 18:16
 Operator : RC/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 92 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020709

Quant Time: Nov 10 22:42:08 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP110924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Nov 08 23:49:48 2024
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 7.417 ng/uL
RT: 3.175 min Scan# 31
Delta R.T. 0.000 min
Lab File: BP023005.D
Acq: 10 Nov 2024 18:16




Instrument :
BNA_P
ClientSampleId :
SSTD020709


Tgt Ion: 88 Resp: 17032
Ion Ratio Lower Upper
88 100
43 27.8 24.2 36.4
58 69.2 59.2 88.8



#5
Pyridine
Concen: 19.697 ng/uL
RT: 3.575 min Scan# 100
Delta R.T. 0.000 min
Lab File: BP023005.D
Acq: 10 Nov 2024 18:16

Tgt Ion: 79 Resp: 116798
Ion Ratio Lower Upper
79 100
52 66.1 59.5 89.3
51 34.0 30.4 45.6









#6
Benzaldehyde
Concen: 25.222 ng/ul
RT: 6.758 min Scan# 641
Delta R.T. 0.000 min
Lab File: BP023005.D
Acq: 10 Nov 2024 18:16




Instrument :
BNA_P
ClientSampleId :
SSTD020709


Tgt Ion: 77 Resp: 102506
Ion Ratio Lower Upper
77 100
105 92.3 72.5 108.7
106 87.7 65.5 98.3



#8
Phenol
Concen: 20.007 ng/ul
RT: 6.816 min Scan# 651
Delta R.T. -0.006 min
Lab File: BP023005.D
Acq: 10 Nov 2024 18:16

Tgt Ion: 94 Resp: 147552
Ion Ratio Lower Upper
94 100
65 32.3 26.2 39.4
66 49.6 40.2 60.4









#10
Bis(2-Chloroethyl)ether
Concen: 20.490 ng/ul
RT: 7.028 min Scan# 61
Delta R.T. -0.006 min
Lab File: BP023005.D
Acq: 10 Nov 2024 18:16




Instrument :
BNA_P
ClientSampleId :
SSTD020709


Tgt Ion: 93 Resp: 113552
Ion Ratio Lower Upper
93 100
63 68.1 60.6 91.0
95 32.5 29.0 43.4



#12
2-Chlorophenol
Concen: 20.981 ng/ul
RT: 7.164 min Scan# 710
Delta R.T. -0.012 min
Lab File: BP023005.D
Acq: 10 Nov 2024 18:16

Tgt Ion:128 Resp: 112667
Ion Ratio Lower Upper
128 100
64 52.4 39.0 58.6
130 32.1 26.4 39.6









#13
2-Methylphenol
Concen: 20.223 ng/ul
RT: 8.040 min Scan# 859
Delta R.T. 0.000 min
Lab File: BP023005.D
Acq: 10 Nov 2024 18:16




Instrument :
BNA_P
ClientSampleId :
SSTD020709


Tgt Ion:108 Resp: 111326
Ion Ratio Lower Upper
108 100
107 92.1 73.3 109.9



#14
2,2'-oxybis(1-Chloropropane)
Concen: 20.763 ng/ul
RT: 8.099 min Scan# 869
Delta R.T. 0.000 min
Lab File: BP023005.D
Acq: 10 Nov 2024 18:16

Tgt Ion: 45 Resp: 130260
Ion Ratio Lower Upper
45 100
77 19.6 18.2 27.2
79 13.7 11.7 17.5









#16
Acetophenone
Concen: 20.350 ng/ul
RT: 8.399 min Scan# 916
Delta R.T. -0.006 min
Lab File: BP023005.D
Acq: 10 Nov 2024 18:16




Instrument :
BNA_P
ClientSampleId :
SSTD020709


Tgt Ion:105 Resp: 198733
Ion Ratio Lower Upper
105 100
77 84.0 69.4 104.0
51 27.0 23.2 34.8



#17
N-Nitroso-di-n-propylamine
Concen: 20.660 ng/ul
RT: 8.381 min Scan# 917
Delta R.T. -0.006 min
Lab File: BP023005.D
Acq: 10 Nov 2024 18:16

Tgt Ion: 70 Resp: 100744
Ion Ratio Lower Upper
70 100
42 51.7 42.3 63.5
101 9.0 7.4 11.2
130 17.8 15.4 23.0









#18
4-Methylphenol
Concen: 20.269 ng/ul
RT: 8.363 min Scan# 914
Delta R.T. 0.000 min
Lab File: BP023005.D
Acq: 10 Nov 2024 18:16




Instrument :
BNA_P
ClientSampleId :
SSTD020709


Tgt Ion:108 Resp: 123247
Ion Ratio Lower Upper
108 100
107 116.8 91.5 137.3



#19
Hexachloroethane
Concen: 20.142 ng/ul
RT: 8.634 min Scan# 960
Delta R.T. -0.012 min
Lab File: BP023005.D
Acq: 10 Nov 2024 18:16

Tgt Ion:117 Resp: 52312
Ion Ratio Lower Upper
117 100
201 130.9 105.0 157.6
199 79.4 65.9 98.9









#22
Nitrobenzene
Concen: 20.743 ng/ul
RT: 8.775 min Scan# 984
Delta R.T. 0.000 min
Lab File: BP023005.D
Acq: 10 Nov 2024 18:16




Instrument :
BNA_P
ClientSampleId :
SSTD020709


Tgt Ion: 77 Resp: 158496
Ion Ratio Lower Upper
77 100
123 39.3 32.2 48.4
65 14.2 10.9 16.3



#23
Isophorone
Concen: 20.993 ng/ul
RT: 9.287 min Scan# 1071
Delta R.T. -0.012 min
Lab File: BP023005.D
Acq: 10 Nov 2024 18:16

Tgt Ion: 82 Resp: 286809
Ion Ratio Lower Upper
82 100
95 8.6 7.6 11.4
138 15.8 13.4 20.2









#25
2-Nitrophenol
Concen: 20.708 ng/ul
RT: 9.475 min Scan# 1103
Delta R.T. 0.000 min
Lab File: BP023005.D
Acq: 10 Nov 2024 18:16




Instrument :
BNA_P
ClientSampleId :
SSTD020709


Tgt Ion:139 Resp: 69388
Ion Ratio Lower Upper
139 100
65 56.1 52.8 79.2
109 45.6 38.9 58.3



#26
2,4-Dimethylphenol
Concen: 20.635 ng/ul
RT: 9.540 min Scan# 1114
Delta R.T. -0.006 min
Lab File: BP023005.D
Acq: 10 Nov 2024 18:16

Tgt Ion:107 Resp: 149310
Ion Ratio Lower Upper
107 100
121 43.7 34.4 51.6
122 71.6 56.7 85.1









#27
Bis(2-Chloroethoxy)methane
Concen: 21.325 ng/ul
RT: 9.757 min Scan# 1151
Delta R.T. -0.012 min
Lab File: BP023005.D
Acq: 10 Nov 2024 18:16




Instrument :
BNA_P
ClientSampleId :
SSTD020709


Tgt Ion: 93 Resp: 160305
Ion Ratio Lower Upper
93 100
95 32.2 27.0 40.6
123 10.7 8.2 12.4



#29
2,4-Dichlorophenol
Concen: 20.912 ng/ul
RT: 10.005 min Scan# 1193
Delta R.T. -0.012 min
Lab File: BP023005.D
Acq: 10 Nov 2024 18:16

Tgt Ion:162 Resp: 137709
Ion Ratio Lower Upper
162 100
164 66.8 51.4 77.0
98 35.2 27.0 40.6









#30
Naphthalene
Concen: 20.142 ng/ul
RT: 10.387 min Scan# 1258
Delta R.T. -0.012 min
Lab File: BP023005.D
Acq: 10 Nov 2024 18:16




Instrument :
BNA_P
ClientSampleId :
SSTD020709


Tgt Ion:128 Resp: 374287
Ion Ratio Lower Upper
128 100
129 10.9 8.7 13.1
127 13.7 10.2 15.2



#32
4-Chloroaniline
Concen: 19.766 ng/ul
RT: 10.505 min Scan# 1278
Delta R.T. -0.012 min
Lab File: BP023005.D
Acq: 10 Nov 2024 18:16

Tgt Ion:127 Resp: 151885
Ion Ratio Lower Upper
127 100
129 33.9 26.4 39.6









#33
Hexachlorobutadiene
Concen: 19.181 ng/ul
RT: 10.663 min Scan# 11
Delta R.T. -0.012 min
Lab File: BP023005.D
Acq: 10 Nov 2024 18:16




Instrument :
BNA_P
ClientSampleId :
SSTD020709


Tgt Ion:225 Resp: 114289
Ion Ratio Lower Upper
225 100
223 61.1 49.7 74.5
227 66.2 51.4 77.0



#34
Caprolactam
Concen: 19.549 ng/ul
RT: 11.293 min Scan# 1412
Delta R.T. -0.017 min
Lab File: BP023005.D
Acq: 10 Nov 2024 18:16

Tgt Ion:113 Resp: 38291
Ion Ratio Lower Upper
113 100
55 158.2 123.0 184.6
56 129.8 99.8 149.8









#35
4-Chloro-3-methylphenol
Concen: 21.078 ng/ul
RT: 11.652 min Scan# 1473
Delta R.T. -0.006 min
Lab File: BP023005.D
Acq: 10 Nov 2024 18:16




Instrument :
BNA_P
ClientSampleId :
SSTD020709


Tgt Ion:107 Resp: 149321
Ion Ratio Lower Upper
107 100
144 24.7 18.8 28.2
142 73.4 59.8 89.8



#36
2-Methylnaphthalene
Concen: 20.574 ng/ul
RT: 11.999 min Scan# 1532
Delta R.T. -0.017 min
Lab File: BP023005.D
Acq: 10 Nov 2024 18:16

Tgt Ion:142 Resp: 288632
Ion Ratio Lower Upper
142 100
141 90.7 73.1 109.7









#37
1-Methylnaphthalene
Concen: 20.506 ng/ul
RT: 12.216 min Scan# 1569
Delta R.T. -0.012 min
Lab File: BP023005.D
Acq: 10 Nov 2024 18:16




Instrument :
BNA_P
ClientSampleId :
SSTD020709


Tgt Ion:142 Resp: 294093
Ion Ratio Lower Upper
142 100
141 93.7 73.8 110.8
116 5.0 4.2 6.4



#39
1,2,4,5-Tetrachlorobenzene
Concen: 20.102 ng/ul
RT: 12.375 min Scan# 1596
Delta R.T. -0.012 min
Lab File: BP023005.D
Acq: 10 Nov 2024 18:16

Tgt Ion:216 Resp: 215729
Ion Ratio Lower Upper
216 100
214 76.9 59.4 89.2
179 20.5 17.3 25.9
108 14.0 11.8 17.6









#40
Hexachlorocyclopentadiene
Concen: 18.998 ng/ul
RT: 12.346 min Scan# 1
Delta R.T. -0.012 min
Lab File: BP023005.D
Acq: 10 Nov 2024 18:16




Instrument :
BNA_P
ClientSampleId :
SSTD020709


Tgt Ion:237 Resp: 107089
Ion Ratio Lower Upper
237 100
235 61.2 48.6 73.0
272 14.4 10.2 15.2



#41
2,4,6-Trichlorophenol
Concen: 20.886 ng/ul
RT: 12.622 min Scan# 1638
Delta R.T. -0.017 min
Lab File: BP023005.D
Acq: 10 Nov 2024 18:16

Tgt Ion:196 Resp: 136490
Ion Ratio Lower Upper
196 100
198 98.2 80.5 120.7
200 31.5 25.2 37.8





#42

2,4,5-Trichlorophenol

Concen: 20.973 ng/ul

RT: 12.710 min Scan# 1

Delta R.T. -0.017 min

Lab File: BP023005.D

Acq: 10 Nov 2024 18:16

Instrument :

BNA_P

ClientSampleId :

SSTD020709





Tgt Ion:196 Resp: 147755

Ion Ratio Lower Upper

196 100

198 95.9 79.6 119.4

200 29.2 25.0 37.6



#43

1,1'-Biphenyl

Concen: 20.649 ng/ul

RT: 13.016 min Scan# 1705

Delta R.T. -0.017 min

Lab File: BP023005.D

Acq: 10 Nov 2024 18:16




Tgt Ion:154 Resp: 420603


Ion Ratio Lower Upper

154 100

153 40.7 32.0 48.0

76 11.6 8.4 12.6









#44
2-Chloronaphthalene
Concen: 20.419 ng/ul
RT: 13.069 min Scan# 1714
Delta R.T. -0.012 min
Lab File: BP023005.D
Acq: 10 Nov 2024 18:16




Instrument :
BNA_P
ClientSampleId :
SSTD020709


Tgt Ion:162 Resp: 340004
Ion Ratio Lower Upper
162 100
164 31.7 26.6 40.0
127 37.2 30.2 45.2



#45
2-Nitroaniline
Concen: 20.980 ng/ul
RT: 13.281 min Scan# 1750
Delta R.T. -0.006 min
Lab File: BP023005.D
Acq: 10 Nov 2024 18:16

Tgt Ion: 65 Resp: 98089
Ion Ratio Lower Upper
65 100
92 69.5 54.2 81.2
138 95.6 65.1 97.7









#47
Dimethylphthalate
Concen: 20.290 ng/ul
RT: 13.657 min Scan# 1814
Delta R.T. -0.006 min
Lab File: BP023005.D
Acq: 10 Nov 2024 18:16




Instrument :
BNA_P
ClientSampleId :
SSTD020709


Tgt Ion:163 Resp: 458459
Ion Ratio Lower Upper
163 100
194 4.4 3.8 5.6
164 9.5 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 21.512 ng/ul
RT: 13.798 min Scan# 1838
Delta R.T. -0.012 min
Lab File: BP023005.D
Acq: 10 Nov 2024 18:16

Tgt Ion:165 Resp: 94548
Ion Ratio Lower Upper
165 100
89 61.6 50.6 75.8
121 20.6 15.8 23.8





#50

Acenaphthylene

Concen: 21.231 ng/ul

RT: 13.916 min Scan# 1858

Delta R.T. -0.017 min

Lab File: BP023005.D

Acq: 10 Nov 2024 18:16

Instrument :

BNA_P

ClientSampleId :

SSTD020709





Tgt Ion:152 Resp: 513037

Ion Ratio Lower Upper

152 100

151 19.7 15.9 23.9

153 12.9 10.2 15.2



#51

3-Nitroaniline

Concen: 21.504 ng/ul

RT: 14.140 min Scan# 1896

Delta R.T. -0.012 min

Lab File: BP023005.D

Acq: 10 Nov 2024 18:16




Tgt Ion:138 Resp: 81407


Ion Ratio Lower Upper

138 100

108 12.8 12.7 19.1

92 130.1 107.8 161.8





#52

Acenaphthene

Concen: 20.404 ng/ul

RT: 14.269 min Scan# 1918

Delta R.T. -0.006 min

Lab File: BP023005.D

Acq: 10 Nov 2024 18:16

Instrument :

BNA_P

ClientSampleId :

SSTD020709





Tgt Ion:153 Resp: 357956

Ion Ratio Lower Upper

153 100

152 48.3 39.6 59.4

154 88.1 70.9 106.3



#53

2,4-Dinitrophenol

Concen: 18.059 ng/ul

RT: 14.345 min Scan# 1931

Delta R.T. -0.017 min

Lab File: BP023005.D

Acq: 10 Nov 2024 18:16




Tgt Ion:184 Resp: 56318


Ion Ratio Lower Upper

184 100

63 65.9 50.1 75.1

154 68.5 58.2 87.4









#55
4-Nitrophenol
Concen: 19.455 ng/ul
RT: 14.475 min Scan# 1953
Delta R.T. -0.006 min
Lab File: BP023005.D
Acq: 10 Nov 2024 18:16




Instrument :
BNA_P
ClientSampleId :
SSTD020709


Tgt Ion:	109	Resp:	78371
Ion Ratio	Lower	Upper	
109	100		
139	92.6	64.6	96.8
65	101.6	80.6	120.8



#56
Dibenzofuran
Concen: 20.185 ng/ul
RT: 14.616 min Scan# 1977
Delta R.T. 0.000 min
Lab File: BP023005.D
Acq: 10 Nov 2024 18:16

Tgt Ion:	168	Resp:	547137
Ion Ratio	Lower	Upper	
168	100		
139	36.8	29.7	44.5









#57
2,4-Dinitrotoluene
Concen: 21.075 ng/ul
RT: 14.598 min Scan# 1976
Delta R.T. -0.006 min
Lab File: BP023005.D
Acq: 10 Nov 2024 18:16




Instrument :
BNA_P
ClientSampleId :
SSTD020709


Tgt Ion:165 Resp: 147630
Ion Ratio Lower Upper
165 100
63 46.7 41.3 61.9
182 3.1 2.5 3.7



#58
2,3,4,6-Tetrachlorophenol
Concen: 20.339 ng/ul
RT: 14.857 min Scan# 2018
Delta R.T. -0.017 min
Lab File: BP023005.D
Acq: 10 Nov 2024 18:16

Tgt Ion:232 Resp: 140064
Ion Ratio Lower Upper
232 100
131 36.3 27.0 40.6
130 2.1 1.9 2.9
166 28.2 21.8 32.6





#59

Diethylphthalate

Concen: 20.330 ng/ul

RT: 15.040 min Scan# 2049

Delta R.T. -0.017 min

Lab File: BP023005.D

Acq: 10 Nov 2024 18:16

Instrument :

BNA_P

ClientSampleId :

SSTD020709





Tgt Ion:149 Resp: 454307

Ion Ratio Lower Upper

149 100

177 23.7 19.7 29.5

150 12.2 9.8 14.6



#61

Fluorene

Concen: 20.353 ng/ul

RT: 15.275 min Scan# 2089

Delta R.T. -0.017 min

Lab File: BP023005.D

Acq: 10 Nov 2024 18:16




Tgt Ion:166 Resp: 451537


Ion Ratio Lower Upper

166 100

165 100.4 79.5 119.3

167 14.1 10.3 15.5









#62
4-Chlorophenyl-phenylether
Concen: 20.000 ng/ul
RT: 15.263 min Scan# 2087
Delta R.T. -0.017 min
Lab File: BP023005.D
Acq: 10 Nov 2024 18:16




Instrument :
BNA_P
ClientSampleId :
SSTD020709


Tgt Ion:204 Resp: 261262
Ion Ratio Lower Upper
204 100
206 32.8 27.2 40.8
141 54.8 44.6 66.8



#63
4-Nitroaniline
Concen: 22.704 ng/ul
RT: 15.310 min Scan# 2095
Delta R.T. -0.012 min
Lab File: BP023005.D
Acq: 10 Nov 2024 18:16

Tgt Ion:138 Resp: 79974
Ion Ratio Lower Upper
138 100
92 59.1 53.9 80.9
108 120.3 106.6 159.8









#66
4,6-Dinitro-2-methylphenol
Concen: 20.174 ng/ul
RT: 15.375 min Scan# 2106
Delta R.T. -0.012 min
Lab File: BP023005.D
Acq: 10 Nov 2024 18:16




Instrument :
BNA_P
ClientSampleId :
SSTD020709


Tgt Ion:198 Resp: 101945
Ion Ratio Lower Upper
198 100
182 4.1 3.8 5.8
77 23.2 20.0 30.0



#67
N-Nitrosodiphenylamine
Concen: 21.349 ng/ul
RT: 15.481 min Scan# 2124
Delta R.T. -0.023 min
Lab File: BP023005.D
Acq: 10 Nov 2024 18:16

Tgt Ion:169 Resp: 389573
Ion Ratio Lower Upper
169 100
168 65.9 53.4 80.0
167 35.6 29.4 44.2









#68
4-Bromophenyl-phenylether
Concen: 20.481 ng/ul
RT: 16.169 min Scan# 21
Delta R.T. -0.023 min
Lab File: BP023005.D
Acq: 10 Nov 2024 18:16




Instrument :
BNA_P
ClientSampleId :
SSTD020709


Tgt Ion:248 Resp: 183148
Ion Ratio Lower Upper
248 100
250 99.7 77.1 115.7
141 64.6 50.7 76.1



#69
Hexachlorobenzene
Concen: 20.449 ng/ul
RT: 16.304 min Scan# 2264
Delta R.T. -0.012 min
Lab File: BP023005.D
Acq: 10 Nov 2024 18:16

Tgt Ion:284 Resp: 230202
Ion Ratio Lower Upper
284 100
142 25.5 21.5 32.3
249 31.3 25.6 38.4









#70
Atrazine
Concen: 20.587 ng/ul
RT: 16.463 min Scan# 21
Delta R.T. -0.017 min
Lab File: BP023005.D
Acq: 10 Nov 2024 18:16




Instrument :
BNA_P
ClientSampleId :
SSTD020709


Tgt Ion:200 Resp: 181588
Ion Ratio Lower Upper
200 100
173 23.6 18.6 27.8
215 47.2 38.3 57.5



#71
Pentachlorophenol
Concen: 19.474 ng/ul
RT: 16.669 min Scan# 2326
Delta R.T. -0.017 min
Lab File: BP023005.D
Acq: 10 Nov 2024 18:16

Tgt Ion:266 Resp: 136105
Ion Ratio Lower Upper
266 100
264 63.1 50.7 76.1
268 63.1 51.9 77.9





#72

Phenanthrene

Concen: 20.732 ng/ul

RT: 17.057 min Scan# 21

Delta R.T. -0.012 min

Lab File: BP023005.D

Acq: 10 Nov 2024 18:16

Instrument :

BNA_P

ClientSampleId :

SSTD020709





Tgt Ion:178 Resp: 772994

Ion Ratio Lower Upper

178 100

179 15.7 12.7 19.1

176 19.1 15.4 23.2



#74

Anthracene

Concen: 20.941 ng/ul

RT: 17.145 min Scan# 2407

Delta R.T. -0.006 min

Lab File: BP023005.D

Acq: 10 Nov 2024 18:16




Tgt Ion:178 Resp: 784992


Ion Ratio Lower Upper

178 100

179 14.9 12.4 18.6

176 18.6 14.5 21.7





#75

1,2,3,4-Tetrachlorobenzene

Concen: 20.971 ng/uL

RT: 12.987 min Scan# 1700

Delta R.T. -0.012 min

Lab File: BP023005.D

Acq: 10 Nov 2024 18:16

Instrument :

BNA_P

ClientSampleId :

SSTD020709





Tgt Ion:216 Resp: 218650

Ion Ratio Lower Upper

216 100

214 78.1 60.2 90.2

218 47.6 38.1 57.1



#76

Pentachlorobenzene

Concen: 20.547 ng/uL

RT: 14.534 min Scan# 1963

Delta R.T. -0.012 min

Lab File: BP023005.D

Acq: 10 Nov 2024 18:16




Tgt Ion:250 Resp: 244608


Ion Ratio Lower Upper

250 100

248 61.5 52.2 78.2

252 63.0 50.1 75.1





#77

Carbazole

Concen: 21.027 ng/ul

RT: 17.439 min Scan# 2456

Delta R.T. -0.006 min

Lab File: BP023005.D

Acq: 10 Nov 2024 18:16

Instrument :

BNA_P

ClientSampleId :

SSTD020709





Tgt Ion:167 Resp: 687680

Ion Ratio Lower Upper

167 100

166 22.2 17.5 26.3

139 12.0 9.6 14.4



#78

Di-n-butylphthalate

Concen: 21.215 ng/ul

RT: 18.022 min Scan# 2556

Delta R.T. -0.012 min

Lab File: BP023005.D

Acq: 10 Nov 2024 18:16




Tgt Ion:149 Resp: 769950


Ion Ratio Lower Upper

149 100

150 8.9 7.2 10.8

104 6.0 5.0 7.4





#80

Fluoranthene

Concen: 21.925 ng/ul

RT: 19.151 min Scan# 21

Delta R.T. -0.012 min

Lab File: BP023005.D

Acq: 10 Nov 2024 18:16

Instrument :

BNA_P

ClientSampleId :

SSTD020709





Tgt Ion:202 Resp: 1002481

Ion Ratio Lower Upper

202 100

101 6.3 5.2 7.8

100 5.0 4.1 6.1



#82

Pyrene

Concen: 21.679 ng/ul

RT: 19.533 min Scan# 2813

Delta R.T. -0.006 min

Lab File: BP023005.D

Acq: 10 Nov 2024 18:16




Tgt Ion:202 Resp: 1064838


Ion Ratio Lower Upper

202 100

101 7.7 6.2 9.2

100 6.3 4.9 7.3





#83

Butylbenzylphthalate

Concen: 22.224 ng/ul

RT: 20.480 min Scan# 2974

Delta R.T. -0.012 min

Lab File: BP023005.D

Acq: 10 Nov 2024 18:16

Instrument :

BNA_P

ClientSampleId :

SSTD020709





Tgt Ion:149 Resp: 359404

Ion Ratio Lower Upper

149 100

91 71.4 55.6 83.4

206 25.2 20.2 30.2



#84

3,3'-Dichlorobenzidine

Concen: 21.651 ng/ul

RT: 21.357 min Scan# 3123

Delta R.T. -0.006 min

Lab File: BP023005.D

Acq: 10 Nov 2024 18:16




Tgt Ion:252 Resp: 418758


Ion Ratio Lower Upper

252 100

254 65.4 52.0 78.0

126 6.5 5.0 7.6





#85

Benzo(a)anthracene

Concen: 21.073 ng/ul

RT: 21.433 min Scan# 3135

Delta R.T. -0.006 min

Lab File: BP023005.D

Acq: 10 Nov 2024 18:16

Instrument :

BNA_P

ClientSampleId :

SSTD020709





Tgt Ion:228 Resp: 1108751

Ion Ratio Lower Upper

228 100

229 19.6 15.5 23.3

226 27.4 21.9 32.9



#86

Bis(2-ethylhexyl)phthalate

Concen: 22.134 ng/ul

RT: 21.357 min Scan# 3123

Delta R.T. -0.012 min

Lab File: BP023005.D

Acq: 10 Nov 2024 18:16




Tgt Ion:149 Resp: 508674


Ion Ratio Lower Upper

149 100

167 28.8 24.5 36.7

279 6.5 5.5 8.3





#87

Chrysene

Concen: 20.955 ng/ul

RT: 21.498 min Scan# 3146

Delta R.T. 0.000 min

Lab File: BP023005.D

Acq: 10 Nov 2024 18:16

Instrument :

BNA_P

ClientSampleId :

SSTD020709





Tgt Ion:228 Resp: 1037414

Ion Ratio Lower Upper

228 100

226 29.7 24.2 36.2

229 19.7 16.1 24.1



#89

Di-n-octyl phthalate

Concen: 20.524 ng/ul




RT: 22.580 min Scan# 3331


Delta R.T. -0.012 min

Lab File: BP023005.D

Acq: 10 Nov 2024 18:16

Tgt Ion:149 Resp: 828649





#90

Benzo(b)fluoranthene

Concen: 21.760 ng/ul

RT: 23.657 min Scan# 3514

Delta R.T. -0.017 min

Lab File: BP023005.D

Acq: 10 Nov 2024 18:16

Instrument :

BNA_P

ClientSampleId :

SSTD020709





Tgt Ion:252 Resp: 1205322

Ion Ratio Lower Upper

252 100

253 22.0 17.4 26.2

125 4.4 3.7 5.5



#91

Benzo(k)fluoranthene

Concen: 22.104 ng/ul

RT: 23.657 min Scan# 3514

Delta R.T. -0.082 min

Lab File: BP023005.D

Acq: 10 Nov 2024 18:16




Tgt Ion:252 Resp: 1205322


Ion Ratio Lower Upper

252 100

253 22.0 17.8 26.8

125 4.4 3.0 4.4









#93
Benzo(a)pyrene
Concen: 21.462 ng/ul
RT: 24.527 min Scan# 3662
Delta R.T. -0.012 min
Lab File: BP023005.D
Acq: 10 Nov 2024 18:16




Instrument :
BNA_P
ClientSampleId :
SSTD020709


Tgt Ion:252 Resp: 1074684
Ion Ratio Lower Upper
252 100
253 22.1 17.6 26.4
125 4.7 4.2 6.2



#94
Indeno(1,2,3-cd)pyrene
Concen: 20.346 ng/ul
RT: 28.315 min Scan# 4306
Delta R.T. -0.029 min
Lab File: BP023005.D
Acq: 10 Nov 2024 18:16

Tgt Ion:276 Resp: 1419424
Ion Ratio Lower Upper
276 100
138 9.4 7.0 10.4
277 24.1 19.8 29.8





#95

Dibenzo(a,h)anthracene

Concen: 20.121 ng/ul

RT: 28.380 min Scan# 41

Delta R.T. -0.023 min

Lab File: BP023005.D

Acq: 10 Nov 2024 18:16

Instrument :

BNA_P

ClientSampleId :

SSTD020709





Tgt Ion:278 Resp: 1139738

Ion Ratio Lower Upper

278 100

139 7.0 5.4 8.0

279 23.2 19.4 29.2



#96

Benzo(g,h,i)perylene

Concen: 16.688 ng/ul

RT: 29.456 min Scan# 4500

Delta R.T. -0.023 min

Lab File: BP023005.D

Acq: 10 Nov 2024 18:16

Tgt Ion:276 Resp: 884276

Ion Ratio Lower Upper

276 100

138 9.1 7.1 10.7

277 24.2 19.4 29.2

