

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111023\
 Data File : BP018032.D
 Acq On : 10 Nov 2023 18:52
 Operator : MA/JU
 Sample : SST08056
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST080621

Quant Time: Nov 10 21:25:04 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP111023.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Nov 10 21:21:36 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.852	152	65984	20.000	ng/ul	0.00
20) Naphthalene-d8	10.675	136	296723	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.528	164	195060	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.310	188	437926	20.000	ng/ul	0.00
79) Chrysene-d12	21.439	240	409003	20.000	ng/ul	0.00
88) Perylene-d12	23.945	264	435104	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.228	96	56388	30.146	ng/uL	0.00
4) Pyridine-d5	3.658	84	409616	83.988	ng/ul	0.00
7) Phenol-d5	7.016	99	551008	86.496	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.205	67	307943	82.258	ng/ul	0.00
11) 2-Chlorophenol-d4	7.375	132	418226	82.905	ng/ul	0.00
15) 4-Methylphenol-d8	8.581	113	451448	86.484	ng/ul	0.00
21) Nitrobenzene-d5	9.063	128	216961	81.158	ng/ul	0.00
24) 2-Nitrophenol-d4	9.769	143	256812	85.416	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.299	165	444434	82.334	ng/ul	0.00
31) 4-Chloroaniline-d4	10.851	131	619730	83.719	ng/ul	0.00
46) Dimethylphthalate-d6	13.951	166	1393824	77.444	ng/ul	0.00
49) Acenaphthylene-d8	14.228	160	1574349	80.313	ng/ul	0.00
54) 4-Nitrophenol-d4	14.787	143	256834	95.499	ng/ul	0.00
60) Fluorene-d10	15.534	176	1163793	78.322	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.686	200	274360	86.662	ng/ul	0.00
73) Anthracene-d10	17.416	188	1862355	78.631	ng/ul	0.00
81) Pyrene-d10	19.663	212	2167437	79.336	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.786	264	2074226	81.985	ng/ul	0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.264	88	60591	29.685	ng/uL	98
5) Pyridine	3.681	79	420144	83.470	ng/ul	95
6) Benzaldehyde	7.016	77	268339	79.523	ng/ul	99
8) Phenol	7.046	94	539003	85.771	ng/ul	99
10) Bis(2-Chloroethyl)ether	7.299	93	399658	81.976	ng/ul	98
12) 2-Chlorophenol	7.405	128	417251	82.380	ng/ul	99
13) 2-Methylphenol	8.305	108	410377	86.428	ng/ul	96
14) 2,2'-oxybis(1-Chloropr...	8.363	45	260680	46.825	ng/ul	100
16) Acetophenone	8.710	105	703732	84.812	ng/ul	98
17) N-Nitroso-di-n-propyla...	8.687	70	375430	82.874	ng/ul	96
18) 4-Methylphenol	8.646	108	445054	85.947	ng/ul	95
19) Hexachloroethane	8.904	117	185427	76.517	ng/ul	97
22) Nitrobenzene	9.104	77	582268	80.665	ng/ul	94
23) Isophorone	9.616	82	1063788	80.763	ng/ul	100
25) 2-Nitrophenol	9.804	139	260470	83.693	ng/ul	96
26) 2,4-Dimethylphenol	9.846	107	521694	78.802	ng/ul	98
27) Bis(2-Chloroethoxy)met...	10.104	93	564793	78.633	ng/ul	99
29) 2,4-Dichlorophenol	10.322	162	432349	83.877	ng/ul	99
30) Naphthalene	10.728	128	1388888	76.968	ng/ul	99
32) 4-Chloroaniline	10.875	127	593498	83.703	ng/ul	97
33) Hexachlorobutadiene	10.940	225	285390	75.419	ng/ul	97
34) Caprolactam	11.728	113	63699	36.731	ng/ul	96
35) 4-Chloro-3-methylphenol	11.987	107	518993	84.641	ng/ul	95
36) 2-Methylnaphthalene	12.340	142	979783	79.088	ng/ul	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111023\
 Data File : BP018032.D
 Acq On : 10 Nov 2023 18:52
 Operator : MA/JU
 Sample : SSTD08056
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD080621

Quant Time: Nov 10 21:25:04 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP111023.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Nov 10 21:21:36 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.563	142	1006199	79.039	ng/ul	97
39) 1,2,4,5-Tetrachloroben...	12.693	216	523628	79.464	ng/ul	97
40) Hexachlorocyclopentadiene	12.628	237	291094	91.519	ng/ul	98
41) 2,4,6-Trichlorophenol	12.957	196	364855	84.404	ng/ul	100
42) 2,4,5-Trichlorophenol	13.034	196	390239	87.103	ng/ul	97
43) 1,1'-Biphenyl	13.357	154	1303003	78.558	ng/ul	99
44) 2-Chloronaphthalene	13.410	162	1049365	79.687	ng/ul	97
45) 2-Nitroaniline	13.663	65	383252	90.188	ng/ul	98
47) Dimethylphthalate	13.998	163	1383033	78.337	ng/ul	99
48) 2,6-Dinitrotoluene	14.157	165	295504	84.210	ng/ul	96
50) Acenaphthylene	14.257	152	1763669	79.689	ng/ul	100
51) 3-Nitroaniline	14.498	138	283006	86.912	ng/ul	96
52) Acenaphthene	14.598	153	1153758	78.617	ng/ul	99
53) 2,4-Dinitrophenol	14.710	184	182333	95.958	ng/ul	97
55) 4-Nitrophenol	14.804	109	319741	89.565	ng/ul	93
56) Dibenzofuran	14.934	168	1581594	78.642	ng/ul	97
57) 2,4-Dinitrotoluene	14.945	165	446254	85.116	ng/ul	96
58) 2,3,4,6-Tetrachlorophenol	15.157	232	336148	83.061	ng/ul	99
59) Diethylphthalate	15.357	149	1427736	77.406	ng/ul	99
61) Fluorene	15.592	166	1350424	79.091	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.581	204	664069	79.366	ng/ul	98
63) 4-Nitroaniline	15.681	138	260310	85.415	ng/ul	94
66) 4,6-Dinitro-2-methylph...	15.704	198	282757	86.072	ng/ul#	96
67) N-Nitrosodiphenylamine	15.810	169	1107772	78.876	ng/ul	100
68) 4-Bromophenyl-phenylether	16.486	248	390109	79.037	ng/ul	96
69) Hexachlorobenzene	16.569	284	442228	80.295	ng/ul	93
70) Atrazine	16.775	200	387134	79.021	ng/ul	98
71) Pentachlorophenol	16.939	266	286646	86.214	ng/ul	97
72) Phenanthrene	17.357	178	2114329	78.633	ng/ul	100
74) Anthracene	17.451	178	2170450	79.099	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.310	216	537171	80.325	ng/uL	100
76) Pentachlorobenzene	14.828	250	519450	79.345	ng/uL	97
77) Carbazole	17.745	167	1934654	81.191	ng/ul	100
78) Di-n-butylphthalate	18.245	149	2372298	78.465	ng/ul	98
80) Fluoranthene	19.333	202	2504426	78.494	ng/ul	99
82) Pyrene	19.698	202	2626581	78.562	ng/ul	98
83) Butylbenzylphthalate	20.557	149	1131557	80.479	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.369	252	815469	83.452	ng/ul	96
85) Benzo(a)anthracene	21.421	228	2614685	79.170	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.298	149	1556246	80.280	ng/ul	99
87) Chrysene	21.480	228	2392020	77.418	ng/ul	98
89) Di-n-octyl phthalate	22.263	149	2661927	81.806	ng/ul	100
90) Benzo(b)fluoranthene	23.168	252	2519830	81.376	ng/ul	98
91) Benzo(k)fluoranthene	23.221	252	2525211	81.055	ng/ul	99
93) Benzo(a)pyrene	23.839	252	2391319	81.776	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.621	276	2882188	82.178	ng/ul	98
95) Dibenzo(a,h)anthracene	26.645	278	2389535	82.744	ng/ul	98
96) Benzo(g,h,i)perylene	27.451	276	2270410	80.935	ng/ul	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111023\
 Data File : BP018032.D
 Acq On : 10 Nov 2023 18:52
 Operator : MA/JU
 Sample : SSTD08056
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD080621

Quant Time: Nov 10 21:25:04 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP111023.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Nov 10 21:21:36 2023
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 29.685 ng/uL
RT: 3.264 min Scan# 30
Delta R.T. -0.006 min
Lab File: BP018032.D
Acq: 10 Nov 2023 18:52




Instrument :
BNA_P
ClientSampleId :
SSTD080621


Tgt Ion:	88	Resp:	60591
Ion	Ratio	Lower	Upper
88	100		
43	32.9	25.4	38.2
58	65.2	53.3	79.9



#5
Pyridine
Concen: 83.470 ng/uL
RT: 3.681 min Scan# 101
Delta R.T. -0.006 min
Lab File: BP018032.D
Acq: 10 Nov 2023 18:52

Tgt Ion:	79	Resp:	420144
Ion	Ratio	Lower	Upper
79	100		
52	64.0	53.9	80.9
51	35.9	32.2	48.2









#6
Benzaldehyde
Concen: 79.523 ng/ul
RT: 7.016 min Scan# 60
Delta R.T. -0.006 min
Lab File: BP018032.D
Acq: 10 Nov 2023 18:52




Instrument :
BNA_P
ClientSampleId :
SSTD080621


Tgt Ion: 77 Resp: 268339
Ion Ratio Lower Upper
77 100
105 89.5 72.7 109.1
106 84.9 67.5 101.3



#8
Phenol
Concen: 85.771 ng/ul
RT: 7.046 min Scan# 673
Delta R.T. 0.000 min
Lab File: BP018032.D
Acq: 10 Nov 2023 18:52

Tgt Ion: 94 Resp: 539003
Ion Ratio Lower Upper
94 100
65 35.1 28.2 42.2
66 49.4 40.3 60.5









#10
Bis(2-Chloroethyl)ether
Concen: 81.976 ng/ul
RT: 7.299 min Scan# 716
Delta R.T. 0.000 min
Lab File: BP018032.D
Acq: 10 Nov 2023 18:52




Instrument :
BNA_P
ClientSampleId :
SSTD080621


Tgt Ion: 93 Resp: 399658
Ion Ratio Lower Upper
93 100
63 74.1 57.8 86.6
95 31.5 25.5 38.3



#12
2-Chlorophenol
Concen: 82.380 ng/ul
RT: 7.405 min Scan# 734
Delta R.T. 0.000 min
Lab File: BP018032.D
Acq: 10 Nov 2023 18:52

Tgt Ion: 128 Resp: 417251
Ion Ratio Lower Upper
128 100
64 49.2 39.2 58.8
130 32.2 26.6 40.0





#13

2-Methylphenol

Concen: 86.428 ng/ul

RT: 8.305 min Scan# 887

Delta R.T. 0.000 min

Lab File: BP018032.D

Acq: 10 Nov 2023 18:52

Instrument :

BNA_P

ClientSampleId :





SSTD080621

Tgt Ion:108 Resp: 410377

Ion Ratio Lower Upper

108 100

107 90.2 75.2 112.8



#14

2,2'-oxybis(1-Chloropropane)

Concen: 46.825 ng/ul

RT: 8.363 min Scan# 897

Delta R.T. 0.000 min

Lab File: BP018032.D

Acq: 10 Nov 2023 18:52




Tgt Ion: 45 Resp: 260680


Ion Ratio Lower Upper

45 100

77 20.0 16.1 24.1

79 13.6 11.1 16.7









#16
Acetophenone
Concen: 84.812 ng/ul
RT: 8.710 min Scan# 910
Delta R.T. 0.000 min
Lab File: BP018032.D
Acq: 10 Nov 2023 18:52




Instrument :
BNA_P
ClientSampleId :
SSTD080621


Tgt Ion:105 Resp: 703732
Ion Ratio Lower Upper
105 100
77 85.4 70.1 105.1
51 28.8 23.4 35.0



#17
N-Nitroso-di-n-propylamine
Concen: 82.874 ng/ul
RT: 8.687 min Scan# 952
Delta R.T. 0.018 min
Lab File: BP018032.D
Acq: 10 Nov 2023 18:52

Tgt Ion: 70 Resp: 375430
Ion Ratio Lower Upper
70 100
42 58.4 43.7 65.5
101 10.1 8.2 12.4
130 20.4 15.2 22.8









#18
4-Methylphenol
Concen: 85.947 ng/ul
RT: 8.646 min Scan# 944
Delta R.T. 0.006 min
Lab File: BP018032.D
Acq: 10 Nov 2023 18:52




Instrument :
BNA_P
ClientSampleId :
SSTD080621


Tgt Ion:108 Resp: 445054
Ion Ratio Lower Upper
108 100
107 115.1 88.2 132.4



#19
Hexachloroethane
Concen: 76.517 ng/ul
RT: 8.904 min Scan# 989
Delta R.T. 0.000 min
Lab File: BP018032.D
Acq: 10 Nov 2023 18:52

Tgt Ion:117 Resp: 185427
Ion Ratio Lower Upper
117 100
201 102.5 84.8 127.2
199 62.8 49.0 73.4









#22
Nitrobenzene
Concen: 80.665 ng/ul
RT: 9.104 min Scan# 1109
Delta R.T. 0.006 min
Lab File: BP018032.D
Acq: 10 Nov 2023 18:52




Instrument :
BNA_P
ClientSampleId :
SSTD080621


Tgt Ion: 77 Resp: 582268
Ion Ratio Lower Upper
77 100
123 39.7 27.8 41.8
65 14.0 11.5 17.3



#23
Isophorone
Concen: 80.763 ng/ul
RT: 9.616 min Scan# 1110
Delta R.T. 0.006 min
Lab File: BP018032.D
Acq: 10 Nov 2023 18:52

Tgt Ion: 82 Resp: 1063788
Ion Ratio Lower Upper
82 100
95 8.4 6.6 9.8
138 16.3 13.1 19.7









#25
2-Nitrophenol
Concen: 83.693 ng/ul
RT: 9.804 min Scan# 1142
Delta R.T. 0.000 min
Lab File: BP018032.D
Acq: 10 Nov 2023 18:52




Instrument :
BNA_P
ClientSampleId :
SSTD080621


Tgt Ion:139 Resp: 260470
Ion Ratio Lower Upper
139 100
65 61.4 51.4 77.0
109 43.7 33.1 49.7



#26
2,4-Dimethylphenol
Concen: 78.802 ng/ul
RT: 9.846 min Scan# 1149
Delta R.T. 0.000 min
Lab File: BP018032.D
Acq: 10 Nov 2023 18:52

Tgt Ion:107 Resp: 521694
Ion Ratio Lower Upper
107 100
121 45.3 37.8 56.8
122 75.8 62.1 93.1









#27
Bis(2-Chloroethoxy)methane
Concen: 78.633 ng/ul
RT: 10.104 min Scan# 1193
Delta R.T. 0.006 min
Lab File: BP018032.D
Acq: 10 Nov 2023 18:52




Instrument :
BNA_P
ClientSampleId :
SSTD080621


Tgt Ion: 93 Resp: 564793
Ion Ratio Lower Upper
93 100
95 31.9 26.2 39.2
123 10.7 8.6 12.8



#29
2,4-Dichlorophenol
Concen: 83.877 ng/ul
RT: 10.322 min Scan# 1230
Delta R.T. 0.000 min
Lab File: BP018032.D
Acq: 10 Nov 2023 18:52

Tgt Ion:162 Resp: 432349
Ion Ratio Lower Upper
162 100
164 64.4 51.8 77.6
98 37.8 28.7 43.1









#30
Naphthalene
Concen: 76.968 ng/ul
RT: 10.728 min Scan# 1298
Delta R.T. 0.006 min
Lab File: BP018032.D
Acq: 10 Nov 2023 18:52




Instrument :
BNA_P
ClientSampleId :
SSTD080621


Tgt Ion:128 Resp: 1388888
Ion Ratio Lower Upper
128 100
129 10.7 9.0 13.4
127 13.4 10.3 15.5



#32
4-Chloroaniline
Concen: 83.703 ng/ul
RT: 10.875 min Scan# 1324
Delta R.T. 0.000 min
Lab File: BP018032.D
Acq: 10 Nov 2023 18:52

Tgt Ion:127 Resp: 593498
Ion Ratio Lower Upper
127 100
129 30.9 26.1 39.1









#33
Hexachlorobutadiene
Concen: 75.419 ng/ul
RT: 10.940 min Scan# 1335
Delta R.T. 0.000 min
Lab File: BP018032.D
Acq: 10 Nov 2023 18:52




Instrument :
BNA_P
ClientSampleId :
SSTD080621


Tgt Ion:225 Resp: 285390
Ion Ratio Lower Upper
225 100
223 64.5 52.5 78.7
227 61.8 51.9 77.9



#34
Caprolactam
Concen: 36.731 ng/ul
RT: 11.728 min Scan# 1469
Delta R.T. 0.006 min
Lab File: BP018032.D
Acq: 10 Nov 2023 18:52

Tgt Ion:113 Resp: 63699
Ion Ratio Lower Upper
113 100
55 150.0 118.3 177.5
56 126.3 95.6 143.4









#35
4-Chloro-3-methylphenol
Concen: 84.641 ng/ul
RT: 11.987 min Scan# 1513
Delta R.T. 0.000 min
Lab File: BP018032.D
Acq: 10 Nov 2023 18:52




Instrument :
BNA_P
ClientSampleId :
SSTD080621


Tgt Ion:107 Resp: 518993
Ion Ratio Lower Upper
107 100
144 23.8 19.1 28.7
142 73.8 54.8 82.2



#36
2-Methylnaphthalene
Concen: 79.088 ng/ul
RT: 12.340 min Scan# 1573
Delta R.T. 0.000 min
Lab File: BP018032.D
Acq: 10 Nov 2023 18:52

Tgt Ion:142 Resp: 979783
Ion Ratio Lower Upper
142 100
141 89.5 71.7 107.5





#37

1-Methylnaphthalene

Concen: 79.039 ng/ul

RT: 12.563 min Scan# 1611

Delta R.T. 0.006 min

Lab File: BP018032.D

Acq: 10 Nov 2023 18:52

Instrument :

BNA_P

ClientSampleId :

SSTD080621





Tgt Ion:142 Resp: 1006199

Ion Ratio Lower Upper

142 100

141 93.8 72.8 109.2

116 4.6 4.0 6.0



#39

1,2,4,5-Tetrachlorobenzene

Concen: 79.464 ng/ul

RT: 12.693 min Scan# 1633

Delta R.T. 0.000 min

Lab File: BP018032.D

Acq: 10 Nov 2023 18:52

Tgt Ion:216 Resp: 523628




Ion Ratio Lower Upper


216 100

214 77.0 64.8 97.2

179 23.3 18.3 27.5

108 19.6 15.7 23.5









#40
Hexachlorocyclopentadiene
Concen: 91.519 ng/ul
RT: 12.628 min Scan# 1622
Delta R.T. 0.000 min
Lab File: BP018032.D
Acq: 10 Nov 2023 18:52




Instrument :
BNA_P
ClientSampleId :
SSTD080621


Tgt Ion:237 Resp: 291094
Ion Ratio Lower Upper
237 100
235 64.1 49.9 74.9
272 11.7 8.9 13.3



#41
2,4,6-Trichlorophenol
Concen: 84.404 ng/ul
RT: 12.957 min Scan# 1678
Delta R.T. 0.000 min
Lab File: BP018032.D
Acq: 10 Nov 2023 18:52

Tgt Ion:196 Resp: 364855
Ion Ratio Lower Upper
196 100
198 98.2 78.9 118.3
200 31.9 25.8 38.8





#42

2,4,5-Trichlorophenol

Concen: 87.103 ng/ul

RT: 13.034 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP018032.D

Acq: 10 Nov 2023 18:52

Instrument :

BNA_P

ClientSampleId :

SSTD080621





Tgt Ion:196 Resp: 390239

Ion Ratio Lower Upper

196 100

198 95.7 78.9 118.3

200 31.3 26.6 40.0



#43

1,1'-Biphenyl

Concen: 78.558 ng/ul

RT: 13.357 min Scan# 1746

Delta R.T. 0.000 min

Lab File: BP018032.D

Acq: 10 Nov 2023 18:52




Tgt Ion:154 Resp: 1303003


Ion Ratio Lower Upper

154 100

153 41.1 32.4 48.6

76 12.4 9.8 14.8









#44
2-Chloronaphthalene
Concen: 79.687 ng/ul
RT: 13.410 min Scan# 1755
Delta R.T. 0.006 min
Lab File: BP018032.D
Acq: 10 Nov 2023 18:52




Instrument :
BNA_P
ClientSampleId :
SSTD080621


Tgt Ion:162 Resp: 1049365
Ion Ratio Lower Upper
162 100
164 32.1 26.9 40.3
127 41.8 35.4 53.2



#45
2-Nitroaniline
Concen: 90.188 ng/ul
RT: 13.663 min Scan# 1798
Delta R.T. 0.006 min
Lab File: BP018032.D
Acq: 10 Nov 2023 18:52

Tgt Ion: 65 Resp: 383252
Ion Ratio Lower Upper
65 100
92 67.2 53.8 80.6
138 96.9 75.1 112.7









#47
Dimethylphthalate
Concen: 78.337 ng/ul
RT: 13.998 min Scan# 1855
Delta R.T. 0.000 min
Lab File: BP018032.D
Acq: 10 Nov 2023 18:52




Instrument :
BNA_P
ClientSampleId :
SSTD080621


Tgt Ion:163 Resp: 1383033
Ion Ratio Lower Upper
163 100
194 3.8 3.5 5.3
164 10.0 8.0 12.0



#48
2,6-Dinitrotoluene
Concen: 84.210 ng/ul
RT: 14.157 min Scan# 1882
Delta R.T. 0.006 min
Lab File: BP018032.D
Acq: 10 Nov 2023 18:52

Tgt Ion:165 Resp: 295504
Ion Ratio Lower Upper
165 100
89 70.1 59.4 89.0
121 23.3 18.2 27.4









#50
Acenaphthylene
Concen: 79.689 ng/ul
RT: 14.257 min Scan# 1899
Delta R.T. 0.000 min
Lab File: BP018032.D
Acq: 10 Nov 2023 18:52




Instrument :
BNA_P
ClientSampleId :
SSTD080621


Tgt Ion:152 Resp: 1763669
Ion Ratio Lower Upper
152 100
151 20.1 16.2 24.2
153 13.3 10.4 15.6



#51
3-Nitroaniline
Concen: 86.912 ng/ul
RT: 14.498 min Scan# 1940
Delta R.T. 0.000 min
Lab File: BP018032.D
Acq: 10 Nov 2023 18:52

Tgt Ion:138 Resp: 283006
Ion Ratio Lower Upper
138 100
108 12.4 8.3 12.5
92 128.0 99.0 148.4









#52
Acenaphthene
Concen: 78.617 ng/ul
RT: 14.598 min Scan# 1956
Delta R.T. 0.006 min
Lab File: BP018032.D
Acq: 10 Nov 2023 18:52




Instrument :
BNA_P
ClientSampleId :
SSTD080621


Tgt Ion:153 Resp: 1153758
Ion Ratio Lower Upper
153 100
152 48.3 39.0 58.6
154 87.6 71.0 106.4



#53
2,4-Dinitrophenol
Concen: 95.958 ng/ul
RT: 14.710 min Scan# 1976
Delta R.T. 0.000 min
Lab File: BP018032.D
Acq: 10 Nov 2023 18:52

Tgt Ion:184 Resp: 182333
Ion Ratio Lower Upper
184 100
63 83.2 67.7 101.5
154 79.7 60.2 90.4









#55
4-Nitrophenol
Concen: 89.565 ng/ul
RT: 14.804 min Scan# 1992
Delta R.T. 0.006 min
Lab File: BP018032.D
Acq: 10 Nov 2023 18:52




Instrument :
BNA_P
ClientSampleId :
SSTD080621


Tgt Ion:109 Resp: 319741
Ion Ratio Lower Upper
109 100
139 89.8 74.9 112.3
65 106.1 93.8 140.6



#56
Dibenzofuran
Concen: 78.642 ng/ul
RT: 14.934 min Scan# 2014
Delta R.T. 0.000 min
Lab File: BP018032.D
Acq: 10 Nov 2023 18:52

Tgt Ion:168 Resp: 1581594
Ion Ratio Lower Upper
168 100
139 44.1 34.0 51.0









#57
2,4-Dinitrotoluene
Concen: 85.116 ng/ul
RT: 14.945 min Scan# 2016
Delta R.T. 0.000 min
Lab File: BP018032.D
Acq: 10 Nov 2023 18:52




Instrument :
BNA_P
ClientSampleId :
SSTD080621


Tgt Ion:165 Resp: 446254
Ion Ratio Lower Upper
165 100
63 63.7 48.6 72.8
182 2.7 2.1 3.1



#58
2,3,4,6-Tetrachlorophenol
Concen: 83.061 ng/ul
RT: 15.157 min Scan# 2052
Delta R.T. 0.000 min
Lab File: BP018032.D
Acq: 10 Nov 2023 18:52

Tgt Ion:232 Resp: 336148
Ion Ratio Lower Upper
232 100
131 54.6 44.1 66.1
130 3.2 2.5 3.7
166 35.7 27.9 41.9









#59
Diethylphthalate
Concen: 77.406 ng/ul
RT: 15.357 min Scan# 2085
Delta R.T. 0.006 min
Lab File: BP018032.D
Acq: 10 Nov 2023 18:52




Instrument :
BNA_P
ClientSampleId :
SSTD080621


Tgt Ion:149 Resp: 1427736
Ion Ratio Lower Upper
149 100
177 20.7 16.3 24.5
150 12.8 10.0 15.0



#61
Fluorene
Concen: 79.091 ng/ul
RT: 15.592 min Scan# 2126
Delta R.T. 0.006 min
Lab File: BP018032.D
Acq: 10 Nov 2023 18:52

Tgt Ion:166 Resp: 1350424
Ion Ratio Lower Upper
166 100
165 98.3 79.0 118.6
167 13.5 10.6 16.0









#62
4-Chlorophenyl-phenylether
Concen: 79.366 ng/ul
RT: 15.581 min Scan# 2124
Delta R.T. 0.000 min
Lab File: BP018032.D
Acq: 10 Nov 2023 18:52




Instrument :
BNA_P
ClientSampleId :
SSTD080621


Tgt Ion:204 Resp: 664069
Ion Ratio Lower Upper
204 100
206 32.4 26.2 39.2
141 67.8 56.0 84.0



#63
4-Nitroaniline
Concen: 85.415 ng/ul
RT: 15.681 min Scan# 2141
Delta R.T. 0.006 min
Lab File: BP018032.D
Acq: 10 Nov 2023 18:52

Tgt Ion:138 Resp: 260310
Ion Ratio Lower Upper
138 100
92 59.7 46.6 70.0
108 140.7 104.7 157.1









#66
4,6-Dinitro-2-methylphenol
Concen: 86.072 ng/ul
RT: 15.704 min Scan# 2145
Delta R.T. 0.006 min
Lab File: BP018032.D
Acq: 10 Nov 2023 18:52




Instrument :
BNA_P
ClientSampleId :
SSTD080621


Tgt Ion:198 Resp: 282757
Ion Ratio Lower Upper
198 100
182 3.2 3.8 5.6#
77 31.7 27.2 40.8



#67
N-Nitrosodiphenylamine
Concen: 78.876 ng/ul
RT: 15.810 min Scan# 2163
Delta R.T. 0.000 min
Lab File: BP018032.D
Acq: 10 Nov 2023 18:52

Tgt Ion:169 Resp: 1107772
Ion Ratio Lower Upper
169 100
168 67.1 53.8 80.8
167 36.8 29.8 44.6









#68
4-Bromophenyl-phenylether
Concen: 79.037 ng/ul
RT: 16.486 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP018032.D
Acq: 10 Nov 2023 18:52




Instrument :
BNA_P
ClientSampleId :
SSTD080621


Tgt Ion:248 Resp: 390109
Ion Ratio Lower Upper
248 100
250 97.8 72.7 109.1
141 89.4 70.4 105.6



#69
Hexachlorobenzene
Concen: 80.295 ng/ul
RT: 16.569 min Scan# 2292
Delta R.T. 0.006 min
Lab File: BP018032.D
Acq: 10 Nov 2023 18:52

Tgt Ion:284 Resp: 442228
Ion Ratio Lower Upper
284 100
142 41.5 38.9 58.3
249 34.3 28.6 43.0





#70

Atrazine

Concen: 79.021 ng/ul

RT: 16.775 min Scan# 21

Delta R.T. 0.006 min

Lab File: BP018032.D

Acq: 10 Nov 2023 18:52

Instrument :

BNA_P

ClientSampleId :

SSTD080621





Tgt Ion:200 Resp: 387134

Ion Ratio Lower Upper

200 100

173 25.3 20.2 30.2

215 45.4 37.8 56.8



#71

Pentachlorophenol

Concen: 86.214 ng/ul

RT: 16.939 min Scan# 2355

Delta R.T. 0.000 min

Lab File: BP018032.D

Acq: 10 Nov 2023 18:52




Tgt Ion:266 Resp: 286646


Ion Ratio Lower Upper

266 100

264 62.1 50.7 76.1

268 62.6 52.4 78.6









#72
Phenanthrene
Concen: 78.633 ng/ul
RT: 17.357 min Scan# 2425
Delta R.T. 0.006 min
Lab File: BP018032.D
Acq: 10 Nov 2023 18:52




Instrument :
BNA_P
ClientSampleId :
SSTD080621


Tgt Ion:178 Resp: 2114329
Ion Ratio Lower Upper
178 100
179 15.1 12.0 18.0
176 19.5 15.5 23.3



#74
Anthracene
Concen: 79.099 ng/ul
RT: 17.451 min Scan# 2442
Delta R.T. 0.006 min
Lab File: BP018032.D
Acq: 10 Nov 2023 18:52

Tgt Ion:178 Resp: 2170450
Ion Ratio Lower Upper
178 100
179 15.3 12.6 19.0
176 18.6 15.4 23.0









#75
1,2,3,4-Tetrachlorobenzene
Concen: 80.325 ng/uL
RT: 13.310 min Scan# 1738
Delta R.T. 0.000 min
Lab File: BP018032.D
Acq: 10 Nov 2023 18:52




Instrument :
BNA_P
ClientSampleId :
SSTD080621


Tgt Ion:	216	Resp:	537171
Ion Ratio	Lower	Upper	
216	100		
214	79.4	64.0	96.0
218	48.8	39.0	58.4



#76
Pentachlorobenzene
Concen: 79.345 ng/uL
RT: 14.828 min Scan# 1996
Delta R.T. 0.006 min
Lab File: BP018032.D
Acq: 10 Nov 2023 18:52

Tgt Ion:	250	Resp:	519450
Ion Ratio	Lower	Upper	
250	100		
248	62.9	52.7	79.1
252	63.9	52.3	78.5





#77

Carbazole

Concen: 81.191 ng/ul

RT: 17.745 min Scan# 2491

Delta R.T. 0.006 min

Lab File: BP018032.D

Acq: 10 Nov 2023 18:52

Instrument :

BNA_P

ClientSampleId :

SSTD080621





Tgt Ion:167 Resp: 1934654

Ion Ratio Lower Upper

167 100

166 21.9 17.6 26.4

139 14.6 11.4 17.0



#78

Di-n-butylphthalate

Concen: 78.465 ng/ul

RT: 18.245 min Scan# 2577

Delta R.T. 0.000 min

Lab File: BP018032.D

Acq: 10 Nov 2023 18:52




Tgt Ion:149 Resp: 2372298


Ion Ratio Lower Upper

149 100

150 9.5 6.7 10.1

104 6.7 5.1 7.7





#80

Fluoranthene

Concen: 78.494 ng/ul

RT: 19.333 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP018032.D

Acq: 10 Nov 2023 18:52

Instrument :

BNA_P

ClientSampleId :

SSTD080621





Tgt Ion:202 Resp: 2504426

Ion Ratio Lower Upper

202 100

101 9.5 7.2 10.8

100 7.2 5.8 8.6



#82

Pyrene

Concen: 78.562 ng/ul

RT: 19.698 min Scan# 2824

Delta R.T. 0.006 min

Lab File: BP018032.D

Acq: 10 Nov 2023 18:52




Tgt Ion:202 Resp: 2626581


Ion Ratio Lower Upper

202 100

101 9.9 8.3 12.5

100 8.4 7.4 11.2





#83

Butylbenzylphthalate

Concen: 80.479 ng/ul

RT: 20.557 min Scan# 2970

Delta R.T. 0.000 min

Lab File: BP018032.D

Acq: 10 Nov 2023 18:52

Instrument :

BNA_P

ClientSampleId :

SSTD080621





Tgt Ion:149 Resp: 1131557

Ion Ratio Lower Upper

149 100

91 77.2 59.6 89.4

206 19.7 14.3 21.5



#84

3,3'-Dichlorobenzidine

Concen: 83.452 ng/ul

RT: 21.369 min Scan# 3108

Delta R.T. 0.006 min

Lab File: BP018032.D

Acq: 10 Nov 2023 18:52




Tgt Ion:252 Resp: 815469


Ion Ratio Lower Upper

252 100

254 64.5 48.9 73.3

126 10.3 9.4 14.0





#85

Benzo(a)anthracene

Concen: 79.170 ng/ul

RT: 21.421 min Scan# 3116

Delta R.T. 0.006 min

Lab File: BP018032.D

Acq: 10 Nov 2023 18:52

Instrument :

BNA_P

ClientSampleId :

SSTD080621





Tgt Ion:228 Resp: 2614685

Ion Ratio Lower Upper

228 100

229 19.5 15.4 23.0

226 28.3 21.4 32.0



#86

Bis(2-ethylhexyl)phthalate

Concen: 80.280 ng/ul

RT: 21.298 min Scan# 3096

Delta R.T. 0.000 min

Lab File: BP018032.D

Acq: 10 Nov 2023 18:52




Tgt Ion:149 Resp: 1556246


Ion Ratio Lower Upper

149 100

167 26.8 21.0 31.4

279 4.2 3.2 4.8





#87

Chrysene

Concen: 77.418 ng/ul

RT: 21.480 min Scan# 3126

Delta R.T. 0.006 min

Lab File: BP018032.D

Acq: 10 Nov 2023 18:52

Instrument :

BNA_P

ClientSampleId :

SSTD080621





Tgt Ion:228 Resp: 2392020

Ion Ratio Lower Upper

228 100

226 30.4 23.1 34.7

229 19.4 15.4 23.0



#89

Di-n-octyl phthalate

Concen: 81.806 ng/ul




RT: 22.263 min Scan# 3260


Delta R.T. 0.006 min

Lab File: BP018032.D

Acq: 10 Nov 2023 18:52

Tgt Ion:149 Resp: 2661927





#90

Benzo(b)fluoranthene

Concen: 81.376 ng/ul

RT: 23.168 min Scan# 3413

Delta R.T. 0.006 min

Lab File: BP018032.D

Acq: 10 Nov 2023 18:52

Instrument :

BNA_P

ClientSampleId :

SSTD080621





Tgt Ion:252 Resp: 2519830

Ion Ratio Lower Upper

252 100

253 22.0 17.0 25.4

125 8.1 7.0 10.6



#91

Benzo(k)fluoranthene

Concen: 81.055 ng/ul

RT: 23.221 min Scan# 3423

Delta R.T. 0.012 min

Lab File: BP018032.D

Acq: 10 Nov 2023 18:52




Tgt Ion:252 Resp: 2525211


Ion Ratio Lower Upper

252 100

253 21.8 17.6 26.4

125 7.9 6.8 10.2





#93

Benzo(a)pyrene

Concen: 81.776 ng/ul

RT: 23.839 min Scan# 3

Delta R.T. 0.012 min

Lab File: BP018032.D

Acq: 10 Nov 2023 18:52

Instrument :

BNA_P

ClientSampleId :

SSTD080621





Tgt Ion:252 Resp: 2391319

Ion Ratio Lower Upper

252 100

253 21.9 17.4 26.0

125 9.2 7.3 10.9



#94

Indeno(1,2,3-cd)pyrene

Concen: 82.178 ng/ul

RT: 26.621 min Scan# 4001

Delta R.T. 0.024 min

Lab File: BP018032.D

Acq: 10 Nov 2023 18:52




Tgt Ion:276 Resp: 2882188


Ion Ratio Lower Upper

276 100

138 20.1 15.6 23.4

277 25.2 19.3 28.9





#95

Dibenzo(a,h)anthracene

Concen: 82.744 ng/ul

RT: 26.645 min Scan# 4001

Delta R.T. 0.024 min

Lab File: BP018032.D

Acq: 10 Nov 2023 18:52

Instrument :

BNA_P

ClientSampleId :

SSTD080621





Tgt Ion:278 Resp: 2389535

Ion Ratio Lower Upper

278 100

139 13.2 11.1 16.7

279 24.4 19.0 28.4



#96

Benzo(g,h,i)perylene

Concen: 80.935 ng/ul

RT: 27.451 min Scan# 4142

Delta R.T. 0.018 min

Lab File: BP018032.D

Acq: 10 Nov 2023 18:52

Tgt Ion:276 Resp: 2270410

Ion Ratio Lower Upper

276 100

138 18.1 13.8 20.6

277 23.1 20.2 30.2

