

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111222\
 Data File : BP012626.D
 Acq On : 12 Nov 2022 19:04
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 14 03:21:00 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Nov 13 22:44:57 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	76	0.00
2	1,4-Dioxane	40.000	40.175	-0.4	77	0.00
3	Pyridine	40.000	42.000	-5.0	79	0.00
4	n-Nitrosodimethylamine	40.000	43.046	-7.6	80	0.00
5 S	2-Fluorophenol	80.000	83.035	-3.8	79	0.00
6	Aniline	40.000	41.354	-3.4	79	0.00
7 S	Phenol-d6	80.000	83.186	-4.0	79	0.00
8	2-Chlorophenol	40.000	40.819	-2.0	78	0.00
9	Benzaldehyde	40.000	40.823	-2.1	83	0.00
10 C	Phenol	40.000	41.336	-3.3	79	0.00
11	bis(2-Chloroethyl)ether	40.000	40.320	-0.8	78	0.00
12	1,3-Dichlorobenzene	40.000	40.229	-0.6	78	0.00
13 C	1,4-Dichlorobenzene	40.000	40.253	-0.6	79	0.00
14	1,2-Dichlorobenzene	40.000	40.486	-1.2	80	0.00
15	Benzyl Alcohol	40.000	43.717	-9.3	81	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	40.903	-2.3	80	0.00
17	2-Methylphenol	40.000	41.420	-3.6	78	0.00
18	Hexachloroethane	40.000	41.346	-3.4	79	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	41.041	-2.6	80	0.00
20	3+4-Methylphenols	40.000	41.891	-4.7	79	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	78	0.00
22	Acetophenone	40.000	40.318	-0.8	80	0.00
23 S	Nitrobenzene-d5	80.000	83.264	-4.1	80	0.00
24	Nitrobenzene	40.000	41.191	-3.0	80	0.00
25	Isophorone	40.000	40.993	-2.5	78	0.00
26 C	2-Nitrophenol	40.000	42.259	-5.6	77	0.00
27	2,4-Dimethylphenol	40.000	40.661	-1.7	78	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.837	0.4	77	0.00
29 C	2,4-Dichlorophenol	40.000	41.219	-3.0	78	0.00
30	1,2,4-Trichlorobenzene	40.000	40.209	-0.5	79	0.00
31	Naphthalene	40.000	39.999	0.0	80	0.00
32	Benzoic acid	40.000	39.563	1.1	77	0.00
33	4-Chloroaniline	40.000	40.782	-2.0	80	0.00
34 C	Hexachlorobutadiene	40.000	40.825	-2.1	80	0.00
35	Caprolactam	40.000	43.021	-7.6	81	0.00
36 C	4-Chloro-3-methylphenol	40.000	41.555	-3.9	80	0.00
37	2-Methylnaphthalene	40.000	39.997	0.0	80	0.00
38	1-Methylnaphthalene	40.000	39.773	0.6	80	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	80	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.139	-0.3	80	0.00
41 P	Hexachlorocyclopentadiene	40.000	38.392	4.0	74	0.00
42 S	2,4,6-Tribromophenol	80.000	87.865	-9.8	86	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.453	-3.6	80	0.00
44	2,4,5-Trichlorophenol	40.000	41.272	-3.2	81	0.00
45 S	2-Fluorobiphenyl	80.000	77.920	2.6	83	0.00
46	1,1'-Biphenyl	40.000	39.594	1.0	81	0.00
47	2-Chloronaphthalene	40.000	39.596	1.0	80	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111222\
 Data File : BP012626.D
 Acq On : 12 Nov 2022 19:04
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 14 03:21:00 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Nov 13 22:44:57 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	43.311	-8.3	82	0.00
49	Acenaphthylene	40.000	40.334	-0.8	82	0.00
50	Dimethylphthalate	40.000	39.996	0.0	82	0.00
51	2,6-Dinitrotoluene	40.000	41.540	-3.8	81	0.00
52 C	Acenaphthene	40.000	40.102	-0.3	82	0.00
53	3-Nitroaniline	40.000	42.470	-6.2	82	0.00
54 P	2,4-Dinitrophenol	40.000	39.064	2.3	80	0.00
55	Dibenzofuran	40.000	39.736	0.7	84	0.00
56 P	4-Nitrophenol	40.000	41.810	-4.5	84	0.00
57	2,4-Dinitrotoluene	40.000	43.049	-7.6	86	0.00
58	Fluorene	40.000	40.002	-0.0	86	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	41.160	-2.9	83	0.00
60	Diethylphthalate	40.000	40.694	-1.7	83	0.00
61	4-Chlorophenyl-phenylether	40.000	39.654	0.9	85	0.00
62	4-Nitroaniline	40.000	41.125	-2.8	86	0.00
63	Azobenzene	40.000	41.026	-2.6	83	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	85	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.632	-1.6	83	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.655	0.9	84	0.00
67	4-Bromophenyl-phenylether	40.000	40.343	-0.9	84	0.00
68	Hexachlorobenzene	40.000	40.387	-1.0	86	0.00
69	Atrazine	40.000	41.564	-3.9	88	0.00
70 C	Pentachlorophenol	40.000	43.018	-7.5	86	0.00
71	Phenanthrene	40.000	39.498	1.3	87	0.00
72	Anthracene	40.000	40.299	-0.7	88	0.00
73	Carbazole	40.000	40.360	-0.9	89	0.00
74	Di-n-butylphthalate	40.000	41.669	-4.2	88	0.00
75 C	Fluoranthene	40.000	40.708	-1.8	93	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	111	0.00
77	Benzidine	40.000	47.999	-20.0	111	0.00
78	Pyrene	40.000	38.997	2.5	95	0.00
79 S	Terphenyl-d14	80.000	74.290	7.1	98	0.00
80	Butylbenzylphthalate	40.000	42.110	-5.3	104	0.00
81	Benzo(a)anthracene	40.000	40.010	-0.0	110	0.00
82	3,3'-Dichlorobenzidine	40.000	43.028	-7.6	119	0.00
83	Chrysene	40.000	40.013	-0.0	111	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	43.128	-7.8	110	0.00
85 c	Di-n-octyl phthalate	40.000	44.270	-10.7	124	0.00
86 I	Perylene-d12	20.000	20.000	0.0	133	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.705	-1.8	125	0.00
88	Benzo(b)fluoranthene	40.000	39.407	1.5	126	0.00
89	Benzo(k)fluoranthene	40.000	40.768	-1.9	136	0.00
90 C	Benzo(a)pyrene	40.000	40.881	-2.2	133	0.00
91	Dibenzo(a,h)anthracene	40.000	40.574	-1.4	126	0.00
92	Benzo(g,h,i)perylene	40.000	39.859	0.4	121	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111222\
Data File : BP012626.D
Acq On : 12 Nov 2022 19:04
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Nov 14 03:21:00 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111222.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Nov 13 22:44:57 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0