

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111423\
 Data File : BP018162.D
 Acq On : 14 Nov 2023 18:58
 Operator : MA/JU
 Sample : 05337-08
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :

Quant Time: Nov 15 00:39:58 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP111023.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Nov 13 15:50:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.787	152	80536	20.000	ng/ul	0.00
20) Naphthalene-d8	10.604	136	326327	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.469	164	202337	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.251	188	453056	20.000	ng/ul	0.00
79) Chrysene-d12	21.380	240	453460	20.000	ng/ul	-0.01
88) Perylene-d12	23.845	264	499720	20.000	ng/ul	-0.02

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.164	96	6899	3.022	ng/uL	0.00
4) Pyridine-d5	3.605	84	155	0.026	ng/ul	0.00
7) Phenol-d5	6.969	99	31219	4.015	ng/ul	0.01
9) Bis-(2-Chloroethyl)eth...	7.140	67	115116	25.194	ng/ul	0.00
11) 2-Chlorophenol-d4	7.310	132	102812	16.698	ng/ul	0.00
15) 4-Methylphenol-d8	8.516	113	56025	8.793	ng/ul	0.00
21) Nitrobenzene-d5	8.993	128	67988	23.125	ng/ul	0.00
24) 2-Nitrophenol-d4	9.710	143	71525	21.474	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.234	165	113618	19.139	ng/ul	0.00
31) 4-Chloroaniline-d4	10.798	131	74251	9.121	ng/ul	0.01
46) Dimethylphthalate-d6	13.892	166	498864	26.721	ng/ul	0.00
49) Acenaphthylene-d8	14.163	160	483930	23.799	ng/ul	0.00
54) 4-Nitrophenol-d4	14.734	143	12	0.004	ng/ul	0.00
60) Fluorene-d10	15.469	176	406566	26.377	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.634	200	58336	17.811	ng/ul	-0.03
73) Anthracene-d10	17.351	188	679922	27.748	ng/ul	0.00
81) Pyrene-d10	19.604	212	873527	28.839	ng/ul	-0.01
92) Benzo(a)pyrene-d12	23.680	264	832012	28.634	ng/ul	-0.02

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111423\
Data File : BP018162.D
Acq On : 14 Nov 2023 18:58
Operator : MA/JU
Sample : 05337-08
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :

Quant Time: Nov 15 00:39:58 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP111023.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Nov 13 15:50:18 2023
Response via : Initial Calibration

