

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111425\
 Data File : BP026137.D
 Acq On : 14 Nov 2025 11:41
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 14 12:01:01 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP102925.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 30 11:51:34 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	130	-0.01
2	1,4-Dioxane	40.000	40.431	-1.1	130	0.00
3	Pyridine	40.000	40.826	-2.1	130	0.00
4	n-Nitrosodimethylamine	40.000	38.485	3.8	124	0.00
5 S	2-Fluorophenol	80.000	84.674	-5.8	133	0.00
6	Aniline	40.000	40.836	-2.1	129	0.00
7 S	Phenol-d6	80.000	85.356	-6.7	133	0.00
8	2-Chlorophenol	40.000	43.946	-9.9	136	0.00
9	Benzaldehyde	40.000	38.393	4.0	123	0.00
10 C	Phenol	40.000	41.691	-4.2	131	0.00
11	bis(2-Chloroethyl)ether	40.000	41.047	-2.6	131	-0.01
12	1,3-Dichlorobenzene	40.000	40.325	-0.8	128	-0.01
13 C	1,4-Dichlorobenzene	40.000	40.540	-1.3	130	-0.01
14	1,2-Dichlorobenzene	40.000	40.419	-1.0	129	0.00
15	Benzyl Alcohol	40.000	43.195	-8.0	137	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	38.167	4.6	121	-0.01
17	2-Methylphenol	40.000	42.530	-6.3	132	0.00
18	Hexachloroethane	40.000	43.216	-8.0	136	-0.02
19 P	n-Nitroso-di-n-propylamine	40.000	42.158	-5.4	128	-0.02
20	3+4-Methylphenols	40.000	42.460	-6.2	133	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	129	-0.01
22	Acetophenone	40.000	40.133	-0.3	127	-0.01
23 S	Nitrobenzene-d5	80.000	90.501	-13.1	139	0.00
24	Nitrobenzene	40.000	43.575	-8.9	134	-0.01
25	Isophorone	40.000	41.601	-4.0	128	0.00
26 C	2-Nitrophenol	40.000	62.670	-56.7#	200	0.00
27	2,4-Dimethylphenol	40.000	41.125	-2.8	127	0.00
28	bis(2-Chloroethoxy)methane	40.000	41.454	-3.6	130	0.00
29 C	2,4-Dichlorophenol	40.000	44.371	-10.9	133	0.00
30	1,2,4-Trichlorobenzene	40.000	40.935	-2.3	129	0.00
31	Naphthalene	40.000	39.927	0.2	126	0.00
32	Benzoic acid	40.000	56.559	-41.4#	203	0.00
33	4-Chloroaniline	40.000	41.492	-3.7	129	0.00
34 C	Hexachlorobutadiene	40.000	41.723	-4.3	131	0.00
35	Caprolactam	40.000	44.549	-11.4	137	0.00
36 C	4-Chloro-3-methylphenol	40.000	44.525	-11.3	133	0.00
37	2-Methylnaphthalene	40.000	40.378	-0.9	125	0.00
38	1-Methylnaphthalene	40.000	40.234	-0.6	125	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	132	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.853	0.4	129	-0.01
41 P	Hexachlorocyclopentadiene	40.000	47.607	-19.0	151	-0.01
42 S	2,4,6-Tribromophenol	80.000	98.380	-23.0	148	0.00
43 C	2,4,6-Trichlorophenol	40.000	46.748	-16.9	141	0.00
44	2,4,5-Trichlorophenol	40.000	44.962	-12.4	136	0.00
45 S	2-Fluorobiphenyl	80.000	77.262	3.4	125	-0.01
46	1,1'-Biphenyl	40.000	38.791	3.0	125	-0.01
47	2-Chloronaphthalene	40.000	39.153	2.1	126	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111425\
 Data File : BP026137.D
 Acq On : 14 Nov 2025 11:41
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 14 12:01:01 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP102925.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 30 11:51:34 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	44.968	-12.4	150	0.00
49	Acenaphthylene	40.000	39.702	0.7	125	0.00
50	Dimethylphthalate	40.000	40.445	-1.1	127	0.00
51	2,6-Dinitrotoluene	40.000	46.531	-16.3	158	0.00
52 C	Acenaphthene	40.000	38.537	3.7	122	0.00
53	3-Nitroaniline	40.000	44.349	-10.9	147	0.00
54 P	2,4-Dinitrophenol	40.000	63.260	-58.1#	241	0.00
55	Dibenzofuran	40.000	38.820	2.9	125	0.00
56 P	4-Nitrophenol	40.000	44.305	-10.8	148	0.01
57	2,4-Dinitrotoluene	40.000	46.960	-17.4	157	0.00
58	Fluorene	40.000	38.645	3.4	125	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	49.047	-22.6	146	0.00
60	Diethylphthalate	40.000	41.119	-2.8	128	0.00
61	4-Chlorophenyl-phenylether	40.000	38.898	2.8	125	0.00
62	4-Nitroaniline	40.000	46.923	-17.3	140	0.00
63	Azobenzene	40.000	39.203	2.0	124	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	128	0.01
65	4,6-Dinitro-2-methylphenol	40.000	60.951	-52.4#	224	0.01
66 c	n-Nitrosodiphenylamine	40.000	39.937	0.2	124	0.00
67	4-Bromophenyl-phenylether	40.000	42.011	-5.0	131	0.00
68	Hexachlorobenzene	40.000	41.942	-4.9	132	0.00
69	Atrazine	40.000	41.598	-4.0	128	0.00
70 C	Pentachlorophenol	40.000	47.621	-19.1	146	0.00
71	Phenanthrene	40.000	39.140	2.1	123	0.00
72	Anthracene	40.000	40.534	-1.3	127	0.00
73	Carbazole	40.000	40.246	-0.6	125	0.01
74	Di-n-butylphthalate	40.000	44.353	-10.9	130	0.00
75 C	Fluoranthene	40.000	40.161	-0.4	124	0.01
76 I	Chrysene-d12	20.000	20.000	0.0	124	0.02
77	Benzydine	40.000	38.776	3.1	118	0.01
78	Pyrene	40.000	41.363	-3.4	122	0.01
79 S	Terphenyl-d14	80.000	80.729	-0.9	122	0.01
80	Butylbenzylphthalate	40.000	48.605	-21.5	154	0.01
81	Benzo(a)anthracene	40.000	40.094	-0.2	121	0.02
82	3,3'-Dichlorobenzidine	40.000	45.374	-13.4	135	0.01
83	Chrysene	40.000	40.478	-1.2	122	0.01
84	Bis(2-ethylhexyl)phthalate	40.000	45.555	-13.9	137	0.00
85 c	Di-n-octyl phthalate	40.000	54.429	-36.1#	157	0.01
86 I	Perylene-d12	20.000	20.000	0.0	138	0.02
87	Indeno(1,2,3-cd)pyrene	40.000	40.841	-2.1	133	0.06
88	Benzo(b)fluoranthene	40.000	39.151	2.1	128	0.02
89	Benzo(k)fluoranthene	40.000	38.901	2.7	127	0.01
90 C	Benzo(a)pyrene	40.000	40.381	-1.0	131	0.02
91	Dibenzo(a,h)anthracene	40.000	40.603	-1.5	132	0.05
92	Benzo(g,h,i)perylene	40.000	40.880	-2.2	134	0.06

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111425\
Data File : BP026137.D
Acq On : 14 Nov 2025 11:41
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Nov 14 12:01:01 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP102925.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 30 11:51:34 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 2