

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111523\
 Data File : BP018211.D
 Acq On : 16 Nov 2023 13:34
 Operator : MA/JU
 Sample : SP6360
 Misc : SFAM Spike
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SP6360

Quant Time: Nov 16 21:39:06 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP111023.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Nov 15 18:13:37 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.781	152	57046	20.000	ng/u1	0.00
20) Naphthalene-d8	10.599	136	231179	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.463	164	156262	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.245	188	356497	20.000	ng/u1	0.00
79) Chrysene-d12	21.380	240	340720	20.000	ng/u1	0.00
88) Perylene-d12	23.851	264	362269	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.164	96	35	0.022	ng/uL	0.00
4) Pyridine-d5	3.611	84	742	0.176	ng/u1	0.02
7) Phenol-d5	6.934	99	14	0.003	ng/u1	-0.02
9) Bis-(2-Chloroethyl)eth...	7.128	67	86	0.027	ng/u1	0.00
11) 2-Chlorophenol-d4	7.299	132	21	0.005	ng/u1	0.00
15) 4-Methylphenol-d8	8.611	113	39264	8.700	ng/u1	0.10
21) Nitrobenzene-d5	8.999	128	52	0.025	ng/u1	0.00
24) 2-Nitrophenol-d4	9.699	143	12	0.005	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.252	165	14001	3.329	ng/u1	0.02
31) 4-Chloroaniline-d4	10.775	131	406	0.070	ng/u1	-0.01
46) Dimethylphthalate-d6	13.893	166	137	0.010	ng/u1	0.00
49) Acenaphthylene-d8	14.169	160	140	0.009	ng/u1	0.00
54) 4-Nitrophenol-d4	14.757	143	69989	32.486	ng/u1	0.02
60) Fluorene-d10	15.481	176	281	0.024	ng/u1	0.01
65) 4,6-Dinitro-2-methylph...	15.645	200	2863	1.111	ng/u1	0.02
73) Anthracene-d10	17.245	188	356497	18.490	ng/u1	-0.10
81) Pyrene-d10	19.604	212	798	0.035	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.692	264	601	0.029	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.193	88	54944	31.136	ng/uL	97
5) Pyridine	3.605	79	347057	79.753	ng/u1	94
6) Benzaldehyde	6.952	77	292486	98.611	ng/u1	98
8) Phenol	6.981	94	432419	79.592	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.228	93	340345	80.748	ng/u1	96
12) 2-Chlorophenol	7.334	128	357007	81.529	ng/u1	98
13) 2-Methylphenol	8.234	108	329856	80.354	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.293	45	221628	46.050	ng/u1	99
16) Acetophenone	8.640	105	560144	78.085	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.605	70	305135	77.910	ng/u1	94
18) 4-Methylphenol	8.575	108	355810	79.478	ng/u1	96
19) Hexachloroethane	8.828	117	173515	82.820	ng/u1	91
22) Nitrobenzene	9.034	77	476472	84.723	ng/u1	96
23) Isophorone	9.540	82	840230	81.877	ng/u1	99
25) 2-Nitrophenol	9.734	139	198865	82.015	ng/u1	97
26) 2,4-Dimethylphenol	9.775	107	444911	86.257	ng/u1	97
27) Bis(2-Chloroethoxy)met...	10.028	93	467678	83.572	ng/u1	98
29) 2,4-Dichlorophenol	10.252	162	337943	84.150	ng/u1	97
30) Naphthalene	10.652	128	1154590	82.125	ng/u1	99
32) 4-Chloroaniline	10.805	127	342875	62.067	ng/u1	99
33) Hexachlorobutadiene	10.863	225	251485	85.301	ng/u1	99
34) Caprolactam	11.652	113	113799	85.300	ng/u1	99
35) 4-Chloro-3-methylphenol	11.922	107	415420	85.981	ng/u1	97
36) 2-Methylnaphthalene	12.269	142	808986	83.815	ng/u1	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111523\
 Data File : BP018211.D
 Acq On : 16 Nov 2023 13:34
 Operator : MA/JU
 Sample : SP6360
 Misc : SFAM Spike
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SP6360

Quant Time: Nov 16 21:39:06 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP111023.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Nov 15 18:13:37 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.487	142	798117	80.468	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.622	216	429719	81.404	ng/ul	98
40) Hexachlorocyclopentadiene	12.557	237	229938	90.241	ng/ul	98
41) 2,4,6-Trichlorophenol	12.893	196	289178	83.507	ng/ul	98
42) 2,4,5-Trichlorophenol	12.969	196	303061	83.156	ng/ul	97
43) 1,1'-Biphenyl	13.293	154	1004243	75.578	ng/ul	99
44) 2-Chloronaphthalene	13.340	162	802178	76.041	ng/ul	99
45) 2-Nitroaniline	13.599	65	302847	88.961	ng/ul	98
47) Dimethylphthalate	13.940	163	1168844	82.643	ng/ul	99
48) 2,6-Dinitrotoluene	14.093	165	241257	85.821	ng/ul	96
50) Acenaphthylene	14.193	152	1440487	81.247	ng/ul	100
51) 3-Nitroaniline	14.440	138	206117	79.016	ng/ul	100
52) Acenaphthene	14.528	153	971240	82.612	ng/ul	99
53) 2,4-Dinitrophenol	14.651	184	107140	69.863	ng/ul	93
55) 4-Nitrophenol	14.746	109	239414	83.715	ng/ul	86
56) Dibenzofuran	14.869	168	1339427	83.136	ng/ul	96
57) 2,4-Dinitrotoluene	14.887	165	370720	88.266	ng/ul	99
58) 2,3,4,6-Tetrachlorophenol	15.093	232	260352	80.305	ng/ul	96
59) Diethylphthalate	15.298	149	1291213	87.385	ng/ul	99
61) Fluorene	15.522	166	1180944	86.338	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.516	204	592710	88.425	ng/ul	99
63) 4-Nitroaniline	15.616	138	237733	96.577	ng/ul	95
66) 4,6-Dinitro-2-methylph...	15.645	198	206123	77.076	ng/ul#	97
67) N-Nitrosodiphenylamine	15.751	169	962171	84.157	ng/ul	99
68) 4-Bromophenyl-phenylether	16.422	248	340171	84.662	ng/ul	94
69) Hexachlorobenzene	16.504	284	372478	83.078	ng/ul	98
70) Atrazine	16.716	200	396774	99.488	ng/ul	98
71) Pentachlorophenol	16.875	266	130801	48.061	ng/ul	95
72) Phenanthrene	17.292	178	1839066	84.018	ng/ul	99
74) Anthracene	17.381	178	1894746	84.824	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.246	216	422320	77.576	ng/ul	98
76) Pentachlorobenzene	14.757	250	425501	79.840	ng/ul	97
77) Carbazole	17.681	167	1706166	87.957	ng/ul	99
78) Di-n-butylphthalate	18.192	149	2337275	94.965	ng/ul	99
80) Fluoranthene	19.275	202	2255485	84.859	ng/ul	99
82) Pyrene	19.634	202	2361444	84.786	ng/ul	99
83) Butylbenzylphthalate	20.504	149	1140175	97.344	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.310	252	674454	82.853	ng/ul	97
85) Benzo(a)anthracene	21.363	228	2337658	84.967	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.245	149	1546619	95.772	ng/ul	99
87) Chrysene	21.416	228	2156509	83.783	ng/ul	98
89) Di-n-octyl phthalate	22.198	149	2662218	98.264	ng/ul	100
90) Benzo(b)fluoranthene	23.086	252	2248125	87.199	ng/ul	98
91) Benzo(k)fluoranthene	23.139	252	2292625	88.385	ng/ul	100
93) Benzo(a)pyrene	23.745	252	2145330	88.113	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.492	276	2473972	84.720	ng/ul	97
95) Dibenzo(a,h)anthracene	26.515	278	2072090	86.178	ng/ul	98
96) Benzo(g,h,i)perylene	27.310	276	1935551	82.870	ng/ul	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111523\
 Data File : BP018211.D
 Acq On : 16 Nov 2023 13:34
 Operator : MA/JU
 Sample : SP6360
 Misc : SFAM Spike
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SP6360

Quant Time: Nov 16 21:39:06 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP111023.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Nov 15 18:13:37 2023
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 31.136 ng/uL
RT: 3.193 min Scan# 11
Delta R.T. 0.000 min
Lab File: BP018211.D
Acq: 16 Nov 2023 13:34




Instrument :
BNA_P
ClientSampleId :
SP6360


Tgt Ion: 88 Resp: 54944
Ion Ratio Lower Upper
88 100
43 35.5 25.4 38.2
58 67.8 53.3 79.9



#5
Pyridine
Concen: 79.753 ng/uL
RT: 3.605 min Scan# 88
Delta R.T. -0.006 min
Lab File: BP018211.D
Acq: 16 Nov 2023 13:34

Tgt Ion: 79 Resp: 347057
Ion Ratio Lower Upper
79 100
52 61.3 53.9 80.9
51 38.7 32.2 48.2









#6
Benzaldehyde
Concen: 98.611 ng/ul
RT: 6.952 min Scan# 61
Delta R.T. -0.006 min
Lab File: BP018211.D
Acq: 16 Nov 2023 13:34




Instrument :
BNA_P
ClientSampleId :
SP6360


Tgt Ion: 77 Resp: 292486
Ion Ratio Lower Upper
77 100
105 87.5 72.7 109.1
106 84.2 67.5 101.3



#8
Phenol
Concen: 79.592 ng/ul
RT: 6.981 min Scan# 662
Delta R.T. 0.000 min
Lab File: BP018211.D
Acq: 16 Nov 2023 13:34

Tgt Ion: 94 Resp: 432419
Ion Ratio Lower Upper
94 100
65 35.7 28.2 42.2
66 49.7 40.3 60.5









#10
 Bis(2-Chloroethyl)ether
 Concen: 80.748 ng/ul
 RT: 7.228 min Scan# 704
 Delta R.T. 0.000 min
 Lab File: BP018211.D
 Acq: 16 Nov 2023 13:34




Instrument :
 BNA_P
 ClientSampleId :
 SP6360


Tgt Ion: 93 Resp: 340345
 Ion Ratio Lower Upper
 93 100
 63 76.7 57.8 86.6
 95 32.1 25.5 38.3



#12
 2-Chlorophenol
 Concen: 81.529 ng/ul
 RT: 7.334 min Scan# 722
 Delta R.T. -0.006 min
 Lab File: BP018211.D
 Acq: 16 Nov 2023 13:34

Tgt Ion:128 Resp: 357007
 Ion Ratio Lower Upper
 128 100
 64 50.8 39.2 58.8
 130 32.3 26.6 40.0





#13

2-Methylphenol

Concen: 80.354 ng/ul

RT: 8.234 min Scan# 81

Delta R.T. -0.006 min

Lab File: BP018211.D

Acq: 16 Nov 2023 13:34

Instrument :

BNA_P

ClientSampleId :





SP6360

Tgt Ion:108 Resp: 329856

Ion Ratio Lower Upper

108 100

107 92.6 75.2 112.8



#14

2,2'-oxybis(1-Chloropropane)

Concen: 46.050 ng/ul

RT: 8.293 min Scan# 885

Delta R.T. -0.006 min

Lab File: BP018211.D

Acq: 16 Nov 2023 13:34




Tgt Ion: 45 Resp: 221628


Ion Ratio Lower Upper

45 100

77 20.7 16.1 24.1

79 14.3 11.1 16.7









#16
Acetophenone
Concen: 78.085 ng/ul
RT: 8.640 min Scan# 944
Delta R.T. 0.000 min
Lab File: BP018211.D
Acq: 16 Nov 2023 13:34




Instrument :
BNA_P
ClientSampleId :
SP6360


Tgt Ion:105 Resp: 560144
Ion Ratio Lower Upper
105 100
77 89.7 70.1 105.1
51 30.7 23.4 35.0



#17
N-Nitroso-di-n-propylamine
Concen: 77.910 ng/ul
RT: 8.605 min Scan# 938
Delta R.T. 0.000 min
Lab File: BP018211.D
Acq: 16 Nov 2023 13:34

Tgt Ion: 70 Resp: 305135
Ion Ratio Lower Upper
70 100
42 60.6 43.7 65.5
101 9.9 8.2 12.4
130 18.7 15.2 22.8









#18
4-Methylphenol
Concen: 79.478 ng/ul
RT: 8.575 min Scan# 911
Delta R.T. 0.000 min
Lab File: BP018211.D
Acq: 16 Nov 2023 13:34




Instrument :
BNA_P
ClientSampleId :
SP6360


Tgt Ion:108 Resp: 355810
Ion Ratio Lower Upper
108 100
107 114.2 88.2 132.4



#19
Hexachloroethane
Concen: 82.820 ng/ul
RT: 8.828 min Scan# 976
Delta R.T. -0.006 min
Lab File: BP018211.D
Acq: 16 Nov 2023 13:34

Tgt Ion:117 Resp: 173515
Ion Ratio Lower Upper
117 100
201 93.3 84.8 127.2
199 58.9 49.0 73.4









#22
Nitrobenzene
Concen: 84.723 ng/ul
RT: 9.034 min Scan# 1011
Delta R.T. 0.000 min
Lab File: BP018211.D
Acq: 16 Nov 2023 13:34




Instrument :
BNA_P
ClientSampleId :
SP6360


Tgt Ion: 77 Resp: 476472
Ion Ratio Lower Upper
77 100
123 37.5 27.8 41.8
65 14.8 11.5 17.3



#23
Isophorone
Concen: 81.877 ng/ul
RT: 9.540 min Scan# 1097
Delta R.T. 0.006 min
Lab File: BP018211.D
Acq: 16 Nov 2023 13:34

Tgt Ion: 82 Resp: 840230
Ion Ratio Lower Upper
82 100
95 8.4 6.6 9.8
138 16.2 13.1 19.7









#25
2-Nitrophenol
Concen: 82.015 ng/ul
RT: 9.734 min Scan# 1131
Delta R.T. 0.000 min
Lab File: BP018211.D
Acq: 16 Nov 2023 13:34




Instrument :
BNA_P
ClientSampleId :
SP6360


Tgt Ion:139 Resp: 198865
Ion Ratio Lower Upper
139 100
65 66.1 51.4 77.0
109 43.9 33.1 49.7



#26
2,4-Dimethylphenol
Concen: 86.257 ng/ul
RT: 9.775 min Scan# 1137
Delta R.T. -0.006 min
Lab File: BP018211.D
Acq: 16 Nov 2023 13:34

Tgt Ion:107 Resp: 444911
Ion Ratio Lower Upper
107 100
121 45.7 37.8 56.8
122 75.2 62.1 93.1









#27
Bis(2-Chloroethoxy)methane
Concen: 83.572 ng/ul
RT: 10.028 min Scan# 1180
Delta R.T. -0.006 min
Lab File: BP018211.D
Acq: 16 Nov 2023 13:34




Instrument :
BNA_P
ClientSampleId :
SP6360


Tgt Ion: 93 Resp: 467678
Ion Ratio Lower Upper
93 100
95 31.6 26.2 39.2
123 10.8 8.6 12.8



#29
2,4-Dichlorophenol
Concen: 84.150 ng/ul
RT: 10.252 min Scan# 1218
Delta R.T. -0.006 min
Lab File: BP018211.D
Acq: 16 Nov 2023 13:34

Tgt Ion:162 Resp: 337943
Ion Ratio Lower Upper
162 100
164 63.7 51.8 77.6
98 38.8 28.7 43.1





#30

Naphthalene

Concen: 82.125 ng/ul

RT: 10.652 min Scan# 11

Delta R.T. 0.000 min

Lab File: BP018211.D

Acq: 16 Nov 2023 13:34

Instrument :

BNA_P

ClientSampleId :

SP6360





Tgt Ion:128 Resp: 1154590

Ion Ratio Lower Upper

128 100

129 10.6 9.0 13.4

127 13.3 10.3 15.5



#32

4-Chloroaniline

Concen: 62.067 ng/ul

RT: 10.805 min Scan# 1312

Delta R.T. -0.006 min

Lab File: BP018211.D




Acq: 16 Nov 2023 13:34


Tgt Ion:127 Resp: 342875

Ion Ratio Lower Upper

127 100

129 32.2 26.1 39.1





#33

Hexachlorobutadiene

Concen: 85.301 ng/ul

RT: 10.863 min Scan# 11

Delta R.T. -0.006 min

Lab File: BP018211.D

Acq: 16 Nov 2023 13:34

Instrument :

BNA_P

ClientSampleId :

SP6360





Tgt Ion:225 Resp: 251485

Ion Ratio Lower Upper

225 100

223 64.9 52.5 78.7

227 64.6 51.9 77.9



#34

Caprolactam

Concen: 85.300 ng/ul

RT: 11.652 min Scan# 1456

Delta R.T. 0.006 min

Lab File: BP018211.D

Acq: 16 Nov 2023 13:34




Tgt Ion:113 Resp: 113799


Ion Ratio Lower Upper

113 100

55 145.7 118.3 177.5

56 119.6 95.6 143.4









#35
4-Chloro-3-methylphenol
Concen: 85.981 ng/ul
RT: 11.922 min Scan# 1502
Delta R.T. 0.000 min
Lab File: BP018211.D
Acq: 16 Nov 2023 13:34




Instrument :
BNA_P
ClientSampleId :
SP6360


Tgt Ion:107 Resp: 415420
Ion Ratio Lower Upper
107 100
144 24.0 19.1 28.7
142 71.7 54.8 82.2



#36
2-Methylnaphthalene
Concen: 83.815 ng/ul
RT: 12.269 min Scan# 1561
Delta R.T. 0.000 min
Lab File: BP018211.D
Acq: 16 Nov 2023 13:34

Tgt Ion:142 Resp: 808986
Ion Ratio Lower Upper
142 100
141 88.1 71.7 107.5









#37
1-Methylnaphthalene
Concen: 80.468 ng/ul
RT: 12.487 min Scan# 1598
Delta R.T. -0.006 min
Lab File: BP018211.D
Acq: 16 Nov 2023 13:34




Instrument :
BNA_P
ClientSampleId :
SP6360


Tgt Ion:142 Resp: 798117
Ion Ratio Lower Upper
142 100
141 93.4 72.8 109.2
116 5.3 4.0 6.0



#39
1,2,4,5-Tetrachlorobenzene
Concen: 81.404 ng/ul
RT: 12.622 min Scan# 1621
Delta R.T. -0.006 min
Lab File: BP018211.D
Acq: 16 Nov 2023 13:34

Tgt Ion:216 Resp: 429719
Ion Ratio Lower Upper
216 100
214 80.3 64.8 97.2
179 24.6 18.3 27.5
108 21.7 15.7 23.5









#40
Hexachlorocyclopentadiene
Concen: 90.241 ng/ul
RT: 12.557 min Scan# 1610
Delta R.T. 0.000 min
Lab File: BP018211.D
Acq: 16 Nov 2023 13:34




Instrument :
BNA_P
ClientSampleId :
SP6360


Tgt Ion:237 Resp: 229938
Ion Ratio Lower Upper
237 100
235 61.0 49.9 74.9
272 11.8 8.9 13.3



#41
2,4,6-Trichlorophenol
Concen: 83.507 ng/ul
RT: 12.893 min Scan# 1667
Delta R.T. -0.006 min
Lab File: BP018211.D
Acq: 16 Nov 2023 13:34

Tgt Ion:196 Resp: 289178
Ion Ratio Lower Upper
196 100
198 96.7 78.9 118.3
200 31.9 25.8 38.8









#42
2,4,5-Trichlorophenol
Concen: 83.156 ng/ul
RT: 12.969 min Scan# 1681
Delta R.T. -0.012 min
Lab File: BP018211.D
Acq: 16 Nov 2023 13:34




Instrument :
BNA_P
ClientSampleId :
SP6360


Tgt Ion:196 Resp: 303061
Ion Ratio Lower Upper
196 100
198 96.5 78.9 118.3
200 30.5 26.6 40.0



#43
1,1'-Biphenyl
Concen: 75.578 ng/ul
RT: 13.293 min Scan# 1735
Delta R.T. 0.000 min
Lab File: BP018211.D
Acq: 16 Nov 2023 13:34

Tgt Ion:154 Resp: 1004243
Ion Ratio Lower Upper
154 100
153 40.4 32.4 48.6
76 11.7 9.8 14.8









#44
2-Chloronaphthalene
Concen: 76.041 ng/ul
RT: 13.340 min Scan# 1743
Delta R.T. 0.000 min
Lab File: BP018211.D
Acq: 16 Nov 2023 13:34




Instrument :
BNA_P
ClientSampleId :
SP6360


Tgt Ion:162 Resp: 802178
Ion Ratio Lower Upper
162 100
164 32.9 26.9 40.3
127 45.0 35.4 53.2



#45
2-Nitroaniline
Concen: 88.961 ng/ul
RT: 13.599 min Scan# 1787
Delta R.T. -0.006 min
Lab File: BP018211.D
Acq: 16 Nov 2023 13:34

Tgt Ion: 65 Resp: 302847
Ion Ratio Lower Upper
65 100
92 63.4 53.8 80.6
138 93.7 75.1 112.7





#47

Dimethylphthalate

Concen: 82.643 ng/ul

RT: 13.940 min Scan# 1845

Delta R.T. 0.006 min

Lab File: BP018211.D

Acq: 16 Nov 2023 13:34

Instrument :

BNA_P

ClientSampleId :

SP6360





Tgt Ion:163 Resp: 1168844

Ion Ratio Lower Upper

163 100

194 3.9 3.5 5.3

164 9.7 8.0 12.0



#48

2,6-Dinitrotoluene

Concen: 85.821 ng/ul

RT: 14.093 min Scan# 1871

Delta R.T. 0.000 min

Lab File: BP018211.D

Acq: 16 Nov 2023 13:34




Tgt Ion:165 Resp: 241257


Ion Ratio Lower Upper

165 100

89 77.5 59.4 89.0

121 24.1 18.2 27.4





#50

Acenaphthylene

Concen: 81.247 ng/ul

RT: 14.193 min Scan# 1888

Delta R.T. 0.000 min

Lab File: BP018211.D

Acq: 16 Nov 2023 13:34

Instrument :

BNA_P

ClientSampleId :

SP6360





Tgt Ion:152 Resp: 1440487

Ion Ratio Lower Upper

152 100

151 20.2 16.2 24.2

153 13.1 10.4 15.6



#51

3-Nitroaniline

Concen: 79.016 ng/ul

RT: 14.440 min Scan# 1930

Delta R.T. -0.006 min

Lab File: BP018211.D

Acq: 16 Nov 2023 13:34




Tgt Ion:138 Resp: 206117


Ion Ratio Lower Upper

138 100

108 11.2 8.3 12.5

92 123.8 99.0 148.4





#52

Acenaphthene

Concen: 82.612 ng/ul

RT: 14.528 min Scan# 1945

Delta R.T. 0.000 min

Lab File: BP018211.D

Acq: 16 Nov 2023 13:34

Instrument :

BNA_P

ClientSampleId :

SP6360





Tgt Ion:153 Resp: 971240

Ion Ratio Lower Upper

153 100

152 47.6 39.0 58.6

154 89.3 71.0 106.4



#53

2,4-Dinitrophenol

Concen: 69.863 ng/ul

RT: 14.651 min Scan# 1966

Delta R.T. -0.023 min

Lab File: BP018211.D

Acq: 16 Nov 2023 13:34




Tgt Ion:184 Resp: 107140


Ion Ratio Lower Upper

184 100

63 87.3 67.7 101.5

154 85.1 60.2 90.4









#55
4-Nitrophenol
Concen: 83.715 ng/ul
RT: 14.746 min Scan# 1982
Delta R.T. -0.006 min
Lab File: BP018211.D
Acq: 16 Nov 2023 13:34




Instrument :
BNA_P
ClientSampleId :
SP6360


Tgt Ion:109 Resp: 239414
Ion Ratio Lower Upper
109 100
139 82.0 74.9 112.3
65 99.2 93.8 140.6



#56
Dibenzofuran
Concen: 83.136 ng/ul
RT: 14.869 min Scan# 2003
Delta R.T. 0.000 min
Lab File: BP018211.D
Acq: 16 Nov 2023 13:34

Tgt Ion:168 Resp: 1339427
Ion Ratio Lower Upper
168 100
139 45.1 34.0 51.0





#57

2,4-Dinitrotoluene

Concen: 88.266 ng/ul

RT: 14.887 min Scan# 2006

Delta R.T. -0.006 min

Lab File: BP018211.D

Acq: 16 Nov 2023 13:34

Instrument :

BNA_P

ClientSampleId :

SP6360





Tgt Ion:165 Resp: 370720

Ion Ratio Lower Upper

165 100

63 59.8 48.6 72.8

182 2.7 2.1 3.1



#58

2,3,4,6-Tetrachlorophenol

Concen: 80.305 ng/ul

RT: 15.093 min Scan# 2041

Delta R.T. -0.006 min

Lab File: BP018211.D

Acq: 16 Nov 2023 13:34

Tgt Ion:232 Resp: 260352




Ion Ratio Lower Upper


232 100

131 58.2 44.1 66.1

130 3.6 2.5 3.7

166 36.5 27.9 41.9









#59
Diethylphthalate
Concen: 87.385 ng/ul
RT: 15.298 min Scan# 2075
Delta R.T. 0.006 min
Lab File: BP018211.D
Acq: 16 Nov 2023 13:34




Instrument :
BNA_P
ClientSampleId :
SP6360


Tgt Ion:149 Resp: 1291213
Ion Ratio Lower Upper
149 100
177 20.9 16.3 24.5
150 12.0 10.0 15.0



#61
Fluorene
Concen: 86.338 ng/ul
RT: 15.522 min Scan# 2114
Delta R.T. 0.000 min
Lab File: BP018211.D
Acq: 16 Nov 2023 13:34

Tgt Ion:166 Resp: 1180944
Ion Ratio Lower Upper
166 100
165 97.9 79.0 118.6
167 13.0 10.6 16.0









#62
4-Chlorophenyl-phenylether
Concen: 88.425 ng/ul
RT: 15.516 min Scan# 2113
Delta R.T. 0.000 min
Lab File: BP018211.D
Acq: 16 Nov 2023 13:34




Instrument :
BNA_P
ClientSampleId :
SP6360


Tgt Ion:204 Resp: 592710
Ion Ratio Lower Upper
204 100
206 32.9 26.2 39.2
141 71.2 56.0 84.0



#63
4-Nitroaniline
Concen: 96.577 ng/ul
RT: 15.616 min Scan# 2130
Delta R.T. 0.000 min
Lab File: BP018211.D
Acq: 16 Nov 2023 13:34

Tgt Ion:138 Resp: 237733
Ion Ratio Lower Upper
138 100
92 60.1 46.6 70.0
108 123.0 104.7 157.1









#66
4,6-Dinitro-2-methylphenol
Concen: 77.076 ng/ul
RT: 15.645 min Scan# 2134
Delta R.T. 0.000 min
Lab File: BP018211.D
Acq: 16 Nov 2023 13:34




Instrument :
BNA_P
ClientSampleId :
SP6360


Tgt Ion:198 Resp: 206123
Ion Ratio Lower Upper
198 100
182 3.5 3.8 5.6#
77 32.6 27.2 40.8



#67
N-Nitrosodiphenylamine
Concen: 84.157 ng/ul
RT: 15.751 min Scan# 2153
Delta R.T. 0.006 min
Lab File: BP018211.D
Acq: 16 Nov 2023 13:34

Tgt Ion:169 Resp: 962171
Ion Ratio Lower Upper
169 100
168 66.4 53.8 80.8
167 36.6 29.8 44.6









#68
4-Bromophenyl-phenylether
Concen: 84.662 ng/ul
RT: 16.422 min Scan# 2267
Delta R.T. 0.000 min
Lab File: BP018211.D
Acq: 16 Nov 2023 13:34




Instrument :
BNA_P
ClientSampleId :
SP6360


Tgt Ion:248 Resp: 340171
Ion Ratio Lower Upper
248 100
250 96.3 72.7 109.1
141 94.8 70.4 105.6



#69
Hexachlorobenzene
Concen: 83.078 ng/ul
RT: 16.504 min Scan# 2281
Delta R.T. 0.000 min
Lab File: BP018211.D
Acq: 16 Nov 2023 13:34

Tgt Ion:284 Resp: 372478
Ion Ratio Lower Upper
284 100
142 46.3 38.9 58.3
249 35.3 28.6 43.0





#70

Atrazine

Concen: 99.488 ng/ul

RT: 16.716 min Scan# 21

Delta R.T. 0.006 min

Lab File: BP018211.D

Acq: 16 Nov 2023 13:34

Instrument :

BNA_P

ClientSampleId :

SP6360





Tgt Ion:200 Resp: 396774

Ion Ratio Lower Upper

200 100

173 26.0 20.2 30.2

215 45.5 37.8 56.8



#71

Pentachlorophenol

Concen: 48.061 ng/ul

RT: 16.875 min Scan# 2344

Delta R.T. -0.006 min

Lab File: BP018211.D

Acq: 16 Nov 2023 13:34




Tgt Ion:266 Resp: 130801


Ion Ratio Lower Upper

266 100

264 58.6 50.7 76.1

268 62.7 52.4 78.6





#72

Phenanthrene

Concen: 84.018 ng/ul

RT: 17.292 min Scan# 2414

Delta R.T. 0.006 min

Lab File: BP018211.D

Acq: 16 Nov 2023 13:34

Instrument :

BNA_P

ClientSampleId :

SP6360





Tgt Ion:178 Resp: 1839066

Ion Ratio Lower Upper

178 100

179 15.2 12.0 18.0

176 19.7 15.5 23.3



#74

Anthracene

Concen: 84.824 ng/ul

RT: 17.381 min Scan# 2430

Delta R.T. 0.000 min

Lab File: BP018211.D

Acq: 16 Nov 2023 13:34




Tgt Ion:178 Resp: 1894746


Ion Ratio Lower Upper

178 100

179 15.4 12.6 19.0

176 19.1 15.4 23.0





#75

1,2,3,4-Tetrachlorobenzene

Concen: 77.576 ng/uL

RT: 13.246 min Scan# 1727

Delta R.T. 0.000 min

Lab File: BP018211.D

Acq: 16 Nov 2023 13:34

Instrument :

BNA_P

ClientSampleId :

SP6360





Tgt Ion:216 Resp: 422320

Ion Ratio Lower Upper

216 100

214 76.7 64.0 96.0

218 48.5 39.0 58.4



#76

Pentachlorobenzene

Concen: 79.840 ng/uL

RT: 14.757 min Scan# 1984

Delta R.T. -0.006 min

Lab File: BP018211.D

Acq: 16 Nov 2023 13:34




Tgt Ion:250 Resp: 425501


Ion Ratio Lower Upper

250 100

248 63.8 52.7 79.1

252 63.2 52.3 78.5





#77

Carbazole

Concen: 87.957 ng/ul

RT: 17.681 min Scan# 2481

Delta R.T. 0.000 min

Lab File: BP018211.D

Acq: 16 Nov 2023 13:34

Instrument :

BNA_P

ClientSampleId :

SP6360





Tgt Ion:167 Resp: 1706166

Ion Ratio Lower Upper

167 100

166 21.6 17.6 26.4

139 14.4 11.4 17.0



#78

Di-n-butylphthalate

Concen: 94.965 ng/ul

RT: 18.192 min Scan# 2568

Delta R.T. 0.006 min

Lab File: BP018211.D

Acq: 16 Nov 2023 13:34




Tgt Ion:149 Resp: 2337275


Ion Ratio Lower Upper

149 100

150 9.0 6.7 10.1

104 6.4 5.1 7.7





#80

Fluoranthene

Concen: 84.859 ng/ul

RT: 19.275 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP018211.D

Acq: 16 Nov 2023 13:34

Instrument :

BNA_P

ClientSampleId :

SP6360





Tgt Ion:202 Resp: 2255485

Ion Ratio Lower Upper

202 100

101 9.3 7.2 10.8

100 7.4 5.8 8.6



#82

Pyrene

Concen: 84.786 ng/ul

RT: 19.634 min Scan# 2813

Delta R.T. 0.000 min

Lab File: BP018211.D

Acq: 16 Nov 2023 13:34




Tgt Ion:202 Resp: 2361444


Ion Ratio Lower Upper

202 100

101 11.1 8.3 12.5

100 9.1 7.4 11.2





#83

Butylbenzylphthalate

Concen: 97.344 ng/ul

RT: 20.504 min Scan# 2961

Delta R.T. 0.000 min

Lab File: BP018211.D

Acq: 16 Nov 2023 13:34

Instrument :

BNA_P

ClientSampleId :

SP6360





Tgt Ion:149 Resp: 1140175

Ion Ratio Lower Upper

149 100

91 78.0 59.6 89.4

206 19.6 14.3 21.5



#84

3,3'-Dichlorobenzidine

Concen: 82.853 ng/ul

RT: 21.310 min Scan# 3098

Delta R.T. 0.000 min

Lab File: BP018211.D

Acq: 16 Nov 2023 13:34




Tgt Ion:252 Resp: 674454


Ion Ratio Lower Upper

252 100

254 63.9 48.9 73.3

126 10.8 9.4 14.0









#85
Benzo(a)anthracene
Concen: 84.967 ng/ul
RT: 21.363 min Scan# 3106
Delta R.T. 0.006 min
Lab File: BP018211.D
Acq: 16 Nov 2023 13:34




Instrument :
BNA_P
ClientSampleId :
SP6360


Tgt Ion:228 Resp: 2337658
Ion Ratio Lower Upper
228 100
229 20.0 15.4 23.0
226 28.0 21.4 32.0



#86
Bis(2-ethylhexyl)phthalate
Concen: 95.772 ng/ul
RT: 21.245 min Scan# 3087
Delta R.T. 0.000 min
Lab File: BP018211.D
Acq: 16 Nov 2023 13:34

Tgt Ion:149 Resp: 1546619
Ion Ratio Lower Upper
149 100
167 26.8 21.0 31.4
279 4.1 3.2 4.8





#87

Chrysene

Concen: 83.783 ng/ul

RT: 21.416 min Scan# 3116

Delta R.T. 0.000 min

Lab File: BP018211.D

Acq: 16 Nov 2023 13:34

Instrument :

BNA_P

ClientSampleId :

SP6360





Tgt Ion:228 Resp: 2156509

Ion Ratio Lower Upper

228 100

226 30.1 23.1 34.7

229 19.6 15.4 23.0



#89

Di-n-octyl phthalate

Concen: 98.264 ng/ul




RT: 22.198 min Scan# 3249


Delta R.T. 0.006 min

Lab File: BP018211.D

Acq: 16 Nov 2023 13:34

Tgt Ion:149 Resp: 2662218





#90

Benzo(b)fluoranthene

Concen: 87.199 ng/ul

RT: 23.086 min Scan# 3400

Delta R.T. 0.006 min

Lab File: BP018211.D

Acq: 16 Nov 2023 13:34

Instrument :

BNA_P

ClientSampleId :

SP6360





Tgt Ion:252 Resp: 2248125

Ion Ratio Lower Upper

252 100

253 22.2 17.0 25.4

125 9.2 7.0 10.6



#91

Benzo(k)fluoranthene

Concen: 88.385 ng/ul

RT: 23.139 min Scan# 3409

Delta R.T. 0.006 min

Lab File: BP018211.D

Acq: 16 Nov 2023 13:34




Tgt Ion:252 Resp: 2292625


Ion Ratio Lower Upper

252 100

253 22.0 17.6 26.4

125 8.9 6.8 10.2





#93

Benzo(a)pyrene

Concen: 88.113 ng/ul

RT: 23.745 min Scan# 3512

Delta R.T. 0.012 min

Lab File: BP018211.D

Acq: 16 Nov 2023 13:34

Instrument :

BNA_P

ClientSampleId :

SP6360





Tgt Ion:252 Resp: 2145330

Ion Ratio Lower Upper

252 100

253 21.4 17.4 26.0

125 9.6 7.3 10.9



#94

Indeno(1,2,3-cd)pyrene

Concen: 84.720 ng/ul

RT: 26.492 min Scan# 3979

Delta R.T. 0.030 min

Lab File: BP018211.D

Acq: 16 Nov 2023 13:34




Tgt Ion:276 Resp: 2473972


Ion Ratio Lower Upper

276 100

138 21.8 15.6 23.4

277 25.2 19.3 28.9





#95

Dibenzo(a,h)anthracene

Concen: 86.178 ng/ul

RT: 26.515 min Scan# 3978

Delta R.T. 0.024 min

Lab File: BP018211.D

Acq: 16 Nov 2023 13:34

Instrument :

BNA_P

ClientSampleId :

SP6360





Tgt Ion:278 Resp: 2072090

Ion Ratio Lower Upper

278 100

139 14.8 11.1 16.7

279 24.7 19.0 28.4



#96

Benzo(g,h,i)perylene

Concen: 82.870 ng/ul

RT: 27.310 min Scan# 4118

Delta R.T. 0.030 min

Lab File: BP018211.D

Acq: 16 Nov 2023 13:34

Tgt Ion:276 Resp: 1935551

Ion Ratio Lower Upper

276 100

138 20.1 13.8 20.6

277 23.9 20.2 30.2

