




Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111621\
Data File : BP007991.D
Acq On : 16 Nov 2021 09:35
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Nov 16 12:18:58 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 15 05:36:49 2021
Response via : Initial Calibration



TIC: BP007990.D\data.ms

(70) Pentachlorophenol (C)




16.887min (-0.006) 6785106.45 ng

response 327004

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	61.60	63.30
264.00	60.90	63.30
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111621\
 Data File : BP007991.D
 Acq On : 16 Nov 2021 09:35
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Nov 16 12:18:58 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 15 05:36:49 2021
 Response via : Initial Calibration



TIC: BP007990.D\data.ms

(77) Benzidine**19.428min (+ 0.001) 425638.96 ng****response 1880847**

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.00	13.92
183.00	11.40	12.60
0.00	0.00	0.00