

Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP111723\  
 Data File : BP018248.D  
 Acq On : 18 Nov 2023 07:36  
 Operator : MA/JU  
 Sample : PB157221BL  
 Misc :  
 ALS Vial : 37 Sample Multiplier: 1

Instrument :  
 BNA\_P  
 ClientSampleId :  
 SBLK221

Quant Time: Nov 19 01:15:13 2023  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\SFAM-EPA-BP111023.MA.M  
 Quant Title : SVOA CALIBRATION  
 QLast Update : Wed Nov 15 18:13:37 2023  
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.781	152	59259	20.000	ng/ul	0.00
20) Naphthalene-d8	10.593	136	244503	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.457	164	153334	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.239	188	360040	20.000	ng/ul	0.00
79) Chrysene-d12	21.375	240	342903	20.000	ng/ul	0.00
88) Perylene-d12	23.851	264	379575	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.158	96	12095	7.200	ng/uL	0.00
4) Pyridine-d5	3.593	84	144910	33.084	ng/ul	0.00
7) Phenol-d5	6.946	99	180116	31.483	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.128	67	116130	34.541	ng/ul	0.00
11) 2-Chlorophenol-d4	7.299	132	155072	34.229	ng/ul	0.00
15) 4-Methylphenol-d8	8.505	113	147888	31.546	ng/ul	0.00
21) Nitrobenzene-d5	8.987	128	73278	33.265	ng/ul	0.00
24) 2-Nitrophenol-d4	9.699	143	79184	31.729	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.228	165	133383	29.988	ng/ul	0.00
31) 4-Chloroaniline-d4	10.781	131	190406	31.215	ng/ul	0.00
46) Dimethylphthalate-d6	13.881	166	523114	36.975	ng/ul	0.00
49) Acenaphthylene-d8	14.157	160	534142	34.663	ng/ul	0.00
54) 4-Nitrophenol-d4	14.734	143	52147	24.666	ng/ul	0.00
60) Fluorene-d10	15.463	176	435622	37.295	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.628	200	73899	28.392	ng/ul	0.00
73) Anthracene-d10	17.339	188	706885	36.302	ng/ul	0.00
81) Pyrene-d10	19.598	212	875126	38.208	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.686	264	847815	38.413	ng/ul	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP111723\  
 Data File : BP018248.D  
 Acq On : 18 Nov 2023 07:36  
 Operator : MA/JU  
 Sample : PB157221BL  
 Misc :  
 ALS Vial : 37 Sample Multiplier: 1

Instrument :  
 BNA\_P  
 ClientSampleId :  
 SBLK221

Quant Time: Nov 19 01:15:13 2023  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\SFAM-EPA-BP111023.MA.M  
 Quant Title : SVOA CALIBRATION  
 QLast Update : Wed Nov 15 18:13:37 2023  
 Response via : Initial Calibration

