

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111820\
 Data File : BP003990.D
 Acq On : 18 Nov 2020 13:01
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 18 14:50:25 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP110320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 12 15:22:47 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	105	0.00
2	1,4-Dioxane	40.000	35.374	11.6	99	0.00
3	Pyridine	40.000	36.171	9.6	100	0.00
4	n-Nitrosodimethylamine	40.000	38.811	3.0	108	0.00
5 S	2-Fluorophenol	80.000	76.023	5.0	105	0.00
6	Aniline	40.000	38.459	3.9	105	0.00
7 S	Phenol-d6	80.000	75.999	5.0	104	0.00
8	2-Chlorophenol	40.000	38.630	3.4	105	0.00
9	Benzaldehyde	40.000	42.356	-5.9	116	0.00
10 C	Phenol	40.000	37.765	5.6	104	0.00
11	bis(2-Chloroethyl)ether	40.000	37.608	6.0	104	0.00
12	1,3-Dichlorobenzene	40.000	37.864	5.3	106	0.00
13 C	1,4-Dichlorobenzene	40.000	38.043	4.9	106	0.00
14	1,2-Dichlorobenzene	40.000	37.990	5.0	105	0.00
15	Benzyl Alcohol	40.000	40.016	-0.0	106	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	37.426	6.4	104	0.00
17	2-Methylphenol	40.000	38.961	2.6	106	0.00
18	Hexachloroethane	40.000	40.139	-0.3	109	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.661	0.8	106	0.00
20	3+4-Methylphenols	40.000	39.637	0.9	107	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	104	0.00
22	Acetophenone	40.000	37.480	6.3	105	0.00
23 S	Nitrobenzene-d5	80.000	74.859	6.4	103	0.00
24	Nitrobenzene	40.000	38.329	4.2	106	0.00
25	Isophorone	40.000	39.622	0.9	108	0.00
26 C	2-Nitrophenol	40.000	45.297	-13.2	119	0.00
27	2,4-Dimethylphenol	40.000	38.338	4.2	107	0.00
28	bis(2-Chloroethoxy)methane	40.000	37.069	7.3	103	0.00
29 C	2,4-Dichlorophenol	40.000	38.550	3.6	105	0.00
30	1,2,4-Trichlorobenzene	40.000	37.811	5.5	106	0.00
31	Naphthalene	40.000	37.157	7.1	105	0.00
32	Benzoic acid	40.000	26.095	34.8#	65	0.02
33	4-Chloroaniline	40.000	37.537	6.2	104	0.00
34 C	Hexachlorobutadiene	40.000	37.914	5.2	107	0.00
35	Caprolactam	40.000	41.608	-4.0	110	0.00
36 C	4-Chloro-3-methylphenol	40.000	38.620	3.5	106	0.00
37	2-Methylnaphthalene	40.000	36.816	8.0	104	0.00
38	1-Methylnaphthalene	40.000	36.630	8.4	104	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	101	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	37.994	5.0	104	0.00
41 P	Hexachlorocyclopentadiene	40.000	35.495	11.3	93	0.00
42 S	2,4,6-Tribromophenol	80.000	76.010	5.0	101	0.00
43 C	2,4,6-Trichlorophenol	40.000	38.756	3.1	103	0.00
44	2,4,5-Trichlorophenol	40.000	38.064	4.8	101	0.00
45 S	2-Fluorobiphenyl	80.000	74.803	6.5	103	0.00
46	1,1'-Biphenyl	40.000	37.532	6.2	104	0.00
47	2-Chloronaphthalene	40.000	37.831	5.4	104	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111820\
 Data File : BP003990.D
 Acq On : 18 Nov 2020 13:01
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 18 14:50:25 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP110320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 12 15:22:47 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	43.392	-8.5	112	0.00
49	Acenaphthylene	40.000	38.209	4.5	104	0.00
50	Dimethylphthalate	40.000	38.173	4.6	105	0.00
51	2,6-Dinitrotoluene	40.000	40.309	-0.8	108	0.00
52 C	Acenaphthene	40.000	37.368	6.6	103	0.00
53	3-Nitroaniline	40.000	39.993	0.0	107	0.00
54 P	2,4-Dinitrophenol	40.000	34.253	14.4	87	0.00
55	Dibenzofuran	40.000	37.179	7.1	103	0.00
56 P	4-Nitrophenol	40.000	35.832	10.4	93	0.02
57	2,4-Dinitrotoluene	40.000	41.367	-3.4	108	0.00
58	Fluorene	40.000	37.240	6.9	104	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	35.999	10.0	96	0.00
60	Diethylphthalate	40.000	38.516	3.7	106	0.00
61	4-Chlorophenyl-phenylether	40.000	37.171	7.1	104	0.00
62	4-Nitroaniline	40.000	41.116	-2.8	109	0.00
63	Azobenzene	40.000	37.884	5.3	103	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	102	0.00
65	4,6-Dinitro-2-methylphenol	40.000	37.941	5.1	105	0.01
66 c	n-Nitrosodiphenylamine	40.000	37.742	5.6	103	0.00
67	4-Bromophenyl-phenylether	40.000	38.262	4.3	104	0.00
68	Hexachlorobenzene	40.000	37.295	6.8	103	0.00
69	Atrazine	40.000	38.840	2.9	103	0.00
70 C	Pentachlorophenol	40.000	37.069	7.3	97	0.01
71	Phenanthrene	40.000	37.081	7.3	103	0.00
72	Anthracene	40.000	37.595	6.0	104	0.00
73	Carbazole	40.000	37.715	5.7	104	0.01
74	Di-n-butylphthalate	40.000	41.285	-3.2	108	0.00
75 C	Fluoranthene	40.000	37.025	7.4	101	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	91	0.00
77	Benzydine	40.000	42.377	-5.9	95	0.00
78	Pyrene	40.000	40.981	-2.5	99	0.01
79 S	Terphenyl-d14	80.000	79.931	0.1	99	0.01
80	Butylbenzylphthalate	40.000	47.338	-18.3	108	0.00
81	Benzo(a)anthracene	40.000	38.427	3.9	94	0.01
82	3,3'-Dichlorobenzidine	40.000	40.118	-0.3	96	0.00
83	Chrysene	40.000	38.099	4.8	94	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	44.723	-11.8	104	0.00
85 c	Di-n-octyl phthalate	40.000	45.708	-14.3	106	0.00
86 I	Perylene-d12	20.000	20.000	0.0	93	0.02
87	Indeno(1,2,3-cd)pyrene	40.000	36.440	8.9	94	0.03
88	Benzo(b)fluoranthene	40.000	35.840	10.4	95	0.01
89	Benzo(k)fluoranthene	40.000	35.471	11.3	91	0.01
90 C	Benzo(a)pyrene	40.000	36.522	8.7	95	0.02
91	Dibenzo(a,h)anthracene	40.000	36.538	8.7	94	0.02
92	Benzo(g,h,i)perylene	40.000	36.580	8.6	95	0.03

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111820\
Data File : BP003990.D
Acq On : 18 Nov 2020 13:01
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Nov 18 14:50:25 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP110320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 12 15:22:47 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0