

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111920\
 Data File : BP004004.D
 Acq On : 19 Nov 2020 14:52
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 19 16:43:28 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP110320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 12 15:22:47 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	108	0.00
2	1,4-Dioxane	40.000	35.080	12.3	101	0.00
3	Pyridine	40.000	35.870	10.3	101	0.00
4	n-Nitrosodimethylamine	40.000	38.346	4.1	109	0.00
5 S	2-Fluorophenol	80.000	75.753	5.3	107	0.00
6	Aniline	40.000	38.023	4.9	106	0.00
7 S	Phenol-d6	80.000	75.540	5.6	106	0.00
8	2-Chlorophenol	40.000	38.635	3.4	108	0.00
9	Benzaldehyde	40.000	41.029	-2.6	116	0.00
10 C	Phenol	40.000	37.867	5.3	107	0.00
11	bis(2-Chloroethyl)ether	40.000	37.359	6.6	105	0.00
12	1,3-Dichlorobenzene	40.000	37.686	5.8	108	0.00
13 C	1,4-Dichlorobenzene	40.000	37.669	5.8	107	0.00
14	1,2-Dichlorobenzene	40.000	37.721	5.7	107	0.00
15	Benzyl Alcohol	40.000	40.342	-0.9	110	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	36.554	8.6	104	0.00
17	2-Methylphenol	40.000	39.156	2.1	109	0.00
18	Hexachloroethane	40.000	39.579	1.1	109	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.183	2.0	107	0.00
20	3+4-Methylphenols	40.000	38.997	2.5	108	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	105	0.00
22	Acetophenone	40.000	37.285	6.8	105	0.00
23 S	Nitrobenzene-d5	80.000	75.475	5.7	105	0.00
24	Nitrobenzene	40.000	38.476	3.8	108	0.00
25	Isophorone	40.000	39.754	0.6	109	0.00
26 C	2-Nitrophenol	40.000	45.816	-14.5	122	0.00
27	2,4-Dimethylphenol	40.000	38.530	3.7	108	0.00
28	bis(2-Chloroethoxy)methane	40.000	37.027	7.4	104	0.00
29 C	2,4-Dichlorophenol	40.000	39.672	0.8	109	0.00
30	1,2,4-Trichlorobenzene	40.000	37.990	5.0	108	0.00
31	Naphthalene	40.000	37.241	6.9	106	0.00
32	Benzoic acid	40.000	29.334	26.7#	78	0.02
33	4-Chloroaniline	40.000	37.312	6.7	104	0.00
34 C	Hexachlorobutadiene	40.000	38.443	3.9	110	0.00
35	Caprolactam	40.000	42.402	-6.0	113	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.016	2.5	108	0.00
37	2-Methylnaphthalene	40.000	37.147	7.1	106	0.00
38	1-Methylnaphthalene	40.000	36.834	7.9	106	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	103	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.084	4.8	106	0.00
41 P	Hexachlorocyclopentadiene	40.000	35.595	11.0	95	0.00
42 S	2,4,6-Tribromophenol	80.000	77.315	3.4	104	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.138	-0.3	108	0.00
44	2,4,5-Trichlorophenol	40.000	38.905	2.7	105	0.00
45 S	2-Fluorobiphenyl	80.000	75.086	6.1	105	0.00
46	1,1'-Biphenyl	40.000	37.599	6.0	106	0.00
47	2-Chloronaphthalene	40.000	37.468	6.3	104	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111920\
 Data File : BP004004.D
 Acq On : 19 Nov 2020 14:52
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 19 16:43:28 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP110320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 12 15:22:47 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	43.299	-8.2	113	0.00
49	Acenaphthylene	40.000	38.277	4.3	106	0.00
50	Dimethylphthalate	40.000	38.094	4.8	106	0.00
51	2,6-Dinitrotoluene	40.000	40.584	-1.5	110	0.00
52 C	Acenaphthene	40.000	37.425	6.4	104	0.00
53	3-Nitroaniline	40.000	40.085	-0.2	108	0.00
54 P	2,4-Dinitrophenol	40.000	36.344	9.1	95	0.00
55	Dibenzofuran	40.000	37.336	6.7	105	0.00
56 P	4-Nitrophenol	40.000	36.425	8.9	95	0.01
57	2,4-Dinitrotoluene	40.000	41.801	-4.5	111	0.00
58	Fluorene	40.000	37.390	6.5	105	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	36.058	9.9	97	0.00
60	Diethylphthalate	40.000	38.699	3.3	108	0.00
61	4-Chlorophenyl-phenylether	40.000	37.219	7.0	106	0.00
62	4-Nitroaniline	40.000	40.499	-1.2	109	0.00
63	Azobenzene	40.000	37.613	6.0	104	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	104	0.00
65	4,6-Dinitro-2-methylphenol	40.000	38.275	4.3	108	0.00
66 c	n-Nitrosodiphenylamine	40.000	37.772	5.6	105	0.00
67	4-Bromophenyl-phenylether	40.000	38.473	3.8	107	0.00
68	Hexachlorobenzene	40.000	37.554	6.1	106	0.00
69	Atrazine	40.000	38.721	3.2	105	0.00
70 C	Pentachlorophenol	40.000	37.803	5.5	101	0.00
71	Phenanthrene	40.000	36.746	8.1	104	0.00
72	Anthracene	40.000	37.516	6.2	105	0.00
73	Carbazole	40.000	37.473	6.3	105	0.00
74	Di-n-butylphthalate	40.000	41.031	-2.6	110	0.00
75 C	Fluoranthene	40.000	36.290	9.3	101	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	92	0.00
77	Benzydine	40.000	35.620	11.0	81	0.00
78	Pyrene	40.000	40.582	-1.5	99	0.00
79 S	Terphenyl-d14	80.000	79.359	0.8	99	0.00
80	Butylbenzylphthalate	40.000	47.179	-17.9	109	0.00
81	Benzo(a)anthracene	40.000	38.525	3.7	95	0.00
82	3,3'-Dichlorobenzidine	40.000	40.433	-1.1	98	0.00
83	Chrysene	40.000	37.936	5.2	94	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	43.832	-9.6	103	0.00
85 c	Di-n-octyl phthalate	40.000	45.421	-13.6	106	0.00
86 I	Perylene-d12	20.000	20.000	0.0	96	0.01
87	Indeno(1,2,3-cd)pyrene	40.000	36.284	9.3	97	0.02
88	Benzo(b)fluoranthene	40.000	35.451	11.4	97	0.00
89	Benzo(k)fluoranthene	40.000	35.359	11.6	93	0.00
90 C	Benzo(a)pyrene	40.000	36.160	9.6	97	0.00
91	Dibenzo(a,h)anthracene	40.000	36.378	9.1	96	0.02
92	Benzo(g,h,i)perylene	40.000	36.698	8.3	98	0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111920\
Data File : BP004004.D
Acq On : 19 Nov 2020 14:52
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Nov 19 16:43:28 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP110320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 12 15:22:47 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0