

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111925\
 Data File : BP026166.D
 Acq On : 19 Nov 2025 20:18
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 20 00:56:20 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 19 01:25:31 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	82	0.00
2	1,4-Dioxane	0.505	0.503	0.4	85	0.00
3	Pyridine	1.461	1.477	-1.1	84	0.00
4	n-Nitrosodimethylamine	0.546	0.549	-0.5	84	0.00
5 S	2-Fluorophenol	1.246	1.254	-0.6	83	0.00
6	Aniline	2.094	2.113	-0.9	82	0.00
7 S	Phenol-d6	1.586	1.618	-2.0	82	0.00
8	2-Chlorophenol	1.418	1.436	-1.3	83	0.00
9	Benzaldehyde	0.831	0.956	-15.0	85	0.00
10 C	Phenol	1.709	1.747	-2.2	82	0.00
11	bis(2-Chloroethyl)ether	1.338	1.361	-1.7	82	0.00
12	1,3-Dichlorobenzene	1.528	1.531	-0.2	83	0.00
13 C	1,4-Dichlorobenzene	1.540	1.547	-0.5	83	0.00
14	1,2-Dichlorobenzene	1.479	1.495	-1.1	83	0.00
15	Benzyl Alcohol	1.162	1.176	-1.2	82	0.00
16	2,2'-oxybis(1-Chloropropane	1.879	1.947	-3.6	84	0.00
17	2-Methylphenol	1.163	1.188	-2.1	82	0.00
18	Hexachloroethane	0.560	0.568	-1.4	83	0.00
19 P	n-Nitroso-di-n-propylamine	0.985	1.022	-3.8	82	0.00
20	3+4-Methylphenols	1.612	1.623	-0.7	82	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	82	-0.01
22	Acetophenone	0.508	0.506	0.4	82	0.00
23 S	Nitrobenzene-d5	0.352	0.355	-0.9	82	0.00
24	Nitrobenzene	0.356	0.359	-0.8	82	0.00
25	Isophorone	0.698	0.696	0.3	81	0.00
26 C	2-Nitrophenol	0.180	0.177	1.7	77	0.00
27	2,4-Dimethylphenol	0.325	0.326	-0.3	82	0.00
28	bis(2-Chloroethoxy)methane	0.437	0.440	-0.7	83	0.00
29 C	2,4-Dichlorophenol	0.325	0.323	0.6	80	0.00
30	1,2,4-Trichlorobenzene	0.331	0.328	0.9	82	0.00
31	Naphthalene	1.081	1.082	-0.1	83	-0.01
32	Benzoic acid	0.274	0.246	10.2	73	-0.01
33	4-Chloroaniline	0.460	0.457	0.7	80	0.00
34 C	Hexachlorobutadiene	0.215	0.211	1.9	81	-0.02
35	Caprolactam	0.116	0.114	1.7	80	0.00
36 C	4-Chloro-3-methylphenol	0.345	0.348	-0.9	81	-0.01
37	2-Methylnaphthalene	0.766	0.765	0.1	82	-0.01
38	1-Methylnaphthalene	0.749	0.752	-0.4	81	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	80	-0.02
40	1,2,4,5-Tetrachlorobenzene	0.605	0.609	-0.7	81	-0.01
41 P	Hexachlorocyclopentadiene	0.396	0.402	-1.5	81	-0.01
42 S	2,4,6-Tribromophenol	0.259	0.260	-0.4	79	0.00
43 C	2,4,6-Trichlorophenol	0.415	0.416	-0.2	79	0.00
44	2,4,5-Trichlorophenol	0.463	0.454	1.9	78	-0.01
45 S	2-Fluorobiphenyl	1.365	1.361	0.3	81	-0.02
46	1,1'-Biphenyl	1.506	1.524	-1.2	82	-0.01
47	2-Chloronaphthalene	1.184	1.185	-0.1	81	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111925\
 Data File : BP026166.D
 Acq On : 19 Nov 2025 20:18
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 20 00:56:20 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 19 01:25:31 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.329	0.337	-2.4	80	-0.01
49	Acenaphthylene	1.953	1.978	-1.3	81	-0.01
50	Dimethylphthalate	1.494	1.506	-0.8	81	-0.01
51	2,6-Dinitrotoluene	0.296	0.306	-3.4	78	-0.01
52 C	Acenaphthene	1.145	1.159	-1.2	81	0.00
53	3-Nitroaniline	0.349	0.350	-0.3	77	-0.02
54 P	2,4-Dinitrophenol	0.179	0.170	5.0	73	-0.01
55	Dibenzofuran	1.773	1.786	-0.7	81	-0.01
56 P	4-Nitrophenol	0.315	0.306	2.9	77	-0.02
57	2,4-Dinitrotoluene	0.442	0.458	-3.6	78	-0.02
58	Fluorene	1.401	1.404	-0.2	81	-0.01
59	2,3,4,6-Tetrachlorophenol	0.393	0.390	0.8	78	-0.02
60	Diethylphthalate	1.505	1.505	0.0	79	-0.01
61	4-Chlorophenyl-phenylether	0.697	0.699	-0.3	81	-0.01
62	4-Nitroaniline	0.362	0.367	-1.4	80	0.00
63	Azobenzene	1.269	1.302	-2.6	81	-0.02
64 I	Phenanthrene-d10	1.000	1.000	0.0	80	0.00
65	4,6-Dinitro-2-methylphenol	0.127	0.123	3.1	74	0.00
66 c	n-Nitrosodiphenylamine	0.619	0.629	-1.6	81	-0.01
67	4-Bromophenyl-phenylether	0.221	0.220	0.5	77	-0.02
68	Hexachlorobenzene	0.257	0.255	0.8	78	-0.02
69	Atrazine	0.233	0.232	0.4	79	0.00
70 C	Pentachlorophenol	0.180	0.176	2.2	77	-0.01
71	Phenanthrene	1.127	1.145	-1.6	82	-0.01
72	Anthracene	1.149	1.147	0.2	79	-0.01
73	Carbazole	1.091	1.101	-0.9	80	0.00
74	Di-n-butylphthalate	1.294	1.294	0.0	77	-0.01
75 C	Fluoranthene	1.369	1.370	-0.1	81	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	83	0.00
77	Benzydine	0.635	0.614	3.3	82	-0.01
78	Pyrene	1.311	1.307	0.3	81	0.00
79 S	Terphenyl-d14	0.981	0.963	1.8	80	-0.01
80	Butylbenzylphthalate	0.578	0.573	0.9	79	-0.01
81	Benzo(a)anthracene	1.374	1.356	1.3	82	0.00
82	3,3'-Dichlorobenzidine	0.500	0.483	3.4	80	-0.02
83	Chrysene	1.295	1.296	-0.1	83	0.00
84	Bis(2-ethylhexyl)phthalate	0.874	0.863	1.3	76	0.00
85 c	Di-n-octyl phthalate	1.529	1.480	3.2	77	-0.01
86 I	Perylene-d12	1.000	1.000	0.0	84	-0.01
87	Indeno(1,2,3-cd)pyrene	1.500	1.526	-1.7	85	-0.03
88	Benzo(b)fluoranthene	1.218	1.234	-1.3	84	-0.02
89	Benzo(k)fluoranthene	1.217	1.211	0.5	84	-0.02
90 C	Benzo(a)pyrene	1.154	1.175	-1.8	85	-0.01
91	Dibenzo(a,h)anthracene	1.228	1.247	-1.5	85	-0.05
92	Benzo(g,h,i)perylene	1.220	1.244	-2.0	86	-0.04

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111925\
Data File : BP026166.D
Acq On : 19 Nov 2025 20:18
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Nov 20 00:56:20 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 19 01:25:31 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0