

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112025\
 Data File : BP026168.D
 Acq On : 20 Nov 2025 10:14
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 20 10:32:59 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 19 01:25:31 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	79	0.00
2	1,4-Dioxane	40.000	41.007	-2.5	84	0.00
3	Pyridine	40.000	40.756	-1.9	82	0.00
4	n-Nitrosodimethylamine	40.000	39.983	0.0	80	0.00
5 S	2-Fluorophenol	80.000	80.597	-0.7	80	0.00
6	Aniline	40.000	40.532	-1.3	79	0.00
7 S	Phenol-d6	80.000	82.159	-2.7	80	-0.01
8	2-Chlorophenol	40.000	40.238	-0.6	79	0.00
9	Benzaldehyde	40.000	44.048	-10.1	79	0.00
10 C	Phenol	40.000	41.015	-2.5	79	-0.01
11	bis(2-Chloroethyl)ether	40.000	40.516	-1.3	79	0.00
12	1,3-Dichlorobenzene	40.000	40.730	-1.8	81	-0.01
13 C	1,4-Dichlorobenzene	40.000	40.590	-1.5	80	0.00
14	1,2-Dichlorobenzene	40.000	40.304	-0.8	80	0.00
15	Benzyl Alcohol	40.000	39.981	0.0	77	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	41.045	-2.6	80	-0.01
17	2-Methylphenol	40.000	40.460	-1.2	78	-0.02
18	Hexachloroethane	40.000	40.209	-0.5	79	-0.01
19 P	n-Nitroso-di-n-propylamine	40.000	40.648	-1.6	77	-0.01
20	3+4-Methylphenols	40.000	39.534	1.2	78	-0.02
21 I	Naphthalene-d8	20.000	20.000	0.0	79	-0.02
22	Acetophenone	40.000	39.444	1.4	78	-0.01
23 S	Nitrobenzene-d5	80.000	80.365	-0.5	78	-0.01
24	Nitrobenzene	40.000	40.088	-0.2	78	-0.01
25	Isophorone	40.000	39.304	1.7	76	-0.01
26 C	2-Nitrophenol	40.000	39.454	1.4	74	0.00
27	2,4-Dimethylphenol	40.000	39.931	0.2	78	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.817	0.5	78	-0.01
29 C	2,4-Dichlorophenol	40.000	40.037	-0.1	77	-0.01
30	1,2,4-Trichlorobenzene	40.000	39.226	1.9	77	0.00
31	Naphthalene	40.000	40.124	-0.3	79	-0.01
32	Benzoic acid	40.000	35.741	10.6	70	-0.02
33	4-Chloroaniline	40.000	40.456	-1.1	77	0.00
34 C	Hexachlorobutadiene	40.000	38.887	2.8	77	-0.02
35	Caprolactam	40.000	39.067	2.3	76	-0.01
36 C	4-Chloro-3-methylphenol	40.000	40.074	-0.2	77	-0.01
37	2-Methylnaphthalene	40.000	39.911	0.2	78	-0.01
38	1-Methylnaphthalene	40.000	40.447	-1.1	78	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	76	-0.02
40	1,2,4,5-Tetrachlorobenzene	40.000	39.639	0.9	77	-0.01
41 P	Hexachlorocyclopentadiene	40.000	39.538	1.2	75	-0.01
42 S	2,4,6-Tribromophenol	80.000	79.142	1.1	74	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.930	0.2	75	0.00
44	2,4,5-Trichlorophenol	40.000	39.967	0.1	76	-0.01
45 S	2-Fluorobiphenyl	80.000	80.454	-0.6	78	-0.02
46	1,1'-Biphenyl	40.000	39.290	1.8	76	0.00
47	2-Chloronaphthalene	40.000	39.923	0.2	77	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112025\
 Data File : BP026168.D
 Acq On : 20 Nov 2025 10:14
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 20 10:32:59 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 19 01:25:31 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.230	-3.1	77	0.00
49	Acenaphthylene	40.000	40.492	-1.2	78	-0.01
50	Dimethylphthalate	40.000	39.840	0.4	76	0.00
51	2,6-Dinitrotoluene	40.000	40.845	-2.1	73	-0.01
52 C	Acenaphthene	40.000	42.920	-7.3	82	0.00
53	3-Nitroaniline	40.000	40.370	-0.9	74	-0.02
54 P	2,4-Dinitrophenol	40.000	36.336	9.2	67	-0.02
55	Dibenzofuran	40.000	40.635	-1.6	78	0.00
56 P	4-Nitrophenol	40.000	38.569	3.6	73	-0.02
57	2,4-Dinitrotoluene	40.000	41.408	-3.5	75	-0.02
58	Fluorene	40.000	40.220	-0.5	78	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.787	0.5	75	-0.01
60	Diethylphthalate	40.000	40.045	-0.1	76	-0.01
61	4-Chlorophenyl-phenylether	40.000	40.226	-0.6	77	0.00
62	4-Nitroaniline	40.000	41.006	-2.5	78	0.00
63	Azobenzene	40.000	41.125	-2.8	77	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	76	-0.01
65	4,6-Dinitro-2-methylphenol	40.000	38.063	4.8	69	-0.01
66 c	n-Nitrosodiphenylamine	40.000	41.372	-3.4	78	-0.02
67	4-Bromophenyl-phenylether	40.000	40.283	-0.7	74	-0.01
68	Hexachlorobenzene	40.000	40.401	-1.0	75	-0.01
69	Atrazine	40.000	39.036	2.4	73	-0.01
70 C	Pentachlorophenol	40.000	39.079	2.3	73	-0.02
71	Phenanthrene	40.000	40.199	-0.5	77	-0.02
72	Anthracene	40.000	40.891	-2.2	77	-0.02
73	Carbazole	40.000	39.628	0.9	75	-0.01
74	Di-n-butylphthalate	40.000	38.494	3.8	70	-0.02
75 C	Fluoranthene	40.000	39.622	0.9	76	-0.01
76 I	Chrysene-d12	20.000	20.000	0.0	77	0.00
77	Benzydine	40.000	37.376	6.6	74	-0.02
78	Pyrene	40.000	40.653	-1.6	77	-0.01
79 S	Terphenyl-d14	80.000	81.143	-1.4	77	-0.02
80	Butylbenzylphthalate	40.000	38.399	4.0	71	0.00
81	Benzo(a)anthracene	40.000	39.831	0.4	77	0.00
82	3,3'-Dichlorobenzidine	40.000	38.758	3.1	75	-0.02
83	Chrysene	40.000	40.192	-0.5	78	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	37.640	5.9	68	0.00
85 c	Di-n-octyl phthalate	40.000	37.056	7.4	69	-0.02
86 I	Perylene-d12	20.000	20.000	0.0	79	-0.02
87	Indeno(1,2,3-cd)pyrene	40.000	40.434	-1.1	79	-0.05
88	Benzo(b)fluoranthene	40.000	40.762	-1.9	79	-0.02
89	Benzo(k)fluoranthene	40.000	39.915	0.2	79	-0.02
90 C	Benzo(a)pyrene	40.000	40.155	-0.4	78	-0.02
91	Dibenzo(a,h)anthracene	40.000	40.534	-1.3	79	-0.06
92	Benzo(g,h,i)perylene	40.000	40.633	-1.6	80	-0.06

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112025\
Data File : BP026168.D
Acq On : 20 Nov 2025 10:14
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Nov 20 10:32:59 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 19 01:25:31 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0