

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112025\
 Data File : BP026168.D
 Acq On : 20 Nov 2025 10:14
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 20 10:32:59 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 19 01:25:31 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	79	0.00
2	1,4-Dioxane	0.505	0.518	-2.6	84	0.00
3	Pyridine	1.461	1.488	-1.8	82	0.00
4	n-Nitrosodimethylamine	0.546	0.546	0.0	80	0.00
5 S	2-Fluorophenol	1.246	1.255	-0.7	80	0.00
6	Aniline	2.094	2.122	-1.3	79	0.00
7 S	Phenol-d6	1.586	1.629	-2.7	80	-0.01
8	2-Chlorophenol	1.418	1.427	-0.6	79	0.00
9	Benzaldehyde	0.831	0.915	-10.1	79	0.00
10 C	Phenol	1.709	1.752	-2.5	79	-0.01
11	bis(2-Chloroethyl)ether	1.338	1.355	-1.3	79	0.00
12	1,3-Dichlorobenzene	1.528	1.556	-1.8	81	-0.01
13 C	1,4-Dichlorobenzene	1.540	1.562	-1.4	80	0.00
14	1,2-Dichlorobenzene	1.479	1.490	-0.7	80	0.00
15	Benzyl Alcohol	1.162	1.162	0.0	77	0.00
16	2,2'-oxybis(1-Chloropropane	1.879	1.928	-2.6	80	-0.01
17	2-Methylphenol	1.163	1.177	-1.2	78	-0.02
18	Hexachloroethane	0.560	0.563	-0.5	79	-0.01
19 P	n-Nitroso-di-n-propylamine	0.985	1.001	-1.6	77	-0.01
20	3+4-Methylphenols	1.612	1.593	1.2	78	-0.02
21 I	Naphthalene-d8	1.000	1.000	0.0	79	-0.02
22	Acetophenone	0.508	0.501	1.4	78	-0.01
23 S	Nitrobenzene-d5	0.352	0.354	-0.6	78	-0.01
24	Nitrobenzene	0.356	0.357	-0.3	78	-0.01
25	Isophorone	0.698	0.686	1.7	76	-0.01
26 C	2-Nitrophenol	0.180	0.178	1.1	74	0.00
27	2,4-Dimethylphenol	0.325	0.324	0.3	78	0.00
28	bis(2-Chloroethoxy)methane	0.437	0.435	0.5	78	-0.01
29 C	2,4-Dichlorophenol	0.325	0.325	0.0	77	-0.01
30	1,2,4-Trichlorobenzene	0.331	0.324	2.1	77	0.00
31	Naphthalene	1.081	1.084	-0.3	79	-0.01
32	Benzoic acid	0.274	0.245	10.6	70	-0.02
33	4-Chloroaniline	0.460	0.465	-1.1	77	0.00
34 C	Hexachlorobutadiene	0.215	0.209	2.8	77	-0.02
35	Caprolactam	0.116	0.114	1.7	76	-0.01
36 C	4-Chloro-3-methylphenol	0.345	0.346	-0.3	77	-0.01
37	2-Methylnaphthalene	0.766	0.765	0.1	78	-0.01
38	1-Methylnaphthalene	0.749	0.757	-1.1	78	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	76	-0.02
40	1,2,4,5-Tetrachlorobenzene	0.605	0.599	1.0	77	-0.01
41 P	Hexachlorocyclopentadiene	0.396	0.391	1.3	75	-0.01
42 S	2,4,6-Tribromophenol	0.259	0.257	0.8	74	0.00
43 C	2,4,6-Trichlorophenol	0.415	0.414	0.2	75	0.00
44	2,4,5-Trichlorophenol	0.463	0.462	0.2	76	-0.01
45 S	2-Fluorobiphenyl	1.365	1.372	-0.5	78	-0.02
46	1,1'-Biphenyl	1.506	1.479	1.8	76	0.00
47	2-Chloronaphthalene	1.184	1.182	0.2	77	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112025\
 Data File : BP026168.D
 Acq On : 20 Nov 2025 10:14
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 20 10:32:59 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 19 01:25:31 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.329	0.339	-3.0	77	0.00
49	Acenaphthylene	1.953	1.977	-1.2	78	-0.01
50	Dimethylphthalate	1.494	1.488	0.4	76	0.00
51	2,6-Dinitrotoluene	0.296	0.302	-2.0	73	-0.01
52 C	Acenaphthene	1.145	1.228	-7.2	82	0.00
53	3-Nitroaniline	0.349	0.352	-0.9	74	-0.02
54 P	2,4-Dinitrophenol	0.179	0.162	9.5	67	-0.02
55	Dibenzofuran	1.773	1.801	-1.6	78	0.00
56 P	4-Nitrophenol	0.315	0.304	3.5	73	-0.02
57	2,4-Dinitrotoluene	0.442	0.458	-3.6	75	-0.02
58	Fluorene	1.401	1.409	-0.6	78	0.00
59	2,3,4,6-Tetrachlorophenol	0.393	0.391	0.5	75	-0.01
60	Diethylphthalate	1.505	1.506	-0.1	76	-0.01
61	4-Chlorophenyl-phenylether	0.697	0.701	-0.6	77	0.00
62	4-Nitroaniline	0.362	0.371	-2.5	78	0.00
63	Azobenzene	1.269	1.304	-2.8	77	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	76	-0.01
65	4,6-Dinitro-2-methylphenol	0.127	0.121	4.7	69	-0.01
66 c	n-Nitrosodiphenylamine	0.619	0.640	-3.4	78	-0.02
67	4-Bromophenyl-phenylether	0.221	0.223	-0.9	74	-0.01
68	Hexachlorobenzene	0.257	0.260	-1.2	75	-0.01
69	Atrazine	0.233	0.228	2.1	73	-0.01
70 C	Pentachlorophenol	0.180	0.176	2.2	73	-0.02
71	Phenanthrene	1.127	1.132	-0.4	77	-0.02
72	Anthracene	1.149	1.175	-2.3	77	-0.02
73	Carbazole	1.091	1.081	0.9	75	-0.01
74	Di-n-butylphthalate	1.294	1.245	3.8	70	-0.02
75 C	Fluoranthene	1.369	1.356	0.9	76	-0.01
76 I	Chrysene-d12	1.000	1.000	0.0	77	0.00
77	Benzydine	0.635	0.593	6.6	74	-0.02
78	Pyrene	1.311	1.333	-1.7	77	-0.01
79 S	Terphenyl-d14	0.981	0.995	-1.4	77	-0.02
80	Butylbenzylphthalate	0.578	0.554	4.2	71	0.00
81	Benzo(a)anthracene	1.374	1.368	0.4	77	0.00
82	3,3'-Dichlorobenzidine	0.500	0.484	3.2	75	-0.02
83	Chrysene	1.295	1.301	-0.5	78	0.00
84	Bis(2-ethylhexyl)phthalate	0.874	0.823	5.8	68	0.00
85 c	Di-n-octyl phthalate	1.529	1.416	7.4	69	-0.02
86 I	Perylene-d12	1.000	1.000	0.0	79	-0.02
87	Indeno(1,2,3-cd)pyrene	1.500	1.516	-1.1	79	-0.05
88	Benzo(b)fluoranthene	1.218	1.241	-1.9	79	-0.02
89	Benzo(k)fluoranthene	1.217	1.214	0.2	79	-0.02
90 C	Benzo(a)pyrene	1.154	1.158	-0.3	78	-0.02
91	Dibenzo(a,h)anthracene	1.228	1.244	-1.3	79	-0.06
92	Benzo(g,h,i)perylene	1.220	1.240	-1.6	80	-0.06

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112025\
Data File : BP026168.D
Acq On : 20 Nov 2025 10:14
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Nov 20 10:32:59 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 19 01:25:31 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0