

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112025\
 Data File : BP026177.D
 Acq On : 20 Nov 2025 17:03
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 20 17:30:26 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 19 01:25:31 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	131	0.00
2	1,4-Dioxane	40.000	42.844	-7.1	145	0.02
3	Pyridine	40.000	39.984	0.0	132	0.02
4	n-Nitrosodimethylamine	40.000	37.670	5.8	124	0.02
5 S	2-Fluorophenol	80.000	80.770	-1.0	132	0.00
6	Aniline	40.000	37.094	7.3	120	0.00
7 S	Phenol-d6	80.000	74.759	6.6	120	-0.01
8	2-Chlorophenol	40.000	38.688	3.3	125	0.00
9	Benzaldehyde	40.000	42.268	-5.7	124	0.00
10 C	Phenol	40.000	37.733	5.7	121	0.00
11	bis(2-Chloroethyl)ether	40.000	38.387	4.0	123	0.00
12	1,3-Dichlorobenzene	40.000	39.899	0.3	131	-0.01
13 C	1,4-Dichlorobenzene	40.000	39.514	1.2	129	0.00
14	1,2-Dichlorobenzene	40.000	39.097	2.3	128	0.00
15	Benzyl Alcohol	40.000	36.086	9.8	115	-0.01
16	2,2'-oxybis(1-Chloropropane	40.000	38.313	4.2	123	-0.01
17	2-Methylphenol	40.000	36.766	8.1	117	-0.01
18	Hexachloroethane	40.000	40.367	-0.9	131	-0.01
19 P	n-Nitroso-di-n-propylamine	40.000	32.663	18.3	103	-0.01
20	3+4-Methylphenols	40.000	35.223	11.9	114	-0.01
21 I	Naphthalene-d8	20.000	20.000	0.0	114	-0.02
22	Acetophenone	40.000	39.427	1.4	113	-0.01
23 S	Nitrobenzene-d5	80.000	81.410	-1.8	115	-0.01
24	Nitrobenzene	40.000	40.854	-2.1	116	-0.01
25	Isophorone	40.000	38.036	4.9	107	-0.01
26 C	2-Nitrophenol	40.000	42.522	-6.3	116	-0.01
27	2,4-Dimethylphenol	40.000	38.314	4.2	108	-0.01
28	bis(2-Chloroethoxy)methane	40.000	39.882	0.3	113	-0.02
29 C	2,4-Dichlorophenol	40.000	39.909	0.2	112	-0.01
30	1,2,4-Trichlorobenzene	40.000	40.931	-2.3	117	0.00
31	Naphthalene	40.000	39.733	0.7	114	-0.02
32	Benzoic acid	40.000	34.465	13.8	98	-0.01
33	4-Chloroaniline	40.000	37.870	5.3	105	-0.01
34 C	Hexachlorobutadiene	40.000	42.129	-5.3	120	-0.02
35	Caprolactam	40.000	32.807	18.0	93	0.00
36 C	4-Chloro-3-methylphenol	40.000	36.232	9.4	101	-0.01
37	2-Methylnaphthalene	40.000	38.336	4.2	109	-0.02
38	1-Methylnaphthalene	40.000	38.373	4.1	108	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	101	-0.02
40	1,2,4,5-Tetrachlorobenzene	40.000	42.196	-5.5	108	-0.02
41 P	Hexachlorocyclopentadiene	40.000	40.105	-0.3	101	-0.02
42 S	2,4,6-Tribromophenol	80.000	75.069	6.2	93	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.741	-4.4	104	0.00
44	2,4,5-Trichlorophenol	40.000	40.301	-0.8	101	-0.01
45 S	2-Fluorobiphenyl	80.000	82.154	-2.7	105	-0.02
46	1,1'-Biphenyl	40.000	40.627	-1.6	103	-0.01
47	2-Chloronaphthalene	40.000	40.761	-1.9	104	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112025\
 Data File : BP026177.D
 Acq On : 20 Nov 2025 17:03
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 20 17:30:26 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 19 01:25:31 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.076	-0.2	98	-0.01
49	Acenaphthylene	40.000	40.315	-0.8	102	-0.01
50	Dimethylphthalate	40.000	38.706	3.2	97	-0.01
51	2,6-Dinitrotoluene	40.000	39.881	0.3	94	-0.01
52 C	Acenaphthene	40.000	39.504	1.2	100	0.00
53	3-Nitroaniline	40.000	37.592	6.0	91	-0.02
54 P	2,4-Dinitrophenol	40.000	36.110	9.7	87	-0.02
55	Dibenzofuran	40.000	39.298	1.8	99	0.00
56 P	4-Nitrophenol	40.000	34.553	13.6	87	-0.02
57	2,4-Dinitrotoluene	40.000	38.620	3.5	92	-0.02
58	Fluorene	40.000	38.159	4.6	97	-0.01
59	2,3,4,6-Tetrachlorophenol	40.000	37.713	5.7	93	-0.02
60	Diethylphthalate	40.000	37.782	5.5	94	-0.01
61	4-Chlorophenyl-phenylether	40.000	38.501	3.7	97	-0.01
62	4-Nitroaniline	40.000	35.657	10.9	89	0.00
63	Azobenzene	40.000	38.469	3.8	95	-0.01
64 I	Phenanthrene-d10	20.000	20.000	0.0	91	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.109	-0.3	87	-0.01
66 c	n-Nitrosodiphenylamine	40.000	41.086	-2.7	93	-0.01
67	4-Bromophenyl-phenylether	40.000	43.357	-8.4	95	0.00
68	Hexachlorobenzene	40.000	41.318	-3.3	91	-0.02
69	Atrazine	40.000	39.100	2.2	87	0.00
70 C	Pentachlorophenol	40.000	40.151	-0.4	90	-0.02
71	Phenanthrene	40.000	40.719	-1.8	93	-0.01
72	Anthracene	40.000	41.094	-2.7	92	-0.01
73	Carbazole	40.000	38.359	4.1	86	0.00
74	Di-n-butylphthalate	40.000	40.282	-0.7	87	-0.01
75 C	Fluoranthene	40.000	38.207	4.5	87	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	85	0.02
77	Benzidine	40.000	31.621	20.9	69	0.00
78	Pyrene	40.000	41.724	-4.3	87	0.00
79 S	Terphenyl-d14	80.000	84.575	-5.7	89	0.00
80	Butylbenzylphthalate	40.000	40.872	-2.2	83	0.00
81	Benzo(a)anthracene	40.000	39.818	0.5	84	0.01
82	3,3'-Dichlorobenzidine	40.000	38.396	4.0	81	0.00
83	Chrysene	40.000	39.931	0.2	85	0.02
84	Bis(2-ethylhexyl)phthalate	40.000	40.921	-2.3	81	0.00
85 c	Di-n-octyl phthalate	40.000	39.525	1.2	81	0.00
86 I	Perylene-d12	20.000	20.000	0.0	84	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	39.331	1.7	82	0.00
88	Benzo(b)fluoranthene	40.000	40.254	-0.6	84	0.00
89	Benzo(k)fluoranthene	40.000	40.454	-1.1	86	0.00
90 C	Benzo(a)pyrene	40.000	39.754	0.6	83	0.00
91	Dibenzo(a,h)anthracene	40.000	39.341	1.6	82	-0.01
92	Benzo(g,h,i)perylene	40.000	39.364	1.6	83	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112025\
Data File : BP026177.D
Acq On : 20 Nov 2025 17:03
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Nov 20 17:30:26 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 19 01:25:31 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0