

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112025\
 Data File : BP026177.D
 Acq On : 20 Nov 2025 17:03
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 20 17:30:26 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 19 01:25:31 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	131	0.00
2	1,4-Dioxane	0.505	0.541	-7.1	145	0.02
3	Pyridine	1.461	1.460	0.1	132	0.02
4	n-Nitrosodimethylamine	0.546	0.514	5.9	124	0.02
5 S	2-Fluorophenol	1.246	1.258	-1.0	132	0.00
6	Aniline	2.094	1.942	7.3	120	0.00
7 S	Phenol-d6	1.586	1.482	6.6	120	-0.01
8	2-Chlorophenol	1.418	1.372	3.2	125	0.00
9	Benzaldehyde	0.831	0.878	-5.7	124	0.00
10 C	Phenol	1.709	1.612	5.7	121	0.00
11	bis(2-Chloroethyl)ether	1.338	1.284	4.0	123	0.00
12	1,3-Dichlorobenzene	1.528	1.524	0.3	131	-0.01
13 C	1,4-Dichlorobenzene	1.540	1.521	1.2	129	0.00
14	1,2-Dichlorobenzene	1.479	1.445	2.3	128	0.00
15	Benzyl Alcohol	1.162	1.048	9.8	115	-0.01
16	2,2'-oxybis(1-Chloropropane	1.879	1.800	4.2	123	-0.01
17	2-Methylphenol	1.163	1.069	8.1	117	-0.01
18	Hexachloroethane	0.560	0.565	-0.9	131	-0.01
19 P	n-Nitroso-di-n-propylamine	0.985	0.805	18.3	103	-0.01
20	3+4-Methylphenols	1.612	1.419	12.0	114	-0.01
21 I	Naphthalene-d8	1.000	1.000	0.0	114	-0.02
22	Acetophenone	0.508	0.501	1.4	113	-0.01
23 S	Nitrobenzene-d5	0.352	0.358	-1.7	115	-0.01
24	Nitrobenzene	0.356	0.364	-2.2	116	-0.01
25	Isophorone	0.698	0.664	4.9	107	-0.01
26 C	2-Nitrophenol	0.180	0.192	-6.7	116	-0.01
27	2,4-Dimethylphenol	0.325	0.311	4.3	108	-0.01
28	bis(2-Chloroethoxy)methane	0.437	0.436	0.2	113	-0.02
29 C	2,4-Dichlorophenol	0.325	0.324	0.3	112	-0.01
30	1,2,4-Trichlorobenzene	0.331	0.338	-2.1	117	0.00
31	Naphthalene	1.081	1.074	0.6	114	-0.02
32	Benzoic acid	0.274	0.236	13.9	98	-0.01
33	4-Chloroaniline	0.460	0.435	5.4	105	-0.01
34 C	Hexachlorobutadiene	0.215	0.227	-5.6	120	-0.02
35	Caprolactam	0.116	0.095	18.1	93	0.00
36 C	4-Chloro-3-methylphenol	0.345	0.313	9.3	101	-0.01
37	2-Methylnaphthalene	0.766	0.734	4.2	109	-0.02
38	1-Methylnaphthalene	0.749	0.719	4.0	108	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	101	-0.02
40	1,2,4,5-Tetrachlorobenzene	0.605	0.638	-5.5	108	-0.02
41 P	Hexachlorocyclopentadiene	0.396	0.397	-0.3	101	-0.02
42 S	2,4,6-Tribromophenol	0.259	0.243	6.2	93	0.00
43 C	2,4,6-Trichlorophenol	0.415	0.433	-4.3	104	0.00
44	2,4,5-Trichlorophenol	0.463	0.466	-0.6	101	-0.01
45 S	2-Fluorobiphenyl	1.365	1.401	-2.6	105	-0.02
46	1,1'-Biphenyl	1.506	1.530	-1.6	103	-0.01
47	2-Chloronaphthalene	1.184	1.207	-1.9	104	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112025\
 Data File : BP026177.D
 Acq On : 20 Nov 2025 17:03
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 20 17:30:26 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 19 01:25:31 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.329	0.330	-0.3	98	-0.01
49	Acenaphthylene	1.953	1.969	-0.8	102	-0.01
50	Dimethylphthalate	1.494	1.445	3.3	97	-0.01
51	2,6-Dinitrotoluene	0.296	0.295	0.3	94	-0.01
52 C	Acenaphthene	1.145	1.131	1.2	100	0.00
53	3-Nitroaniline	0.349	0.328	6.0	91	-0.02
54 P	2,4-Dinitrophenol	0.179	0.161	10.1	87	-0.02
55	Dibenzofuran	1.773	1.742	1.7	99	0.00
56 P	4-Nitrophenol	0.315	0.272	13.7	87	-0.02
57	2,4-Dinitrotoluene	0.442	0.427	3.4	92	-0.02
58	Fluorene	1.401	1.337	4.6	97	-0.01
59	2,3,4,6-Tetrachlorophenol	0.393	0.370	5.9	93	-0.02
60	Diethylphthalate	1.505	1.421	5.6	94	-0.01
61	4-Chlorophenyl-phenylether	0.697	0.671	3.7	97	-0.01
62	4-Nitroaniline	0.362	0.323	10.8	89	0.00
63	Azobenzene	1.269	1.220	3.9	95	-0.01
64 I	Phenanthrene-d10	1.000	1.000	0.0	91	0.00
65	4,6-Dinitro-2-methylphenol	0.127	0.127	0.0	87	-0.01
66 c	n-Nitrosodiphenylamine	0.619	0.635	-2.6	93	-0.01
67	4-Bromophenyl-phenylether	0.221	0.240	-8.6	95	0.00
68	Hexachlorobenzene	0.257	0.265	-3.1	91	-0.02
69	Atrazine	0.233	0.228	2.1	87	0.00
70 C	Pentachlorophenol	0.180	0.180	0.0	90	-0.02
71	Phenanthrene	1.127	1.147	-1.8	93	-0.01
72	Anthracene	1.149	1.181	-2.8	92	-0.01
73	Carbazole	1.091	1.047	4.0	86	0.00
74	Di-n-butylphthalate	1.294	1.303	-0.7	87	-0.01
75 C	Fluoranthene	1.369	1.308	4.5	87	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	85	0.02
77	Benzidine	0.635	0.502	20.9	69	0.00
78	Pyrene	1.311	1.368	-4.3	87	0.00
79 S	Terphenyl-d14	0.981	1.037	-5.7	89	0.00
80	Butylbenzylphthalate	0.578	0.590	-2.1	83	0.00
81	Benzo(a)anthracene	1.374	1.367	0.5	84	0.01
82	3,3'-Dichlorobenzidine	0.500	0.479	4.2	81	0.00
83	Chrysene	1.295	1.292	0.2	85	0.02
84	Bis(2-ethylhexyl)phthalate	0.874	0.894	-2.3	81	0.00
85 c	Di-n-octyl phthalate	1.529	1.510	1.2	81	0.00
86 I	Perylene-d12	1.000	1.000	0.0	84	0.00
87	Indeno(1,2,3-cd)pyrene	1.500	1.475	1.7	82	0.00
88	Benzo(b)fluoranthene	1.218	1.226	-0.7	84	0.00
89	Benzo(k)fluoranthene	1.217	1.230	-1.1	86	0.00
90 C	Benzo(a)pyrene	1.154	1.147	0.6	83	0.00
91	Dibenzo(a,h)anthracene	1.228	1.208	1.6	82	-0.01
92	Benzo(g,h,i)perylene	1.220	1.201	1.6	83	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112025\
Data File : BP026177.D
Acq On : 20 Nov 2025 17:03
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Nov 20 17:30:26 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 19 01:25:31 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0