

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112125\
 Data File : BP026179.D
 Acq On : 21 Nov 2025 10:27
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 21 10:48:26 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 19 01:25:31 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	128	-0.01
2	1,4-Dioxane	40.000	42.473	-6.2	141	0.02
3	Pyridine	40.000	40.004	-0.0	130	0.02
4	n-Nitrosodimethylamine	40.000	37.405	6.5	121	0.02
5 S	2-Fluorophenol	80.000	80.314	-0.4	129	0.00
6	Aniline	40.000	36.549	8.6	116	0.00
7 S	Phenol-d6	80.000	74.386	7.0	117	-0.01
8	2-Chlorophenol	40.000	38.410	4.0	122	0.00
9	Benzaldehyde	40.000	41.579	-3.9	120	0.00
10 C	Phenol	40.000	37.237	6.9	117	0.00
11	bis(2-Chloroethyl)ether	40.000	38.075	4.8	120	0.00
12	1,3-Dichlorobenzene	40.000	40.023	-0.1	129	-0.01
13 C	1,4-Dichlorobenzene	40.000	39.970	0.1	128	0.00
14	1,2-Dichlorobenzene	40.000	39.544	1.1	127	0.00
15	Benzyl Alcohol	40.000	35.178	12.1	110	-0.01
16	2,2'-oxybis(1-Chloropropane	40.000	37.881	5.3	119	0.00
17	2-Methylphenol	40.000	36.527	8.7	114	-0.01
18	Hexachloroethane	40.000	40.786	-2.0	130	-0.01
19 P	n-Nitroso-di-n-propylamine	40.000	32.833	17.9	101	0.00
20	3+4-Methylphenols	40.000	34.982	12.5	111	-0.01
21 I	Naphthalene-d8	20.000	20.000	0.0	113	-0.02
22	Acetophenone	40.000	39.447	1.4	112	-0.01
23 S	Nitrobenzene-d5	80.000	80.851	-1.1	113	-0.01
24	Nitrobenzene	40.000	40.414	-1.0	113	0.00
25	Isophorone	40.000	37.516	6.2	104	-0.02
26 C	2-Nitrophenol	40.000	40.227	-0.6	108	0.00
27	2,4-Dimethylphenol	40.000	37.571	6.1	105	-0.01
28	bis(2-Chloroethoxy)methane	40.000	38.966	2.6	109	-0.01
29 C	2,4-Dichlorophenol	40.000	39.288	1.8	109	-0.02
30	1,2,4-Trichlorobenzene	40.000	40.624	-1.6	115	-0.01
31	Naphthalene	40.000	39.727	0.7	113	-0.01
32	Benzoic acid	40.000	32.357	19.1	90	-0.01
33	4-Chloroaniline	40.000	37.029	7.4	102	0.00
34 C	Hexachlorobutadiene	40.000	42.320	-5.8	119	-0.02
35	Caprolactam	40.000	31.877	20.3	89	0.00
36 C	4-Chloro-3-methylphenol	40.000	35.898	10.3	99	-0.01
37	2-Methylnaphthalene	40.000	38.022	4.9	106	0.00
38	1-Methylnaphthalene	40.000	37.339	6.7	103	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	97	-0.01
40	1,2,4,5-Tetrachlorobenzene	40.000	43.614	-9.0	107	-0.01
41 P	Hexachlorocyclopentadiene	40.000	43.319	-8.3	105	-0.01
42 S	2,4,6-Tribromophenol	80.000	76.475	4.4	91	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.218	-3.0	99	-0.01
44	2,4,5-Trichlorophenol	40.000	41.026	-2.6	99	-0.01
45 S	2-Fluorobiphenyl	80.000	82.531	-3.2	102	-0.01
46	1,1'-Biphenyl	40.000	41.563	-3.9	102	0.00
47	2-Chloronaphthalene	40.000	41.975	-4.9	103	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112125\
 Data File : BP026179.D
 Acq On : 21 Nov 2025 10:27
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 21 10:48:26 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 19 01:25:31 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	39.553	1.1	94	-0.01
49	Acenaphthylene	40.000	40.576	-1.4	99	-0.01
50	Dimethylphthalate	40.000	38.807	3.0	94	0.00
51	2,6-Dinitrotoluene	40.000	40.256	-0.6	92	0.00
52 C	Acenaphthene	40.000	39.883	0.3	97	0.00
53	3-Nitroaniline	40.000	37.793	5.5	88	0.00
54 P	2,4-Dinitrophenol	40.000	33.914	15.2	79	-0.02
55	Dibenzofuran	40.000	39.513	1.2	96	-0.01
56 P	4-Nitrophenol	40.000	34.120	14.7	83	-0.01
57	2,4-Dinitrotoluene	40.000	38.432	3.9	88	-0.01
58	Fluorene	40.000	38.623	3.4	95	-0.01
59	2,3,4,6-Tetrachlorophenol	40.000	38.351	4.1	91	-0.01
60	Diethylphthalate	40.000	37.658	5.9	90	0.00
61	4-Chlorophenyl-phenylether	40.000	38.899	2.8	95	0.00
62	4-Nitroaniline	40.000	35.674	10.8	86	0.00
63	Azobenzene	40.000	39.057	2.4	93	-0.01
64 I	Phenanthrene-d10	20.000	20.000	0.0	90	-0.01
65	4,6-Dinitro-2-methylphenol	40.000	38.365	4.1	82	0.00
66 c	n-Nitrosodiphenylamine	40.000	41.199	-3.0	92	0.00
67	4-Bromophenyl-phenylether	40.000	42.123	-5.3	91	0.00
68	Hexachlorobenzene	40.000	41.286	-3.2	90	-0.01
69	Atrazine	40.000	38.564	3.6	85	0.00
70 C	Pentachlorophenol	40.000	40.422	-1.1	89	-0.02
71	Phenanthrene	40.000	40.237	-0.6	91	-0.01
72	Anthracene	40.000	40.423	-1.1	89	0.00
73	Carbazole	40.000	38.193	4.5	85	0.00
74	Di-n-butylphthalate	40.000	39.389	1.5	84	0.00
75 C	Fluoranthene	40.000	38.178	4.6	86	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	84	0.00
77	Benzidine	40.000	35.644	10.9	77	0.00
78	Pyrene	40.000	41.688	-4.2	86	0.00
79 S	Terphenyl-d14	80.000	82.492	-3.1	86	0.00
80	Butylbenzylphthalate	40.000	40.027	-0.1	80	0.00
81	Benzo(a)anthracene	40.000	39.436	1.4	82	0.00
82	3,3'-Dichlorobenzidine	40.000	37.482	6.3	78	0.00
83	Chrysene	40.000	39.678	0.8	84	0.01
84	Bis(2-ethylhexyl)phthalate	40.000	39.537	1.2	78	0.00
85 c	Di-n-octyl phthalate	40.000	38.653	3.4	78	0.00
86 I	Perylene-d12	20.000	20.000	0.0	82	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.201	-0.5	82	-0.03
88	Benzo(b)fluoranthene	40.000	39.957	0.1	81	-0.02
89	Benzo(k)fluoranthene	40.000	41.130	-2.8	85	-0.01
90 C	Benzo(a)pyrene	40.000	40.378	-0.9	82	0.00
91	Dibenzo(a,h)anthracene	40.000	40.578	-1.4	83	-0.03
92	Benzo(g,h,i)perylene	40.000	40.224	-0.6	82	-0.03

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112125\
Data File : BP026179.D
Acq On : 21 Nov 2025 10:27
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Nov 21 10:48:26 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 19 01:25:31 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0