

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112125\
 Data File : BP026179.D
 Acq On : 21 Nov 2025 10:27
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 21 10:48:26 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 19 01:25:31 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	128	-0.01
2	1,4-Dioxane	0.505	0.537	-6.3	141	0.02
3	Pyridine	1.461	1.461	0.0	130	0.02
4	n-Nitrosodimethylamine	0.546	0.510	6.6	121	0.02
5 S	2-Fluorophenol	1.246	1.251	-0.4	129	0.00
6	Aniline	2.094	1.914	8.6	116	0.00
7 S	Phenol-d6	1.586	1.475	7.0	117	-0.01
8	2-Chlorophenol	1.418	1.362	3.9	122	0.00
9	Benzaldehyde	0.831	0.864	-4.0	120	0.00
10 C	Phenol	1.709	1.591	6.9	117	0.00
11	bis(2-Chloroethyl)ether	1.338	1.274	4.8	120	0.00
12	1,3-Dichlorobenzene	1.528	1.529	-0.1	129	-0.01
13 C	1,4-Dichlorobenzene	1.540	1.539	0.1	128	0.00
14	1,2-Dichlorobenzene	1.479	1.462	1.1	127	0.00
15	Benzyl Alcohol	1.162	1.022	12.0	110	-0.01
16	2,2'-oxybis(1-Chloropropane	1.879	1.780	5.3	119	0.00
17	2-Methylphenol	1.163	1.062	8.7	114	-0.01
18	Hexachloroethane	0.560	0.571	-2.0	130	-0.01
19 P	n-Nitroso-di-n-propylamine	0.985	0.809	17.9	101	0.00
20	3+4-Methylphenols	1.612	1.410	12.5	111	-0.01
21 I	Naphthalene-d8	1.000	1.000	0.0	113	-0.02
22	Acetophenone	0.508	0.501	1.4	112	-0.01
23 S	Nitrobenzene-d5	0.352	0.356	-1.1	113	-0.01
24	Nitrobenzene	0.356	0.360	-1.1	113	0.00
25	Isophorone	0.698	0.655	6.2	104	-0.02
26 C	2-Nitrophenol	0.180	0.181	-0.6	108	0.00
27	2,4-Dimethylphenol	0.325	0.305	6.2	105	-0.01
28	bis(2-Chloroethoxy)methane	0.437	0.426	2.5	109	-0.01
29 C	2,4-Dichlorophenol	0.325	0.319	1.8	109	-0.02
30	1,2,4-Trichlorobenzene	0.331	0.336	-1.5	115	-0.01
31	Naphthalene	1.081	1.074	0.6	113	-0.01
32	Benzoic acid	0.274	0.221	19.3	90	-0.01
33	4-Chloroaniline	0.460	0.426	7.4	102	0.00
34 C	Hexachlorobutadiene	0.215	0.228	-6.0	119	-0.02
35	Caprolactam	0.116	0.093	19.8	89	0.00
36 C	4-Chloro-3-methylphenol	0.345	0.310	10.1	99	-0.01
37	2-Methylnaphthalene	0.766	0.728	5.0	106	0.00
38	1-Methylnaphthalene	0.749	0.699	6.7	103	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	97	-0.01
40	1,2,4,5-Tetrachlorobenzene	0.605	0.659	-8.9	107	-0.01
41 P	Hexachlorocyclopentadiene	0.396	0.429	-8.3	105	-0.01
42 S	2,4,6-Tribromophenol	0.259	0.248	4.2	91	0.00
43 C	2,4,6-Trichlorophenol	0.415	0.427	-2.9	99	-0.01
44	2,4,5-Trichlorophenol	0.463	0.475	-2.6	99	-0.01
45 S	2-Fluorobiphenyl	1.365	1.408	-3.2	102	-0.01
46	1,1'-Biphenyl	1.506	1.565	-3.9	102	0.00
47	2-Chloronaphthalene	1.184	1.243	-5.0	103	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112125\
 Data File : BP026179.D
 Acq On : 21 Nov 2025 10:27
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 21 10:48:26 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 19 01:25:31 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.329	0.326	0.9	94	-0.01
49	Acenaphthylene	1.953	1.982	-1.5	99	-0.01
50	Dimethylphthalate	1.494	1.449	3.0	94	0.00
51	2,6-Dinitrotoluene	0.296	0.298	-0.7	92	0.00
52 C	Acenaphthene	1.145	1.141	0.3	97	0.00
53	3-Nitroaniline	0.349	0.329	5.7	88	0.00
54 P	2,4-Dinitrophenol	0.179	0.151	15.6	79	-0.02
55	Dibenzofuran	1.773	1.751	1.2	96	-0.01
56 P	4-Nitrophenol	0.315	0.269	14.6	83	-0.01
57	2,4-Dinitrotoluene	0.442	0.425	3.8	88	-0.01
58	Fluorene	1.401	1.353	3.4	95	-0.01
59	2,3,4,6-Tetrachlorophenol	0.393	0.376	4.3	91	-0.01
60	Diethylphthalate	1.505	1.417	5.8	90	0.00
61	4-Chlorophenyl-phenylether	0.697	0.678	2.7	95	0.00
62	4-Nitroaniline	0.362	0.323	10.8	86	0.00
63	Azobenzene	1.269	1.239	2.4	93	-0.01
64 I	Phenanthrene-d10	1.000	1.000	0.0	90	-0.01
65	4,6-Dinitro-2-methylphenol	0.127	0.122	3.9	82	0.00
66 c	n-Nitrosodiphenylamine	0.619	0.637	-2.9	92	0.00
67	4-Bromophenyl-phenylether	0.221	0.233	-5.4	91	0.00
68	Hexachlorobenzene	0.257	0.265	-3.1	90	-0.01
69	Atrazine	0.233	0.225	3.4	85	0.00
70 C	Pentachlorophenol	0.180	0.182	-1.1	89	-0.02
71	Phenanthrene	1.127	1.133	-0.5	91	-0.01
72	Anthracene	1.149	1.161	-1.0	89	0.00
73	Carbazole	1.091	1.042	4.5	85	0.00
74	Di-n-butylphthalate	1.294	1.274	1.5	84	0.00
75 C	Fluoranthene	1.369	1.307	4.5	86	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	84	0.00
77	Benzydine	0.635	0.566	10.9	77	0.00
78	Pyrene	1.311	1.367	-4.3	86	0.00
79 S	Terphenyl-d14	0.981	1.011	-3.1	86	0.00
80	Butylbenzylphthalate	0.578	0.578	0.0	80	0.00
81	Benzo(a)anthracene	1.374	1.354	1.5	82	0.00
82	3,3'-Dichlorobenzidine	0.500	0.468	6.4	78	0.00
83	Chrysene	1.295	1.284	0.8	84	0.01
84	Bis(2-ethylhexyl)phthalate	0.874	0.864	1.1	78	0.00
85 c	Di-n-octyl phthalate	1.529	1.477	3.4	78	0.00
86 I	Perylene-d12	1.000	1.000	0.0	82	0.00
87	Indeno(1,2,3-cd)pyrene	1.500	1.507	-0.5	82	-0.03
88	Benzo(b)fluoranthene	1.218	1.217	0.1	81	-0.02
89	Benzo(k)fluoranthene	1.217	1.251	-2.8	85	-0.01
90 C	Benzo(a)pyrene	1.154	1.165	-1.0	82	0.00
91	Dibenzo(a,h)anthracene	1.228	1.246	-1.5	83	-0.03
92	Benzo(g,h,i)perylene	1.220	1.227	-0.6	82	-0.03

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112125\
Data File : BP026179.D
Acq On : 21 Nov 2025 10:27
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Nov 21 10:48:26 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 19 01:25:31 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0