

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112125\
 Data File : BP026188.D
 Acq On : 21 Nov 2025 17:52
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 21 18:11:18 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 19 01:25:31 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	79	-0.01
2	1,4-Dioxane	0.505	0.519	-2.8	84	0.00
3	Pyridine	1.461	1.449	0.8	79	0.00
4	n-Nitrosodimethylamine	0.546	0.528	3.3	77	0.00
5 S	2-Fluorophenol	1.246	1.283	-3.0	81	-0.01
6	Aniline	2.094	1.976	5.6	73	0.00
7 S	Phenol-d6	1.586	1.555	2.0	76	-0.02
8	2-Chlorophenol	1.418	1.404	1.0	77	-0.01
9	Benzaldehyde	0.831	0.911	-9.6	78	-0.01
10 C	Phenol	1.709	1.648	3.6	74	-0.01
11	bis(2-Chloroethyl)ether	1.338	1.280	4.3	74	0.00
12	1,3-Dichlorobenzene	1.528	1.526	0.1	79	-0.02
13 C	1,4-Dichlorobenzene	1.540	1.525	1.0	78	0.00
14	1,2-Dichlorobenzene	1.479	1.461	1.2	78	0.00
15	Benzyl Alcohol	1.162	1.093	5.9	73	-0.01
16	2,2'-oxybis(1-Chloropropane	1.879	1.777	5.4	73	-0.02
17	2-Methylphenol	1.163	1.090	6.3	72	-0.02
18	Hexachloroethane	0.560	0.567	-1.2	79	-0.01
19 P	n-Nitroso-di-n-propylamine	0.985	0.908	7.8	70	-0.01
20	3+4-Methylphenols	1.612	1.477	8.4	72	-0.01
21 I	Naphthalene-d8	1.000	1.000	0.0	71	-0.01
22	Acetophenone	0.508	0.509	-0.2	72	-0.02
23 S	Nitrobenzene-d5	0.352	0.363	-3.1	73	-0.01
24	Nitrobenzene	0.356	0.364	-2.2	72	-0.01
25	Isophorone	0.698	0.671	3.9	67	-0.01
26 C	2-Nitrophenol	0.180	0.188	-4.4	71	-0.02
27	2,4-Dimethylphenol	0.325	0.302	7.1	65	-0.01
28	bis(2-Chloroethoxy)methane	0.437	0.426	2.5	69	-0.01
29 C	2,4-Dichlorophenol	0.325	0.322	0.9	69	0.00
30	1,2,4-Trichlorobenzene	0.331	0.339	-2.4	73	0.00
31	Naphthalene	1.081	1.075	0.6	71	-0.01
32	Benzoic acid	0.274	0.249	9.1	64	-0.02
33	4-Chloroaniline	0.460	0.443	3.7	67	-0.02
34 C	Hexachlorobutadiene	0.215	0.224	-4.2	74	-0.02
35	Caprolactam	0.116	0.104	10.3	63	-0.01
36 C	4-Chloro-3-methylphenol	0.345	0.331	4.1	67	-0.01
37	2-Methylnaphthalene	0.766	0.743	3.0	68	-0.01
38	1-Methylnaphthalene	0.749	0.730	2.5	68	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	66	-0.02
40	1,2,4,5-Tetrachlorobenzene	0.605	0.630	-4.1	69	-0.02
41 P	Hexachlorocyclopentadiene	0.396	0.402	-1.5	67	-0.01
42 S	2,4,6-Tribromophenol	0.259	0.259	0.0	64	-0.02
43 C	2,4,6-Trichlorophenol	0.415	0.422	-1.7	66	0.00
44	2,4,5-Trichlorophenol	0.463	0.459	0.9	64	0.00
45 S	2-Fluorobiphenyl	1.365	1.404	-2.9	69	-0.02
46	1,1'-Biphenyl	1.506	1.501	0.3	66	-0.01
47	2-Chloronaphthalene	1.184	1.193	-0.8	67	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112125\
 Data File : BP026188.D
 Acq On : 21 Nov 2025 17:52
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 21 18:11:18 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 19 01:25:31 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.329	0.330	-0.3	64	-0.01
49	Acenaphthylene	1.953	1.955	-0.1	66	-0.01
50	Dimethylphthalate	1.494	1.430	4.3	63	-0.01
51	2,6-Dinitrotoluene	0.296	0.296	0.0	61	-0.02
52 C	Acenaphthene	1.145	1.141	0.3	66	-0.01
53	3-Nitroaniline	0.349	0.333	4.6	60	-0.02
54 P	2,4-Dinitrophenol	0.179	0.160	10.6	56	-0.02
55	Dibenzofuran	1.773	1.749	1.4	65	-0.01
56 P	4-Nitrophenol	0.315	0.276	12.4	57	-0.02
57	2,4-Dinitrotoluene	0.442	0.430	2.7	60	-0.02
58	Fluorene	1.401	1.384	1.2	66	-0.01
59	2,3,4,6-Tetrachlorophenol	0.393	0.381	3.1	62	-0.01
60	Diethylphthalate	1.505	1.426	5.2	61	-0.02
61	4-Chlorophenyl-phenylether	0.697	0.688	1.3	65	-0.02
62	4-Nitroaniline	0.362	0.338	6.6	61	-0.02
63	Azobenzene	1.269	1.247	1.7	63	-0.02
64 I	Phenanthrene-d10	1.000	1.000	0.0	64	-0.01
65	4,6-Dinitro-2-methylphenol	0.127	0.119	6.3	57	-0.02
66 c	n-Nitrosodiphenylamine	0.619	0.622	-0.5	64	-0.02
67	4-Bromophenyl-phenylether	0.221	0.224	-1.4	63	-0.02
68	Hexachlorobenzene	0.257	0.259	-0.8	63	-0.02
69	Atrazine	0.233	0.230	1.3	62	-0.02
70 C	Pentachlorophenol	0.180	0.178	1.1	62	-0.02
71	Phenanthrene	1.127	1.120	0.6	64	-0.02
72	Anthracene	1.149	1.146	0.3	63	-0.02
73	Carbazole	1.091	1.066	2.3	62	-0.01
74	Di-n-butylphthalate	1.294	1.308	-1.1	62	-0.02
75 C	Fluoranthene	1.369	1.279	6.6	60	-0.01
76 I	Chrysene-d12	1.000	1.000	0.0	65	0.00
77	Benzydine	0.635	0.474	25.4#	50#	-0.01
78	Pyrene	1.311	1.295	1.2	63	0.00
79 S	Terphenyl-d14	0.981	0.991	-1.0	65	0.00
80	Butylbenzylphthalate	0.578	0.571	1.2	61	0.00
81	Benzo(a)anthracene	1.374	1.358	1.2	64	0.00
82	3,3'-Dichlorobenzidine	0.500	0.490	2.0	63	0.00
83	Chrysene	1.295	1.276	1.5	64	0.00
84	Bis(2-ethylhexyl)phthalate	0.874	0.861	1.5	60	0.00
85 c	Di-n-octyl phthalate	1.529	1.517	0.8	62	0.00
86 I	Perylene-d12	1.000	1.000	0.0	68	0.01
87	Indeno(1,2,3-cd)pyrene	1.500	1.492	0.5	67	-0.01
88	Benzo(b)fluoranthene	1.218	1.214	0.3	67	0.00
89	Benzo(k)fluoranthene	1.217	1.192	2.1	67	0.00
90 C	Benzo(a)pyrene	1.154	1.144	0.9	66	0.00
91	Dibenzo(a,h)anthracene	1.228	1.217	0.9	67	0.00
92	Benzo(g,h,i)perylene	1.220	1.196	2.0	66	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112125\
Data File : BP026188.D
Acq On : 21 Nov 2025 17:52
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Nov 21 18:11:18 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 19 01:25:31 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range SPCC's out = 0 CCC's out = 0