

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112223\
 Data File : BP018279.D
 Acq On : 22 Nov 2023 17:30
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020687

Quant Time: Nov 22 23:18:19 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP112223.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Nov 22 14:09:55 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.869	152	325604	20.000	ng/ul	-0.01
20) Naphthalene-d8	10.675	136	1419653	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.510	164	891534	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.263	188	1715707	20.000	ng/ul	0.00
79) Chrysene-d12	21.369	240	959212	20.000	ng/ul	0.00
88) Perylene-d12	23.792	264	999417	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.270	96	69667	8.014	ng/uL	0.00
4) Pyridine-d5	3.681	84	498168	20.852	ng/ul	0.00
7) Phenol-d5	7.028	99	615714	20.691	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.205	67	377691	21.226	ng/ul	-0.01
11) 2-Chlorophenol-d4	7.399	132	492256	19.703	ng/ul	0.00
15) 4-Methylphenol-d8	8.575	113	508256	21.165	ng/ul	-0.01
21) Nitrobenzene-d5	9.028	128	237620	20.607	ng/ul	-0.01
24) 2-Nitrophenol-d4	9.752	143	219336	20.596	ng/ul	-0.01
28) 2,4-Dichlorophenol-d3	10.293	165	512394	19.944	ng/ul	0.00
31) 4-Chloroaniline-d4	10.810	131	744561	21.487	ng/ul	0.00
46) Dimethylphthalate-d6	13.928	166	1558717	19.649	ng/ul	0.00
49) Acenaphthylene-d8	14.204	160	1814347	20.072	ng/ul	0.00
54) 4-Nitrophenol-d4	14.692	143	205013	18.720	ng/ul	0.00
60) Fluorene-d10	15.504	176	1298226	19.566	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.616	200	127335	17.434	ng/ul	0.00
73) Anthracene-d10	17.363	188	1823260	19.882	ng/ul	0.00
81) Pyrene-d10	19.598	212	1687738	22.108	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.633	264	1137334	19.629	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.305	88	74756	8.019	ng/uL	95
5) Pyridine	3.699	79	501756	20.844	ng/ul	99
6) Benzaldehyde	7.005	77	377999	24.005	ng/ul	100
8) Phenol	7.052	94	609641	20.886	ng/ul	100
10) Bis(2-Chloroethyl)ether	7.299	93	512611	21.083	ng/ul	98
12) 2-Chlorophenol	7.428	128	495650	19.774	ng/ul	98
13) 2-Methylphenol	8.310	108	478229	21.034	ng/ul	98
14) 2,2'-oxybis(1-Chloropr...	8.405	45	747851	21.638	ng/ul	99
16) Acetophenone	8.693	105	803426	21.864	ng/ul	99
17) N-Nitroso-di-n-propyla...	8.687	70	407014	19.828	ng/ul	99
18) 4-Methylphenol	8.640	108	515904	21.249	ng/ul	99
19) Hexachloroethane	8.952	117	207182	19.793	ng/ul	97
22) Nitrobenzene	9.075	77	549603	20.345	ng/ul	98
23) Isophorone	9.599	82	1179809	19.595	ng/ul	99
25) 2-Nitrophenol	9.787	139	246899	20.606	ng/ul	97
26) 2,4-Dimethylphenol	9.852	107	536272	19.935	ng/ul	98
27) Bis(2-Chloroethoxy)met...	10.093	93	724511	19.924	ng/ul	99
29) 2,4-Dichlorophenol	10.316	162	491371	19.966	ng/ul	99
30) Naphthalene	10.728	128	1705454	19.863	ng/ul	99
32) 4-Chloroaniline	10.834	127	703238	21.507	ng/ul	100
33) Hexachlorobutadiene	11.016	225	329673	19.641	ng/ul	99
34) Caprolactam	11.587	113	141845	20.001	ng/ul	99
35) 4-Chloro-3-methylphenol	11.957	107	512432	19.475	ng/ul	97
36) 2-Methylnaphthalene	12.340	142	1187862	19.548	ng/ul	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112223\
 Data File : BP018279.D
 Acq On : 22 Nov 2023 17:30
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SSTD020687


Quant Time: Nov 22 23:18:19 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP112223.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Nov 22 14:09:55 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.557	142	1202865	19.665	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.704	216	637307	20.240	ng/ul	98
40) Hexachlorocyclopentadiene	12.687	237	354253	19.865	ng/ul	98
41) 2,4,6-Trichlorophenol	12.940	196	388398	20.540	ng/ul	97
42) 2,4,5-Trichlorophenol	13.004	196	416476	20.368	ng/ul	99
43) 1,1'-Biphenyl	13.351	154	1575058	20.277	ng/ul	100
44) 2-Chloronaphthalene	13.387	162	1234846	20.085	ng/ul	100
45) 2-Nitroaniline	13.587	65	271919	20.555	ng/ul	99
47) Dimethylphthalate	13.975	163	1533568	19.758	ng/ul	99
48) 2,6-Dinitrotoluene	14.087	165	263794	20.158	ng/ul	97
50) Acenaphthylene	14.234	152	2023015	20.078	ng/ul	99
51) 3-Nitroaniline	14.410	138	267874	20.602	ng/ul	98
52) Acenaphthene	14.575	153	1308012	19.873	ng/ul	99
53) 2,4-Dinitrophenol	14.616	184	74042	16.218	ng/ul	99
55) 4-Nitrophenol	14.710	109	159396	19.646	ng/ul	100
56) Dibenzofuran	14.910	168	1792594	19.921	ng/ul	100
57) 2,4-Dinitrotoluene	14.869	165	355790	20.119	ng/ul	99
58) 2,3,4,6-Tetrachlorophenol	15.134	232	317474	19.770	ng/ul	99
59) Diethylphthalate	15.345	149	1500901	19.249	ng/ul	99
61) Fluorene	15.563	166	1452838	19.798	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.563	204	732558	19.659	ng/ul	98
63) 4-Nitroaniline	15.575	138	245911	20.040	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.628	198	146073	18.033	ng/ul	94
67) N-Nitrosodiphenylamine	15.775	169	1207663	20.757	ng/ul	99
68) 4-Bromophenyl-phenylether	16.457	248	445895	20.584	ng/ul	99
69) Hexachlorobenzene	16.563	284	488848	20.186	ng/ul	96
70) Atrazine	16.728	200	373571	21.032	ng/ul	99
71) Pentachlorophenol	16.904	266	231872	18.635	ng/ul	97
72) Phenanthrene	17.304	178	2117920	20.295	ng/ul	99
74) Anthracene	17.398	178	2125129	20.174	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.310	216	637002	21.600	ng/ul	99
76) Pentachlorobenzene	14.828	250	609826	21.288	ng/ul	98
77) Carbazole	17.663	167	1759353	20.615	ng/ul	99
78) Di-n-butylphthalate	18.239	149	2285639	19.255	ng/ul	100
80) Fluoranthene	19.275	202	2063578	22.627	ng/ul	99
82) Pyrene	19.627	202	2053520	22.361	ng/ul	98
83) Butylbenzylphthalate	20.527	149	647050	19.039	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.286	252	444222	19.148	ng/ul	99
85) Benzo(a)anthracene	21.351	228	1560173	19.958	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.310	149	914266	17.930	ng/ul	99
87) Chrysene	21.404	228	1444036	20.299	ng/ul	99
89) Di-n-octyl phthalate	22.257	149	1292045	17.560	ng/ul	100
90) Benzo(b)fluoranthene	23.051	252	1359369	18.726	ng/ul	99
91) Benzo(k)fluoranthene	23.104	252	1407678	19.479	ng/ul	100
93) Benzo(a)pyrene	23.686	252	1329637	19.917	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.345	276	1711105	22.167	ng/ul#	95
95) Dibenzo(a,h)anthracene	26.380	278	1411879	22.342	ng/ul	99
96) Benzo(g,h,i)perylene	27.121	276	1419580	22.707	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Nov 22 23:18:19 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP112223.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Nov 22 14:09:55 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 8.019 ng/uL
RT: 3.305 min Scan# 31
Delta R.T. -0.006 min
Lab File: BP018279.D
Acq: 22 Nov 2023 17:30




Instrument :
BNA_P
ClientSampleId :
SSTD020687


Tgt Ion: 88 Resp: 74756
Ion Ratio Lower Upper
88 100
43 29.5 26.9 40.3
58 71.3 59.6 89.4



#5
Pyridine
Concen: 20.844 ng/uL
RT: 3.699 min Scan# 104
Delta R.T. -0.012 min
Lab File: BP018279.D
Acq: 22 Nov 2023 17:30

Tgt Ion: 79 Resp: 501756
Ion Ratio Lower Upper
79 100
52 73.5 57.7 86.5
51 35.3 27.8 41.8









#6
Benzaldehyde
Concen: 24.005 ng/ul
RT: 7.005 min Scan# 60
Delta R.T. -0.012 min
Lab File: BP018279.D
Acq: 22 Nov 2023 17:30




Instrument :
BNA_P
ClientSampleId :
SSTD020687


Tgt Ion: 77 Resp: 377999
Ion Ratio Lower Upper
77 100
105 102.2 81.5 122.3
106 98.0 77.9 116.9



#8
Phenol
Concen: 20.886 ng/ul
RT: 7.052 min Scan# 674
Delta R.T. -0.012 min
Lab File: BP018279.D
Acq: 22 Nov 2023 17:30

Tgt Ion: 94 Resp: 609641
Ion Ratio Lower Upper
94 100
65 26.4 21.2 31.8
66 34.8 28.0 42.0









#10
Bis(2-Chloroethyl)ether
Concen: 21.083 ng/ul
RT: 7.299 min Scan# 716
Delta R.T. -0.012 min
Lab File: BP018279.D
Acq: 22 Nov 2023 17:30




Instrument :
BNA_P
ClientSampleId :
SSTD020687


Tgt Ion: 93 Resp: 512611
Ion Ratio Lower Upper
93 100
63 74.8 58.2 87.4
95 32.8 25.6 38.4



#12
2-Chlorophenol
Concen: 19.774 ng/ul
RT: 7.428 min Scan# 738
Delta R.T. -0.012 min
Lab File: BP018279.D
Acq: 22 Nov 2023 17:30

Tgt Ion:128 Resp: 495650
Ion Ratio Lower Upper
128 100
64 45.0 34.8 52.2
130 32.5 25.5 38.3





#13

2-Methylphenol

Concen: 21.034 ng/ul

RT: 8.310 min Scan# 890

Delta R.T. -0.012 min

Lab File: BP018279.D

Acq: 22 Nov 2023 17:30

Instrument :

BNA_P

ClientSampleId :





SSTD020687

Tgt Ion:108 Resp: 478229

Ion Ratio Lower Upper

108 100

107 90.3 70.9 106.3



#14

2,2'-oxybis(1-Chloropropane)

Concen: 21.638 ng/ul

RT: 8.405 min Scan# 904

Delta R.T. -0.012 min

Lab File: BP018279.D

Acq: 22 Nov 2023 17:30




Tgt Ion: 45 Resp: 747851


Ion Ratio Lower Upper

45 100

77 14.0 11.4 17.2

79 10.4 8.4 12.6









#16
Acetophenone
Concen: 21.864 ng/ul
RT: 8.693 min Scan# 91
Delta R.T. -0.012 min
Lab File: BP018279.D
Acq: 22 Nov 2023 17:30




Instrument :
BNA_P
ClientSampleId :
SSTD020687


Tgt Ion:105 Resp: 803426
Ion Ratio Lower Upper
105 100
77 75.1 59.5 89.3
51 26.1 20.5 30.7



#17
N-Nitroso-di-n-propylamine
Concen: 19.828 ng/ul
RT: 8.687 min Scan# 952
Delta R.T. -0.012 min
Lab File: BP018279.D
Acq: 22 Nov 2023 17:30

Tgt Ion: 70 Resp: 407014
Ion Ratio Lower Upper
70 100
42 57.4 46.8 70.2
101 9.9 8.4 12.6
130 23.3 18.5 27.7









#18
4-Methylphenol
Concen: 21.249 ng/ul
RT: 8.640 min Scan# 946
Delta R.T. -0.012 min
Lab File: BP018279.D
Acq: 22 Nov 2023 17:30




Instrument :
BNA_P
ClientSampleId :
SSTD020687


Tgt Ion:108 Resp: 515904
Ion Ratio Lower Upper
108 100
107 112.9 89.5 134.3



#19
Hexachloroethane
Concen: 19.793 ng/ul
RT: 8.952 min Scan# 997
Delta R.T. -0.012 min
Lab File: BP018279.D
Acq: 22 Nov 2023 17:30

Tgt Ion:117 Resp: 207182
Ion Ratio Lower Upper
117 100
201 103.4 85.6 128.4
199 65.6 54.2 81.4









#22
Nitrobenzene
Concen: 20.345 ng/ul
RT: 9.075 min Scan# 1107
Delta R.T. -0.006 min
Lab File: BP018279.D
Acq: 22 Nov 2023 17:30




Instrument :
BNA_P
ClientSampleId :
SSTD020687


Tgt Ion: 77 Resp: 549603
Ion Ratio Lower Upper
77 100
123 48.9 37.7 56.5
65 13.3 10.5 15.7



#23
Isophorone
Concen: 19.595 ng/ul
RT: 9.599 min Scan# 1107
Delta R.T. -0.012 min
Lab File: BP018279.D
Acq: 22 Nov 2023 17:30

Tgt Ion: 82 Resp: 1179809
Ion Ratio Lower Upper
82 100
95 7.0 5.8 8.6
138 18.1 14.9 22.3









#25
2-Nitrophenol
Concen: 20.606 ng/ul
RT: 9.787 min Scan# 1139
Delta R.T. -0.006 min
Lab File: BP018279.D
Acq: 22 Nov 2023 17:30




Instrument :
BNA_P
ClientSampleId :
SSTD020687


Tgt Ion:139 Resp: 246899
Ion Ratio Lower Upper
139 100
65 44.4 38.3 57.5
109 29.5 23.4 35.0



#26
2,4-Dimethylphenol
Concen: 19.935 ng/ul
RT: 9.852 min Scan# 1150
Delta R.T. -0.006 min
Lab File: BP018279.D
Acq: 22 Nov 2023 17:30

Tgt Ion:107 Resp: 536272
Ion Ratio Lower Upper
107 100
121 52.0 40.0 60.0
122 88.8 70.4 105.6









#27
Bis(2-Chloroethoxy)methane
Concen: 19.924 ng/ul
RT: 10.093 min Scan# 1193
Delta R.T. -0.012 min
Lab File: BP018279.D
Acq: 22 Nov 2023 17:30




Instrument :
BNA_P
ClientSampleId :
SSTD020687


Tgt Ion: 93 Resp: 724511
Ion Ratio Lower Upper
93 100
95 32.0 26.0 39.0
123 11.0 9.1 13.7



#29
2,4-Dichlorophenol
Concen: 19.966 ng/ul
RT: 10.316 min Scan# 1229
Delta R.T. -0.006 min
Lab File: BP018279.D
Acq: 22 Nov 2023 17:30

Tgt Ion:162 Resp: 491371
Ion Ratio Lower Upper
162 100
164 65.4 53.0 79.6
98 34.5 27.7 41.5









#30
Naphthalene
Concen: 19.863 ng/ul
RT: 10.728 min Scan# 1299
Delta R.T. -0.006 min
Lab File: BP018279.D
Acq: 22 Nov 2023 17:30




Instrument :
BNA_P
ClientSampleId :
SSTD020687


Tgt Ion:128 Resp: 1705454
Ion Ratio Lower Upper
128 100
129 10.7 8.9 13.3
127 13.1 10.4 15.6



#32
4-Chloroaniline
Concen: 21.507 ng/ul
RT: 10.834 min Scan# 1317
Delta R.T. -0.006 min
Lab File: BP018279.D
Acq: 22 Nov 2023 17:30

Tgt Ion:127 Resp: 703238
Ion Ratio Lower Upper
127 100
129 32.3 25.8 38.8





#33

Hexachlorobutadiene

Concen: 19.641 ng/ul

RT: 11.016 min Scan# 11

Delta R.T. -0.006 min

Lab File: BP018279.D

Acq: 22 Nov 2023 17:30

Instrument :

BNA_P

ClientSampleId :

SSTD020687





Tgt Ion:225 Resp: 329673

Ion Ratio Lower Upper

225 100

223 63.7 50.2 75.4

227 64.0 50.4 75.6



#34

Caprolactam

Concen: 20.001 ng/ul

RT: 11.587 min Scan# 1445

Delta R.T. -0.006 min

Lab File: BP018279.D

Acq: 22 Nov 2023 17:30




Tgt Ion:113 Resp: 141845


Ion Ratio Lower Upper

113 100

55 160.3 129.0 193.4

56 117.2 94.8 142.2









#35
4-Chloro-3-methylphenol
Concen: 19.475 ng/ul
RT: 11.957 min Scan# 1508
Delta R.T. -0.000 min
Lab File: BP018279.D
Acq: 22 Nov 2023 17:30




Instrument :
BNA_P
ClientSampleId :
SSTD020687


Tgt Ion:107 Resp: 512432
Ion Ratio Lower Upper
107 100
144 27.4 21.6 32.4
142 85.6 65.4 98.0



#36
2-Methylnaphthalene
Concen: 19.548 ng/ul
RT: 12.340 min Scan# 1573
Delta R.T. -0.000 min
Lab File: BP018279.D
Acq: 22 Nov 2023 17:30

Tgt Ion:142 Resp: 1187862
Ion Ratio Lower Upper
142 100
141 89.4 71.0 106.4









#37
1-Methylnaphthalene
Concen: 19.665 ng/ul
RT: 12.557 min Scan# 1610
Delta R.T. -0.006 min
Lab File: BP018279.D
Acq: 22 Nov 2023 17:30




Instrument :
BNA_P
ClientSampleId :
SSTD020687


Tgt Ion:142 Resp: 1202865
Ion Ratio Lower Upper
142 100
141 91.8 73.9 110.9
116 3.6 3.0 4.4



#39
1,2,4,5-Tetrachlorobenzene
Concen: 20.240 ng/ul
RT: 12.704 min Scan# 1635
Delta R.T. -0.000 min
Lab File: BP018279.D
Acq: 22 Nov 2023 17:30

Tgt Ion:216 Resp: 637307
Ion Ratio Lower Upper
216 100
214 77.5 63.3 94.9
179 20.5 17.3 25.9
108 15.7 13.4 20.0









#40
Hexachlorocyclopentadiene
Concen: 19.865 ng/ul
RT: 12.687 min Scan# 1632
Delta R.T. -0.006 min
Lab File: BP018279.D
Acq: 22 Nov 2023 17:30




Instrument :
BNA_P
ClientSampleId :
SSTD020687


Tgt Ion:237 Resp: 354253
Ion Ratio Lower Upper
237 100
235 62.0 50.7 76.1
272 12.7 10.3 15.5



#41
2,4,6-Trichlorophenol
Concen: 20.540 ng/ul
RT: 12.940 min Scan# 1675
Delta R.T. -0.006 min
Lab File: BP018279.D
Acq: 22 Nov 2023 17:30

Tgt Ion:196 Resp: 388398
Ion Ratio Lower Upper
196 100
198 93.7 78.0 117.0
200 30.1 24.6 37.0





#42

2,4,5-Trichlorophenol

Concen: 20.368 ng/ul

RT: 13.004 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP018279.D

Acq: 22 Nov 2023 17:30

Instrument :

BNA_P

ClientSampleId :

SSTD020687





Tgt Ion:196 Resp: 416476

Ion Ratio Lower Upper

196 100

198 95.2 76.7 115.1

200 31.7 24.6 36.8



#43

1,1'-Biphenyl

Concen: 20.277 ng/ul

RT: 13.351 min Scan# 1745

Delta R.T. -0.000 min

Lab File: BP018279.D

Acq: 22 Nov 2023 17:30




Tgt Ion:154 Resp: 1575058


Ion Ratio Lower Upper

154 100

153 40.3 32.2 48.2

76 12.3 9.9 14.9









#44
2-Chloronaphthalene
Concen: 20.085 ng/ul
RT: 13.387 min Scan# 1751
Delta R.T. -0.000 min
Lab File: BP018279.D
Acq: 22 Nov 2023 17:30




Instrument :
BNA_P
ClientSampleId :
SSTD020687


Tgt Ion:162 Resp: 1234846
Ion Ratio Lower Upper
162 100
164 32.9 26.2 39.4
127 38.2 30.6 45.8



#45
2-Nitroaniline
Concen: 20.555 ng/ul
RT: 13.587 min Scan# 1785
Delta R.T. -0.000 min
Lab File: BP018279.D
Acq: 22 Nov 2023 17:30

Tgt Ion: 65 Resp: 271919
Ion Ratio Lower Upper
65 100
92 70.1 57.4 86.2
138 114.1 90.6 136.0









#47
Dimethylphthalate
Concen: 19.758 ng/ul
RT: 13.975 min Scan# 1851
Delta R.T. -0.006 min
Lab File: BP018279.D
Acq: 22 Nov 2023 17:30




Instrument :
BNA_P
ClientSampleId :
SSTD020687


Tgt Ion:163 Resp: 1533568
Ion Ratio Lower Upper
163 100
194 4.1 3.1 4.7
164 10.1 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 20.158 ng/ul
RT: 14.087 min Scan# 1870
Delta R.T. -0.000 min
Lab File: BP018279.D
Acq: 22 Nov 2023 17:30

Tgt Ion:165 Resp: 263794
Ion Ratio Lower Upper
165 100
89 54.1 45.8 68.6
121 21.9 17.8 26.8









#50
Acenaphthylene
Concen: 20.078 ng/ul
RT: 14.234 min Scan# 1895
Delta R.T. -0.000 min
Lab File: BP018279.D
Acq: 22 Nov 2023 17:30




Instrument :
BNA_P
ClientSampleId :
SSTD020687


Tgt Ion:152 Resp: 2023015
Ion Ratio Lower Upper
152 100
151 19.7 15.7 23.5
153 13.1 10.8 16.2



#51
3-Nitroaniline
Concen: 20.602 ng/ul
RT: 14.410 min Scan# 1925
Delta R.T. -0.000 min
Lab File: BP018279.D
Acq: 22 Nov 2023 17:30

Tgt Ion:138 Resp: 267874
Ion Ratio Lower Upper
138 100
108 15.5 13.2 19.8
92 110.7 90.2 135.4









#52
Acenaphthene
Concen: 19.873 ng/ul
RT: 14.575 min Scan# 1953
Delta R.T. -0.000 min
Lab File: BP018279.D
Acq: 22 Nov 2023 17:30




Instrument :
BNA_P
ClientSampleId :
SSTD020687


Tgt Ion:153 Resp: 1308012
Ion Ratio Lower Upper
153 100
152 47.4 37.8 56.8
154 89.7 71.2 106.8



#53
2,4-Dinitrophenol
Concen: 16.218 ng/ul
RT: 14.616 min Scan# 1960
Delta R.T. -0.000 min
Lab File: BP018279.D
Acq: 22 Nov 2023 17:30

Tgt Ion:184 Resp: 74042
Ion Ratio Lower Upper
184 100
63 51.4 41.8 62.6
154 58.4 47.3 70.9









#55
4-Nitrophenol
Concen: 19.646 ng/ul
RT: 14.710 min Scan# 1976
Delta R.T. -0.000 min
Lab File: BP018279.D
Acq: 22 Nov 2023 17:30




Instrument :
BNA_P
ClientSampleId :
SSTD020687


Tgt Ion:109 Resp: 159396
Ion Ratio Lower Upper
109 100
139 142.0 113.0 169.4
65 124.2 99.6 149.4



#56
Dibenzofuran
Concen: 19.921 ng/ul
RT: 14.910 min Scan# 2010
Delta R.T. -0.000 min
Lab File: BP018279.D
Acq: 22 Nov 2023 17:30

Tgt Ion:168 Resp: 1792594
Ion Ratio Lower Upper
168 100
139 36.3 29.0 43.4





#57

2,4-Dinitrotoluene

Concen: 20.119 ng/ul

RT: 14.869 min Scan# 2003

Delta R.T. -0.000 min

Lab File: BP018279.D

Acq: 22 Nov 2023 17:30

Instrument :

BNA_P

ClientSampleId :

SSTD020687





Tgt Ion:165 Resp: 355790

Ion Ratio Lower Upper

165 100

63 39.5 32.1 48.1

182 2.6 2.0 3.0



#58

2,3,4,6-Tetrachlorophenol

Concen: 19.770 ng/ul

RT: 15.134 min Scan# 2048

Delta R.T. -0.000 min

Lab File: BP018279.D

Acq: 22 Nov 2023 17:30

Tgt Ion:232 Resp: 317474




Ion Ratio Lower Upper


232 100

131 43.9 34.1 51.1

130 2.1 1.8 2.6

166 30.6 24.8 37.2









#59
Diethylphthalate
Concen: 19.249 ng/ul
RT: 15.345 min Scan# 2084
Delta R.T. -0.000 min
Lab File: BP018279.D
Acq: 22 Nov 2023 17:30




Instrument :
BNA_P
ClientSampleId :
SSTD020687


Tgt Ion:149 Resp: 1500901
Ion Ratio Lower Upper
149 100
177 22.2 18.2 27.2
150 12.4 10.2 15.2



#61
Fluorene
Concen: 19.798 ng/ul
RT: 15.563 min Scan# 2121
Delta R.T. -0.000 min
Lab File: BP018279.D
Acq: 22 Nov 2023 17:30

Tgt Ion:166 Resp: 1452838
Ion Ratio Lower Upper
166 100
165 98.3 77.5 116.3
167 13.3 10.6 16.0









#62
4-Chlorophenyl-phenylether
Concen: 19.659 ng/ul
RT: 15.563 min Scan# 2121
Delta R.T. -0.000 min
Lab File: BP018279.D
Acq: 22 Nov 2023 17:30




Instrument :
BNA_P
ClientSampleId :
SSTD020687


Tgt Ion:204 Resp: 732558
Ion Ratio Lower Upper
204 100
206 33.7 26.2 39.2
141 57.1 44.8 67.2



#63
4-Nitroaniline
Concen: 20.040 ng/ul
RT: 15.575 min Scan# 2123
Delta R.T. -0.000 min
Lab File: BP018279.D
Acq: 22 Nov 2023 17:30

Tgt Ion:138 Resp: 245911
Ion Ratio Lower Upper
138 100
92 50.2 39.3 58.9
108 77.8 61.8 92.6









#66
4,6-Dinitro-2-methylphenol
Concen: 18.033 ng/ul
RT: 15.628 min Scan# 2132
Delta R.T. -0.006 min
Lab File: BP018279.D
Acq: 22 Nov 2023 17:30




Instrument :
BNA_P
ClientSampleId :
SSTD020687


Tgt Ion:198 Resp: 146073
Ion Ratio Lower Upper
198 100
182 4.3 3.0 4.4
77 24.5 17.3 25.9



#67
N-Nitrosodiphenylamine
Concen: 20.757 ng/ul
RT: 15.775 min Scan# 2157
Delta R.T. -0.000 min
Lab File: BP018279.D
Acq: 22 Nov 2023 17:30

Tgt Ion:169 Resp: 1207663
Ion Ratio Lower Upper
169 100
168 66.2 53.1 79.7
167 36.7 28.8 43.2









#68
4-Bromophenyl-phenylether
Concen: 20.584 ng/ul
RT: 16.457 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP018279.D
Acq: 22 Nov 2023 17:30




Instrument :
BNA_P
ClientSampleId :
SSTD020687


Tgt Ion:248 Resp: 445895
Ion Ratio Lower Upper
248 100
250 98.7 77.4 116.2
141 70.0 56.0 84.0



#69
Hexachlorobenzene
Concen: 20.186 ng/ul
RT: 16.563 min Scan# 2291
Delta R.T. 0.006 min
Lab File: BP018279.D
Acq: 22 Nov 2023 17:30

Tgt Ion:284 Resp: 488848
Ion Ratio Lower Upper
284 100
142 33.2 28.9 43.3
249 31.1 26.0 39.0





#70

Atrazine

Concen: 21.032 ng/ul

RT: 16.728 min Scan# 21

Delta R.T. -0.000 min

Lab File: BP018279.D

Acq: 22 Nov 2023 17:30

Instrument :

BNA_P

ClientSampleId :

SSTD020687





Tgt Ion:200 Resp: 373571

Ion Ratio Lower Upper

200 100

173 27.3 21.1 31.7

215 47.6 38.3 57.5



#71

Pentachlorophenol

Concen: 18.635 ng/ul

RT: 16.904 min Scan# 2349

Delta R.T. -0.000 min

Lab File: BP018279.D

Acq: 22 Nov 2023 17:30




Tgt Ion:266 Resp: 231872


Ion Ratio Lower Upper

266 100

264 63.6 50.1 75.1

268 65.8 50.1 75.1





#72

Phenanthrene

Concen: 20.295 ng/ul

RT: 17.304 min Scan# 2417

Delta R.T. -0.000 min

Lab File: BP018279.D

Acq: 22 Nov 2023 17:30

Instrument :

BNA_P

ClientSampleId :

SSTD020687





Tgt Ion:178 Resp: 2117920

Ion Ratio Lower Upper

178 100

179 15.5 12.4 18.6

176 19.4 15.1 22.7



#74

Anthracene

Concen: 20.174 ng/ul

RT: 17.398 min Scan# 2433

Delta R.T. -0.000 min

Lab File: BP018279.D

Acq: 22 Nov 2023 17:30




Tgt Ion:178 Resp: 2125129


Ion Ratio Lower Upper

178 100

179 15.3 12.4 18.6

176 18.8 15.2 22.8





#75

1,2,3,4-Tetrachlorobenzene

Concen: 21.600 ng/uL

RT: 13.310 min Scan# 1738

Delta R.T. -0.000 min

Lab File: BP018279.D

Acq: 22 Nov 2023 17:30

Instrument :

BNA_P

ClientSampleId :

SSTD020687





Tgt Ion:216 Resp: 637002

Ion Ratio Lower Upper

216 100

214 78.6 62.0 93.0

218 47.8 38.6 57.8



#76

Pentachlorobenzene

Concen: 21.288 ng/uL

RT: 14.828 min Scan# 1996

Delta R.T. -0.000 min

Lab File: BP018279.D

Acq: 22 Nov 2023 17:30




Tgt Ion:250 Resp: 609826


Ion Ratio Lower Upper

250 100

248 63.7 51.0 76.6

252 61.7 51.5 77.3





#77

Carbazole

Concen: 20.615 ng/ul

RT: 17.663 min Scan# 2478

Delta R.T. -0.000 min

Lab File: BP018279.D

Acq: 22 Nov 2023 17:30

Instrument :

BNA_P

ClientSampleId :

SSTD020687





Tgt Ion:167 Resp: 1759353

Ion Ratio Lower Upper

167 100

166 21.1 17.5 26.3

139 12.6 10.0 15.0



#78

Di-n-butylphthalate

Concen: 19.255 ng/ul

RT: 18.239 min Scan# 2576

Delta R.T. -0.000 min

Lab File: BP018279.D

Acq: 22 Nov 2023 17:30




Tgt Ion:149 Resp: 2285639


Ion Ratio Lower Upper

149 100

150 9.1 7.4 11.2

104 4.6 3.8 5.6





#80

Fluoranthene

Concen: 22.627 ng/ul

RT: 19.275 min Scan# 21

Delta R.T. -0.000 min

Lab File: BP018279.D

Acq: 22 Nov 2023 17:30

Instrument :

BNA_P

ClientSampleId :

SSTD020687





Tgt Ion:202 Resp: 2063578

Ion Ratio Lower Upper

202 100

101 11.3 8.6 13.0

100 8.5 6.7 10.1



#82

Pyrene

Concen: 22.361 ng/ul

RT: 19.627 min Scan# 2812

Delta R.T. -0.000 min

Lab File: BP018279.D

Acq: 22 Nov 2023 17:30




Tgt Ion:202 Resp: 2053520


Ion Ratio Lower Upper

202 100

101 13.7 10.2 15.4

100 11.0 8.4 12.6





#83

Butylbenzylphthalate

Concen: 19.039 ng/ul

RT: 20.527 min Scan# 2964

Delta R.T. 0.006 min

Lab File: BP018279.D

Acq: 22 Nov 2023 17:30

Instrument :

BNA_P

ClientSampleId :

SSTD020687





Tgt Ion:149 Resp: 647050

Ion Ratio Lower Upper

149 100

91 65.1 53.7 80.5

206 21.3 16.4 24.6



#84

3,3'-Dichlorobenzidine

Concen: 19.148 ng/ul

RT: 21.286 min Scan# 3094

Delta R.T. -0.000 min

Lab File: BP018279.D

Acq: 22 Nov 2023 17:30




Tgt Ion:252 Resp: 444222


Ion Ratio Lower Upper

252 100

254 64.3 51.1 76.7

126 11.5 8.6 12.8





#85

Benzo(a)anthracene

Concen: 19.958 ng/ul

RT: 21.351 min Scan# 3105

Delta R.T. 0.006 min

Lab File: BP018279.D

Acq: 22 Nov 2023 17:30

Instrument :

BNA_P

ClientSampleId :

SSTD020687





Tgt Ion:228 Resp: 1560173

Ion Ratio Lower Upper

228 100

229 19.4 16.2 24.2

226 27.3 22.2 33.2



#86

Bis(2-ethylhexyl)phthalate

Concen: 17.930 ng/ul

RT: 21.310 min Scan# 3098

Delta R.T. 0.006 min

Lab File: BP018279.D

Acq: 22 Nov 2023 17:30




Tgt Ion:149 Resp: 914266


Ion Ratio Lower Upper

149 100

167 27.2 22.1 33.1

279 4.4 3.4 5.0





#87

Chrysene

Concen: 20.299 ng/ul

RT: 21.404 min Scan# 3114

Delta R.T. 0.006 min

Lab File: BP018279.D

Acq: 22 Nov 2023 17:30

Instrument :

BNA_P

ClientSampleId :

SSTD020687





Tgt Ion:228 Resp: 1444036

Ion Ratio Lower Upper

228 100

226 30.1 24.2 36.4

229 19.7 16.1 24.1



#89

Di-n-octyl phthalate

Concen: 17.560 ng/ul




RT: 22.257 min Scan# 3259


Delta R.T. 0.006 min

Lab File: BP018279.D

Acq: 22 Nov 2023 17:30

Tgt Ion:149 Resp: 1292045





#90

Benzo(b)fluoranthene

Concen: 18.726 ng/ul

RT: 23.051 min Scan# 31

Delta R.T. 0.006 min

Lab File: BP018279.D

Acq: 22 Nov 2023 17:30

Instrument :

BNA_P

ClientSampleId :

SSTD020687





Tgt Ion:252 Resp: 1359369

Ion Ratio Lower Upper

252 100

253 21.5 17.7 26.5

125 10.5 8.4 12.6



#91

Benzo(k)fluoranthene

Concen: 19.479 ng/ul

RT: 23.104 min Scan# 3403

Delta R.T. 0.006 min

Lab File: BP018279.D

Acq: 22 Nov 2023 17:30




Tgt Ion:252 Resp: 1407678


Ion Ratio Lower Upper

252 100

253 22.0 17.7 26.5

125 10.1 8.0 12.0





#93

Benzo(a)pyrene

Concen: 19.917 ng/ul

RT: 23.686 min Scan# 3501

Delta R.T. 0.006 min

Lab File: BP018279.D

Acq: 22 Nov 2023 17:30

Instrument :

BNA_P

ClientSampleId :

SSTD020687





Tgt Ion:252 Resp: 1329637

Ion Ratio Lower Upper

252 100

253 21.9 17.7 26.5

125 11.2 8.8 13.2



#94

Indeno(1,2,3-cd)pyrene

Concen: 22.167 ng/ul

RT: 26.345 min Scan# 3954

Delta R.T. 0.018 min

Lab File: BP018279.D

Acq: 22 Nov 2023 17:30




Tgt Ion:276 Resp: 1711105


Ion Ratio Lower Upper

276 100

138 26.7 17.5 26.3#

277 23.9 19.1 28.7





#95

Dibenzo(a,h)anthracene

Concen: 22.342 ng/ul

RT: 26.380 min Scan# 3958

Delta R.T. 0.012 min

Lab File: BP018279.D

Acq: 22 Nov 2023 17:30

Instrument :

BNA_P

ClientSampleId :

SSTD020687





Tgt Ion:278 Resp: 1411879

Ion Ratio Lower Upper

278 100

139 17.1 13.3 19.9

279 24.2 19.0 28.4



#96

Benzo(g,h,i)perylene

Concen: 22.707 ng/ul

RT: 27.121 min Scan# 4086

Delta R.T. 0.012 min

Lab File: BP018279.D

Acq: 22 Nov 2023 17:30

Tgt Ion:276 Resp: 1419580

Ion Ratio Lower Upper

276 100

138 22.2 17.7 26.5

277 23.8 18.6 28.0

