

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112223\
 Data File : BP018318.D
 Acq On : 23 Nov 2023 16:41
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020689

Quant Time: Nov 23 23:53:06 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP112223.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Nov 22 14:09:55 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.869	152	422823	20.000	ng/u1	-0.01
20) Naphthalene-d8	10.675	136	1786689	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.504	164	1098376	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.251	188	2051744	20.000	ng/u1	-0.01
79) Chrysene-d12	21.357	240	1259331	20.000	ng/u1	0.00
88) Perylene-d12	23.780	264	1478152	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.275	96	89223	7.904	ng/uL	0.00
4) Pyridine-d5	3.687	84	617592	19.907	ng/u1	0.00
7) Phenol-d5	7.028	99	770752	19.946	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.205	67	460404	19.925	ng/u1	-0.01
11) 2-Chlorophenol-d4	7.399	132	628800	19.381	ng/u1	0.00
15) 4-Methylphenol-d8	8.581	113	626234	20.082	ng/u1	0.00
21) Nitrobenzene-d5	9.034	128	303797	20.934	ng/u1	0.00
24) 2-Nitrophenol-d4	9.751	143	285958	21.336	ng/u1	-0.01
28) 2,4-Dichlorophenol-d3	10.287	165	652561	20.181	ng/u1	-0.01
31) 4-Chloroaniline-d4	10.810	131	916322	21.012	ng/u1	0.00
46) Dimethylphthalate-d6	13.922	166	1896039	19.400	ng/u1	-0.01
49) Acenaphthylene-d8	14.198	160	2213076	19.873	ng/u1	0.00
54) 4-Nitrophenol-d4	14.692	143	253620	18.797	ng/u1	0.00
60) Fluorene-d10	15.498	176	1595678	19.520	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.610	200	167583	19.187	ng/u1	0.00
73) Anthracene-d10	17.351	188	2192681	19.994	ng/u1	-0.01
81) Pyrene-d10	19.592	212	2012967	20.084	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.621	264	1673470	19.528	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.311	88	93786	7.748	ng/uL#	90
5) Pyridine	3.711	79	618564	19.788	ng/u1	95
6) Benzaldehyde	7.010	77	458203	22.408	ng/u1	98
8) Phenol	7.058	94	758463	20.010	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.299	93	640297	20.279	ng/u1	99
12) 2-Chlorophenol	7.434	128	631889	19.413	ng/u1	98
13) 2-Methylphenol	8.316	108	596167	20.193	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.405	45	858441	19.127	ng/u1	97
16) Acetophenone	8.699	105	987919	20.703	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.687	70	481669	18.070	ng/u1	95
18) 4-Methylphenol	8.646	108	651639	20.669	ng/u1	97
19) Hexachloroethane	8.952	117	261046	19.204	ng/u1	98
22) Nitrobenzene	9.075	77	680856	20.026	ng/u1	98
23) Isophorone	9.604	82	1395615	18.417	ng/u1	98
25) 2-Nitrophenol	9.787	139	323115	21.427	ng/u1	95
26) 2,4-Dimethylphenol	9.846	107	657477	19.420	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.093	93	889731	19.442	ng/u1	99
29) 2,4-Dichlorophenol	10.316	162	627348	20.255	ng/u1	98
30) Naphthalene	10.722	128	2148465	19.882	ng/u1	100
32) 4-Chloroaniline	10.834	127	868640	21.108	ng/u1	99
33) Hexachlorobutadiene	11.010	225	432514	20.474	ng/u1	98
34) Caprolactam	11.593	113	167534	18.770	ng/u1	92
35) 4-Chloro-3-methylphenol	11.951	107	628165	18.969	ng/u1	97
36) 2-Methylnaphthalene	12.334	142	1488597	19.464	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112223\
 Data File : BP018318.D
 Acq On : 23 Nov 2023 16:41
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020689

Quant Time: Nov 23 23:53:06 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP112223.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Nov 22 14:09:55 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.551	142	1507084	19.577	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.698	216	826096	21.295	ng/ul	98
40) Hexachlorocyclopentadiene	12.681	237	439927	20.023	ng/ul	99
41) 2,4,6-Trichlorophenol	12.940	196	488756	20.980	ng/ul	98
42) 2,4,5-Trichlorophenol	13.004	196	534229	21.207	ng/ul	99
43) 1,1'-Biphenyl	13.345	154	1959747	20.478	ng/ul	99
44) 2-Chloronaphthalene	13.381	162	1551071	20.478	ng/ul	99
45) 2-Nitroaniline	13.581	65	328820	20.176	ng/ul	92
47) Dimethylphthalate	13.969	163	1877477	19.633	ng/ul	99
48) 2,6-Dinitrotoluene	14.081	165	336715	20.885	ng/ul	95
50) Acenaphthylene	14.228	152	2487871	20.042	ng/ul	99
51) 3-Nitroaniline	14.404	138	336475	21.005	ng/ul	94
52) Acenaphthene	14.569	153	1616258	19.932	ng/ul	99
53) 2,4-Dinitrophenol	14.610	184	99773	17.738	ng/ul	98
55) 4-Nitrophenol	14.704	109	184879	18.496	ng/ul	97
56) Dibenzofuran	14.904	168	2218745	20.014	ng/ul	97
57) 2,4-Dinitrotoluene	14.863	165	442249	20.299	ng/ul	94
58) 2,3,4,6-Tetrachlorophenol	15.122	232	399551	20.196	ng/ul	97
59) Diethylphthalate	15.334	149	1785040	18.582	ng/ul	99
61) Fluorene	15.551	166	1755685	19.420	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.551	204	911355	19.852	ng/ul	99
63) 4-Nitroaniline	15.569	138	307619	20.348	ng/ul	95
66) 4,6-Dinitro-2-methylph...	15.622	198	189038	19.515	ng/ul	98
67) N-Nitrosodiphenylamine	15.763	169	1452808	20.881	ng/ul	99
68) 4-Bromophenyl-phenylether	16.445	248	526952	20.342	ng/ul	99
69) Hexachlorobenzene	16.551	284	614890	21.232	ng/ul	95
70) Atrazine	16.722	200	434990	20.479	ng/ul	100
71) Pentachlorophenol	16.898	266	292695	19.670	ng/ul	99
72) Phenanthrene	17.292	178	2486350	19.923	ng/ul	99
74) Anthracene	17.386	178	2527497	20.064	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.304	216	824153	23.369	ng/ul	99
76) Pentachlorobenzene	14.822	250	777126	22.685	ng/ul	98
77) Carbazole	17.657	167	2055599	20.142	ng/ul	99
78) Di-n-butylphthalate	18.222	149	2505212	17.648	ng/ul	100
80) Fluoranthene	19.263	202	2416902	20.185	ng/ul	100
82) Pyrene	19.616	202	2415413	20.033	ng/ul	100
83) Butylbenzylphthalate	20.510	149	715841	16.043	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.280	252	583996	19.173	ng/ul	98
85) Benzo(a)anthracene	21.339	228	2027115	19.751	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.292	149	1057263	15.793	ng/ul	100
87) Chrysene	21.392	228	1872359	20.048	ng/ul	99
89) Di-n-octyl phthalate	22.239	149	1575885	14.481	ng/ul	100
90) Benzo(b)fluoranthene	23.039	252	2011515	18.735	ng/ul	99
91) Benzo(k)fluoranthene	23.086	252	2019843	18.898	ng/ul	99
93) Benzo(a)pyrene	23.674	252	1963563	19.887	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.327	276	2543645	22.280	ng/ul	99
95) Dibenzo(a,h)anthracene	26.368	278	2110735	22.583	ng/ul	98
96) Benzo(g,h,i)perylene	27.109	276	2098506	22.696	ng/ul	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112223\
 Data File : BP018318.D
 Acq On : 23 Nov 2023 16:41
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020689

Quant Time: Nov 23 23:53:06 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP112223.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Nov 22 14:09:55 2023
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 7.748 ng/uL
RT: 3.311 min Scan# 31
Delta R.T. -0.000 min
Lab File: BP018318.D
Acq: 23 Nov 2023 16:41




Instrument :
BNA_P
ClientSampleId :
SSTD020689


Tgt Ion: 88 Resp: 93786
Ion Ratio Lower Upper
88 100
43 25.2 26.9 40.3#
58 67.4 59.6 89.4



#5
Pyridine
Concen: 19.788 ng/uL
RT: 3.711 min Scan# 106
Delta R.T. -0.000 min
Lab File: BP018318.D
Acq: 23 Nov 2023 16:41

Tgt Ion: 79 Resp: 618564
Ion Ratio Lower Upper
79 100
52 67.4 57.7 86.5
51 32.3 27.8 41.8









#6
Benzaldehyde
Concen: 22.408 ng/ul
RT: 7.010 min Scan# 60
Delta R.T. -0.006 min
Lab File: BP018318.D
Acq: 23 Nov 2023 16:41




Instrument :
BNA_P
ClientSampleId :
SSTD020689


Tgt Ion: 77 Resp: 458203
Ion Ratio Lower Upper
77 100
105 103.9 81.5 122.3
106 98.8 77.9 116.9



#8
Phenol
Concen: 20.010 ng/ul
RT: 7.058 min Scan# 675
Delta R.T. -0.006 min
Lab File: BP018318.D
Acq: 23 Nov 2023 16:41

Tgt Ion: 94 Resp: 758463
Ion Ratio Lower Upper
94 100
65 25.8 21.2 31.8
66 33.7 28.0 42.0









#10
Bis(2-Chloroethyl)ether
Concen: 20.279 ng/ul
RT: 7.299 min Scan# 716
Delta R.T. -0.012 min
Lab File: BP018318.D
Acq: 23 Nov 2023 16:41




Instrument :
BNA_P
ClientSampleId :
SSTD020689


Tgt Ion: 93 Resp: 640297
Ion Ratio Lower Upper
93 100
63 72.4 58.2 87.4
95 33.1 25.6 38.4



#12
2-Chlorophenol
Concen: 19.413 ng/ul
RT: 7.434 min Scan# 739
Delta R.T. -0.006 min
Lab File: BP018318.D
Acq: 23 Nov 2023 16:41

Tgt Ion: 128 Resp: 631889
Ion Ratio Lower Upper
128 100
64 41.3 34.8 52.2
130 32.4 25.5 38.3





#13

2-Methylphenol

Concen: 20.193 ng/ul

RT: 8.316 min Scan# 888

Delta R.T. -0.006 min

Lab File: BP018318.D

Acq: 23 Nov 2023 16:41

Instrument :

BNA_P

ClientSampleId :





SSTD020689

Tgt Ion:108 Resp: 596167

Ion Ratio Lower Upper

108 100

107 88.9 70.9 106.3



#14

2,2'-oxybis(1-Chloropropane)

Concen: 19.127 ng/ul

RT: 8.405 min Scan# 904

Delta R.T. -0.012 min

Lab File: BP018318.D

Acq: 23 Nov 2023 16:41




Tgt Ion: 45 Resp: 858441


Ion Ratio Lower Upper

45 100

77 15.7 11.4 17.2

79 11.3 8.4 12.6









#16
Acetophenone
Concen: 20.703 ng/ul
RT: 8.699 min Scan# 91
Delta R.T. -0.006 min
Lab File: BP018318.D
Acq: 23 Nov 2023 16:41




Instrument :
BNA_P
ClientSampleId :
SSTD020689


Tgt Ion:105 Resp: 987919
Ion Ratio Lower Upper
105 100
77 72.5 59.5 89.3
51 23.1 20.5 30.7



#17
N-Nitroso-di-n-propylamine
Concen: 18.070 ng/ul
RT: 8.687 min Scan# 952
Delta R.T. -0.012 min
Lab File: BP018318.D
Acq: 23 Nov 2023 16:41

Tgt Ion: 70 Resp: 481669
Ion Ratio Lower Upper
70 100
42 53.5 46.8 70.2
101 10.4 8.4 12.6
130 23.9 18.5 27.7









#18
4-Methylphenol
Concen: 20.669 ng/ul
RT: 8.646 min Scan# 944
Delta R.T. -0.006 min
Lab File: BP018318.D
Acq: 23 Nov 2023 16:41




Instrument :
BNA_P
ClientSampleId :
SSTD020689


Tgt Ion:108 Resp: 651639
Ion Ratio Lower Upper
108 100
107 108.4 89.5 134.3



#19
Hexachloroethane
Concen: 19.204 ng/ul
RT: 8.952 min Scan# 997
Delta R.T. -0.012 min
Lab File: BP018318.D
Acq: 23 Nov 2023 16:41

Tgt Ion:117 Resp: 261046
Ion Ratio Lower Upper
117 100
201 105.9 85.6 128.4
199 65.9 54.2 81.4









#22
Nitrobenzene
Concen: 20.026 ng/ul
RT: 9.075 min Scan# 1108
Delta R.T. -0.006 min
Lab File: BP018318.D
Acq: 23 Nov 2023 16:41




Instrument :
BNA_P
ClientSampleId :
SSTD020689


Tgt Ion: 77 Resp: 680856
Ion Ratio Lower Upper
77 100
123 48.4 37.7 56.5
65 13.1 10.5 15.7



#23
Isophorone
Concen: 18.417 ng/ul
RT: 9.604 min Scan# 1108
Delta R.T. -0.006 min
Lab File: BP018318.D
Acq: 23 Nov 2023 16:41

Tgt Ion: 82 Resp: 1395615
Ion Ratio Lower Upper
82 100
95 6.7 5.8 8.6
138 19.4 14.9 22.3









#25
2-Nitrophenol
Concen: 21.427 ng/ul
RT: 9.787 min Scan# 1139
Delta R.T. -0.006 min
Lab File: BP018318.D
Acq: 23 Nov 2023 16:41




Instrument :
BNA_P
ClientSampleId :
SSTD020689


Tgt Ion:139 Resp: 323115
Ion Ratio Lower Upper
139 100
65 43.6 38.3 57.5
109 27.3 23.4 35.0



#26
2,4-Dimethylphenol
Concen: 19.420 ng/ul
RT: 9.846 min Scan# 1149
Delta R.T. -0.012 min
Lab File: BP018318.D
Acq: 23 Nov 2023 16:41

Tgt Ion:107 Resp: 657477
Ion Ratio Lower Upper
107 100
121 52.4 40.0 60.0
122 89.0 70.4 105.6





#27

Bis(2-Chloroethoxy)methane

Concen: 19.442 ng/ul

RT: 10.093 min Scan# 1191

Delta R.T. -0.012 min

Lab File: BP018318.D

Acq: 23 Nov 2023 16:41

Instrument :

BNA_P

ClientSampleId :

SSTD020689





Tgt Ion: 93 Resp: 889731

Ion Ratio Lower Upper

93 100

95 31.9 26.0 39.0

123 12.0 9.1 13.7



#29

2,4-Dichlorophenol

Concen: 20.255 ng/ul

RT: 10.316 min Scan# 1229

Delta R.T. -0.006 min

Lab File: BP018318.D

Acq: 23 Nov 2023 16:41




Tgt Ion:162 Resp: 627348


Ion Ratio Lower Upper

162 100

164 64.8 53.0 79.6

98 33.0 27.7 41.5









#30
Naphthalene
Concen: 19.882 ng/ul
RT: 10.722 min Scan# 1299
Delta R.T. -0.012 min
Lab File: BP018318.D
Acq: 23 Nov 2023 16:41




Instrument :
BNA_P
ClientSampleId :
SSTD020689


Tgt Ion:128 Resp: 2148465
Ion Ratio Lower Upper
128 100
129 11.1 8.9 13.3
127 13.0 10.4 15.6



#32
4-Chloroaniline
Concen: 21.108 ng/ul
RT: 10.834 min Scan# 1317
Delta R.T. -0.006 min
Lab File: BP018318.D
Acq: 23 Nov 2023 16:41

Tgt Ion:127 Resp: 868640
Ion Ratio Lower Upper
127 100
129 32.9 25.8 38.8





#33

Hexachlorobutadiene

Concen: 20.474 ng/ul

RT: 11.010 min Scan# 11

Delta R.T. -0.012 min

Lab File: BP018318.D

Acq: 23 Nov 2023 16:41

Instrument :

BNA_P

ClientSampleId :

SSTD020689





Tgt Ion:225 Resp: 432514

Ion Ratio Lower Upper

225 100

223 62.4 50.2 75.4

227 64.9 50.4 75.6



#34

Caprolactam

Concen: 18.770 ng/ul

RT: 11.593 min Scan# 1446

Delta R.T. -0.000 min

Lab File: BP018318.D

Acq: 23 Nov 2023 16:41




Tgt Ion:113 Resp: 167534


Ion Ratio Lower Upper

113 100

55 147.7 129.0 193.4

56 111.8 94.8 142.2









#35
4-Chloro-3-methylphenol
Concen: 18.969 ng/ul
RT: 11.951 min Scan# 1507
Delta R.T. -0.006 min
Lab File: BP018318.D
Acq: 23 Nov 2023 16:41




Instrument :
BNA_P
ClientSampleId :
SSTD020689


Tgt Ion:107 Resp: 628165
Ion Ratio Lower Upper
107 100
144 27.5 21.6 32.4
142 84.8 65.4 98.0



#36
2-Methylnaphthalene
Concen: 19.464 ng/ul
RT: 12.334 min Scan# 1572
Delta R.T. -0.006 min
Lab File: BP018318.D
Acq: 23 Nov 2023 16:41

Tgt Ion:142 Resp: 1488597
Ion Ratio Lower Upper
142 100
141 89.0 71.0 106.4





#37

1-Methylnaphthalene

Concen: 19.577 ng/ul

RT: 12.551 min Scan# 1

Delta R.T. -0.012 min

Lab File: BP018318.D

Acq: 23 Nov 2023 16:41

Instrument :

BNA_P

ClientSampleId :

SSTD020689





Tgt Ion:142 Resp: 1507084

Ion Ratio Lower Upper

142 100

141 91.8 73.9 110.9

116 3.4 3.0 4.4



#39

1,2,4,5-Tetrachlorobenzene

Concen: 21.295 ng/ul

RT: 12.698 min Scan# 1634

Delta R.T. -0.006 min

Lab File: BP018318.D

Acq: 23 Nov 2023 16:41

Tgt Ion:216 Resp: 826096




Ion Ratio Lower Upper


216 100

214 78.5 63.3 94.9

179 20.2 17.3 25.9

108 14.8 13.4 20.0









#40
Hexachlorocyclopentadiene
Concen: 20.023 ng/ul
RT: 12.681 min Scan# 1631
Delta R.T. -0.012 min
Lab File: BP018318.D
Acq: 23 Nov 2023 16:41




Instrument :
BNA_P
ClientSampleId :
SSTD020689


Tgt Ion:237 Resp: 439927
Ion Ratio Lower Upper
237 100
235 63.8 50.7 76.1
272 13.1 10.3 15.5



#41
2,4,6-Trichlorophenol
Concen: 20.980 ng/ul
RT: 12.940 min Scan# 1675
Delta R.T. -0.006 min
Lab File: BP018318.D
Acq: 23 Nov 2023 16:41

Tgt Ion:196 Resp: 488756
Ion Ratio Lower Upper
196 100
198 95.3 78.0 117.0
200 31.3 24.6 37.0





#42

2,4,5-Trichlorophenol

Concen: 21.207 ng/ul

RT: 13.004 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP018318.D

Acq: 23 Nov 2023 16:41

Instrument :

BNA_P

ClientSampleId :

SSTD020689





Tgt Ion:196 Resp: 534229

Ion Ratio Lower Upper

196 100

198 97.2 76.7 115.1

200 30.7 24.6 36.8



#43

1,1'-Biphenyl

Concen: 20.478 ng/ul

RT: 13.345 min Scan# 1744

Delta R.T. -0.006 min

Lab File: BP018318.D

Acq: 23 Nov 2023 16:41




Tgt Ion:154 Resp: 1959747


Ion Ratio Lower Upper

154 100

153 39.7 32.2 48.2

76 11.7 9.9 14.9





#44

2-Chloronaphthalene

Concen: 20.478 ng/ul

RT: 13.381 min Scan# 1751

Delta R.T. -0.006 min

Lab File: BP018318.D

Acq: 23 Nov 2023 16:41

Instrument :

BNA_P

ClientSampleId :

SSTD020689





Tgt Ion:162 Resp: 1551071

Ion Ratio Lower Upper

162 100

164 33.0 26.2 39.4

127 36.8 30.6 45.8



#45

2-Nitroaniline

Concen: 20.176 ng/ul

RT: 13.581 min Scan# 1784

Delta R.T. -0.006 min

Lab File: BP018318.D

Acq: 23 Nov 2023 16:41




Tgt Ion: 65 Resp: 328820


Ion Ratio Lower Upper

65 100

92 76.1 57.4 86.2

138 124.3 90.6 136.0









#47
Dimethylphthalate
Concen: 19.633 ng/ul
RT: 13.969 min Scan# 1869
Delta R.T. -0.012 min
Lab File: BP018318.D
Acq: 23 Nov 2023 16:41




Instrument :
BNA_P
ClientSampleId :
SSTD020689


Tgt Ion:163 Resp: 1877477
Ion Ratio Lower Upper
163 100
194 4.0 3.1 4.7
164 10.3 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 20.885 ng/ul
RT: 14.081 min Scan# 1869
Delta R.T. -0.006 min
Lab File: BP018318.D
Acq: 23 Nov 2023 16:41

Tgt Ion:165 Resp: 336715
Ion Ratio Lower Upper
165 100
89 52.3 45.8 68.6
121 21.8 17.8 26.8









#50
Acenaphthylene
Concen: 20.042 ng/ul
RT: 14.228 min Scan# 1894
Delta R.T. -0.006 min
Lab File: BP018318.D
Acq: 23 Nov 2023 16:41




Instrument :
BNA_P
ClientSampleId :
SSTD020689


Tgt Ion:152 Resp: 2487871
Ion Ratio Lower Upper
152 100
151 20.1 15.7 23.5
153 13.3 10.8 16.2



#51
3-Nitroaniline
Concen: 21.005 ng/ul
RT: 14.404 min Scan# 1924
Delta R.T. -0.006 min
Lab File: BP018318.D
Acq: 23 Nov 2023 16:41

Tgt Ion:138 Resp: 336475
Ion Ratio Lower Upper
138 100
108 13.4 13.2 19.8
92 107.1 90.2 135.4









#52
Acenaphthene
Concen: 19.932 ng/ul
RT: 14.569 min Scan# 1952
Delta R.T. -0.006 min
Lab File: BP018318.D
Acq: 23 Nov 2023 16:41




Instrument :
BNA_P
ClientSampleId :
SSTD020689


Tgt Ion:153 Resp: 1616258
Ion Ratio Lower Upper
153 100
152 48.3 37.8 56.8
154 89.2 71.2 106.8



#53
2,4-Dinitrophenol
Concen: 17.738 ng/ul
RT: 14.610 min Scan# 1959
Delta R.T. -0.006 min
Lab File: BP018318.D
Acq: 23 Nov 2023 16:41

Tgt Ion:184 Resp: 99773
Ion Ratio Lower Upper
184 100
63 50.5 41.8 62.6
154 57.9 47.3 70.9









#55
4-Nitrophenol
Concen: 18.496 ng/ul
RT: 14.704 min Scan# 1976
Delta R.T. -0.006 min
Lab File: BP018318.D
Acq: 23 Nov 2023 16:41




Instrument :
BNA_P
ClientSampleId :
SSTD020689


Tgt Ion:109 Resp: 184879
Ion Ratio Lower Upper
109 100
139 147.6 113.0 169.4
65 123.2 99.6 149.4



#56
Dibenzofuran
Concen: 20.014 ng/ul
RT: 14.904 min Scan# 2009
Delta R.T. -0.006 min
Lab File: BP018318.D
Acq: 23 Nov 2023 16:41

Tgt Ion:168 Resp: 2218745
Ion Ratio Lower Upper
168 100
139 34.7 29.0 43.4









#57
2,4-Dinitrotoluene
Concen: 20.299 ng/ul
RT: 14.863 min Scan# 2003
Delta R.T. -0.006 min
Lab File: BP018318.D
Acq: 23 Nov 2023 16:41




Instrument :
BNA_P
ClientSampleId :
SSTD020689


Tgt Ion:165 Resp: 442249
Ion Ratio Lower Upper
165 100
63 36.1 32.1 48.1
182 2.8 2.0 3.0



#58
2,3,4,6-Tetrachlorophenol
Concen: 20.196 ng/ul
RT: 15.122 min Scan# 2046
Delta R.T. -0.012 min
Lab File: BP018318.D
Acq: 23 Nov 2023 16:41

Tgt Ion:232 Resp: 399551
Ion Ratio Lower Upper
232 100
131 44.6 34.1 51.1
130 2.4 1.8 2.6
166 32.0 24.8 37.2









#59
Diethylphthalate
Concen: 18.582 ng/ul
RT: 15.334 min Scan# 2082
Delta R.T. -0.012 min
Lab File: BP018318.D
Acq: 23 Nov 2023 16:41




Instrument :
BNA_P
ClientSampleId :
SSTD020689


Tgt Ion:149 Resp: 1785040
Ion Ratio Lower Upper
149 100
177 22.2 18.2 27.2
150 12.4 10.2 15.2



#61
Fluorene
Concen: 19.420 ng/ul
RT: 15.551 min Scan# 2119
Delta R.T. -0.012 min
Lab File: BP018318.D
Acq: 23 Nov 2023 16:41

Tgt Ion:166 Resp: 1755685
Ion Ratio Lower Upper
166 100
165 97.8 77.5 116.3
167 13.5 10.6 16.0









#62
4-Chlorophenyl-phenylether
Concen: 19.852 ng/ul
RT: 15.551 min Scan# 2119
Delta R.T. -0.012 min
Lab File: BP018318.D
Acq: 23 Nov 2023 16:41




Instrument :
BNA_P
ClientSampleId :
SSTD020689


Tgt Ion:204 Resp: 911355
Ion Ratio Lower Upper
204 100
206 33.1 26.2 39.2
141 55.5 44.8 67.2



#63
4-Nitroaniline
Concen: 20.348 ng/ul
RT: 15.569 min Scan# 2122
Delta R.T. -0.006 min
Lab File: BP018318.D
Acq: 23 Nov 2023 16:41

Tgt Ion:138 Resp: 307619
Ion Ratio Lower Upper
138 100
92 46.2 39.3 58.9
108 72.2 61.8 92.6









#66
4,6-Dinitro-2-methylphenol
Concen: 19.515 ng/ul
RT: 15.622 min Scan# 2132
Delta R.T. -0.012 min
Lab File: BP018318.D
Acq: 23 Nov 2023 16:41




Instrument :
BNA_P
ClientSampleId :
SSTD020689


Tgt Ion:198 Resp: 189038
Ion Ratio Lower Upper
198 100
182 4.3 3.0 4.4
77 22.5 17.3 25.9



#67
N-Nitrosodiphenylamine
Concen: 20.881 ng/ul
RT: 15.763 min Scan# 2155
Delta R.T. -0.012 min
Lab File: BP018318.D
Acq: 23 Nov 2023 16:41

Tgt Ion:169 Resp: 1452808
Ion Ratio Lower Upper
169 100
168 66.6 53.1 79.7
167 36.8 28.8 43.2









#68
4-Bromophenyl-phenylether
Concen: 20.342 ng/ul
RT: 16.445 min Scan# 21
Delta R.T. -0.012 min
Lab File: BP018318.D
Acq: 23 Nov 2023 16:41




Instrument :
BNA_P
ClientSampleId :
SSTD020689


Tgt Ion:248 Resp: 526952
Ion Ratio Lower Upper
248 100
250 97.7 77.4 116.2
141 70.9 56.0 84.0



#69
Hexachlorobenzene
Concen: 21.232 ng/ul
RT: 16.551 min Scan# 2289
Delta R.T. -0.006 min
Lab File: BP018318.D
Acq: 23 Nov 2023 16:41

Tgt Ion:284 Resp: 614890
Ion Ratio Lower Upper
284 100
142 32.5 28.9 43.3
249 30.7 26.0 39.0





#70

Atrazine

Concen: 20.479 ng/ul

RT: 16.722 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP018318.D

Acq: 23 Nov 2023 16:41

Instrument :

BNA_P

ClientSampleId :

SSTD020689





Tgt Ion:200 Resp: 434990

Ion Ratio Lower Upper

200 100

173 26.3 21.1 31.7

215 48.3 38.3 57.5



#71

Pentachlorophenol

Concen: 19.670 ng/ul

RT: 16.898 min Scan# 2348

Delta R.T. -0.006 min

Lab File: BP018318.D

Acq: 23 Nov 2023 16:41




Tgt Ion:266 Resp: 292695


Ion Ratio Lower Upper

266 100

264 62.8 50.1 75.1

268 64.5 50.1 75.1





#72

Phenanthrene

Concen: 19.923 ng/ul

RT: 17.292 min Scan# 2415

Delta R.T. -0.012 min

Lab File: BP018318.D

Acq: 23 Nov 2023 16:41

Instrument :

BNA_P

ClientSampleId :

SSTD020689





Tgt Ion:178 Resp: 2486350

Ion Ratio Lower Upper

178 100

179 15.6 12.4 18.6

176 19.3 15.1 22.7



#74

Anthracene

Concen: 20.064 ng/ul

RT: 17.386 min Scan# 2431

Delta R.T. -0.012 min

Lab File: BP018318.D

Acq: 23 Nov 2023 16:41




Tgt Ion:178 Resp: 2527497


Ion Ratio Lower Upper

178 100

179 15.3 12.4 18.6

176 18.7 15.2 22.8









#75
1,2,3,4-Tetrachlorobenzene
Concen: 23.369 ng/uL
RT: 13.304 min Scan# 1737
Delta R.T. -0.006 min
Lab File: BP018318.D
Acq: 23 Nov 2023 16:41




Instrument :
BNA_P
ClientSampleId :
SSTD020689


Tgt Ion:216 Resp: 824153
Ion Ratio Lower Upper
216 100
214 78.6 62.0 93.0
218 48.1 38.6 57.8



#76
Pentachlorobenzene
Concen: 22.685 ng/uL
RT: 14.822 min Scan# 1995
Delta R.T. -0.006 min
Lab File: BP018318.D
Acq: 23 Nov 2023 16:41

Tgt Ion:250 Resp: 777126
Ion Ratio Lower Upper
250 100
248 62.0 51.0 76.6
252 62.7 51.5 77.3





#77

Carbazole

Concen: 20.142 ng/ul

RT: 17.657 min Scan# 2478

Delta R.T. -0.006 min

Lab File: BP018318.D

Acq: 23 Nov 2023 16:41

Instrument :

BNA_P

ClientSampleId :

SSTD020689





Tgt Ion:167 Resp: 2055599

Ion Ratio Lower Upper

167 100

166 21.4 17.5 26.3

139 12.5 10.0 15.0



#78

Di-n-butylphthalate

Concen: 17.648 ng/ul

RT: 18.222 min Scan# 2573

Delta R.T. -0.018 min

Lab File: BP018318.D

Acq: 23 Nov 2023 16:41




Tgt Ion:149 Resp: 2505212


Ion Ratio Lower Upper

149 100

150 9.5 7.4 11.2

104 4.7 3.8 5.6





#80

Fluoranthene

Concen: 20.185 ng/ul

RT: 19.263 min Scan# 21

Delta R.T. -0.012 min

Lab File: BP018318.D

Acq: 23 Nov 2023 16:41

Instrument :

BNA_P

ClientSampleId :

SSTD020689





Tgt Ion:202 Resp: 2416902

Ion Ratio Lower Upper

202 100

101 10.9 8.6 13.0

100 8.1 6.7 10.1



#82

Pyrene

Concen: 20.033 ng/ul

RT: 19.616 min Scan# 2810

Delta R.T. -0.012 min

Lab File: BP018318.D

Acq: 23 Nov 2023 16:41




Tgt Ion:202 Resp: 2415413


Ion Ratio Lower Upper

202 100

101 12.8 10.2 15.4

100 10.6 8.4 12.6





#83

Butylbenzylphthalate

Concen: 16.043 ng/ul

RT: 20.510 min Scan# 2965

Delta R.T. -0.012 min

Lab File: BP018318.D

Acq: 23 Nov 2023 16:41

Instrument :

BNA_P

ClientSampleId :

SSTD020689





Tgt Ion:149 Resp: 715841

Ion Ratio Lower Upper

149 100

91 67.8 53.7 80.5

206 20.9 16.4 24.6



#84

3,3'-Dichlorobenzidine

Concen: 19.173 ng/ul

RT: 21.280 min Scan# 3093

Delta R.T. -0.006 min

Lab File: BP018318.D

Acq: 23 Nov 2023 16:41




Tgt Ion:252 Resp: 583996


Ion Ratio Lower Upper

252 100

254 65.2 51.1 76.7

126 10.5 8.6 12.8





#85

Benzo(a)anthracene

Concen: 19.751 ng/ul

RT: 21.339 min Scan# 3103

Delta R.T. -0.006 min

Lab File: BP018318.D

Acq: 23 Nov 2023 16:41

Instrument :

BNA_P

ClientSampleId :

SSTD020689





Tgt Ion:228 Resp: 2027115

Ion Ratio Lower Upper

228 100

229 19.3 16.2 24.2

226 27.0 22.2 33.2



#86

Bis(2-ethylhexyl)phthalate

Concen: 15.793 ng/ul

RT: 21.292 min Scan# 3095

Delta R.T. -0.012 min

Lab File: BP018318.D

Acq: 23 Nov 2023 16:41




Tgt Ion:149 Resp: 1057263


Ion Ratio Lower Upper

149 100

167 27.4 22.1 33.1

279 4.3 3.4 5.0





#87

Chrysene

Concen: 20.048 ng/ul

RT: 21.392 min Scan# 3112

Delta R.T. -0.006 min

Lab File: BP018318.D

Acq: 23 Nov 2023 16:41

Instrument :

BNA_P

ClientSampleId :

SSTD020689





Tgt Ion:228 Resp: 1872359

Ion Ratio Lower Upper

228 100

226 30.0 24.2 36.4

229 19.7 16.1 24.1



#89

Di-n-octyl phthalate

Concen: 14.481 ng/ul




RT: 22.239 min Scan# 3256


Delta R.T. -0.012 min

Lab File: BP018318.D

Acq: 23 Nov 2023 16:41

Tgt Ion:149 Resp: 1575885









#90
Benzo(b)fluoranthene
Concen: 18.735 ng/ul
RT: 23.039 min Scan# 31
Delta R.T. -0.006 min
Lab File: BP018318.D
Acq: 23 Nov 2023 16:41




Instrument :
BNA_P
ClientSampleId :
SSTD020689


Tgt Ion:252 Resp: 2011515
Ion Ratio Lower Upper
252 100
253 21.7 17.7 26.5
125 10.0 8.4 12.6



#91
Benzo(k)fluoranthene
Concen: 18.898 ng/ul
RT: 23.086 min Scan# 3400
Delta R.T. -0.012 min
Lab File: BP018318.D
Acq: 23 Nov 2023 16:41

Tgt Ion:252 Resp: 2019843
Ion Ratio Lower Upper
252 100
253 21.6 17.7 26.5
125 9.9 8.0 12.0





#93

Benzo(a)pyrene

Concen: 19.887 ng/ul

RT: 23.674 min Scan# 3

Delta R.T. -0.006 min

Lab File: BP018318.D

Acq: 23 Nov 2023 16:41

Instrument :

BNA_P

ClientSampleId :

SSTD020689





Tgt Ion:252 Resp: 1963563

Ion Ratio Lower Upper

252 100

253 22.4 17.7 26.5

125 10.7 8.8 13.2



#94

Indeno(1,2,3-cd)pyrene

Concen: 22.280 ng/ul

RT: 26.327 min Scan# 3951

Delta R.T. -0.000 min

Lab File: BP018318.D

Acq: 23 Nov 2023 16:41




Tgt Ion:276 Resp: 2543645


Ion Ratio Lower Upper

276 100

138 21.4 17.5 26.3

277 24.8 19.1 28.7





#95

Dibenzo(a,h)anthracene

Concen: 22.583 ng/ul

RT: 26.368 min Scan# 3958

Delta R.T. -0.000 min

Lab File: BP018318.D

Acq: 23 Nov 2023 16:41

Instrument :

BNA_P

ClientSampleId :

SSTD020689





Tgt Ion:278 Resp: 2110735

Ion Ratio Lower Upper

278 100

139 14.7 13.3 19.9

279 24.2 19.0 28.4



#96

Benzo(g,h,i)perylene

Concen: 22.696 ng/ul

RT: 27.109 min Scan# 4084

Delta R.T. -0.000 min

Lab File: BP018318.D

Acq: 23 Nov 2023 16:41

Tgt Ion:276 Resp: 2098506

Ion Ratio Lower Upper

276 100

138 21.5 17.7 26.5

277 23.5 18.6 28.0

