

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112223\
 Data File : BP018299.D
 Acq On : 23 Nov 2023 05:23
 Operator : MA/JU
 Sample : PB157005BL
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK005

Quant Time: Nov 23 05:54:31 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP112223.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Nov 22 14:09:55 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.875	152	451590	20.000	ng/u1	0.00
20) Naphthalene-d8	10.675	136	1899021	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.504	164	1265409	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.257	188	2586553	20.000	ng/u1	0.00
79) Chrysene-d12	21.357	240	1600268	20.000	ng/u1	0.00
88) Perylene-d12	23.774	264	1629162	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.275	96	60067	4.982	ng/uL	0.00
4) Pyridine-d5	3.687	84	835530	25.216	ng/u1	0.00
7) Phenol-d5	7.034	99	1025911	24.858	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.210	67	633319	25.662	ng/u1	0.00
11) 2-Chlorophenol-d4	7.399	132	845111	24.389	ng/u1	0.00
15) 4-Methylphenol-d8	8.581	113	841371	25.262	ng/u1	0.00
21) Nitrobenzene-d5	9.034	128	409876	26.573	ng/u1	0.00
24) 2-Nitrophenol-d4	9.757	143	394592	27.700	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.293	165	832095	24.212	ng/u1	0.00
31) 4-Chloroaniline-d4	10.810	131	1293077	27.897	ng/u1	0.00
46) Dimethylphthalate-d6	13.922	166	2962011	26.306	ng/u1	-0.01
49) Acenaphthylene-d8	14.198	160	3254766	25.369	ng/u1	0.00
54) 4-Nitrophenol-d4	14.692	143	347303	22.343	ng/u1	0.00
60) Fluorene-d10	15.498	176	2429705	25.799	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.610	200	216964	19.705	ng/u1	0.00
73) Anthracene-d10	17.351	188	3571866	25.836	ng/u1	-0.01
81) Pyrene-d10	19.592	212	3493665	27.431	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.621	264	2428002	25.706	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112223\
Data File : BP018299.D
Acq On : 23 Nov 2023 05:23
Operator : MA/JU
Sample : PB157005BL
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SBLK005

Quant Time: Nov 23 05:54:31 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP112223.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Nov 22 14:09:55 2023
Response via : Initial Calibration

