

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112423\
 Data File : BP018335.D
 Acq On : 24 Nov 2023 19:34
 Operator : MA/JU
 Sample : 05264-14
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DCKS5

Quant Time: Nov 24 22:02:20 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP112223.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Nov 24 22:02:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.869	152	387519	20.000	ng/ul	0.00
20) Naphthalene-d8	10.669	136	1494020	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.498	164	788513	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.251	188	1385134	20.000	ng/ul	0.00
79) Chrysene-d12	21.351	240	1265894	20.000	ng/ul	0.00
88) Perylene-d12	23.780	264	1559329	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.275	96	35850	3.465	ng/uL	0.00
4) Pyridine-d5	3.693	84	209406	7.365	ng/ul	0.00
7) Phenol-d5	7.028	99	157266	4.441	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.205	67	513887	24.265	ng/ul	0.00
11) 2-Chlorophenol-d4	7.399	132	515814	17.347	ng/ul	0.00
15) 4-Methylphenol-d8	8.575	113	307954	10.775	ng/ul	0.00
21) Nitrobenzene-d5	9.028	128	310468	25.585	ng/ul	0.00
24) 2-Nitrophenol-d4	9.751	143	262288	23.404	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.287	165	580019	21.452	ng/ul	0.00
31) 4-Chloroaniline-d4	10.804	131	632308	17.339	ng/ul	0.00
46) Dimethylphthalate-d6	13.916	166	1825239	26.014	ng/ul	0.00
49) Acenaphthylene-d8	14.192	160	2116207	26.471	ng/ul	0.00
54) 4-Nitrophenol-d4	14.686	143	27474	2.836	ng/ul	0.00
60) Fluorene-d10	15.492	176	1530291	26.076	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.604	200	108180	18.347	ng/ul	0.00
73) Anthracene-d10	17.345	188	1879012	25.380	ng/ul	0.00
81) Pyrene-d10	19.586	212	2387311	23.695	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.621	264	2699793	29.864	ng/ul	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112423\
Data File : BP018335.D
Acq On : 24 Nov 2023 19:34
Operator : MA/JU
Sample : 05264-14
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DCKS5

Quant Time: Nov 24 22:02:20 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP112223.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Nov 24 22:02:00 2023
Response via : Initial Calibration

