

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112423\
 Data File : BP018387.D
 Acq On : 26 Nov 2023 02:38
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 71 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :

Quant Time: Nov 27 00:06:58 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP112223.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Nov 24 22:02:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.864	152	498678	20.000	ng/u1	0.00
20) Naphthalene-d8	10.663	136	2054237	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.498	164	1241880	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.245	188	2403377	20.000	ng/u1	0.00
79) Chrysene-d12	21.357	240	1632435	20.000	ng/u1	0.00
88) Perylene-d12	23.804	264	1842566	20.000	ng/u1	0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.270	96	97337	7.311	ng/uL	0.00
4) Pyridine-d5	3.687	84	682355	18.649	ng/u1	0.00
7) Phenol-d5	7.028	99	903607	19.827	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.199	67	507494	18.622	ng/u1	0.00
11) 2-Chlorophenol-d4	7.393	132	756129	19.761	ng/u1	0.00
15) 4-Methylphenol-d8	8.575	113	743036	20.203	ng/u1	0.00
21) Nitrobenzene-d5	9.028	128	366563	21.969	ng/u1	0.00
24) 2-Nitrophenol-d4	9.746	143	371335	24.098	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.281	165	788744	21.216	ng/u1	0.00
31) 4-Chloroaniline-d4	10.805	131	1036015	20.662	ng/u1	0.00
46) Dimethylphthalate-d6	13.916	166	2166483	19.606	ng/u1	0.00
49) Acenaphthylene-d8	14.193	160	2503028	19.879	ng/u1	0.00
54) 4-Nitrophenol-d4	14.693	143	292227	19.156	ng/u1	0.00
60) Fluorene-d10	15.487	176	1800979	19.485	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.604	200	224593	21.952	ng/u1	0.00
73) Anthracene-d10	17.345	188	2537762	19.755	ng/u1	0.00
81) Pyrene-d10	19.586	212	2503328	19.268	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.639	264	2121966	19.864	ng/u1	0.02
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.305	88	104013	7.285	ng/uL#	90
5) Pyridine	3.705	79	682716	18.519	ng/u1	97
6) Benzaldehyde	7.005	77	508347	21.078	ng/u1	94
8) Phenol	7.058	94	884741	19.791	ng/u1	96
10) Bis(2-Chloroethyl)ether	7.293	93	721042	19.363	ng/u1	97
12) 2-Chlorophenol	7.428	128	758346	19.754	ng/u1	95
13) 2-Methylphenol	8.311	108	698946	20.073	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.393	45	888705	16.790	ng/u1	96
16) Acetophenone	8.693	105	1099785	19.542	ng/u1	95
17) N-Nitroso-di-n-propyla...	8.681	70	500174	15.910	ng/u1	93
18) 4-Methylphenol	8.640	108	745166	20.040	ng/u1	99
19) Hexachloroethane	8.946	117	298823	18.640	ng/u1	94
22) Nitrobenzene	9.069	77	770191	19.703	ng/u1	94
23) Isophorone	9.599	82	1499171	17.207	ng/u1	97
25) 2-Nitrophenol	9.781	139	405342	23.379	ng/u1	91
26) 2,4-Dimethylphenol	9.840	107	760774	19.544	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.087	93	985595	18.731	ng/u1	100
29) 2,4-Dichlorophenol	10.310	162	756195	21.235	ng/u1	98
30) Naphthalene	10.716	128	2435813	19.605	ng/u1	100
32) 4-Chloroaniline	10.828	127	980605	20.725	ng/u1	99
33) Hexachlorobutadiene	11.005	225	534982	22.027	ng/u1	98
34) Caprolactam	11.599	113	193446	18.851	ng/u1	87
35) 4-Chloro-3-methylphenol	11.952	107	704609	18.506	ng/u1	94
36) 2-Methylnaphthalene	12.322	142	1690430	19.225	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112423\
 Data File : BP018387.D
 Acq On : 26 Nov 2023 02:38
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 71 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :

Quant Time: Nov 27 00:06:58 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP112223.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Nov 24 22:02:00 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.546	142	1687273	19.063	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.693	216	988172	22.530	ng/ul	97
40) Hexachlorocyclopentadiene	12.669	237	470496	18.940	ng/ul	99
41) 2,4,6-Trichlorophenol	12.928	196	599329	22.753	ng/ul	98
42) 2,4,5-Trichlorophenol	12.999	196	629561	22.103	ng/ul	98
43) 1,1'-Biphenyl	13.334	154	2204271	20.372	ng/ul	99
44) 2-Chloronaphthalene	13.375	162	1761646	20.570	ng/ul	98
45) 2-Nitroaniline	13.575	65	368149	19.979	ng/ul	91
47) Dimethylphthalate	13.963	163	2118270	19.592	ng/ul	99
48) 2,6-Dinitrotoluene	14.075	165	406494	22.300	ng/ul	91
50) Acenaphthylene	14.216	152	2778640	19.798	ng/ul	100
51) 3-Nitroaniline	14.404	138	396602	21.898	ng/ul#	91
52) Acenaphthene	14.563	153	1787468	19.496	ng/ul	99
53) 2,4-Dinitrophenol	14.604	184	126046	19.820	ng/ul	97
55) 4-Nitrophenol	14.704	109	203625	18.017	ng/ul	94
56) Dibenzofuran	14.898	168	2483120	19.810	ng/ul	97
57) 2,4-Dinitrotoluene	14.857	165	549098	22.291	ng/ul	90
58) 2,3,4,6-Tetrachlorophenol	15.122	232	496459	22.194	ng/ul	93
59) Diethylphthalate	15.328	149	2042653	18.807	ng/ul	99
61) Fluorene	15.545	166	1952690	19.103	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.545	204	1048174	20.194	ng/ul	96
63) 4-Nitroaniline	15.563	138	352198	20.605	ng/ul	96
66) 4,6-Dinitro-2-methylph...	15.622	198	246201	21.697	ng/ul	95
67) N-Nitrosodiphenylamine	15.757	169	1646384	20.201	ng/ul	99
68) 4-Bromophenyl-phenylether	16.440	248	641869	21.153	ng/ul	96
69) Hexachlorobenzene	16.545	284	751406	22.150	ng/ul	93
70) Atrazine	16.716	200	536973	21.582	ng/ul	97
71) Pentachlorophenol	16.892	266	381412	21.882	ng/ul	98
72) Phenanthrene	17.292	178	2881627	19.712	ng/ul	100
74) Anthracene	17.381	178	2907221	19.702	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.299	216	978023	23.675	ng/ul	99
76) Pentachlorobenzene	14.810	250	926304	23.083	ng/ul	100
77) Carbazole	17.651	167	2372956	19.849	ng/ul	98
78) Di-n-butylphthalate	18.216	149	3214296	19.330	ng/ul	100
80) Fluoranthene	19.263	202	2992696	19.282	ng/ul	97
82) Pyrene	19.616	202	2993870	19.156	ng/ul	97
83) Butylbenzylphthalate	20.510	149	1060438	18.334	ng/ul	95
84) 3,3'-Dichlorobenzidine	21.280	252	809389	20.500	ng/ul	99
85) Benzo(a)anthracene	21.339	228	2598374	19.531	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.298	149	1555797	17.929	ng/ul	99
87) Chrysene	21.392	228	2396087	19.792	ng/ul	99
89) Di-n-octyl phthalate	22.263	149	2220697	16.371	ng/ul	100
90) Benzo(b)fluoranthene	23.051	252	2499086	18.673	ng/ul	99
91) Benzo(k)fluoranthene	23.104	252	2547857	19.123	ng/ul	99
93) Benzo(a)pyrene	23.692	252	2424844	19.701	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.398	276	3017566	21.203	ng/ul	97
95) Dibenzo(a,h)anthracene	26.445	278	2487850	21.353	ng/ul	98
96) Benzo(g,h,i)perylene	27.198	276	2478694	21.506	ng/ul	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112423\
 Data File : BP018387.D
 Acq On : 26 Nov 2023 02:38
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 71 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :

Quant Time: Nov 27 00:06:58 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP112223.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Nov 24 22:02:00 2023
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 7.285 ng/uL
RT: 3.305 min Scan# 31
Delta R.T. -0.006 min
Lab File: BP018387.D
Acq: 26 Nov 2023 02:38




Instrument :
BNA_P
ClientSampleId :


Tgt Ion: 88 Resp: 104013
Ion Ratio Lower Upper
88 100
43 25.1 26.9 40.3#
58 67.5 59.6 89.4



#5
Pyridine
Concen: 18.519 ng/uL
RT: 3.705 min Scan# 105
Delta R.T. 0.000 min
Lab File: BP018387.D
Acq: 26 Nov 2023 02:38

Tgt Ion: 79 Resp: 682716
Ion Ratio Lower Upper
79 100
52 70.4 57.7 86.5
51 32.7 27.8 41.8









#6
Benzaldehyde
Concen: 21.078 ng/ul
RT: 7.005 min Scan# 666
Delta R.T. 0.000 min
Lab File: BP018387.D
Acq: 26 Nov 2023 02:38




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:	77	Resp:	508347
Ion	Ratio	Lower	Upper
77	100		
105	108.6	81.5	122.3
106	103.5	77.9	116.9



#8
Phenol
Concen: 19.791 ng/ul
RT: 7.058 min Scan# 675
Delta R.T. 0.000 min
Lab File: BP018387.D
Acq: 26 Nov 2023 02:38

Tgt Ion:	94	Resp:	884741
Ion	Ratio	Lower	Upper
94	100		
65	24.5	21.2	31.8
66	32.5	28.0	42.0









#10
Bis(2-Chloroethyl)ether
Concen: 19.363 ng/ul
RT: 7.293 min Scan# 715
Delta R.T. -0.006 min
Lab File: BP018387.D
Acq: 26 Nov 2023 02:38




Instrument :
BNA_P
ClientSampleId :


Tgt Ion: 93 Resp: 721042
Ion Ratio Lower Upper
93 100
63 69.9 58.2 87.4
95 32.7 25.6 38.4



#12
2-Chlorophenol
Concen: 19.754 ng/ul
RT: 7.428 min Scan# 738
Delta R.T. 0.000 min
Lab File: BP018387.D
Acq: 26 Nov 2023 02:38

Tgt Ion:128 Resp: 758346
Ion Ratio Lower Upper
128 100
64 39.0 34.8 52.2
130 32.6 25.5 38.3





#13

2-Methylphenol

Concen: 20.073 ng/ul

RT: 8.311 min Scan# 8

Delta R.T. 0.000 min

Lab File: BP018387.D

Acq: 26 Nov 2023 02:38

Instrument :

BNA_P





ClientSampleId :

Tgt Ion:108 Resp: 698946

Ion Ratio Lower Upper

108 100

107 87.8 70.9 106.3



#14

2,2'-oxybis(1-Chloropropane)

Concen: 16.790 ng/ul

RT: 8.393 min Scan# 902

Delta R.T. -0.012 min

Lab File: BP018387.D

Acq: 26 Nov 2023 02:38




Tgt Ion: 45 Resp: 888705


Ion Ratio Lower Upper

45 100

77 15.6 11.4 17.2

79 12.1 8.4 12.6









#16
Acetophenone
Concen: 19.542 ng/ul
RT: 8.693 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP018387.D
Acq: 26 Nov 2023 02:38




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:105 Resp: 1099785
Ion Ratio Lower Upper
105 100
77 70.8 59.5 89.3
51 22.5 20.5 30.7



#17
N-Nitroso-di-n-propylamine
Concen: 15.910 ng/ul
RT: 8.681 min Scan# 951
Delta R.T. -0.006 min
Lab File: BP018387.D
Acq: 26 Nov 2023 02:38

Tgt Ion: 70 Resp: 500174
Ion Ratio Lower Upper
70 100
42 51.3 46.8 70.2
101 10.8 8.4 12.6
130 24.7 18.5 27.7









#18
4-Methylphenol
Concen: 20.040 ng/ul
RT: 8.640 min Scan# 944
Delta R.T. 0.000 min
Lab File: BP018387.D
Acq: 26 Nov 2023 02:38




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:108 Resp: 745166
Ion Ratio Lower Upper
108 100
107 111.3 89.5 134.3



#19
Hexachloroethane
Concen: 18.640 ng/ul
RT: 8.946 min Scan# 996
Delta R.T. -0.006 min
Lab File: BP018387.D
Acq: 26 Nov 2023 02:38

Tgt Ion:117 Resp: 298823
Ion Ratio Lower Upper
117 100
201 114.6 85.6 128.4
199 71.2 54.2 81.4









#22
Nitrobenzene
Concen: 19.703 ng/ul
RT: 9.069 min Scan# 1017
Delta R.T. -0.006 min
Lab File: BP018387.D
Acq: 26 Nov 2023 02:38




Instrument :
BNA_P
ClientSampleId :


Tgt Ion: 77 Resp: 770191
Ion Ratio Lower Upper
77 100
123 52.0 37.7 56.5
65 12.4 10.5 15.7



#23
Isophorone
Concen: 17.207 ng/ul
RT: 9.599 min Scan# 1107
Delta R.T. 0.000 min
Lab File: BP018387.D
Acq: 26 Nov 2023 02:38

Tgt Ion: 82 Resp: 1499171
Ion Ratio Lower Upper
82 100
95 7.0 5.8 8.6
138 20.3 14.9 22.3









#25
2-Nitrophenol
Concen: 23.379 ng/ul
RT: 9.781 min Scan# 1138
Delta R.T. 0.000 min
Lab File: BP018387.D
Acq: 26 Nov 2023 02:38




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:139 Resp: 405342
Ion Ratio Lower Upper
139 100
65 39.7 38.3 57.5
109 27.1 23.4 35.0



#26
2,4-Dimethylphenol
Concen: 19.544 ng/ul
RT: 9.840 min Scan# 1148
Delta R.T. -0.006 min
Lab File: BP018387.D
Acq: 26 Nov 2023 02:38

Tgt Ion:107 Resp: 760774
Ion Ratio Lower Upper
107 100
121 51.6 40.0 60.0
122 88.5 70.4 105.6









#27
 Bis(2-Chloroethoxy)methane
 Concen: 18.731 ng/ul
 RT: 10.087 min Scan# 1190
 Delta R.T. 0.000 min
 Lab File: BP018387.D
 Acq: 26 Nov 2023 02:38




Instrument :
 BNA_P
 ClientSampleId :


Tgt Ion: 93 Resp: 985595
 Ion Ratio Lower Upper
 93 100
 95 32.5 26.0 39.0
 123 11.9 9.1 13.7



#29
 2,4-Dichlorophenol
 Concen: 21.235 ng/ul
 RT: 10.310 min Scan# 1228
 Delta R.T. -0.006 min
 Lab File: BP018387.D
 Acq: 26 Nov 2023 02:38

Tgt Ion:162 Resp: 756195
 Ion Ratio Lower Upper
 162 100
 164 65.5 53.0 79.6
 98 32.1 27.7 41.5









#30
Naphthalene
Concen: 19.605 ng/ul
RT: 10.716 min Scan# 1297
Delta R.T. -0.006 min
Lab File: BP018387.D
Acq: 26 Nov 2023 02:38




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:128 Resp: 2435813
Ion Ratio Lower Upper
128 100
129 10.8 8.9 13.3
127 12.9 10.4 15.6



#32
4-Chloroaniline
Concen: 20.725 ng/ul
RT: 10.828 min Scan# 1316
Delta R.T. 0.000 min
Lab File: BP018387.D
Acq: 26 Nov 2023 02:38


Tgt Ion:127 Resp: 980605
Ion Ratio Lower Upper
127 100
129 31.9 25.8 38.8








#33
Hexachlorobutadiene
Concen: 22.027 ng/ul
RT: 11.005 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP018387.D
Acq: 26 Nov 2023 02:38

Instrument :
BNA_P
ClientSampleId :







Tgt Ion:225 Resp: 534982
Ion Ratio Lower Upper
225 100
223 61.8 50.2 75.4
227 64.8 50.4 75.6



#34
Caprolactam
Concen: 18.851 ng/ul
RT: 11.599 min Scan# 1447
Delta R.T. 0.012 min
Lab File: BP018387.D
Acq: 26 Nov 2023 02:38

Tgt Ion:113 Resp: 193446
Ion Ratio Lower Upper
113 100
55 141.7 129.0 193.4
56 105.8 94.8 142.2









#35
4-Chloro-3-methylphenol
Concen: 18.506 ng/ul
RT: 11.952 min Scan# 11
Delta R.T. 0.000 min
Lab File: BP018387.D
Acq: 26 Nov 2023 02:38




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:107 Resp: 704609
Ion Ratio Lower Upper
107 100
144 28.8 21.6 32.4
142 88.2 65.4 98.0



#36
2-Methylnaphthalene
Concen: 19.225 ng/ul
RT: 12.322 min Scan# 1570
Delta R.T. -0.006 min
Lab File: BP018387.D
Acq: 26 Nov 2023 02:38

Tgt Ion:142 Resp: 1690430
Ion Ratio Lower Upper
142 100
141 89.0 71.0 106.4





#37

1-Methylnaphthalene

Concen: 19.063 ng/ul

RT: 12.546 min Scan# 1608

Delta R.T. -0.006 min

Lab File: BP018387.D

Acq: 26 Nov 2023 02:38

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:142 Resp: 1687273

Ion Ratio Lower Upper

142 100

141 92.4 73.9 110.9

116 3.6 3.0 4.4



#39

1,2,4,5-Tetrachlorobenzene

Concen: 22.530 ng/ul

RT: 12.693 min Scan# 1633

Delta R.T. 0.000 min

Lab File: BP018387.D

Acq: 26 Nov 2023 02:38

Tgt Ion:216 Resp: 988172




Ion Ratio Lower Upper


216 100

214 77.2 63.3 94.9

179 19.7 17.3 25.9

108 13.7 13.4 20.0





#40

Hexachlorocyclopentadiene

Concen: 18.940 ng/ul

RT: 12.669 min Scan# 1

Delta R.T. -0.012 min

Lab File: BP018387.D

Acq: 26 Nov 2023 02:38

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:237 Resp: 470496

Ion Ratio Lower Upper

237 100

235 63.5 50.7 76.1

272 13.8 10.3 15.5



#41

2,4,6-Trichlorophenol

Concen: 22.753 ng/ul

RT: 12.928 min Scan# 1673

Delta R.T. -0.006 min

Lab File: BP018387.D

Acq: 26 Nov 2023 02:38




Tgt Ion:196 Resp: 599329


Ion Ratio Lower Upper

196 100

198 95.4 78.0 117.0

200 30.3 24.6 37.0





#42

2,4,5-Trichlorophenol

Concen: 22.103 ng/ul

RT: 12.999 min Scan# 1

Delta R.T. 0.000 min


Lab File: BP018387.D

Acq: 26 Nov 2023 02:38

Instrument :

BNA_P

ClientSampleId :






Tgt Ion:196 Resp: 629561

Ion Ratio Lower Upper

196 100

198 98.0 76.7 115.1

200 31.1 24.6 36.8



#43

1,1'-Biphenyl

Concen: 20.372 ng/ul

RT: 13.334 min Scan# 1742

Delta R.T. -0.006 min

Lab File: BP018387.D

Acq: 26 Nov 2023 02:38




Tgt Ion:154 Resp: 2204271


Ion Ratio Lower Upper

154 100

153 40.0 32.2 48.2

76 11.1 9.9 14.9





#44

2-Chloronaphthalene

Concen: 20.570 ng/ul

RT: 13.375 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP018387.D

Acq: 26 Nov 2023 02:38

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:162 Resp: 1761646

Ion Ratio Lower Upper

162 100

164 32.2 26.2 39.4

127 36.1 30.6 45.8



#45

2-Nitroaniline

Concen: 19.979 ng/ul

RT: 13.575 min Scan# 1783

Delta R.T. -0.006 min

Lab File: BP018387.D

Acq: 26 Nov 2023 02:38




Tgt Ion: 65 Resp: 368149


Ion Ratio Lower Upper

65 100

92 75.9 57.4 86.2

138 126.3 90.6 136.0









#47
Dimethylphthalate
Concen: 19.592 ng/ul
RT: 13.963 min Scan# 1849
Delta R.T. 0.000 min
Lab File: BP018387.D
Acq: 26 Nov 2023 02:38




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:163 Resp: 2118270
Ion Ratio Lower Upper
163 100
194 4.4 3.1 4.7
164 10.0 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 22.300 ng/ul
RT: 14.075 min Scan# 1868
Delta R.T. -0.006 min
Lab File: BP018387.D
Acq: 26 Nov 2023 02:38

Tgt Ion:165 Resp: 406494
Ion Ratio Lower Upper
165 100
89 49.5 45.8 68.6
121 19.7 17.8 26.8









#50
Acenaphthylene
Concen: 19.798 ng/ul
RT: 14.216 min Scan# 1892
Delta R.T. -0.006 min
Lab File: BP018387.D
Acq: 26 Nov 2023 02:38




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:152 Resp: 2778640
Ion Ratio Lower Upper
152 100
151 19.5 15.7 23.5
153 13.4 10.8 16.2



#51
3-Nitroaniline
Concen: 21.898 ng/ul
RT: 14.404 min Scan# 1924
Delta R.T. 0.000 min
Lab File: BP018387.D
Acq: 26 Nov 2023 02:38

Tgt Ion:138 Resp: 396602
Ion Ratio Lower Upper
138 100
108 12.2 13.2 19.8#
92 103.9 90.2 135.4









#52
Acenaphthene
Concen: 19.496 ng/ul
RT: 14.563 min Scan# 1951
Delta R.T. 0.000 min
Lab File: BP018387.D
Acq: 26 Nov 2023 02:38




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:153 Resp: 1787468
Ion Ratio Lower Upper
153 100
152 47.6 37.8 56.8
154 89.8 71.2 106.8



#53
2,4-Dinitrophenol
Concen: 19.820 ng/ul
RT: 14.604 min Scan# 1958
Delta R.T. 0.000 min
Lab File: BP018387.D
Acq: 26 Nov 2023 02:38

Tgt Ion:184 Resp: 126046
Ion Ratio Lower Upper
184 100
63 48.5 41.8 62.6
154 57.8 47.3 70.9









#55
4-Nitrophenol
Concen: 18.017 ng/ul
RT: 14.704 min Scan# 1975
Delta R.T. 0.000 min
Lab File: BP018387.D
Acq: 26 Nov 2023 02:38




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:109 Resp: 203625
Ion Ratio Lower Upper
109 100
139 149.8 113.0 169.4
65 119.3 99.6 149.4



#56
Dibenzofuran
Concen: 19.810 ng/ul
RT: 14.898 min Scan# 2008
Delta R.T. 0.000 min
Lab File: BP018387.D
Acq: 26 Nov 2023 02:38

Tgt Ion:168 Resp: 2483120
Ion Ratio Lower Upper
168 100
139 34.5 29.0 43.4





#57

2,4-Dinitrotoluene

Concen: 22.291 ng/ul

RT: 14.857 min Scan# 2001

Delta R.T. -0.006 min

Lab File: BP018387.D

Acq: 26 Nov 2023 02:38

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:165 Resp: 549098

Ion Ratio Lower Upper

165 100

63 33.4 32.1 48.1

182 2.8 2.0 3.0



#58

2,3,4,6-Tetrachlorophenol

Concen: 22.194 ng/ul

RT: 15.122 min Scan# 2046

Delta R.T. 0.000 min

Lab File: BP018387.D

Acq: 26 Nov 2023 02:38

Tgt Ion:232 Resp: 496459




Ion Ratio Lower Upper


232 100

131 37.2 34.1 51.1

130 2.0 1.8 2.6

166 28.3 24.8 37.2









#59
Diethylphthalate
Concen: 18.807 ng/ul
RT: 15.328 min Scan# 2081
Delta R.T. -0.006 min
Lab File: BP018387.D
Acq: 26 Nov 2023 02:38




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:149 Resp: 2042653
Ion Ratio Lower Upper
149 100
177 23.2 18.2 27.2
150 12.4 10.2 15.2



#61
Fluorene
Concen: 19.103 ng/ul
RT: 15.545 min Scan# 2118
Delta R.T. -0.006 min
Lab File: BP018387.D
Acq: 26 Nov 2023 02:38

Tgt Ion:166 Resp: 1952690
Ion Ratio Lower Upper
166 100
165 99.1 77.5 116.3
167 13.5 10.6 16.0









#62
4-Chlorophenyl-phenylether
Concen: 20.194 ng/ul
RT: 15.545 min Scan# 2118
Delta R.T. -0.006 min
Lab File: BP018387.D
Acq: 26 Nov 2023 02:38




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:204 Resp: 1048174
Ion Ratio Lower Upper
204 100
206 33.1 26.2 39.2
141 51.9 44.8 67.2



#63
4-Nitroaniline
Concen: 20.605 ng/ul
RT: 15.563 min Scan# 2121
Delta R.T. 0.000 min
Lab File: BP018387.D
Acq: 26 Nov 2023 02:38

Tgt Ion:138 Resp: 352198
Ion Ratio Lower Upper
138 100
92 47.0 39.3 58.9
108 73.2 61.8 92.6









#66
4,6-Dinitro-2-methylphenol
Concen: 21.697 ng/ul
RT: 15.622 min Scan# 2132
Delta R.T. 0.000 min
Lab File: BP018387.D
Acq: 26 Nov 2023 02:38




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:198 Resp: 246201
Ion Ratio Lower Upper
198 100
182 3.8 3.0 4.4
77 19.1 17.3 25.9



#67
N-Nitrosodiphenylamine
Concen: 20.201 ng/ul
RT: 15.757 min Scan# 2154
Delta R.T. -0.006 min
Lab File: BP018387.D
Acq: 26 Nov 2023 02:38

Tgt Ion:169 Resp: 1646384
Ion Ratio Lower Upper
169 100
168 66.8 53.1 79.7
167 36.3 28.8 43.2









#68
4-Bromophenyl-phenylether
Concen: 21.153 ng/ul
RT: 16.440 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP018387.D
Acq: 26 Nov 2023 02:38




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:248 Resp: 641869
Ion Ratio Lower Upper
248 100
250 96.2 77.4 116.2
141 62.1 56.0 84.0



#69
Hexachlorobenzene
Concen: 22.150 ng/ul
RT: 16.545 min Scan# 2288
Delta R.T. -0.006 min
Lab File: BP018387.D
Acq: 26 Nov 2023 02:38

Tgt Ion:284 Resp: 751406
Ion Ratio Lower Upper
284 100
142 29.7 28.9 43.3
249 30.7 26.0 39.0





#70

Atrazine

Concen: 21.582 ng/ul

RT: 16.716 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP018387.D

Acq: 26 Nov 2023 02:38

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:200 Resp: 536973

Ion Ratio Lower Upper

200 100

173 25.3 21.1 31.7

215 50.0 38.3 57.5



#71

Pentachlorophenol

Concen: 21.882 ng/ul

RT: 16.892 min Scan# 2347

Delta R.T. 0.000 min

Lab File: BP018387.D

Acq: 26 Nov 2023 02:38




Tgt Ion:266 Resp: 381412


Ion Ratio Lower Upper

266 100

264 63.2 50.1 75.1

268 64.7 50.1 75.1





#72

Phenanthrene

Concen: 19.712 ng/ul

RT: 17.292 min Scan# 2416

Delta R.T. 0.000 min

Lab File: BP018387.D

Acq: 26 Nov 2023 02:38

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:178 Resp: 2881627

Ion Ratio Lower Upper

178 100

179 15.5 12.4 18.6

176 19.1 15.1 22.7



#74

Anthracene

Concen: 19.702 ng/ul

RT: 17.381 min Scan# 2430

Delta R.T. -0.006 min

Lab File: BP018387.D

Acq: 26 Nov 2023 02:38




Tgt Ion:178 Resp: 2907221


Ion Ratio Lower Upper

178 100

179 15.7 12.4 18.6

176 18.8 15.2 22.8





#75

1,2,3,4-Tetrachlorobenzene

Concen: 23.675 ng/uL

RT: 13.299 min Scan# 1736

Delta R.T. 0.000 min

Lab File: BP018387.D

Acq: 26 Nov 2023 02:38

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:216 Resp: 978023

Ion Ratio Lower Upper

216 100

214 76.8 62.0 93.0

218 48.4 38.6 57.8



#76

Pentachlorobenzene

Concen: 23.083 ng/uL

RT: 14.810 min Scan# 1993

Delta R.T. -0.006 min

Lab File: BP018387.D

Acq: 26 Nov 2023 02:38




Tgt Ion:250 Resp: 926304


Ion Ratio Lower Upper

250 100

248 63.4 51.0 76.6

252 64.4 51.5 77.3









#77
 Carbazole
 Concen: 19.849 ng/ul
 RT: 17.651 min Scan# 2477
 Delta R.T. 0.000 min
 Lab File: BP018387.D
 Acq: 26 Nov 2023 02:38




Instrument :
 BNA_P
 ClientSampleId :


Tgt Ion:167 Resp: 2372956
 Ion Ratio Lower Upper
 167 100
 166 20.8 17.5 26.3
 139 12.2 10.0 15.0



#78
 Di-n-butylphthalate
 Concen: 19.330 ng/ul
 RT: 18.216 min Scan# 2572
 Delta R.T. -0.006 min
 Lab File: BP018387.D
 Acq: 26 Nov 2023 02:38

Tgt Ion:149 Resp: 3214296
 Ion Ratio Lower Upper
 149 100
 150 9.3 7.4 11.2
 104 4.4 3.8 5.6





#80

Fluoranthene

Concen: 19.282 ng/ul

RT: 19.263 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP018387.D

Acq: 26 Nov 2023 02:38

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:202 Resp: 2992696

Ion Ratio Lower Upper

202 100

101 9.8 8.6 13.0

100 7.4 6.7 10.1



#82

Pyrene

Concen: 19.156 ng/ul

RT: 19.616 min Scan# 2810

Delta R.T. 0.000 min

Lab File: BP018387.D

Acq: 26 Nov 2023 02:38




Tgt Ion:202 Resp: 2993870


Ion Ratio Lower Upper

202 100

101 11.3 10.2 15.4

100 9.6 8.4 12.6





#83

Butylbenzylphthalate

Concen: 18.334 ng/ul

RT: 20.510 min Scan# 2963

Delta R.T. 0.000 min

Lab File: BP018387.D

Acq: 26 Nov 2023 02:38

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:149 Resp: 1060438

Ion Ratio Lower Upper

149 100

91 63.2 53.7 80.5

206 22.6 16.4 24.6



#84

3,3'-Dichlorobenzidine

Concen: 20.500 ng/ul

RT: 21.280 min Scan# 3093

Delta R.T. 0.006 min

Lab File: BP018387.D

Acq: 26 Nov 2023 02:38




Tgt Ion:252 Resp: 809389


Ion Ratio Lower Upper

252 100

254 63.7 51.1 76.7

126 9.8 8.6 12.8





#85

Benzo(a)anthracene

Concen: 19.531 ng/ul

RT: 21.339 min Scan# 3104

Delta R.T. 0.000 min

Lab File: BP018387.D

Acq: 26 Nov 2023 02:38

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:228 Resp: 2598374

Ion Ratio Lower Upper

228 100

229 19.8 16.2 24.2

226 27.4 22.2 33.2



#86

Bis(2-ethylhexyl)phthalate

Concen: 17.929 ng/ul

RT: 21.298 min Scan# 3096

Delta R.T. 0.006 min

Lab File: BP018387.D

Acq: 26 Nov 2023 02:38




Tgt Ion:149 Resp: 1555797


Ion Ratio Lower Upper

149 100

167 28.1 22.1 33.1

279 4.8 3.4 5.0





#87

Chrysene

Concen: 19.792 ng/ul

RT: 21.392 min Scan# 3113

Delta R.T. 0.000 min

Lab File: BP018387.D

Acq: 26 Nov 2023 02:38

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:228 Resp: 2396087

Ion Ratio Lower Upper

228 100

226 29.8 24.2 36.4

229 19.4 16.1 24.1



#89

Di-n-octyl phthalate

Concen: 16.371 ng/ul




RT: 22.263 min Scan# 3260


Delta R.T. 0.018 min

Lab File: BP018387.D

Acq: 26 Nov 2023 02:38

Tgt Ion:149 Resp: 2220697





#90

Benzo(b)fluoranthene

Concen: 18.673 ng/ul

RT: 23.051 min Scan# 31

Delta R.T. 0.012 min

Lab File: BP018387.D

Acq: 26 Nov 2023 02:38

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:252 Resp: 2499086

Ion Ratio Lower Upper

252 100

253 22.1 17.7 26.5

125 9.1 8.4 12.6



#91

Benzo(k)fluoranthene

Concen: 19.123 ng/ul

RT: 23.104 min Scan# 3403

Delta R.T. 0.018 min

Lab File: BP018387.D

Acq: 26 Nov 2023 02:38




Tgt Ion:252 Resp: 2547857


Ion Ratio Lower Upper

252 100

253 21.7 17.7 26.5

125 9.2 8.0 12.0





#93

Benzo(a)pyrene

Concen: 19.701 ng/ul

RT: 23.692 min Scan# 3

Delta R.T. 0.018 min

Lab File: BP018387.D

Acq: 26 Nov 2023 02:38

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:252 Resp: 2424844

Ion Ratio Lower Upper

252 100

253 22.1 17.7 26.5

125 10.2 8.8 13.2



#94

Indeno(1,2,3-cd)pyrene

Concen: 21.203 ng/ul

RT: 26.398 min Scan# 3963

Delta R.T. 0.053 min

Lab File: BP018387.D

Acq: 26 Nov 2023 02:38




Tgt Ion:276 Resp: 3017566


Ion Ratio Lower Upper

276 100

138 19.4 17.5 26.3

277 24.2 19.1 28.7





#95

Dibenzo(a,h)anthracene

Concen: 21.353 ng/ul

RT: 26.445 min Scan# 3962

Delta R.T. 0.065 min

Lab File: BP018387.D

Acq: 26 Nov 2023 02:38

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:278 Resp: 2487850

Ion Ratio Lower Upper

278 100

139 14.7 13.3 19.9

279 23.9 19.0 28.4



#96

Benzo(g,h,i)perylene

Concen: 21.506 ng/ul

RT: 27.198 min Scan# 4099

Delta R.T. 0.071 min

Lab File: BP018387.D

Acq: 26 Nov 2023 02:38

Tgt Ion:276 Resp: 2478694

Ion Ratio Lower Upper

276 100

138 20.8 17.7 26.5

277 24.5 18.6 28.0

