

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112423\
 Data File : BP018352.D
 Acq On : 25 Nov 2023 05:45
 Operator : MA/JU
 Sample : PB157098BL
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK098

Quant Time: Nov 27 00:12:13 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP112223.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Nov 24 22:02:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.869	152	418781	20.000	ng/u1	0.00
20) Naphthalene-d8	10.669	136	1536047	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.498	164	774632	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.245	188	1432355	20.000	ng/u1	0.00
79) Chrysene-d12	21.351	240	1468882	20.000	ng/u1	0.00
88) Perylene-d12	23.786	264	1759313	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.270	96	64287	5.750	ng/uL	0.00
4) Pyridine-d5	3.687	84	696259	22.659	ng/u1	0.00
7) Phenol-d5	7.028	99	975389	25.485	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.205	67	587297	25.662	ng/u1	0.00
11) 2-Chlorophenol-d4	7.393	132	829782	25.823	ng/u1	0.00
15) 4-Methylphenol-d8	8.575	113	729510	23.619	ng/u1	0.00
21) Nitrobenzene-d5	9.028	128	356341	28.561	ng/u1	0.00
24) 2-Nitrophenol-d4	9.752	143	320089	27.780	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.281	165	691847	24.888	ng/u1	0.00
31) 4-Chloroaniline-d4	10.805	131	849828	22.667	ng/u1	0.00
46) Dimethylphthalate-d6	13.916	166	1763344	25.583	ng/u1	0.00
49) Acenaphthylene-d8	14.193	160	2115577	26.937	ng/u1	0.00
54) 4-Nitrophenol-d4	14.687	143	214882	22.583	ng/u1	0.00
60) Fluorene-d10	15.493	176	1506189	26.126	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.604	200	126241	20.704	ng/u1	0.00
73) Anthracene-d10	17.345	188	2188300	28.583	ng/u1	0.00
81) Pyrene-d10	19.586	212	2587328	22.132	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.627	264	2933246	28.758	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112423\
Data File : BP018352.D
Acq On : 25 Nov 2023 05:45
Operator : MA/JU
Sample : PB157098BL
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SBLK098

Quant Time: Nov 27 00:12:13 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP112223.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Nov 24 22:02:00 2023
Response via : Initial Calibration

