

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112520\
 Data File : BP004108.D
 Acq On : 25 Nov 2020 21:21
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 26 00:32:43 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270E-BP110320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 23 16:42:41 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	58	0.00
2	1,4-Dioxane	40.000	37.434	6.4	58	-0.02
3	Pyridine	40.000	40.053	-0.1	60	-0.01
4	n-Nitrosodimethylamine	40.000	41.002	-2.5	62	-0.01
5 S	2-Fluorophenol	80.000	82.119	-2.6	62	0.00
6	Aniline	40.000	42.183	-5.5	63	0.00
7 S	Phenol-d6	80.000	83.757	-4.7	63	0.00
8	2-Chlorophenol	40.000	41.638	-4.1	62	0.00
9	Benzaldehyde	40.000	39.960	0.1	61	0.00
10 C	Phenol	40.000	41.329	-3.3	63	0.00
11	bis(2-Chloroethyl)ether	40.000	39.864	0.3	60	0.00
12	1,3-Dichlorobenzene	40.000	40.392	-1.0	62	0.00
13 C	1,4-Dichlorobenzene	40.000	40.458	-1.1	62	0.00
14	1,2-Dichlorobenzene	40.000	40.731	-1.8	62	0.00
15	Benzyl Alcohol	40.000	42.913	-7.3	63	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	38.439	3.9	59	0.00
17	2-Methylphenol	40.000	42.763	-6.9	64	0.00
18	Hexachloroethane	40.000	42.181	-5.5	63	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	42.742	-6.9	63	0.00
20	3+4-Methylphenols	40.000	43.497	-8.7	64	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	58	0.00
22	Acetophenone	40.000	39.984	0.0	62	0.00
23 S	Nitrobenzene-d5	80.000	79.725	0.3	61	0.00
24	Nitrobenzene	40.000	40.967	-2.4	63	0.00
25	Isophorone	40.000	42.584	-6.5	64	0.00
26 C	2-Nitrophenol	40.000	48.311	-20.8#	71	0.00
27	2,4-Dimethylphenol	40.000	41.265	-3.2	64	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.743	0.6	62	0.00
29 C	2,4-Dichlorophenol	40.000	42.392	-6.0	64	0.00
30	1,2,4-Trichlorobenzene	40.000	40.602	-1.5	64	0.00
31	Naphthalene	40.000	40.261	-0.7	63	0.00
32	Benzoic acid	40.000	30.565	23.6	45	0.00
33	4-Chloroaniline	40.000	41.744	-4.4	64	0.00
34 C	Hexachlorobutadiene	40.000	40.066	-0.2	63	0.00
35	Caprolactam	40.000	50.384	-26.0#	74	0.00
36 C	4-Chloro-3-methylphenol	40.000	43.514	-8.8	67	0.00
37	2-Methylnaphthalene	40.000	40.822	-2.1	64	0.00
38	1-Methylnaphthalene	40.000	40.915	-2.3	65	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	59	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.208	-0.5	65	0.00
41 P	Hexachlorocyclopentadiene	40.000	29.932	25.2#	46	0.00
42 S	2,4,6-Tribromophenol	80.000	88.702	-10.9	69	0.00
43 C	2,4,6-Trichlorophenol	40.000	42.155	-5.4	66	0.00
44	2,4,5-Trichlorophenol	40.000	41.467	-3.7	65	0.00
45 S	2-Fluorobiphenyl	80.000	80.254	-0.3	65	0.00
46	1,1'-Biphenyl	40.000	39.669	0.8	64	0.00
47	2-Chloronaphthalene	40.000	40.557	-1.4	65	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112520\
 Data File : BP004108.D
 Acq On : 25 Nov 2020 21:21
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 26 00:32:43 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270E-BP110320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 23 16:42:41 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	46.853	-17.1	71	0.00
49	Acenaphthylene	40.000	41.677	-4.2	67	0.00
50	Dimethylphthalate	40.000	42.034	-5.1	68	0.00
51	2,6-Dinitrotoluene	40.000	44.384	-11.0	70	0.00
52 C	Acenaphthene	40.000	40.291	-0.7	65	0.00
53	3-Nitroaniline	40.000	44.344	-10.9	69	0.00
54 P	2,4-Dinitrophenol	40.000	33.152	17.1	49	0.00
55	Dibenzofuran	40.000	40.869	-2.2	67	0.00
56 P	4-Nitrophenol	40.000	41.619	-4.0	63	0.00
57	2,4-Dinitrotoluene	40.000	47.571	-18.9	73	0.00
58	Fluorene	40.000	41.700	-4.3	68	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.873	0.3	62	0.00
60	Diethylphthalate	40.000	42.723	-6.8	69	0.00
61	4-Chlorophenyl-phenylether	40.000	40.885	-2.2	67	0.00
62	4-Nitroaniline	40.000	47.844	-19.6	75	0.00
63	Azobenzene	40.000	40.467	-1.2	65	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	64	0.00
65	4,6-Dinitro-2-methylphenol	40.000	37.732	5.7	66	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.755	0.6	69	0.00
67	4-Bromophenyl-phenylether	40.000	39.921	0.2	69	0.00
68	Hexachlorobenzene	40.000	39.981	0.0	70	0.00
69	Atrazine	40.000	40.636	-1.6	68	0.00
70 C	Pentachlorophenol	40.000	40.892	-2.2	67	0.00
71	Phenanthrene	40.000	40.358	-0.9	71	0.00
72	Anthracene	40.000	41.041	-2.6	72	0.00
73	Carbazole	40.000	42.447	-6.1	74	0.00
74	Di-n-butylphthalate	40.000	43.987	-10.0	73	0.00
75 C	Fluoranthene	40.000	42.896	-7.2	74	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	71	0.00
77	Benzydine	40.000	35.142	12.1	61	0.00
78	Pyrene	40.000	39.511	1.2	74	0.00
79 S	Terphenyl-d14	80.000	78.488	1.9	75	0.00
80	Butylbenzylphthalate	40.000	45.829	-14.6	81	0.00
81	Benzo(a)anthracene	40.000	40.912	-2.3	78	0.00
82	3,3'-Dichlorobenzidine	40.000	43.039	-7.6	80	0.00
83	Chrysene	40.000	40.806	-2.0	78	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	42.857	-7.1	78	0.00
85 c	Di-n-octyl phthalate	40.000	44.563	-11.4	80	0.00
86 I	Perylene-d12	20.000	20.000	0.0	80	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	42.767	-6.9	95	0.00
88	Benzo(b)fluoranthene	40.000	38.367	4.1	87	0.00
89	Benzo(k)fluoranthene	40.000	36.124	9.7	79	0.00
90 C	Benzo(a)pyrene	40.000	38.970	2.6	86	0.00
91	Dibenzo(a,h)anthracene	40.000	42.814	-7.0	94	0.00
92	Benzo(g,h,i)perylene	40.000	43.719	-9.3	97	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112520\
Data File : BP004108.D
Acq On : 25 Nov 2020 21:21
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Nov 26 00:32:43 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270E-BP110320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 23 16:42:41 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 1