





Method Name: Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP112719.M
Calibration Table Last Updated: Wed Nov 27 17:38:46 2019

2,4-Dinitrophenol


Response Ratio



Resp Ratio = $1.44 \times 10^{-1} \times \text{Amt} - 5.10 \times 10^{-2}$
Coef of Det (r^2) = 0.996 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP112719.M
Calibration Table Last Updated: Wed Nov 27 17:38:46 2019

4,6-Dinitro-2-methylphenol



Resp Ratio = $9.87 \times 10^{-2} \times \text{Amt} - 3.10 \times 10^{-2}$
Coef of Det (r^2) = 0.998 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP112719.M
Calibration Table Last Updated: Wed Nov 27 17:38:46 2019