


Data Path : Z:\SVOASRV\HPCHEM1\BNA P\DATA\BP112719\
 Data File : BP001083.D
 Acq On : 27 Nov 2019 11:46
 Operator : JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP112719.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Nov 27 17:38:46 2019



AutoFind: Scans 2385, 2386, 2387; Background Corrected with Scan 2376

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	37.0	42085	PASS
68	69	0.00	2	1.7	772	PASS
69	198	0.00	100	40.2	45814	PASS
70	69	0.00	2	0.6	258	PASS
127	198	10	80	44.8	51013	PASS
197	198	0.00	2	0.2	235	PASS
198	198	100	100	100.0	113861	PASS
199	198	5	9	6.9	7840	PASS
275	198	10	60	23.6	26826	PASS
365	198	1	100	2.7	3102	PASS
441	198	0.01	100	12.3	13982	PASS
442	198	50	100	81.2	92432	PASS
443	442	15	24	18.9	17515	PASS