

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112723\
 Data File : BP018398.D
 Acq On : 27 Nov 2023 13:49
 Operator : MA/JU
 Sample : 05328-13
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 A49W2

Quant Time: Nov 27 21:55:42 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP112223.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Nov 27 21:50:01 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.863	152	480077	20.000	ng/u1	0.00
20) Naphthalene-d8	10.663	136	1826536	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.493	164	985853	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.245	188	1692266	20.000	ng/u1	0.00
79) Chrysene-d12	21.375	240	1608414	20.000	ng/u1	0.00
88) Perylene-d12	23.874	264	1944847	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.270	96	49462	3.859	ng/uL	0.00
4) Pyridine-d5	3.687	84	285171	8.096	ng/u1	0.00
7) Phenol-d5	7.022	99	173551	3.956	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.193	67	530703	20.228	ng/u1	0.00
11) 2-Chlorophenol-d4	7.393	132	557992	15.147	ng/u1	0.00
15) 4-Methylphenol-d8	8.569	113	333322	9.414	ng/u1	0.00
21) Nitrobenzene-d5	9.022	128	338893	22.843	ng/u1	0.00
24) 2-Nitrophenol-d4	9.746	143	297900	21.742	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.281	165	610931	18.482	ng/u1	0.00
31) 4-Chloroaniline-d4	10.799	131	822231	18.443	ng/u1	0.00
46) Dimethylphthalate-d6	13.910	166	2028334	23.122	ng/u1	0.00
49) Acenaphthylene-d8	14.187	160	2232009	22.330	ng/u1	0.00
54) 4-Nitrophenol-d4	14.687	143	28484	2.352	ng/u1	0.00
60) Fluorene-d10	15.487	176	1627351	22.179	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.598	200	134944	18.732	ng/u1	0.00
73) Anthracene-d10	17.345	188	2268811	25.083	ng/u1	0.00
81) Pyrene-d10	19.592	212	2577583	20.136	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.710	264	2951415	26.176	ng/u1	0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112723\
Data File : BP018398.D
Acq On : 27 Nov 2023 13:49
Operator : MA/JU
Sample : 05328-13
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
A49W2

Quant Time: Nov 27 21:55:42 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP112223.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Nov 27 21:50:01 2023
Response via : Initial Calibration

