

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112723\
 Data File : BP018416.D
 Acq On : 28 Nov 2023 00:34
 Operator : MA/JU
 Sample : PB157281BL
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK281

Quant Time: Nov 28 03:18:05 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP112223.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Nov 28 02:41:10 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.857	152	502008	20.000	ng/u1	0.00
20) Naphthalene-d8	10.657	136	1989131	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.492	164	1136747	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.245	188	2075776	20.000	ng/u1	0.00
79) Chrysene-d12	21.386	240	1954342	20.000	ng/u1	0.00
88) Perylene-d12	23.921	264	2295277	20.000	ng/u1	0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.269	96	87625	6.538	ng/uL	0.00
4) Pyridine-d5	3.687	84	1150913	31.246	ng/u1	0.00
7) Phenol-d5	7.022	99	1413997	30.820	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.193	67	824606	30.057	ng/u1	0.00
11) 2-Chlorophenol-d4	7.387	132	1208491	31.373	ng/u1	0.00
15) 4-Methylphenol-d8	8.569	113	1132297	30.582	ng/u1	0.00
21) Nitrobenzene-d5	9.022	128	570523	35.312	ng/u1	0.00
24) 2-Nitrophenol-d4	9.740	143	575228	38.551	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.275	165	1131877	31.442	ng/u1	0.00
31) 4-Chloroaniline-d4	10.798	131	1603734	33.031	ng/u1	0.00
46) Dimethylphthalate-d6	13.910	166	3227032	31.904	ng/u1	0.00
49) Acenaphthylene-d8	14.187	160	3718564	32.265	ng/u1	0.00
54) 4-Nitrophenol-d4	14.687	143	398681	28.552	ng/u1	0.00
60) Fluorene-d10	15.486	176	2658022	31.418	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.598	200	261278	29.568	ng/u1	0.00
73) Anthracene-d10	17.345	188	3798424	34.236	ng/u1	0.00
81) Pyrene-d10	19.592	212	4263739	27.412	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.751	264	4715720	35.438	ng/u1	0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112723\
Data File : BP018416.D
Acq On : 28 Nov 2023 00:34
Operator : MA/JU
Sample : PB157281BL
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SBLK281

Quant Time: Nov 28 03:18:05 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP112223.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Nov 28 02:41:10 2023
Response via : Initial Calibration

