

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112723\
 Data File : BP018410.D
 Acq On : 27 Nov 2023 20:37
 Operator : MA/JU
 Sample : 05332-14
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 YCG66

Quant Time: Nov 27 22:30:54 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP112223.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Nov 27 21:50:01 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.857	152	454829	20.000	ng/u1	0.00
20) Naphthalene-d8	10.657	136	1694468	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.492	164	892523	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.245	188	1743000	20.000	ng/u1	0.00
79) Chrysene-d12	21.380	240	1799326	20.000	ng/u1	0.00
88) Perylene-d12	23.898	264	2096380	20.000	ng/u1	0.03
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.269	96	41432	3.412	ng/uL	0.00
4) Pyridine-d5	3.687	84	213561	6.399	ng/u1	0.00
7) Phenol-d5	7.022	99	193528	4.656	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.193	67	586287	23.587	ng/u1	0.00
11) 2-Chlorophenol-d4	7.387	132	631775	18.102	ng/u1	0.00
15) 4-Methylphenol-d8	8.569	113	368209	10.977	ng/u1	0.00
21) Nitrobenzene-d5	9.022	128	379568	27.579	ng/u1	0.00
24) 2-Nitrophenol-d4	9.740	143	329268	25.905	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.275	165	650307	21.206	ng/u1	0.00
31) 4-Chloroaniline-d4	10.798	131	698485	16.888	ng/u1	0.00
46) Dimethylphthalate-d6	13.910	166	2034666	25.620	ng/u1	0.00
49) Acenaphthylene-d8	14.187	160	2346930	25.935	ng/u1	0.00
54) 4-Nitrophenol-d4	14.686	143	38234	3.487	ng/u1	0.00
60) Fluorene-d10	15.486	176	1736404	26.140	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.598	200	172229	23.212	ng/u1	0.00
73) Anthracene-d10	17.345	188	2968321	31.862	ng/u1	0.00
81) Pyrene-d10	19.592	212	3644708	25.451	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.733	264	4013134	33.019	ng/u1	0.04

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112723\
Data File : BP018410.D
Acq On : 27 Nov 2023 20:37
Operator : MA/JU
Sample : 05332-14
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
YCG66

Quant Time: Nov 27 22:30:54 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP112223.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Nov 27 21:50:01 2023
Response via : Initial Calibration

