

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112725\
 Data File : BP026195.D
 Acq On : 26 Nov 2025 19:55
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 27 00:24:52 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 19 01:25:31 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	81	-0.01
2	1,4-Dioxane	0.505	0.478	5.3	79	0.00
3	Pyridine	1.461	1.361	6.8	76	0.00
4	n-Nitrosodimethylamine	0.546	0.510	6.6	76	0.00
5 S	2-Fluorophenol	1.246	1.165	6.5	76	-0.01
6	Aniline	2.094	1.991	4.9	76	0.00
7 S	Phenol-d6	1.586	1.532	3.4	76	-0.01
8	2-Chlorophenol	1.418	1.400	1.3	79	-0.01
9	Benzaldehyde	0.831	0.923	-11.1	81	0.00
10 C	Phenol	1.709	1.648	3.6	76	-0.01
11	bis(2-Chloroethyl)ether	1.338	1.290	3.6	77	0.00
12	1,3-Dichlorobenzene	1.528	1.518	0.7	80	-0.02
13 C	1,4-Dichlorobenzene	1.540	1.530	0.6	80	0.00
14	1,2-Dichlorobenzene	1.479	1.492	-0.9	81	0.00
15	Benzyl Alcohol	1.162	1.132	2.6	77	-0.01
16	2,2'-oxybis(1-Chloropropane	1.879	1.843	1.9	78	-0.02
17	2-Methylphenol	1.163	1.133	2.6	77	-0.02
18	Hexachloroethane	0.560	0.563	-0.5	81	-0.01
19 P	n-Nitroso-di-n-propylamine	0.985	0.987	-0.2	78	-0.01
20	3+4-Methylphenols	1.612	1.560	3.2	78	-0.02
21 I	Naphthalene-d8	1.000	1.000	0.0	81	-0.02
22	Acetophenone	0.508	0.491	3.3	78	-0.01
23 S	Nitrobenzene-d5	0.352	0.341	3.1	77	-0.02
24	Nitrobenzene	0.356	0.347	2.5	78	-0.02
25	Isophorone	0.698	0.691	1.0	78	-0.02
26 C	2-Nitrophenol	0.180	0.183	-1.7	78	-0.02
27	2,4-Dimethylphenol	0.325	0.291	10.5	72	-0.01
28	bis(2-Chloroethoxy)methane	0.437	0.426	2.5	78	-0.02
29 C	2,4-Dichlorophenol	0.325	0.331	-1.8	81	-0.02
30	1,2,4-Trichlorobenzene	0.331	0.339	-2.4	83	-0.01
31	Naphthalene	1.081	1.084	-0.3	81	-0.02
32	Benzoic acid	0.274	0.246	10.2	72	-0.02
33	4-Chloroaniline	0.460	0.448	2.6	77	-0.01
34 C	Hexachlorobutadiene	0.215	0.226	-5.1	85	-0.02
35	Caprolactam	0.116	0.113	2.6	78	-0.01
36 C	4-Chloro-3-methylphenol	0.345	0.350	-1.4	80	-0.01
37	2-Methylnaphthalene	0.766	0.770	-0.5	80	-0.02
38	1-Methylnaphthalene	0.749	0.765	-2.1	81	-0.01
39 I	Acenaphthene-d10	1.000	1.000	0.0	82	-0.02
40	1,2,4,5-Tetrachlorobenzene	0.605	0.607	-0.3	83	-0.03
41 P	Hexachlorocyclopentadiene	0.396	0.320	19.2	66	-0.02
42 S	2,4,6-Tribromophenol	0.259	0.277	-6.9	86	-0.02
43 C	2,4,6-Trichlorophenol	0.415	0.421	-1.4	82	-0.01
44	2,4,5-Trichlorophenol	0.463	0.471	-1.7	83	-0.02
45 S	2-Fluorobiphenyl	1.365	1.390	-1.8	85	-0.02
46	1,1'-Biphenyl	1.506	1.488	1.2	82	-0.02
47	2-Chloronaphthalene	1.184	1.172	1.0	82	-0.03

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112725\
 Data File : BP026195.D
 Acq On : 26 Nov 2025 19:55
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 27 00:24:52 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 19 01:25:31 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.329	0.316	4.0	77	-0.01
49	Acenaphthylene	1.953	1.937	0.8	81	-0.02
50	Dimethylphthalate	1.494	1.482	0.8	81	-0.02
51	2,6-Dinitrotoluene	0.296	0.308	-4.1	80	-0.02
52 C	Acenaphthene	1.145	1.221	-6.6	88	-0.01
53	3-Nitroaniline	0.349	0.334	4.3	76	-0.03
54 P	2,4-Dinitrophenol	0.179	0.167	6.7	74	-0.02
55	Dibenzofuran	1.773	1.776	-0.2	82	-0.01
56 P	4-Nitrophenol	0.315	0.290	7.9	75	-0.03
57	2,4-Dinitrotoluene	0.442	0.458	-3.6	80	-0.02
58	Fluorene	1.401	1.420	-1.4	84	-0.01
59	2,3,4,6-Tetrachlorophenol	0.393	0.401	-2.0	82	-0.02
60	Diethylphthalate	1.505	1.487	1.2	80	-0.02
61	4-Chlorophenyl-phenylether	0.697	0.724	-3.9	86	-0.02
62	4-Nitroaniline	0.362	0.332	8.3	74	-0.01
63	Azobenzene	1.269	1.251	1.4	79	-0.02
64 I	Phenanthrene-d10	1.000	1.000	0.0	82	-0.02
65	4,6-Dinitro-2-methylphenol	0.127	0.123	3.1	75	-0.02
66 c	n-Nitrosodiphenylamine	0.619	0.627	-1.3	82	-0.02
67	4-Bromophenyl-phenylether	0.221	0.235	-6.3	84	-0.02
68	Hexachlorobenzene	0.257	0.272	-5.8	84	-0.02
69	Atrazine	0.233	0.232	0.4	80	-0.02
70 C	Pentachlorophenol	0.180	0.187	-3.9	84	-0.02
71	Phenanthrene	1.127	1.136	-0.8	83	-0.02
72	Anthracene	1.149	1.169	-1.7	82	-0.03
73	Carbazole	1.091	1.067	2.2	79	-0.02
74	Di-n-butylphthalate	1.294	1.311	-1.3	79	-0.03
75 C	Fluoranthene	1.369	1.379	-0.7	83	-0.02
76 I	Chrysene-d12	1.000	1.000	0.0	85	-0.02
77	Benzidine	0.635	0.577	9.1	80	-0.02
78	Pyrene	1.311	1.314	-0.2	84	-0.02
79 S	Terphenyl-d14	0.981	1.000	-1.9	86	-0.02
80	Butylbenzylphthalate	0.578	0.569	1.6	80	-0.02
81	Benzo(a)anthracene	1.374	1.351	1.7	83	-0.02
82	3,3'-Dichlorobenzidine	0.500	0.483	3.4	82	-0.04
83	Chrysene	1.295	1.284	0.8	85	-0.02
84	Bis(2-ethylhexyl)phthalate	0.874	0.864	1.1	79	-0.02
85 c	Di-n-octyl phthalate	1.529	1.486	2.8	80	-0.04
86 I	Perylene-d12	1.000	1.000	0.0	87	-0.05
87	Indeno(1,2,3-cd)pyrene	1.500	1.513	-0.9	87	-0.10
88	Benzo(b)fluoranthene	1.218	1.205	1.1	85	-0.04
89	Benzo(k)fluoranthene	1.217	1.260	-3.5	91	-0.04
90 C	Benzo(a)pyrene	1.154	1.165	-1.0	87	-0.05
91	Dibenzo(a,h)anthracene	1.228	1.248	-1.6	88	-0.09
92	Benzo(g,h,i)perylene	1.220	1.242	-1.8	89	-0.12

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112725\
Data File : BP026195.D
Acq On : 26 Nov 2025 19:55
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Nov 27 00:24:52 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 19 01:25:31 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0