

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112725\
 Data File : BP026195.D
 Acq On : 26 Nov 2025 19:55
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 27 00:24:52 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 19 01:25:31 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	81	-0.01
2	1,4-Dioxane	40.000	37.802	5.5	79	0.00
3	Pyridine	40.000	37.283	6.8	76	0.00
4	n-Nitrosodimethylamine	40.000	37.376	6.6	76	0.00
5 S	2-Fluorophenol	80.000	74.825	6.5	76	-0.01
6	Aniline	40.000	38.017	5.0	76	0.00
7 S	Phenol-d6	80.000	77.256	3.4	76	-0.01
8	2-Chlorophenol	40.000	39.495	1.3	79	-0.01
9	Benzaldehyde	40.000	44.435	-11.1	81	0.00
10 C	Phenol	40.000	38.583	3.5	76	-0.01
11	bis(2-Chloroethyl)ether	40.000	38.577	3.6	77	0.00
12	1,3-Dichlorobenzene	40.000	39.731	0.7	80	-0.02
13 C	1,4-Dichlorobenzene	40.000	39.748	0.6	80	0.00
14	1,2-Dichlorobenzene	40.000	40.364	-0.9	81	0.00
15	Benzyl Alcohol	40.000	38.956	2.6	77	-0.01
16	2,2'-oxybis(1-Chloropropane	40.000	39.233	1.9	78	-0.02
17	2-Methylphenol	40.000	38.955	2.6	77	-0.02
18	Hexachloroethane	40.000	40.197	-0.5	81	-0.01
19 P	n-Nitroso-di-n-propylamine	40.000	40.067	-0.2	78	-0.01
20	3+4-Methylphenols	40.000	38.705	3.2	78	-0.02
21 I	Naphthalene-d8	20.000	20.000	0.0	81	-0.02
22	Acetophenone	40.000	38.641	3.4	78	-0.01
23 S	Nitrobenzene-d5	80.000	77.561	3.0	77	-0.02
24	Nitrobenzene	40.000	38.970	2.6	78	-0.02
25	Isophorone	40.000	39.576	1.1	78	-0.02
26 C	2-Nitrophenol	40.000	40.543	-1.4	78	-0.02
27	2,4-Dimethylphenol	40.000	35.830	10.4	72	-0.01
28	bis(2-Chloroethoxy)methane	40.000	38.987	2.5	78	-0.02
29 C	2,4-Dichlorophenol	40.000	40.727	-1.8	81	-0.02
30	1,2,4-Trichlorobenzene	40.000	40.993	-2.5	83	-0.01
31	Naphthalene	40.000	40.097	-0.2	81	-0.02
32	Benzoic acid	40.000	36.010	10.0	72	-0.02
33	4-Chloroaniline	40.000	38.988	2.5	77	-0.01
34 C	Hexachlorobutadiene	40.000	41.982	-5.0	85	-0.02
35	Caprolactam	40.000	39.011	2.5	78	-0.01
36 C	4-Chloro-3-methylphenol	40.000	40.520	-1.3	80	-0.01
37	2-Methylnaphthalene	40.000	40.194	-0.5	80	-0.02
38	1-Methylnaphthalene	40.000	40.850	-2.1	81	-0.01
39 I	Acenaphthene-d10	20.000	20.000	0.0	82	-0.02
40	1,2,4,5-Tetrachlorobenzene	40.000	40.119	-0.3	83	-0.03
41 P	Hexachlorocyclopentadiene	40.000	32.366	19.1	66	-0.02
42 S	2,4,6-Tribromophenol	80.000	85.317	-6.6	86	-0.02
43 C	2,4,6-Trichlorophenol	40.000	40.665	-1.7	82	-0.01
44	2,4,5-Trichlorophenol	40.000	40.705	-1.8	83	-0.02
45 S	2-Fluorobiphenyl	80.000	81.471	-1.8	85	-0.02
46	1,1'-Biphenyl	40.000	39.526	1.2	82	-0.02
47	2-Chloronaphthalene	40.000	39.580	1.1	82	-0.03

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112725\
 Data File : BP026195.D
 Acq On : 26 Nov 2025 19:55
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 27 00:24:52 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 19 01:25:31 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	38.408	4.0	77	-0.01
49	Acenaphthylene	40.000	39.671	0.8	81	-0.02
50	Dimethylphthalate	40.000	39.679	0.8	81	-0.02
51	2,6-Dinitrotoluene	40.000	41.588	-4.0	80	-0.02
52 C	Acenaphthene	40.000	42.653	-6.6	88	-0.01
53	3-Nitroaniline	40.000	38.355	4.1	76	-0.03
54 P	2,4-Dinitrophenol	40.000	37.384	6.5	74	-0.02
55	Dibenzofuran	40.000	40.066	-0.2	82	-0.01
56 P	4-Nitrophenol	40.000	36.736	8.2	75	-0.03
57	2,4-Dinitrotoluene	40.000	41.473	-3.7	80	-0.02
58	Fluorene	40.000	40.533	-1.3	84	-0.01
59	2,3,4,6-Tetrachlorophenol	40.000	40.849	-2.1	82	-0.02
60	Diethylphthalate	40.000	39.521	1.2	80	-0.02
61	4-Chlorophenyl-phenylether	40.000	41.559	-3.9	86	-0.02
62	4-Nitroaniline	40.000	36.696	8.3	74	-0.01
63	Azobenzene	40.000	39.436	1.4	79	-0.02
64 I	Phenanthrene-d10	20.000	20.000	0.0	82	-0.02
65	4,6-Dinitro-2-methylphenol	40.000	38.788	3.0	75	-0.02
66 c	n-Nitrosodiphenylamine	40.000	40.548	-1.4	82	-0.02
67	4-Bromophenyl-phenylether	40.000	42.431	-6.1	84	-0.02
68	Hexachlorobenzene	40.000	42.306	-5.8	84	-0.02
69	Atrazine	40.000	39.834	0.4	80	-0.02
70 C	Pentachlorophenol	40.000	41.521	-3.8	84	-0.02
71	Phenanthrene	40.000	40.321	-0.8	83	-0.02
72	Anthracene	40.000	40.680	-1.7	82	-0.03
73	Carbazole	40.000	39.094	2.3	79	-0.02
74	Di-n-butylphthalate	40.000	40.536	-1.3	79	-0.03
75 C	Fluoranthene	40.000	40.307	-0.8	83	-0.02
76 I	Chrysene-d12	20.000	20.000	0.0	85	-0.02
77	Benzydine	40.000	36.355	9.1	80	-0.02
78	Pyrene	40.000	40.066	-0.2	84	-0.02
79 S	Terphenyl-d14	80.000	81.579	-2.0	86	-0.02
80	Butylbenzylphthalate	40.000	39.436	1.4	80	-0.02
81	Benzo(a)anthracene	40.000	39.342	1.6	83	-0.02
82	3,3'-Dichlorobenzidine	40.000	38.686	3.3	82	-0.04
83	Chrysene	40.000	39.673	0.8	85	-0.02
84	Bis(2-ethylhexyl)phthalate	40.000	39.539	1.2	79	-0.02
85 c	Di-n-octyl phthalate	40.000	38.883	2.8	80	-0.04
86 I	Perylene-d12	20.000	20.000	0.0	87	-0.05
87	Indeno(1,2,3-cd)pyrene	40.000	40.361	-0.9	87	-0.10
88	Benzo(b)fluoranthene	40.000	39.570	1.1	85	-0.04
89	Benzo(k)fluoranthene	40.000	41.433	-3.6	91	-0.04
90 C	Benzo(a)pyrene	40.000	40.405	-1.0	87	-0.05
91	Dibenzo(a,h)anthracene	40.000	40.655	-1.6	88	-0.09
92	Benzo(g,h,i)perylene	40.000	40.717	-1.8	89	-0.12

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112725\
Data File : BP026195.D
Acq On : 26 Nov 2025 19:55
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Nov 27 00:24:52 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 19 01:25:31 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0