

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112922\
 Data File : BP012844.D
 Acq On : 29 Nov 2022 10:39
 Operator : CG/JU
 Sample : PB149186BS
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :

Quant Time: Nov 29 21:56:08 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP112522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun Nov 27 22:55:08 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.016	152	471402	20.000	ng/u1	0.00
20) Naphthalene-d8	10.828	136	2151466	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.651	164	1460866	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.404	188	3209323	20.000	ng/u1	0.00
79) Chrysene-d12	21.492	240	3236965	20.000	ng/u1	0.00
88) Perylene-d12	23.998	264	2961400	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.393	96	60386	5.489	ng/uL	0.00
4) Pyridine-d5	3.816	84	857419	26.045	ng/u1	0.00
7) Phenol-d5	7.163	99	1189839	30.716	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.334	67	724205	30.341	ng/u1	0.00
11) 2-Chlorophenol-d4	7.540	132	917341	31.079	ng/u1	0.00
15) 4-Methylphenol-d8	8.716	113	950508	30.455	ng/u1	0.00
21) Nitrobenzene-d5	9.181	128	439008	33.339	ng/u1	0.00
24) 2-Nitrophenol-d4	9.904	143	485676	34.735	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.445	165	998451	30.605	ng/u1	0.00
31) 4-Chloroaniline-d4	10.963	131	1307456	29.266	ng/u1	0.00
46) Dimethylphthalate-d6	14.057	166	3117969	31.008	ng/u1	0.00
49) Acenaphthylene-d8	14.345	160	3541021	30.181	ng/u1	0.00
54) 4-Nitrophenol-d4	14.833	143	585891	32.681	ng/u1	0.00
60) Fluorene-d10	15.639	176	2760735	31.350	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.751	200	500930	32.657	ng/u1	0.00
73) Anthracene-d10	17.504	188	4357944	30.848	ng/u1	0.00
81) Pyrene-d10	19.727	212	5233557	27.825	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.839	264	4616357	31.736	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.428	88	121503	10.284	ng/uL	97
5) Pyridine	3.834	79	853689	26.502	ng/u1	96
6) Benzaldehyde	7.146	77	657907	34.785	ng/u1	99
8) Phenol	7.187	94	1191006	29.322	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.428	93	952856	28.632	ng/u1	99
12) 2-Chlorophenol	7.575	128	935500	29.137	ng/u1	100
13) 2-Methylphenol	8.451	108	908350	28.774	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.545	45	1352897	28.184	ng/u1	99
16) Acetophenone	8.840	105	1482411	29.968	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.828	70	721861	29.623	ng/u1	97
18) 4-Methylphenol	8.781	108	1008891	29.019	ng/u1	98
19) Hexachloroethane	9.104	117	353616	28.720	ng/u1	99
22) Nitrobenzene	9.222	77	1050365	30.413	ng/u1	99
23) Isophorone	9.751	82	2097392	27.547	ng/u1	100
25) 2-Nitrophenol	9.934	139	530632	31.617	ng/u1	98
26) 2,4-Dimethylphenol	9.992	107	1002057	26.869	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.234	93	1308504	28.032	ng/u1	99
29) 2,4-Dichlorophenol	10.469	162	973124	28.794	ng/u1	99
30) Naphthalene	10.881	128	3190726	28.322	ng/u1	100
32) 4-Chloroaniline	10.986	127	1277983	27.854	ng/u1	100
33) Hexachlorobutadiene	11.169	225	617679	27.439	ng/u1	98
34) Caprolactam	11.757	113	183522	17.684	ng/u1	96
35) 4-Chloro-3-methylphenol	12.098	107	1028804	30.061	ng/u1	99
36) 2-Methylnaphthalene	12.481	142	2215023	28.552	ng/u1	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112922\
 Data File : BP012844.D
 Acq On : 29 Nov 2022 10:39
 Operator : CG/JU
 Sample : PB149186BS
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :

Quant Time: Nov 29 21:56:08 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP112522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun Nov 27 22:55:08 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.698	142	2249262	28.929	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.845	216	1274736	27.364	ng/ul	99
40) Hexachlorocyclopentadiene	12.828	237	555808	24.762	ng/ul	99
41) 2,4,6-Trichlorophenol	13.081	196	834512	28.630	ng/ul	100
42) 2,4,5-Trichlorophenol	13.151	196	910017	29.042	ng/ul	99
43) 1,1'-Biphenyl	13.486	154	2991789	28.113	ng/ul	99
44) 2-Chloronaphthalene	13.528	162	2390008	28.073	ng/ul	100
45) 2-Nitroaniline	13.728	65	602973	36.199	ng/ul	98
47) Dimethylphthalate	14.104	163	3083394	29.253	ng/ul	99
48) 2,6-Dinitrotoluene	14.222	165	577231	34.993	ng/ul	98
50) Acenaphthylene	14.375	152	3694366	28.797	ng/ul	100
51) 3-Nitroaniline	14.551	138	588845	30.823	ng/ul	98
52) Acenaphthene	14.716	153	2581505	28.582	ng/ul	99
53) 2,4-Dinitrophenol	14.751	184	305041	28.777	ng/ul	96
55) 4-Nitrophenol	14.851	109	402729	30.920	ng/ul	94
56) Dibenzofuran	15.045	168	3643858	29.269	ng/ul	100
57) 2,4-Dinitrotoluene	15.004	165	866717	35.605	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	15.269	232	821400	30.414	ng/ul	99
59) Diethylphthalate	15.469	149	3020310	29.602	ng/ul	99
61) Fluorene	15.698	166	3000076	30.049	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.692	204	1534052	29.614	ng/ul	99
63) 4-Nitroaniline	15.716	138	599606	36.825	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.769	198	516574	31.375	ng/ul	95
67) N-Nitrosodiphenylamine	15.904	169	2605295	28.564	ng/ul	99
68) 4-Bromophenyl-phenylether	16.586	248	956573	30.284	ng/ul	99
69) Hexachlorobenzene	16.704	284	1157998	28.764	ng/ul	97
70) Atrazine	16.857	200	917901	29.332	ng/ul	100
71) Pentachlorophenol	17.045	266	694940	28.109	ng/ul	99
72) Phenanthrene	17.445	178	5003286	29.251	ng/ul	99
74) Anthracene	17.539	178	4974121	29.395	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.451	216	1270298	25.603	ng/uL	99
76) Pentachlorobenzene	14.963	250	1262869	24.486	ng/uL	99
77) Carbazole	17.804	167	4560108	31.855	ng/ul	100
78) Di-n-butylphthalate	18.351	149	5084004	30.001	ng/ul	100
80) Fluoranthene	19.404	202	6007480	25.678	ng/ul	99
82) Pyrene	19.757	202	6211660	25.671	ng/ul	100
83) Butylbenzylphthalate	20.627	149	2041873	36.061	ng/ul	92
84) 3,3'-Dichlorobenzidine	21.398	252	1853588	31.182	ng/ul	99
85) Benzo(a)anthracene	21.474	228	6095227	27.225	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.386	149	2939099	33.942	ng/ul	99
87) Chrysene	21.527	228	5718359	26.416	ng/ul	100
89) Di-n-octyl phthalate	22.351	149	4699008	27.786	ng/ul	100
90) Benzo(b)fluoranthene	23.233	252	5904612	26.925	ng/ul	100
91) Benzo(k)fluoranthene	23.280	252	5792537	26.462	ng/ul	99
93) Benzo(a)pyrene	23.892	252	5030983	27.899	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.633	276	5485567	31.531	ng/ul	99
95) Dibenzo(a,h)anthracene	26.650	278	4844894	30.690	ng/ul	99
96) Benzo(g,h,i)perylene	27.450	276	4627361	30.923	ng/ul	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112922\
Data File : BP012844.D
Acq On : 29 Nov 2022 10:39
Operator : CG/JU
Sample : PB149186BS
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :

Quant Time: Nov 29 21:56:08 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP112522.M
Quant Title : SVOA CALIBRATION
QLast Update : Sun Nov 27 22:55:08 2022
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 10.284 ng/uL
RT: 3.428 min Scan# 41
Delta R.T. -0.006 min
Lab File: BP012844.D
Acq: 29 Nov 2022 10:39




Instrument :
BNA_P
ClientSampleId :


Tgt Ion: 88 Resp: 121503
Ion Ratio Lower Upper
88 100
43 29.5 22.6 34.0
58 77.4 59.6 89.4



#5
Pyridine
Concen: 26.502 ng/uL
RT: 3.834 min Scan# 110
Delta R.T. -0.006 min
Lab File: BP012844.D
Acq: 29 Nov 2022 10:39

Tgt Ion: 79 Resp: 853689
Ion Ratio Lower Upper
79 100
52 77.9 58.9 88.3
51 34.8 27.3 40.9









#6
Benzaldehyde
Concen: 34.785 ng/ul
RT: 7.146 min Scan# 61
Delta R.T. -0.006 min
Lab File: BP012844.D
Acq: 29 Nov 2022 10:39




Instrument :
BNA_P
ClientSampleId :


Tgt Ion: 77 Resp: 657907
Ion Ratio Lower Upper
77 100
105 100.6 81.4 122.2
106 98.5 77.4 116.0



#8
Phenol
Concen: 29.322 ng/ul
RT: 7.187 min Scan# 680
Delta R.T. -0.006 min
Lab File: BP012844.D
Acq: 29 Nov 2022 10:39

Tgt Ion: 94 Resp: 1191006
Ion Ratio Lower Upper
94 100
65 26.8 20.9 31.3
66 32.8 26.0 39.0









#10
Bis(2-Chloroethyl)ether
Concen: 28.632 ng/ul
RT: 7.428 min Scan# 721
Delta R.T. -0.006 min
Lab File: BP012844.D
Acq: 29 Nov 2022 10:39




Instrument :
BNA_P
ClientSampleId :


Tgt Ion: 93 Resp: 952856
Ion Ratio Lower Upper
93 100
63 75.5 59.7 89.5
95 32.4 25.8 38.6



#12
2-Chlorophenol
Concen: 29.137 ng/ul
RT: 7.575 min Scan# 746
Delta R.T. -0.000 min
Lab File: BP012844.D
Acq: 29 Nov 2022 10:39

Tgt Ion:128 Resp: 935500
Ion Ratio Lower Upper
128 100
64 44.4 35.6 53.4
130 32.4 26.0 39.0





#13

2-Methylphenol

Concen: 28.774 ng/ul

RT: 8.451 min Scan# 895

Delta R.T. -0.006 min

Lab File: BP012844.D

Acq: 29 Nov 2022 10:39

Instrument :

BNA_P





ClientSampleId :

Tgt Ion:108 Resp: 908350

Ion Ratio Lower Upper

108 100

107 88.5 69.3 103.9



#14

2,2'-oxybis(1-Chloropropane)

Concen: 28.184 ng/ul

RT: 8.545 min Scan# 911

Delta R.T. -0.000 min

Lab File: BP012844.D

Acq: 29 Nov 2022 10:39




Tgt Ion: 45 Resp: 1352897


Ion Ratio Lower Upper

45 100

77 13.7 11.3 16.9

79 10.8 8.6 12.8









#16
Acetophenone
Concen: 29.968 ng/ul
RT: 8.840 min Scan# 962
Delta R.T. -0.006 min
Lab File: BP012844.D
Acq: 29 Nov 2022 10:39




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:105 Resp: 1482411
Ion Ratio Lower Upper
105 100
77 73.3 59.7 89.5
51 25.7 20.9 31.3



#17
N-Nitroso-di-n-propylamine
Concen: 29.623 ng/ul
RT: 8.828 min Scan# 959
Delta R.T. -0.006 min
Lab File: BP012844.D
Acq: 29 Nov 2022 10:39

Tgt Ion: 70 Resp: 721861
Ion Ratio Lower Upper
70 100
42 55.8 46.1 69.1
101 9.5 7.4 11.2
130 21.5 15.8 23.6









#18
4-Methylphenol
Concen: 29.019 ng/ul
RT: 8.781 min Scan# 91
Delta R.T. -0.006 min
Lab File: BP012844.D
Acq: 29 Nov 2022 10:39




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:108 Resp: 1008891
Ion Ratio Lower Upper
108 100
107 110.4 90.2 135.2



#19
Hexachloroethane
Concen: 28.720 ng/ul
RT: 9.104 min Scan# 1006
Delta R.T. -0.006 min
Lab File: BP012844.D
Acq: 29 Nov 2022 10:39

Tgt Ion:117 Resp: 353616
Ion Ratio Lower Upper
117 100
201 105.9 85.7 128.5
199 67.0 52.9 79.3









#22
Nitrobenzene
Concen: 30.413 ng/ul
RT: 9.222 min Scan# 1027
Delta R.T. -0.006 min
Lab File: BP012844.D
Acq: 29 Nov 2022 10:39




Instrument :
BNA_P
ClientSampleId :


Tgt Ion: 77 Resp: 1050365
Ion Ratio Lower Upper
77 100
123 48.2 38.2 57.2
65 12.8 10.2 15.4



#23
Isophorone
Concen: 27.547 ng/ul
RT: 9.751 min Scan# 1116
Delta R.T. -0.006 min
Lab File: BP012844.D
Acq: 29 Nov 2022 10:39

Tgt Ion: 82 Resp: 2097392
Ion Ratio Lower Upper
82 100
95 6.6 5.2 7.8
138 18.3 14.5 21.7









#25
2-Nitrophenol
Concen: 31.617 ng/ul
RT: 9.934 min Scan# 1147
Delta R.T. -0.006 min
Lab File: BP012844.D
Acq: 29 Nov 2022 10:39




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:139 Resp: 530632
Ion Ratio Lower Upper
139 100
65 48.7 39.5 59.3
109 25.4 21.4 32.0



#26
2,4-Dimethylphenol
Concen: 26.869 ng/ul
RT: 9.992 min Scan# 1157
Delta R.T. -0.006 min
Lab File: BP012844.D
Acq: 29 Nov 2022 10:39

Tgt Ion:107 Resp: 1002057
Ion Ratio Lower Upper
107 100
121 52.0 41.8 62.6
122 89.2 72.3 108.5









#27
 Bis(2-Chloroethoxy)methane
 Concen: 28.032 ng/ul
 RT: 10.234 min Scan# 1198
 Delta R.T. -0.006 min
 Lab File: BP012844.D
 Acq: 29 Nov 2022 10:39




Instrument :
 BNA_P
 ClientSampleId :


Tgt Ion: 93 Resp: 1308504
 Ion Ratio Lower Upper
 93 100
 95 31.9 26.1 39.1
 123 11.5 9.0 13.6



#29
 2,4-Dichlorophenol
 Concen: 28.794 ng/ul
 RT: 10.469 min Scan# 1238
 Delta R.T. -0.006 min
 Lab File: BP012844.D
 Acq: 29 Nov 2022 10:39

Tgt Ion:162 Resp: 973124
 Ion Ratio Lower Upper
 162 100
 164 65.2 52.9 79.3
 98 34.0 26.6 40.0









#30
Naphthalene
Concen: 28.322 ng/ul
RT: 10.881 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP012844.D
Acq: 29 Nov 2022 10:39




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:128 Resp: 3190726
Ion Ratio Lower Upper
128 100
129 11.1 8.7 13.1
127 13.1 10.4 15.6



#32
4-Chloroaniline
Concen: 27.854 ng/ul
RT: 10.986 min Scan# 1326
Delta R.T. -0.006 min
Lab File: BP012844.D
Acq: 29 Nov 2022 10:39


Tgt Ion:127 Resp: 1277983
Ion Ratio Lower Upper
127 100
129 31.9 25.7 38.5








#33
Hexachlorobutadiene
Concen: 27.439 ng/ul
RT: 11.169 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP012844.D
Acq: 29 Nov 2022 10:39

Instrument :
BNA_P
ClientSampleId :







Tgt Ion:225 Resp: 617679
Ion Ratio Lower Upper
225 100
223 64.2 49.6 74.4
227 64.6 51.0 76.6



#34
Caprolactam
Concen: 17.684 ng/ul
RT: 11.757 min Scan# 1457
Delta R.T. -0.000 min
Lab File: BP012844.D
Acq: 29 Nov 2022 10:39

Tgt Ion:113 Resp: 183522
Ion Ratio Lower Upper
113 100
55 176.0 145.9 218.9
56 129.8 107.1 160.7









#35
4-Chloro-3-methylphenol
Concen: 30.061 ng/ul
RT: 12.098 min Scan# 1515
Delta R.T. -0.006 min
Lab File: BP012844.D
Acq: 29 Nov 2022 10:39




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:107 Resp: 1028804
Ion Ratio Lower Upper
107 100
144 28.1 21.4 32.2
142 84.2 66.6 100.0



#36
2-Methylnaphthalene
Concen: 28.552 ng/ul
RT: 12.481 min Scan# 1580
Delta R.T. -0.006 min
Lab File: BP012844.D
Acq: 29 Nov 2022 10:39

Tgt Ion:142 Resp: 2215023
Ion Ratio Lower Upper
142 100
141 89.6 70.2 105.2









#37
1-Methylnaphthalene
Concen: 28.929 ng/ul
RT: 12.698 min Scan# 1617
Delta R.T. -0.006 min
Lab File: BP012844.D
Acq: 29 Nov 2022 10:39




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:142 Resp: 2249262
Ion Ratio Lower Upper
142 100
141 90.6 73.3 109.9
116 3.5 2.9 4.3



#39
1,2,4,5-Tetrachlorobenzene
Concen: 27.364 ng/ul
RT: 12.845 min Scan# 1642
Delta R.T. -0.006 min
Lab File: BP012844.D
Acq: 29 Nov 2022 10:39

Tgt Ion:216 Resp: 1274736
Ion Ratio Lower Upper
216 100
214 78.7 63.0 94.6
179 19.5 16.2 24.4
108 14.7 12.7 19.1





#40

Hexachlorocyclopentadiene

Concen: 24.762 ng/ul

RT: 12.828 min Scan# 1639

Delta R.T. -0.006 min

Lab File: BP012844.D

Acq: 29 Nov 2022 10:39

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:237 Resp: 555808

Ion Ratio Lower Upper

237 100

235 61.7 50.2 75.4

272 12.6 10.2 15.2



#41

2,4,6-Trichlorophenol

Concen: 28.630 ng/ul

RT: 13.081 min Scan# 1682

Delta R.T. -0.006 min

Lab File: BP012844.D

Acq: 29 Nov 2022 10:39




Tgt Ion:196 Resp: 834512


Ion Ratio Lower Upper

196 100

198 95.6 76.5 114.7

200 31.2 25.3 37.9









#42
2,4,5-Trichlorophenol
Concen: 29.042 ng/ul
RT: 13.151 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP012844.D
Acq: 29 Nov 2022 10:39




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:196 Resp: 910017
Ion Ratio Lower Upper
196 100
198 95.9 77.6 116.4
200 31.0 24.2 36.4



#43
1,1'-Biphenyl
Concen: 28.113 ng/ul
RT: 13.486 min Scan# 1751
Delta R.T. -0.006 min
Lab File: BP012844.D
Acq: 29 Nov 2022 10:39

Tgt Ion:154 Resp: 2991789
Ion Ratio Lower Upper
154 100
153 40.2 32.0 48.0
76 12.1 10.2 15.4





#44

2-Chloronaphthalene

Concen: 28.073 ng/ul

RT: 13.528 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP012844.D

Acq: 29 Nov 2022 10:39

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:162 Resp: 2390008

Ion Ratio Lower Upper

162 100

164 33.0 26.5 39.7

127 36.2 28.8 43.2



#45

2-Nitroaniline

Concen: 36.199 ng/ul

RT: 13.728 min Scan# 1792

Delta R.T. -0.006 min

Lab File: BP012844.D

Acq: 29 Nov 2022 10:39




Tgt Ion: 65 Resp: 602973


Ion Ratio Lower Upper

65 100

92 72.5 57.8 86.8

138 114.0 94.6 142.0









#47
Dimethylphthalate
Concen: 29.253 ng/ul
RT: 14.104 min Scan# 1856
Delta R.T. -0.006 min
Lab File: BP012844.D
Acq: 29 Nov 2022 10:39




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:163 Resp: 3083394
Ion Ratio Lower Upper
163 100
194 4.2 3.4 5.0
164 10.1 7.8 11.8



#48
2,6-Dinitrotoluene
Concen: 34.993 ng/ul
RT: 14.222 min Scan# 1876
Delta R.T. -0.006 min
Lab File: BP012844.D
Acq: 29 Nov 2022 10:39

Tgt Ion:165 Resp: 577231
Ion Ratio Lower Upper
165 100
89 51.6 42.7 64.1
121 20.4 16.9 25.3









#50
Acenaphthylene
Concen: 28.797 ng/ul
RT: 14.375 min Scan# 1902
Delta R.T. -0.000 min
Lab File: BP012844.D
Acq: 29 Nov 2022 10:39




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:152 Resp: 3694366
Ion Ratio Lower Upper
152 100
151 19.9 15.9 23.9
153 13.0 10.7 16.1



#51
3-Nitroaniline
Concen: 30.823 ng/ul
RT: 14.551 min Scan# 1932
Delta R.T. -0.000 min
Lab File: BP012844.D
Acq: 29 Nov 2022 10:39

Tgt Ion:138 Resp: 588845
Ion Ratio Lower Upper
138 100
108 10.8 9.9 14.9
92 111.1 90.2 135.2









#52
Acenaphthene
Concen: 28.582 ng/ul
RT: 14.716 min Scan# 1960
Delta R.T. -0.000 min
Lab File: BP012844.D
Acq: 29 Nov 2022 10:39




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:153 Resp: 2581505
Ion Ratio Lower Upper
153 100
152 48.2 38.2 57.4
154 89.0 69.8 104.6



#53
2,4-Dinitrophenol
Concen: 28.777 ng/ul
RT: 14.751 min Scan# 1966
Delta R.T. -0.006 min
Lab File: BP012844.D
Acq: 29 Nov 2022 10:39

Tgt Ion:184 Resp: 305041
Ion Ratio Lower Upper
184 100
63 57.6 48.1 72.1
154 56.2 47.4 71.2









#55
4-Nitrophenol
Concen: 30.920 ng/ul
RT: 14.851 min Scan# 1982
Delta R.T. -0.000 min
Lab File: BP012844.D
Acq: 29 Nov 2022 10:39




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:109 Resp: 402729
Ion Ratio Lower Upper
109 100
139 153.7 119.2 178.8
65 129.1 110.7 166.1



#56
Dibenzofuran
Concen: 29.269 ng/ul
RT: 15.045 min Scan# 2016
Delta R.T. -0.006 min
Lab File: BP012844.D
Acq: 29 Nov 2022 10:39

Tgt Ion:168 Resp: 3643858
Ion Ratio Lower Upper
168 100
139 34.4 27.7 41.5









#57
2,4-Dinitrotoluene
Concen: 35.605 ng/ul
RT: 15.004 min Scan# 2009
Delta R.T. -0.006 min
Lab File: BP012844.D
Acq: 29 Nov 2022 10:39




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:165 Resp: 866717
Ion Ratio Lower Upper
165 100
63 39.0 31.0 46.4
182 2.7 2.0 3.0



#58
2,3,4,6-Tetrachlorophenol
Concen: 30.414 ng/ul
RT: 15.269 min Scan# 2054
Delta R.T. -0.006 min
Lab File: BP012844.D
Acq: 29 Nov 2022 10:39

Tgt Ion:232 Resp: 821400
Ion Ratio Lower Upper
232 100
131 39.1 32.0 48.0
130 2.0 1.5 2.3
166 28.2 23.2 34.8









#59
Diethylphthalate
Concen: 29.602 ng/ul
RT: 15.469 min Scan# 2088
Delta R.T. -0.006 min
Lab File: BP012844.D
Acq: 29 Nov 2022 10:39




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:149 Resp: 3020310
Ion Ratio Lower Upper
149 100
177 23.4 18.1 27.1
150 12.3 9.8 14.8



#61
Fluorene
Concen: 30.049 ng/ul
RT: 15.698 min Scan# 2127
Delta R.T. -0.006 min
Lab File: BP012844.D
Acq: 29 Nov 2022 10:39

Tgt Ion:166 Resp: 3000076
Ion Ratio Lower Upper
166 100
165 97.1 78.2 117.2
167 13.4 10.9 16.3









#62
4-Chlorophenyl-phenylether
Concen: 29.614 ng/ul
RT: 15.692 min Scan# 2126
Delta R.T. -0.006 min
Lab File: BP012844.D
Acq: 29 Nov 2022 10:39




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:204 Resp: 1534052
Ion Ratio Lower Upper
204 100
206 32.9 26.0 39.0
141 49.7 40.7 61.1



#63
4-Nitroaniline
Concen: 36.825 ng/ul
RT: 15.716 min Scan# 2130
Delta R.T. -0.000 min
Lab File: BP012844.D
Acq: 29 Nov 2022 10:39

Tgt Ion:138 Resp: 599606
Ion Ratio Lower Upper
138 100
92 48.2 38.6 57.8
108 69.4 56.4 84.6









#66
4,6-Dinitro-2-methylphenol
Concen: 31.375 ng/ul
RT: 15.769 min Scan# 2139
Delta R.T. -0.006 min
Lab File: BP012844.D
Acq: 29 Nov 2022 10:39




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:198 Resp: 516574
Ion Ratio Lower Upper
198 100
182 4.3 4.1 6.1
77 20.8 18.9 28.3



#67
N-Nitrosodiphenylamine
Concen: 28.564 ng/ul
RT: 15.904 min Scan# 2162
Delta R.T. -0.006 min
Lab File: BP012844.D
Acq: 29 Nov 2022 10:39

Tgt Ion:169 Resp: 2605295
Ion Ratio Lower Upper
169 100
168 66.1 53.2 79.8
167 36.1 29.6 44.4









#68
4-Bromophenyl-phenylether
Concen: 30.284 ng/ul
RT: 16.586 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP012844.D
Acq: 29 Nov 2022 10:39




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:248 Resp: 956573
Ion Ratio Lower Upper
248 100
250 97.3 78.6 118.0
141 62.4 48.7 73.1



#69
Hexachlorobenzene
Concen: 28.764 ng/ul
RT: 16.704 min Scan# 2298
Delta R.T. -0.006 min
Lab File: BP012844.D
Acq: 29 Nov 2022 10:39

Tgt Ion:284 Resp: 1157998
Ion Ratio Lower Upper
284 100
142 28.8 25.0 37.6
249 29.9 24.6 36.8





#70

Atrazine

Concen: 29.332 ng/ul

RT: 16.857 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP012844.D

Acq: 29 Nov 2022 10:39

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:200 Resp: 917901

Ion Ratio Lower Upper

200 100

173 25.7 20.3 30.5

215 50.0 40.1 60.1



#71

Pentachlorophenol

Concen: 28.109 ng/ul

RT: 17.045 min Scan# 2356

Delta R.T. -0.006 min

Lab File: BP012844.D

Acq: 29 Nov 2022 10:39




Tgt Ion:266 Resp: 694940


Ion Ratio Lower Upper

266 100

264 64.1 50.7 76.1

268 64.2 51.8 77.8









#72
Phenanthrene
Concen: 29.251 ng/ul
RT: 17.445 min Scan# 2425
Delta R.T. -0.006 min
Lab File: BP012844.D
Acq: 29 Nov 2022 10:39




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:178 Resp: 5003286
Ion Ratio Lower Upper
178 100
179 15.8 12.3 18.5
176 19.5 15.4 23.0



#74
Anthracene
Concen: 29.395 ng/ul
RT: 17.539 min Scan# 2440
Delta R.T. -0.006 min
Lab File: BP012844.D
Acq: 29 Nov 2022 10:39

Tgt Ion:178 Resp: 4974121
Ion Ratio Lower Upper
178 100
179 15.6 12.4 18.6
176 18.8 14.8 22.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 25.603 ng/uL
RT: 13.451 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP012844.D
Acq: 29 Nov 2022 10:39




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:216 Resp: 1270298
Ion Ratio Lower Upper
216 100
214 79.7 62.3 93.5
218 48.8 39.1 58.7



#76
Pentachlorobenzene
Concen: 24.486 ng/uL
RT: 14.963 min Scan# 2002
Delta R.T. -0.006 min
Lab File: BP012844.D
Acq: 29 Nov 2022 10:39

Tgt Ion:250 Resp: 1262869
Ion Ratio Lower Upper
250 100
248 64.3 50.6 76.0
252 63.8 50.5 75.7









#77
 Carbazole
 Concen: 31.855 ng/ul
 RT: 17.804 min Scan# 2485
 Delta R.T. -0.006 min
 Lab File: BP012844.D
 Acq: 29 Nov 2022 10:39




Instrument :
 BNA_P
 ClientSampleId :


Tgt Ion:167 Resp: 4560108
 Ion Ratio Lower Upper
 167 100
 166 21.4 17.1 25.7
 139 12.1 9.8 14.8



#78
 Di-n-butylphthalate
 Concen: 30.001 ng/ul
 RT: 18.351 min Scan# 2578
 Delta R.T. -0.006 min
 Lab File: BP012844.D
 Acq: 29 Nov 2022 10:39

Tgt Ion:149 Resp: 5084004
 Ion Ratio Lower Upper
 149 100
 150 9.1 7.2 10.8
 104 4.4 3.7 5.5









#80
Fluoranthene
Concen: 25.678 ng/ul
RT: 19.404 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP012844.D
Acq: 29 Nov 2022 10:39




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:202 Resp: 6007480
Ion Ratio Lower Upper
202 100
101 10.7 8.7 13.1
100 8.0 6.9 10.3



#82
Pyrene
Concen: 25.671 ng/ul
RT: 19.757 min Scan# 2817
Delta R.T. -0.006 min
Lab File: BP012844.D
Acq: 29 Nov 2022 10:39

Tgt Ion:202 Resp: 6211660
Ion Ratio Lower Upper
202 100
101 12.9 10.3 15.5
100 10.7 8.6 13.0





#83

Butylbenzylphthalate

Concen: 36.061 ng/ul

RT: 20.627 min Scan# 2965

Delta R.T. -0.006 min

Lab File: BP012844.D

Acq: 29 Nov 2022 10:39

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:149 Resp: 2041873

Ion Ratio Lower Upper

149 100

91 65.3 58.2 87.2

206 22.7 16.5 24.7



#84

3,3'-Dichlorobenzidine

Concen: 31.182 ng/ul

RT: 21.398 min Scan# 3096

Delta R.T. -0.006 min

Lab File: BP012844.D

Acq: 29 Nov 2022 10:39




Tgt Ion:252 Resp: 1853588


Ion Ratio Lower Upper

252 100

254 64.6 52.3 78.5

126 10.0 7.6 11.4





#85

Benzo(a)anthracene

Concen: 27.225 ng/ul

RT: 21.474 min Scan# 3109

Delta R.T. -0.006 min

Lab File: BP012844.D

Acq: 29 Nov 2022 10:39

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:228 Resp: 6095227

Ion Ratio Lower Upper

228 100

229 20.0 16.2 24.2

226 27.3 22.2 33.2



#86

Bis(2-ethylhexyl)phthalate

Concen: 33.942 ng/ul

RT: 21.386 min Scan# 3094

Delta R.T. -0.012 min

Lab File: BP012844.D

Acq: 29 Nov 2022 10:39




Tgt Ion:149 Resp: 2939099


Ion Ratio Lower Upper

149 100

167 29.0 22.7 34.1

279 4.6 3.8 5.6





#87

Chrysene

Concen: 26.416 ng/ul

RT: 21.527 min Scan# 3119

Delta R.T. -0.006 min

Lab File: BP012844.D

Acq: 29 Nov 2022 10:39

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:228 Resp: 5718359

Ion Ratio Lower Upper

228 100

226 30.5 24.3 36.5

229 20.3 15.9 23.9



#89

Di-n-octyl phthalate

Concen: 27.786 ng/ul




RT: 22.351 min Scan# 3258


Delta R.T. -0.012 min

Lab File: BP012844.D

Acq: 29 Nov 2022 10:39

Tgt Ion:149 Resp: 4699008





#90

Benzo(b)fluoranthene

Concen: 26.925 ng/ul

RT: 23.233 min Scan# 3409

Delta R.T. -0.006 min

Lab File: BP012844.D

Acq: 29 Nov 2022 10:39

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:252 Resp: 5904612

Ion Ratio Lower Upper

252 100

253 22.2 17.6 26.4

125 9.2 7.4 11.2



#91

Benzo(k)fluoranthene

Concen: 26.462 ng/ul

RT: 23.280 min Scan# 3416

Delta R.T. -0.012 min

Lab File: BP012844.D

Acq: 29 Nov 2022 10:39




Tgt Ion:252 Resp: 5792537


Ion Ratio Lower Upper

252 100

253 22.2 17.8 26.8

125 9.3 7.8 11.8





#93

Benzo(a)pyrene

Concen: 27.899 ng/ul

RT: 23.892 min Scan# 3

Delta R.T. -0.006 min

Lab File: BP012844.D

Acq: 29 Nov 2022 10:39

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:252 Resp: 5030983

Ion Ratio Lower Upper

252 100

253 22.2 17.6 26.4

125 10.1 8.5 12.7



#94

Indeno(1,2,3-cd)pyrene

Concen: 31.531 ng/ul

RT: 26.633 min Scan# 3986

Delta R.T. -0.018 min

Lab File: BP012844.D

Acq: 29 Nov 2022 10:39




Tgt Ion:276 Resp: 5485567


Ion Ratio Lower Upper

276 100

138 22.0 18.3 27.5

277 25.3 19.8 29.8









#95
Dibenzo(a,h)anthracene
Concen: 30.690 ng/ul
RT: 26.650 min Scan# 3991
Delta R.T. -0.024 min
Lab File: BP012844.D
Acq: 29 Nov 2022 10:39

Instrument :
BNA_P
ClientSampleId :

Tgt Ion:278 Resp: 4844894
Ion Ratio Lower Upper
278 100
139 16.0 12.6 18.8
279 24.0 19.4 29.2



#96
Benzo(g,h,i)perylene
Concen: 30.923 ng/ul
RT: 27.450 min Scan# 4125
Delta R.T. -0.006 min
Lab File: BP012844.D
Acq: 29 Nov 2022 10:39

Tgt Ion:276 Resp: 4627361
Ion Ratio Lower Upper
276 100
138 19.5 16.7 25.1
277 23.4 19.1 28.7

