

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112922\
 Data File : BP012871.D
 Acq On : 30 Nov 2022 02:01
 Operator : CG/JU
 Sample : PB149233BS
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS233

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 11/30/2022
 Supervised By :mohammad ahmed 11/30/2022

Quant Time: Nov 30 03:29:07 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP112522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Nov 29 22:53:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.010	152	630208	20.000	ng/u1	0.00
20) Naphthalene-d8	10.828	136	2841139	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.645	164	1870204	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.404	188	3874640	20.000	ng/u1	0.00
79) Chrysene-d12	21.486	240	2897130	20.000	ng/u1	0.00
88) Perylene-d12	23.998	264	2292314	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.393	96	88325	6.006	ng/uL	0.00
4) Pyridine-d5	3.817	84	971791	22.080	ng/u1	0.00
7) Phenol-d5	7.163	99	1713184	33.082	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.334	67	1010098	31.654	ng/u1	0.00
11) 2-Chlorophenol-d4	7.540	132	1332471	33.768	ng/u1	0.00
15) 4-Methylphenol-d8	8.716	113	1384458	33.181	ng/u1	0.00
21) Nitrobenzene-d5	9.181	128	655770	37.711	ng/u1	0.00
24) 2-Nitrophenol-d4	9.905	143	740244	40.091	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.440	165	1452211	33.708	ng/u1	0.00
31) 4-Chloroaniline-d4	10.957	131	1027357	17.414	ng/u1	0.00
46) Dimethylphthalate-d6	14.051	166	4194112	32.581	ng/u1	0.00
49) Acenaphthylene-d8	14.340	160	4888729	32.548	ng/u1	0.00
54) 4-Nitrophenol-d4	14.834	143	754693	32.883	ng/u1	0.00
60) Fluorene-d10	15.639	176	3669707	32.551	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.751	200	702297	37.923	ng/u1	0.00
73) Anthracene-d10	17.504	188	5542706	32.497	ng/u1	0.00
81) Pyrene-d10	19.727	212	5843189	34.711	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.827	264	3752560	33.327	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.428	88	181549	11.494	ng/uL	99
5) Pyridine	3.840	79	1159048	26.914	ng/u1	98
6) Benzaldehyde	7.146	77	894368m	35.371	ng/u1	
8) Phenol	7.193	94	1756710	32.351	ng/u1	100
10) Bis(2-Chloroethyl)ether	7.428	93	1387589	31.188	ng/u1	99
12) 2-Chlorophenol	7.569	128	1382233	32.202	ng/u1	99
13) 2-Methylphenol	8.452	108	1356380	32.139	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.546	45	1952516	30.425	ng/u1	99
16) Acetophenone	8.840	105	2120627	32.067	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.828	70	1048792	32.194	ng/u1	97
18) 4-Methylphenol	8.781	108	1500593	32.285	ng/u1	98
19) Hexachloroethane	9.099	117	528825	32.127	ng/u1	97
22) Nitrobenzene	9.222	77	1532762	33.608	ng/u1	98
23) Isophorone	9.752	82	3091837	30.751	ng/u1	99
25) 2-Nitrophenol	9.934	139	816153	36.825	ng/u1	96
26) 2,4-Dimethylphenol	9.993	107	1535618	31.181	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.234	93	1924031	31.212	ng/u1	99
29) 2,4-Dichlorophenol	10.469	162	1445307	32.384	ng/u1	100
30) Naphthalene	10.881	128	4585250	30.821	ng/u1	100
32) 4-Chloroaniline	10.987	127	1570054	25.913	ng/u1	100
33) Hexachlorobutadiene	11.163	225	933082	31.389	ng/u1	99
34) Caprolactam	11.763	113	447446m	32.649	ng/u1	
35) 4-Chloro-3-methylphenol	12.099	107	1490280	32.975	ng/u1	99
36) 2-Methylnaphthalene	12.481	142	3214996	31.382	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112922\
 Data File : BP012871.D
 Acq On : 30 Nov 2022 02:01
 Operator : CG/JU
 Sample : PB149233BS
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS233


Quant Time: Nov 30 03:29:07 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP112522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Nov 29 22:53:34 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 11/30/2022
 Supervised By :mohammad ahmed 11/30/2022


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.698	142	3220914	31.370	ng/u1	99
39) 1,2,4,5-Tetrachloroben...	12.846	216	1862002	31.222	ng/u1	98
40) Hexachlorocyclopentadiene	12.828	237	933633	32.490	ng/u1	99
41) 2,4,6-Trichlorophenol	13.081	196	1211412	32.464	ng/u1	98
42) 2,4,5-Trichlorophenol	13.151	196	1328916	33.128	ng/u1	99
43) 1,1'-Biphenyl	13.481	154	4264263	31.300	ng/u1	100
44) 2-Chloronaphthalene	13.528	162	3388574	31.090	ng/u1	100
45) 2-Nitroaniline	13.728	65	872993	40.938	ng/u1	99
47) Dimethylphthalate	14.104	163	4249883	31.495	ng/u1	100
48) 2,6-Dinitrotoluene	14.222	165	844356	39.983	ng/u1	95
50) Acenaphthylene	14.369	152	5155170	31.388	ng/u1	100
51) 3-Nitroaniline	14.551	138	825694	33.760	ng/u1	93
52) Acenaphthene	14.710	153	3593463	31.078	ng/u1	98
53) 2,4-Dinitrophenol	14.751	184	488627	36.008	ng/u1	93
55) 4-Nitrophenol	14.851	109	518336	31.086	ng/u1	94
56) Dibenzofuran	15.045	168	5024277	31.524	ng/u1	99
57) 2,4-Dinitrotoluene	15.004	165	1205123	38.671	ng/u1	98
58) 2,3,4,6-Tetrachlorophenol	15.269	232	1130725	32.704	ng/u1	97
59) Diethylphthalate	15.463	149	4150464	31.775	ng/u1	99
61) Fluorene	15.698	166	4018980	31.444	ng/u1	99
62) 4-Chlorophenyl-phenyle...	15.687	204	2112718	31.859	ng/u1	98
63) 4-Nitroaniline	15.716	138	826281	39.639	ng/u1	99
66) 4,6-Dinitro-2-methylph...	15.769	198	730497	36.749	ng/u1#	93
67) N-Nitrosodiphenylamine	15.904	169	3526346	32.023	ng/u1	100
68) 4-Bromophenyl-phenylether	16.587	248	1340115	35.141	ng/u1	99
69) Hexachlorobenzene	16.704	284	1567552	32.251	ng/u1	97
70) Atrazine	16.851	200	572362	15.149	ng/u1	97
71) Pentachlorophenol	17.045	266	910552	30.506	ng/u1	99
72) Phenanthrene	17.445	178	6448468	31.227	ng/u1	99
74) Anthracene	17.539	178	6381249	31.235	ng/u1	99
75) 1,2,3,4-Tetrachloroben...	13.445	216	1861039	31.069	ng/uL	99
76) Pentachlorobenzene	14.963	250	1793681	28.806	ng/uL	99
77) Carbazole	17.804	167	5599688	32.400	ng/u1	99
78) Di-n-butylphthalate	18.345	149	6737451	32.931	ng/u1	99
80) Fluoranthene	19.404	202	7053222	33.685	ng/u1	98
82) Pyrene	19.757	202	7112042	32.840	ng/u1	98
83) Butylbenzylphthalate	20.622	149	2446083	48.267	ng/u1	93
84) 3,3'-Dichlorobenzidine	21.398	252	1593601	29.953	ng/u1	99
85) Benzo(a)anthracene	21.469	228	5894047	29.414	ng/u1	100
86) Bis(2-ethylhexyl)phtha...	21.380	149	3207981	41.392	ng/u1	100
87) Chrysene	21.527	228	5529272	28.539	ng/u1	100
89) Di-n-octyl phthalate	22.345	149	4538418	34.670	ng/u1	100
90) Benzo(b)fluoranthene	23.227	252	4927836	29.030	ng/u1	99
91) Benzo(k)fluoranthene	23.274	252	4873705	28.763	ng/u1	99
93) Benzo(a)pyrene	23.886	252	4222410	30.250	ng/u1	99
94) Indeno(1,2,3-cd)pyrene	26.627	276	5070381	37.651	ng/u1	100
95) Dibenzo(a,h)anthracene	26.645	278	4463848	36.530	ng/u1	99
96) Benzo(g,h,i)perylene	27.433	276	4427951	38.227	ng/u1	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 11.494 ng/uL
RT: 3.428 min Scan# 41
Delta R.T. 0.000 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01




Instrument :
BNA_P
ClientSampleId :
SLCS233




Tgt Ion: 88 Resp: 181549
Ion Ratio Lower Upper
88 100
43 27.7 22.6 34.0
58 73.4 59.6 89.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022




#5
Pyridine
Concen: 26.914 ng/uL
RT: 3.840 min Scan# 111
Delta R.T. 0.000 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01




Tgt Ion: 79 Resp: 1159048
Ion Ratio Lower Upper
79 100
52 71.8 58.9 88.3
51 35.0 27.3 40.9





#6
Benzaldehyde
Concen: 35.371 ng/ul m
RT: 7.146 min Scan# 61
Delta R.T. 0.000 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01




Instrument :
BNA_P
ClientSampleId :
SLCS233




Tgt Ion: 77 Resp: 894368
Ion Ratio Lower Upper
77 100
105 102.1 81.4 122.2
106 98.0 77.4 116.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022




#8
Phenol
Concen: 32.351 ng/ul
RT: 7.193 min Scan# 681
Delta R.T. 0.006 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01




Tgt Ion: 94 Resp: 1756710
Ion Ratio Lower Upper
94 100
65 25.9 20.9 31.3
66 32.4 26.0 39.0





#10
Bis(2-Chloroethyl)ether
Concen: 31.188 ng/ul
RT: 7.428 min Scan# 721
Delta R.T. 0.000 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01




Instrument :
BNA_P
ClientSampleId :
SLCS233



Tgt Ion: 93 Resp: 1387589
Ion Ratio Lower Upper
93 100
63 74.2 59.7 89.5
95 31.9 25.8 38.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022



#12
2-Chlorophenol
Concen: 32.202 ng/ul
RT: 7.569 min Scan# 745
Delta R.T. 0.000 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01

Tgt Ion:128 Resp: 1382233
Ion Ratio Lower Upper
128 100
64 45.5 35.6 53.4
130 32.9 26.0 39.0





#13

2-Methylphenol

Concen: 32.139 ng/ul

RT: 8.452 min Scan# 895

Delta R.T. 0.000 min

Lab File: BP012871.D

Acq: 30 Nov 2022 02:01

Instrument :

BNA_P

ClientSampleId :

SLCS233

Tgt Ion:108 Resp: 1356380

Ion Ratio Lower Upper





108 100

107 88.0 69.3 103.9

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022



#14

2,2'-oxybis(1-Chloropropane)

Concen: 30.425 ng/ul

RT: 8.546 min Scan# 911

Delta R.T. 0.000 min

Lab File: BP012871.D

Acq: 30 Nov 2022 02:01




Tgt Ion: 45 Resp: 1952516


Ion Ratio Lower Upper

45 100

77 14.1 11.3 16.9


79 11.3 8.6 12.8





#16
Acetophenone
Concen: 32.067 ng/ul
RT: 8.840 min Scan# 90
Delta R.T. 0.000 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01




Instrument :
BNA_P
ClientSampleId :
SLCS233




Tgt Ion:105 Resp: 212062
Ion Ratio Lower Upper
105 100
77 73.3 59.7 89.5
51 24.8 20.9 31.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022




#17
N-Nitroso-di-n-propylamine
Concen: 32.194 ng/ul
RT: 8.828 min Scan# 959
Delta R.T. 0.006 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01




Tgt Ion: 70 Resp: 1048792
Ion Ratio Lower Upper
70 100
42 55.8 46.1 69.1
101 10.1 7.4 11.2
130 22.3 15.8 23.6





#18
4-Methylphenol
Concen: 32.285 ng/ul
RT: 8.781 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01




Instrument :
BNA_P
ClientSampleId :
SLCS233



Tgt Ion:108 Resp: 1500593
Ion Ratio Lower Upper
108 100
107 110.9 90.2 135.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022



#19
Hexachloroethane
Concen: 32.127 ng/ul
RT: 9.099 min Scan# 1005
Delta R.T. -0.006 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01


Tgt Ion:117 Resp: 528825
Ion Ratio Lower Upper
117 100
201 109.9 85.7 128.5
199 69.1 52.9 79.3





#22
Nitrobenzene
Concen: 33.608 ng/ul
RT: 9.222 min Scan# 1026
Delta R.T. 0.000 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01




Instrument :
BNA_P
ClientSampleId :
SLCS233




Tgt Ion: 77 Resp: 1532762
Ion Ratio Lower Upper
77 100
123 49.1 38.2 57.2
65 13.0 10.2 15.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022




#23
Isophorone
Concen: 30.751 ng/ul
RT: 9.752 min Scan# 1116
Delta R.T. 0.000 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01



Tgt Ion: 82 Resp: 3091837
Ion Ratio Lower Upper
82 100
95 6.7 5.2 7.8
138 18.7 14.5 21.7









#25
2-Nitrophenol
Concen: 36.825 ng/ul
RT: 9.934 min Scan# 1147
Delta R.T. 0.000 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01

Instrument :
BNA_P
ClientSampleId :
SLCS233

Tgt Ion:139 Resp: 816153
Ion Ratio Lower Upper
139 100
65 45.8 39.5 59.3
109 25.6 21.4 32.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022



#26
2,4-Dimethylphenol
Concen: 31.181 ng/ul
RT: 9.993 min Scan# 1157
Delta R.T. 0.000 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01


Tgt Ion:107 Resp: 1535618
Ion Ratio Lower Upper
107 100
121 52.4 41.8 62.6
122 89.1 72.3 108.5





#27
Bis(2-Chloroethoxy)methane
Concen: 31.212 ng/ul
RT: 10.234 min Scan# 1198
Delta R.T. 0.000 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01




Instrument :
BNA_P
ClientSampleId :
SLCS233



Tgt Ion: 93 Resp: 1924031
Ion Ratio Lower Upper
93 100
95 32.8 26.1 39.1
123 11.7 9.0 13.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022



#29
2,4-Dichlorophenol
Concen: 32.384 ng/ul
RT: 10.469 min Scan# 1238
Delta R.T. 0.000 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01


Tgt Ion:162 Resp: 1445307
Ion Ratio Lower Upper
162 100
164 65.6 52.9 79.3
98 33.2 26.6 40.0





#30
Naphthalene
Concen: 30.821 ng/ul
RT: 10.881 min Scan# 1308
Delta R.T. 0.000 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01




Instrument :
BNA_P
ClientSampleId :
SLCS233



Tgt Ion:128 Resp: 4585250
Ion Ratio Lower Upper
128 100
129 11.0 8.7 13.1
127 13.1 10.4 15.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022



#32
4-Chloroaniline
Concen: 25.913 ng/ul
RT: 10.987 min Scan# 1326
Delta R.T. 0.000 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01


Tgt Ion:127 Resp: 1570054
Ion Ratio Lower Upper
127 100
129 32.1 25.7 38.5





#33
Hexachlorobutadiene
Concen: 31.389 ng/ul
RT: 11.163 min Scan# 1356
Delta R.T. 0.000 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01




Instrument :
BNA_P
ClientSampleId :
SLCS233



Tgt Ion:225 Resp: 933082
Ion Ratio Lower Upper
225 100
223 63.5 49.6 74.4
227 63.7 51.0 76.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022



#34
Caprolactam
Concen: 32.649 ng/ul m
RT: 11.763 min Scan# 1458
Delta R.T. 0.006 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01


Tgt Ion:113 Resp: 447446
Ion Ratio Lower Upper
113 100
55 166.8 145.9 218.9
56 121.8 107.1 160.7





#35
4-Chloro-3-methylphenol
Concen: 32.975 ng/ul
RT: 12.099 min Scan# 1515
Delta R.T. 0.000 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01




Instrument :
BNA_P
ClientSampleId :
SLCS233




Tgt Ion:107 Resp: 1490280
Ion Ratio Lower Upper
107 100
144 27.6 21.4 32.2
142 84.5 66.6 100.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022




#36
2-Methylnaphthalene
Concen: 31.382 ng/ul
RT: 12.481 min Scan# 1580
Delta R.T. 0.000 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01




Tgt Ion:142 Resp: 3214996
Ion Ratio Lower Upper
142 100
141 88.4 70.2 105.2





#37
1-Methylnaphthalene
Concen: 31.370 ng/ul
RT: 12.698 min Scan# 1617
Delta R.T. 0.000 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01




Instrument :
BNA_P
ClientSampleId :
SLCS233




Tgt Ion:142 Resp: 3220914
Ion Ratio Lower Upper
142 100
141 90.5 73.3 109.9
116 3.4 2.9 4.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022




#39
1,2,4,5-Tetrachlorobenzene
Concen: 31.222 ng/ul
RT: 12.846 min Scan# 1642
Delta R.T. 0.000 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01




Tgt Ion:216 Resp: 1862002
Ion Ratio Lower Upper
216 100
214 78.2 63.0 94.6
179 18.9 16.2 24.4
108 13.8 12.7 19.1





#40
Hexachlorocyclopentadiene
Concen: 32.490 ng/ul
RT: 12.828 min Scan# 1639
Delta R.T. 0.000 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01




Instrument :
BNA_P
ClientSampleId :
SLCS233




Tgt Ion:237 Resp: 933633
Ion Ratio Lower Upper
237 100
235 62.4 50.2 75.4
272 13.3 10.2 15.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022




#41
2,4,6-Trichlorophenol
Concen: 32.464 ng/ul
RT: 13.081 min Scan# 1682
Delta R.T. 0.000 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01




Tgt Ion:196 Resp: 1211412
Ion Ratio Lower Upper
196 100
198 97.3 76.5 114.7
200 31.2 25.3 37.9





#42
2,4,5-Trichlorophenol
Concen: 33.128 ng/ul
RT: 13.151 min Scan# 1693
Delta R.T. 0.006 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01




Instrument :
BNA_P
ClientSampleId :
SLCS233




Tgt Ion:196 Resp: 1328916
Ion Ratio Lower Upper
196 100
198 97.5 77.6 116.4
200 31.8 24.2 36.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022




#43
1,1'-Biphenyl
Concen: 31.300 ng/ul
RT: 13.481 min Scan# 1750
Delta R.T. 0.000 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01




Tgt Ion:154 Resp: 4264263
Ion Ratio Lower Upper
154 100
153 39.8 32.0 48.0
76 12.8 10.2 15.4





#44
2-Chloronaphthalene
Concen: 31.090 ng/ul
RT: 13.528 min Scan# 1758
Delta R.T. 0.000 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01




Instrument :
BNA_P
ClientSampleId :
SLCS233



Tgt Ion:162 Resp: 3388574
Ion Ratio Lower Upper
162 100
164 32.9 26.5 39.7
127 35.6 28.8 43.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022



#45
2-Nitroaniline
Concen: 40.938 ng/ul
RT: 13.728 min Scan# 1792
Delta R.T. 0.000 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01


Tgt Ion: 65 Resp: 872993
Ion Ratio Lower Upper
65 100
92 73.8 57.8 86.8
138 119.7 94.6 142.0





#47
Dimethylphthalate
Concen: 31.495 ng/ul
RT: 14.104 min Scan# 1856
Delta R.T. 0.006 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01




Instrument :
BNA_P
ClientSampleId :
SLCS233



Tgt Ion:163 Resp: 4249883
Ion Ratio Lower Upper
163 100
194 4.2 3.4 5.0
164 10.0 7.8 11.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022



#48
2,6-Dinitrotoluene
Concen: 39.983 ng/ul
RT: 14.222 min Scan# 1876
Delta R.T. 0.000 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01


Tgt Ion:165 Resp: 844356
Ion Ratio Lower Upper
165 100
89 49.7 42.7 64.1
121 19.3 16.9 25.3





#50
Acenaphthylene
Concen: 31.388 ng/ul
RT: 14.369 min Scan# 1901
Delta R.T. 0.000 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01




Instrument :
BNA_P
ClientSampleId :
SLCS233




Tgt Ion:152 Resp: 5155170
Ion Ratio Lower Upper
152 100
151 19.9 15.9 23.9
153 13.4 10.7 16.1

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022




#51
3-Nitroaniline
Concen: 33.760 ng/ul
RT: 14.551 min Scan# 1932
Delta R.T. 0.006 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01




Tgt Ion:138 Resp: 825694
Ion Ratio Lower Upper
138 100
108 10.2 9.9 14.9
92 105.5 90.2 135.2





#52
Acenaphthene
Concen: 31.078 ng/ul
RT: 14.710 min Scan# 1959
Delta R.T. 0.000 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01




Instrument :
BNA_P
ClientSampleId :
SLCS233



Tgt Ion:153 Resp: 3593463
Ion Ratio Lower Upper
153 100
152 48.3 38.2 57.4
154 89.8 69.8 104.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022



#53
2,4-Dinitrophenol
Concen: 36.008 ng/ul
RT: 14.751 min Scan# 1966
Delta R.T. 0.000 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01


Tgt Ion:184 Resp: 488627
Ion Ratio Lower Upper
184 100
63 54.1 48.1 72.1
154 54.6 47.4 71.2





#55
4-Nitrophenol
Concen: 31.086 ng/ul
RT: 14.851 min Scan# 1982
Delta R.T. 0.006 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01




Instrument :
BNA_P
ClientSampleId :
SLCS233



Tgt Ion:109 Resp: 518336
Ion Ratio Lower Upper
109 100
139 156.2 119.2 178.8
65 131.9 110.7 166.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022



#56
Dibenzofuran
Concen: 31.524 ng/ul
RT: 15.045 min Scan# 2016
Delta R.T. 0.000 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01


Tgt Ion:168 Resp: 5024277
Ion Ratio Lower Upper
168 100
139 34.2 27.7 41.5





#57
2,4-Dinitrotoluene
Concen: 38.671 ng/ul
RT: 15.004 min Scan# 2009
Delta R.T. 0.000 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01




Instrument :
BNA_P
ClientSampleId :
SLCS233




Tgt Ion:165 Resp: 1205123
Ion Ratio Lower Upper
165 100
63 37.2 31.0 46.4
182 2.6 2.0 3.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022




#58
2,3,4,6-Tetrachlorophenol
Concen: 32.704 ng/ul
RT: 15.269 min Scan# 2054
Delta R.T. 0.000 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01




Tgt Ion:232 Resp: 1130725
Ion Ratio Lower Upper
232 100
131 37.9 32.0 48.0
130 1.8 1.5 2.3
166 27.8 23.2 34.8





#59
Diethylphthalate
Concen: 31.775 ng/ul
RT: 15.463 min Scan# 2087
Delta R.T. 0.000 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01




Instrument :
BNA_P
ClientSampleId :
SLCS233




Tgt Ion:149 Resp: 4150464
Ion Ratio Lower Upper
149 100
177 23.6 18.1 27.1
150 12.5 9.8 14.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022




#61
Fluorene
Concen: 31.444 ng/ul
RT: 15.698 min Scan# 2127
Delta R.T. 0.000 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01



Tgt Ion:166 Resp: 4018980
Ion Ratio Lower Upper
166 100
165 97.1 78.2 117.2
167 13.4 10.9 16.3









#62
4-Chlorophenyl-phenylether
Concen: 31.859 ng/ul
RT: 15.687 min Scan# 2112718
Delta R.T. 0.000 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01

Instrument :
BNA_P
ClientSampleId :
SLCS233

Tgt Ion:204 Resp: 2112718
Ion Ratio Lower Upper
204 100
206 33.1 26.0 39.0
141 49.6 40.7 61.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022



#63
4-Nitroaniline
Concen: 39.639 ng/ul
RT: 15.716 min Scan# 2130
Delta R.T. 0.006 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01


Tgt Ion:138 Resp: 826281
Ion Ratio Lower Upper
138 100
92 48.4 38.6 57.8
108 69.1 56.4 84.6





#66
4,6-Dinitro-2-methylphenol
Concen: 36.749 ng/ul
RT: 15.769 min Scan# 2138
Delta R.T. 0.006 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01




Instrument :
BNA_P
ClientSampleId :
SLCS233



Tgt Ion:198 Resp: 730497
Ion Ratio Lower Upper
198 100
182 3.7 4.1 6.1
77 20.1 18.9 28.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022



#67
N-Nitrosodiphenylamine
Concen: 32.023 ng/ul
RT: 15.904 min Scan# 2162
Delta R.T. 0.000 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01


Tgt Ion:169 Resp: 3526346
Ion Ratio Lower Upper
169 100
168 66.5 53.2 79.8
167 36.3 29.6 44.4





#68
4-Bromophenyl-phenylether
Concen: 35.141 ng/ul
RT: 16.587 min Scan# 2278
Delta R.T. 0.000 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01




Instrument :
BNA_P
ClientSampleId :
SLCS233



Tgt Ion:248 Resp: 1340115
Ion Ratio Lower Upper
248 100
250 98.1 78.6 118.0
141 59.8 48.7 73.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022



#69
Hexachlorobenzene
Concen: 32.251 ng/ul
RT: 16.704 min Scan# 2298
Delta R.T. 0.000 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01

Tgt Ion:284 Resp: 1567552
Ion Ratio Lower Upper
284 100
142 28.1 25.0 37.6
249 30.1 24.6 36.8





#70

Atrazine

Concen: 15.149 ng/ul

RT: 16.851 min Scan# 2323

Delta R.T. 0.000 min

Lab File: BP012871.D

Acq: 30 Nov 2022 02:01

Instrument :

BNA_P

ClientSampleId :

SLCS233

Tgt Ion:200 Resp: 572362

Ion Ratio Lower Upper

200 100





173 24.8 20.3 30.5

215 47.0 40.1 60.1

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022



#71

Pentachlorophenol

Concen: 30.506 ng/ul

RT: 17.045 min Scan# 2356

Delta R.T. 0.000 min

Lab File: BP012871.D

Acq: 30 Nov 2022 02:01




Tgt Ion:266 Resp: 910552


Ion Ratio Lower Upper

266 100

264 64.0 50.7 76.1

268 64.3 51.8 77.8





#72

Phenanthrene

Concen: 31.227 ng/ul

RT: 17.445 min Scan# 2424

Delta R.T. 0.000 min

Lab File: BP012871.D

Acq: 30 Nov 2022 02:01

Instrument :

BNA_P

ClientSampleId :

SLCS233

Tgt Ion:178 Resp: 6448468

Ion Ratio Lower Upper

178 100





179 15.7 12.3 18.5

176 19.4 15.4 23.0

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022



#74

Anthracene

Concen: 31.235 ng/ul

RT: 17.539 min Scan# 2440

Delta R.T. 0.000 min

Lab File: BP012871.D

Acq: 30 Nov 2022 02:01




Tgt Ion:178 Resp: 6381249


Ion Ratio Lower Upper

178 100

179 15.9 12.4 18.6

176 19.0 14.8 22.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 31.069 ng/uL
RT: 13.445 min Scan# 1744
Delta R.T. -0.006 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01

Instrument :
BNA_P
ClientSampleId :
SLCS233

Tgt Ion:216 Resp: 1861039
Ion Ratio Lower Upper
216 100
214 78.6 62.3 93.5
218 48.3 39.1 58.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022



#76
Pentachlorobenzene
Concen: 28.806 ng/uL
RT: 14.963 min Scan# 2002
Delta R.T. 0.000 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01


Tgt Ion:250 Resp: 1793681
Ion Ratio Lower Upper
250 100
248 63.2 50.6 76.0
252 63.8 50.5 75.7





#77
 Carbazole
 Concen: 32.400 ng/ul
 RT: 17.804 min Scan# 2485
 Delta R.T. 0.000 min
 Lab File: BP012871.D
 Acq: 30 Nov 2022 02:01




Instrument :
 BNA_P
 ClientSampleId :
 SLCS233




Tgt Ion:167 Resp: 5599688
 Ion Ratio Lower Upper
 167 100
 166 21.8 17.1 25.7
 139 11.9 9.8 14.8

Manual Integrations
 APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2022
 Supervised By :mohammad ahmed 11/30/2022




#78
 Di-n-butylphthalate
 Concen: 32.931 ng/ul
 RT: 18.345 min Scan# 2577
 Delta R.T. -0.006 min
 Lab File: BP012871.D
 Acq: 30 Nov 2022 02:01




Tgt Ion:149 Resp: 6737451
 Ion Ratio Lower Upper
 149 100
 150 9.5 7.2 10.8
 104 4.6 3.7 5.5





#80
Fluoranthene
Concen: 33.685 ng/u1
RT: 19.404 min Scan# 2757
Delta R.T. 0.000 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01




Instrument :
BNA_P
ClientSampleId :
SLCS233




Tgt Ion:202 Resp: 705322
Ion Ratio Lower Upper
202 100
101 10.3 8.7 13.1
100 7.8 6.9 10.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022




#82
Pyrene
Concen: 32.840 ng/u1
RT: 19.757 min Scan# 2817
Delta R.T. 0.000 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01




Tgt Ion:202 Resp: 7112042
Ion Ratio Lower Upper
202 100
101 12.4 10.3 15.5
100 10.0 8.6 13.0





#83
Butylbenzylphthalate
Concen: 48.267 ng/ul
RT: 20.622 min Scan# 2964
Delta R.T. 0.000 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01




Instrument :
BNA_P
ClientSampleId :
SLCS233




Tgt Ion:149 Resp: 2446083
Ion Ratio Lower Upper
149 100
91 66.8 58.2 87.2
206 22.9 16.5 24.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022




#84
3,3'-Dichlorobenzidine
Concen: 29.953 ng/ul
RT: 21.398 min Scan# 3096
Delta R.T. 0.000 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01




Tgt Ion:252 Resp: 1593601
Ion Ratio Lower Upper
252 100
254 64.9 52.3 78.5
126 9.8 7.6 11.4





#85
Benzo(a)anthracene
Concen: 29.414 ng/ul
RT: 21.469 min Scan# 3108
Delta R.T. 0.000 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01




Instrument :
BNA_P
ClientSampleId :
SLCS233




Tgt Ion:228 Resp: 5894047
Ion Ratio Lower Upper
228 100
229 20.1 16.2 24.2
226 27.5 22.2 33.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022




#86
Bis(2-ethylhexyl)phthalate
Concen: 41.392 ng/ul
RT: 21.380 min Scan# 3093
Delta R.T. -0.006 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01




Tgt Ion:149 Resp: 3207981
Ion Ratio Lower Upper
149 100
167 28.3 22.7 34.1
279 4.5 3.8 5.6





#87
Chrysene
Concen: 28.539 ng/ul
RT: 21.527 min Scan# 3118
Delta R.T. 0.000 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01




Instrument :
BNA_P
ClientSampleId :
SLCS233



Tgt Ion:228 Resp: 5529272
Ion Ratio Lower Upper
228 100
226 30.6 24.3 36.5
229 20.2 15.9 23.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022



#89
Di-n-octyl phthalate
Concen: 34.670 ng/ul
RT: 22.345 min Scan# 3257
Delta R.T. 0.000 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01


Tgt Ion:149 Resp: 4538418





#90
Benzo(b)fluoranthene
Concen: 29.030 ng/ul
RT: 23.227 min Scan# 3407
Delta R.T. 0.000 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01




Instrument :
BNA_P
ClientSampleId :
SLCS233




Tgt Ion:252 Resp: 4927836
Ion Ratio Lower Upper
252 100
253 22.4 17.6 26.4
125 9.2 7.4 11.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022




#91
Benzo(k)fluoranthene
Concen: 28.763 ng/ul
RT: 23.274 min Scan# 3415
Delta R.T. -0.006 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01




Tgt Ion:252 Resp: 4873705
Ion Ratio Lower Upper
252 100
253 22.0 17.8 26.8
125 9.3 7.8 11.8





#93
Benzo(a)pyrene
Concen: 30.250 ng/ul
RT: 23.886 min Scan# 3519
Delta R.T. 0.000 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01




Instrument :
BNA_P
ClientSampleId :
SLCS233



Tgt Ion:252 Resp: 4222410
Ion Ratio Lower Upper
252 100
253 22.4 17.6 26.4
125 10.2 8.5 12.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022



#94
Indeno(1,2,3-cd)pyrene
Concen: 37.651 ng/ul
RT: 26.627 min Scan# 3985
Delta R.T. 0.000 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01


Tgt Ion:276 Resp: 5070381
Ion Ratio Lower Upper
276 100
138 22.6 18.3 27.5
277 24.7 19.8 29.8





#95
Dibenzo(a,h)anthracene
Concen: 36.530 ng/ul
RT: 26.645 min Scan# 3989
Delta R.T. -0.006 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01




Instrument :
BNA_P
ClientSampleId :
SLCS233




Tgt Ion:278 Resp: 4463848
Ion Ratio Lower Upper
278 100
139 15.1 12.6 18.8
279 24.2 19.4 29.2

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022



#96
Benzo(g,h,i)perylene
Concen: 38.227 ng/ul
RT: 27.433 min Scan# 4122
Delta R.T. -0.006 min
Lab File: BP012871.D
Acq: 30 Nov 2022 02:01



Tgt Ion:276 Resp: 4427951
Ion Ratio Lower Upper
276 100
138 20.1 16.7 25.1
277 24.3 19.1 28.7

