

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112922\
 Data File : BP012872.D
 Acq On : 30 Nov 2022 02:35
 Operator : CG/JU
 Sample : PB149234BS
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS234

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 11/30/2022
 Supervised By :mohammad ahmed 11/30/2022

Quant Time: Nov 30 03:29:21 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP112522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Nov 29 22:53:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.010	152	541361	20.000	ng/u1	0.00
20) Naphthalene-d8	10.828	136	2488340	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.645	164	1678912	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.404	188	3587469	20.000	ng/u1	0.00
79) Chrysene-d12	21.486	240	3065346	20.000	ng/u1	0.00
88) Perylene-d12	23.992	264	2408542	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.393	96	94202	7.457	ng/uL	0.00
4) Pyridine-d5	3.817	84	1118029	29.572	ng/u1	0.00
7) Phenol-d5	7.163	99	1868468	42.002	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.334	67	1082812	39.502	ng/u1	0.00
11) 2-Chlorophenol-d4	7.540	132	1442426	42.553	ng/u1	0.00
15) 4-Methylphenol-d8	8.716	113	1520349	42.418	ng/u1	0.00
21) Nitrobenzene-d5	9.181	128	727638	47.776	ng/u1	0.00
24) 2-Nitrophenol-d4	9.904	143	828858	51.254	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.440	165	1611253	42.702	ng/u1	0.00
31) 4-Chloroaniline-d4	10.963	131	1064314	20.598	ng/u1	0.00
46) Dimethylphthalate-d6	14.051	166	4711977	40.774	ng/u1	0.00
49) Acenaphthylene-d8	14.340	160	5393362	39.999	ng/u1	0.00
54) 4-Nitrophenol-d4	14.834	143	877749	42.602	ng/u1	0.00
60) Fluorene-d10	15.639	176	4128325	40.791	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.757	200	827295	48.249	ng/u1	0.00
73) Anthracene-d10	17.504	188	6316207	39.997	ng/u1	0.00
81) Pyrene-d10	19.727	212	7009076	39.352	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.833	264	5043070	42.627	ng/u1	0.00
Target Compounds					Qvalue	
2) 1,4-Dioxane	3.428	88	193145	14.236	ng/uL	99
5) Pyridine	3.840	79	1347101	36.415	ng/u1	98
6) Benzaldehyde	7.146	77	903349m	41.590	ng/u1	
8) Phenol	7.193	94	1908580	40.916	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.428	93	1502414	39.311	ng/u1	100
12) 2-Chlorophenol	7.569	128	1506353	40.853	ng/u1	100
13) 2-Methylphenol	8.452	108	1481504	40.866	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.546	45	2094998	38.003	ng/u1	99
16) Acetophenone	8.840	105	2298692	40.464	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.828	70	1144679	40.904	ng/u1	96
18) 4-Methylphenol	8.781	108	1636230	40.981	ng/u1	99
19) Hexachloroethane	9.105	117	569423	40.270	ng/u1	93
22) Nitrobenzene	9.222	77	1663220	41.638	ng/u1	98
23) Isophorone	9.752	82	3384432	38.434	ng/u1	98
25) 2-Nitrophenol	9.934	139	904737	46.609	ng/u1	96
26) 2,4-Dimethylphenol	9.993	107	1704483	39.517	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.234	93	2106697	39.021	ng/u1	100
29) 2,4-Dichlorophenol	10.469	162	1591212	40.708	ng/u1	100
30) Naphthalene	10.881	128	4982970	38.243	ng/u1	100
32) 4-Chloroaniline	10.987	127	1684650	31.747	ng/u1	98
33) Hexachlorobutadiene	11.163	225	1017101	39.066	ng/u1	99
34) Caprolactam	11.769	113	514841m	42.893	ng/u1	
35) 4-Chloro-3-methylphenol	12.098	107	1663607	42.029	ng/u1	99
36) 2-Methylnaphthalene	12.481	142	3539729	39.450	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112922\
 Data File : BP012872.D
 Acq On : 30 Nov 2022 02:35
 Operator : CG/JU
 Sample : PB149234BS
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS234

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 11/30/2022
 Supervised By :mohammad ahmed 11/30/2022

Quant Time: Nov 30 03:29:21 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP112522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Nov 29 22:53:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.698	142	3536762	39.330	ng/u1	100
39) 1,2,4,5-Tetrachloroben...	12.845	216	2043241	38.165	ng/u1	99
40) Hexachlorocyclopentadiene	12.828	237	1051930	40.778	ng/u1	99
41) 2,4,6-Trichlorophenol	13.081	196	1356387	40.491	ng/u1	99
42) 2,4,5-Trichlorophenol	13.151	196	1488812	41.343	ng/u1	99
43) 1,1'-Biphenyl	13.481	154	4675953	38.232	ng/u1	100
44) 2-Chloronaphthalene	13.528	162	3741574	38.241	ng/u1	99
45) 2-Nitroaniline	13.728	65	990223	51.726	ng/u1	98
47) Dimethylphthalate	14.104	163	4766684	39.349	ng/u1	99
48) 2,6-Dinitrotoluene	14.222	165	977090	51.540	ng/u1	96
50) Acenaphthylene	14.369	152	5732871	38.883	ng/u1	100
51) 3-Nitroaniline	14.551	138	935996	42.631	ng/u1	95
52) Acenaphthene	14.710	153	3976919	38.313	ng/u1	98
53) 2,4-Dinitrophenol	14.751	184	580961	47.690	ng/u1	94
55) 4-Nitrophenol	14.851	109	603670	40.329	ng/u1	95
56) Dibenzofuran	15.045	168	5585489	39.038	ng/u1	99
57) 2,4-Dinitrotoluene	15.004	165	1391322	49.733	ng/u1	98
58) 2,3,4,6-Tetrachlorophenol	15.269	232	1274645	41.067	ng/u1	97
59) Diethylphthalate	15.469	149	4675214	39.871	ng/u1	98
61) Fluorene	15.698	166	4439912	38.695	ng/u1	99
62) 4-Chlorophenyl-phenyle...	15.687	204	2359110	39.627	ng/u1	98
63) 4-Nitroaniline	15.716	138	980863	52.416	ng/u1	98
66) 4,6-Dinitro-2-methylph...	15.769	198	851485	46.264	ng/u1#	94
67) N-Nitrosodiphenylamine	15.904	169	3969761	38.936	ng/u1	99
68) 4-Bromophenyl-phenylether	16.586	248	1515626	42.925	ng/u1	99
69) Hexachlorobenzene	16.704	284	1794263	39.870	ng/u1	97
70) Atrazine	16.851	200	611214	17.473	ng/u1	99
71) Pentachlorophenol	17.045	266	1066580	38.594	ng/u1	99
72) Phenanthrene	17.445	178	7343742	38.409	ng/u1	99
74) Anthracene	17.539	178	7306407	38.627	ng/u1	99
75) 1,2,3,4-Tetrachloroben...	13.445	216	2055548	37.063	ng/uL	99
76) Pentachlorobenzene	14.963	250	2001799	34.722	ng/uL	99
77) Carbazole	17.804	167	6547611	40.918	ng/u1	100
78) Di-n-butylphthalate	18.345	149	7818999	41.276	ng/u1	99
80) Fluoranthene	19.404	202	8422737	38.018	ng/u1	98
82) Pyrene	19.757	202	8524301	37.201	ng/u1	98
83) Butylbenzylphthalate	20.622	149	3174223	59.198	ng/u1	92
84) 3,3'-Dichlorobenzidine	21.398	252	2088382	37.099	ng/u1	100
85) Benzo(a)anthracene	21.469	228	7679972	36.224	ng/u1	100
86) Bis(2-ethylhexyl)phtha...	21.380	149	4418663	53.885	ng/u1	100
87) Chrysene	21.527	228	7174841	35.000	ng/u1	99
89) Di-n-octyl phthalate	22.345	149	6776705	49.270	ng/u1	100
90) Benzo(b)fluoranthene	23.227	252	6647662	37.272	ng/u1	99
91) Benzo(k)fluoranthene	23.280	252	6658620	37.401	ng/u1	99
93) Benzo(a)pyrene	23.886	252	5581148	38.054	ng/u1	99
94) Indeno(1,2,3-cd)pyrene	26.627	276	6046420	42.732	ng/u1	98
95) Dibenzo(a,h)anthracene	26.645	278	5323583	41.463	ng/u1	98
96) Benzo(g,h,i)perylene	27.433	276	5144745	42.272	ng/u1	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP112922\
 Data File : BP012872.D
 Acq On : 30 Nov 2022 02:35
 Operator : CG/JU
 Sample : PB149234BS
 Misc :
 ALS Vial : 31 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SLCS234

Quant Time: Nov 30 03:29:21 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP112522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Nov 29 22:53:34 2022
 Response via : Initial Calibration

Manual Integrations APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2022
 Supervised By :mohammad ahmed 11/30/2022





#2
1,4-Dioxane
Concen: 14.236 ng/uL
RT: 3.428 min Scan# 41
Delta R.T. 0.000 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35




Instrument :
BNA_P
ClientSampleId :
SLCS234




Tgt Ion: 88 Resp: 193145
Ion Ratio Lower Upper
88 100
43 28.4 22.6 34.0
58 73.0 59.6 89.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022




#5
Pyridine
Concen: 36.415 ng/uL
RT: 3.840 min Scan# 111
Delta R.T. 0.000 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35




Tgt Ion: 79 Resp: 1347101
Ion Ratio Lower Upper
79 100
52 71.3 58.9 88.3
51 34.0 27.3 40.9





#6
Benzaldehyde
Concen: 41.590 ng/ul m
RT: 7.146 min Scan# 61
Delta R.T. 0.000 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35




Instrument :
BNA_P
ClientSampleId :
SLCS234




Tgt Ion: 77 Resp: 903349
Ion Ratio Lower Upper
77 100
105 102.2 81.4 122.2
106 98.3 77.4 116.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022




#8
Phenol
Concen: 40.916 ng/ul
RT: 7.193 min Scan# 681
Delta R.T. 0.006 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35




Tgt Ion: 94 Resp: 1908580
Ion Ratio Lower Upper
94 100
65 26.1 20.9 31.3
66 31.6 26.0 39.0





#10
Bis(2-Chloroethyl)ether
Concen: 39.311 ng/ul
RT: 7.428 min Scan# 721
Delta R.T. 0.000 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35




Instrument :
BNA_P
ClientSampleId :
SLCS234




Tgt Ion: 93 Resp: 1502414
Ion Ratio Lower Upper
93 100
63 74.6 59.7 89.5
95 32.9 25.8 38.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022




#12
2-Chlorophenol
Concen: 40.853 ng/ul
RT: 7.569 min Scan# 745
Delta R.T. 0.000 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35



Tgt Ion:128 Resp: 1506353
Ion Ratio Lower Upper
128 100
64 44.5 35.6 53.4
130 32.6 26.0 39.0





#13

2-Methylphenol

Concen: 40.866 ng/ul

RT: 8.452 min Scan# 895

Delta R.T. 0.000 min

Lab File: BP012872.D

Acq: 30 Nov 2022 02:35

Instrument :

BNA_P

ClientSampleId :

SLCS234

Tgt Ion:108 Resp: 1481504

Ion Ratio Lower Upper





108 100

107 88.0 69.3 103.9

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022



#14

2,2'-oxybis(1-Chloropropane)

Concen: 38.003 ng/ul

RT: 8.546 min Scan# 911

Delta R.T. 0.000 min

Lab File: BP012872.D

Acq: 30 Nov 2022 02:35




Tgt Ion: 45 Resp: 2094998


Ion Ratio Lower Upper

45 100

77 14.2 11.3 16.9


79 11.1 8.6 12.8





#16
Acetophenone
Concen: 40.464 ng/ul
RT: 8.840 min Scan# 90
Delta R.T. 0.000 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35




Instrument :
BNA_P
ClientSampleId :
SLCS234




Tgt Ion:105 Resp: 2298692
Ion Ratio Lower Upper
105 100
77 72.9 59.7 89.5
51 25.3 20.9 31.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022




#17
N-Nitroso-di-n-propylamine
Concen: 40.904 ng/ul
RT: 8.828 min Scan# 959
Delta R.T. 0.006 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35




Tgt Ion: 70 Resp: 1144679
Ion Ratio Lower Upper
70 100
42 55.2 46.1 69.1
101 10.1 7.4 11.2
130 22.1 15.8 23.6





#18
4-Methylphenol
Concen: 40.981 ng/ul
RT: 8.781 min Scan# 951
Delta R.T. 0.000 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35




Instrument :
BNA_P
ClientSampleId :
SLCS234



Tgt Ion:108 Resp: 1636230
Ion Ratio Lower Upper
108 100
107 113.7 90.2 135.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022



#19
Hexachloroethane
Concen: 40.270 ng/ul
RT: 9.105 min Scan# 1006
Delta R.T. 0.000 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35


Tgt Ion:117 Resp: 569423
Ion Ratio Lower Upper
117 100
201 114.9 85.7 128.5
199 71.7 52.9 79.3





#22
Nitrobenzene
Concen: 41.638 ng/ul
RT: 9.222 min Scan# 1026
Delta R.T. 0.000 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35




Instrument :
BNA_P
ClientSampleId :
SLCS234



Tgt Ion: 77 Resp: 1663220
Ion Ratio Lower Upper
77 100
123 49.1 38.2 57.2
65 12.8 10.2 15.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022



#23
Isophorone
Concen: 38.434 ng/ul
RT: 9.752 min Scan# 1116
Delta R.T. 0.000 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35

Tgt Ion: 82 Resp: 3384432
Ion Ratio Lower Upper
82 100
95 6.8 5.2 7.8
138 18.9 14.5 21.7









#25
2-Nitrophenol
Concen: 46.609 ng/ul
RT: 9.934 min Scan# 1147
Delta R.T. 0.000 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35

Instrument :
BNA_P
ClientSampleId :
SLCS234

Tgt Ion:139 Resp: 90473
Ion Ratio Lower Upper
139 100
65 45.9 39.5 59.3
109 26.1 21.4 32.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022



#26
2,4-Dimethylphenol
Concen: 39.517 ng/ul
RT: 9.993 min Scan# 1157
Delta R.T. 0.000 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35


Tgt Ion:107 Resp: 1704483
Ion Ratio Lower Upper
107 100
121 52.9 41.8 62.6
122 91.3 72.3 108.5





#27
Bis(2-Chloroethoxy)methane
Concen: 39.021 ng/ul
RT: 10.234 min Scan# 1198
Delta R.T. 0.000 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35




Instrument :
BNA_P
ClientSampleId :
SLCS234



Tgt Ion: 93 Resp: 2106697
Ion Ratio Lower Upper
93 100
95 32.5 26.1 39.1
123 11.6 9.0 13.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022



#29
2,4-Dichlorophenol
Concen: 40.708 ng/ul
RT: 10.469 min Scan# 1238
Delta R.T. 0.000 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35


Tgt Ion:162 Resp: 1591212
Ion Ratio Lower Upper
162 100
164 65.9 52.9 79.3
98 33.6 26.6 40.0





#30
Naphthalene
Concen: 38.243 ng/ul
RT: 10.881 min Scan# 1308
Delta R.T. 0.000 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35




Instrument :
BNA_P
ClientSampleId :
SLCS234



Tgt Ion:128 Resp: 4982970
Ion Ratio Lower Upper
128 100
129 10.9 8.7 13.1
127 13.1 10.4 15.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022



#32
4-Chloroaniline
Concen: 31.747 ng/ul
RT: 10.987 min Scan# 1326
Delta R.T. 0.000 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35


Tgt Ion:127 Resp: 1684650
Ion Ratio Lower Upper
127 100
129 33.0 25.7 38.5





#33
Hexachlorobutadiene
Concen: 39.066 ng/ul
RT: 11.163 min Scan# 1356
Delta R.T. 0.000 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35




Instrument :
BNA_P
ClientSampleId :
SLCS234



Tgt Ion:225 Resp: 1017101
Ion Ratio Lower Upper
225 100
223 63.1 49.6 74.4
227 64.8 51.0 76.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022



#34
Caprolactam
Concen: 42.893 ng/ul m
RT: 11.769 min Scan# 1459
Delta R.T. 0.012 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35


Tgt Ion:113 Resp: 514841
Ion Ratio Lower Upper
113 100
55 165.4 145.9 218.9
56 121.0 107.1 160.7





#35
4-Chloro-3-methylphenol
Concen: 42.029 ng/ul
RT: 12.098 min Scan# 1515
Delta R.T. 0.000 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35




Instrument :
BNA_P
ClientSampleId :
SLCS234




Tgt Ion:107 Resp: 1663607
Ion Ratio Lower Upper
107 100
144 27.5 21.4 32.2
142 84.4 66.6 100.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022




#36
2-Methylnaphthalene
Concen: 39.450 ng/ul
RT: 12.481 min Scan# 1580
Delta R.T. 0.000 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35




Tgt Ion:142 Resp: 3539729
Ion Ratio Lower Upper
142 100
141 88.4 70.2 105.2





#37
1-Methylnaphthalene
Concen: 39.330 ng/ul
RT: 12.698 min Scan# 1617
Delta R.T. 0.000 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35




Instrument :
BNA_P
ClientSampleId :
SLCS234




Tgt Ion:142 Resp: 3536762
Ion Ratio Lower Upper
142 100
141 91.3 73.3 109.9
116 3.6 2.9 4.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022




#39
1,2,4,5-Tetrachlorobenzene
Concen: 38.165 ng/ul
RT: 12.845 min Scan# 1642
Delta R.T. 0.000 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35




Tgt Ion:216 Resp: 2043241
Ion Ratio Lower Upper
216 100
214 78.3 63.0 94.6
179 19.5 16.2 24.4
108 13.9 12.7 19.1





#40
Hexachlorocyclopentadiene
Concen: 40.778 ng/ul
RT: 12.828 min Scan# 1639
Delta R.T. 0.000 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35




Instrument :
BNA_P
ClientSampleId :
SLCS234




Tgt Ion:237 Resp: 1051930
Ion Ratio Lower Upper
237 100
235 63.5 50.2 75.4
272 13.1 10.2 15.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022




#41
2,4,6-Trichlorophenol
Concen: 40.491 ng/ul
RT: 13.081 min Scan# 1682
Delta R.T. 0.000 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35




Tgt Ion:196 Resp: 1356387
Ion Ratio Lower Upper
196 100
198 96.3 76.5 114.7
200 31.4 25.3 37.9





#42
2,4,5-Trichlorophenol
Concen: 41.343 ng/ul
RT: 13.151 min Scan# 1694
Delta R.T. 0.006 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35




Instrument :
BNA_P
ClientSampleId :
SLCS234




Tgt Ion:196 Resp: 1488812
Ion Ratio Lower Upper
196 100
198 97.0 77.6 116.4
200 31.6 24.2 36.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022




#43
1,1'-Biphenyl
Concen: 38.232 ng/ul
RT: 13.481 min Scan# 1750
Delta R.T. 0.000 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35




Tgt Ion:154 Resp: 4675953
Ion Ratio Lower Upper
154 100
153 39.8 32.0 48.0
76 12.4 10.2 15.4





#44
2-Chloronaphthalene
Concen: 38.241 ng/ul
RT: 13.528 min Scan# 1758
Delta R.T. 0.000 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35




Instrument :
BNA_P
ClientSampleId :
SLCS234



Tgt Ion:162 Resp: 3741574
Ion Ratio Lower Upper
162 100
164 33.8 26.5 39.7
127 35.7 28.8 43.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022



#45
2-Nitroaniline
Concen: 51.726 ng/ul
RT: 13.728 min Scan# 1792
Delta R.T. 0.000 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35


Tgt Ion: 65 Resp: 990223
Ion Ratio Lower Upper
65 100
92 73.0 57.8 86.8
138 121.4 94.6 142.0





#47
Dimethylphthalate
Concen: 39.349 ng/ul
RT: 14.104 min Scan# 1856
Delta R.T. 0.006 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35




Instrument :
BNA_P
ClientSampleId :
SLCS234



Tgt Ion:163 Resp: 4766684
Ion Ratio Lower Upper
163 100
194 4.2 3.4 5.0
164 10.2 7.8 11.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022



#48
2,6-Dinitrotoluene
Concen: 51.540 ng/ul
RT: 14.222 min Scan# 1876
Delta R.T. 0.000 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35


Tgt Ion:165 Resp: 977090
Ion Ratio Lower Upper
165 100
89 50.0 42.7 64.1
121 19.4 16.9 25.3





#50
Acenaphthylene
Concen: 38.883 ng/ul
RT: 14.369 min Scan# 1901
Delta R.T. 0.000 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35




Instrument :
BNA_P
ClientSampleId :
SLCS234




Tgt Ion:152 Resp: 573287
Ion Ratio Lower Upper
152 100
151 20.0 15.9 23.9
153 13.6 10.7 16.1

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022




#51
3-Nitroaniline
Concen: 42.631 ng/ul
RT: 14.551 min Scan# 1932
Delta R.T. 0.006 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35




Tgt Ion:138 Resp: 935996
Ion Ratio Lower Upper
138 100
108 10.1 9.9 14.9
92 107.3 90.2 135.2





#52
Acenaphthene
Concen: 38.313 ng/ul
RT: 14.710 min Scan# 1959
Delta R.T. 0.000 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35




Instrument :
BNA_P
ClientSampleId :
SLCS234




Tgt Ion:153 Resp: 3976919
Ion Ratio Lower Upper
153 100
152 47.6 38.2 57.4
154 89.3 69.8 104.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022




#53
2,4-Dinitrophenol
Concen: 47.690 ng/ul
RT: 14.751 min Scan# 1966
Delta R.T. 0.000 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35




Tgt Ion:184 Resp: 580961
Ion Ratio Lower Upper
184 100
63 54.5 48.1 72.1
154 56.3 47.4 71.2





#55
4-Nitrophenol
Concen: 40.329 ng/ul
RT: 14.851 min Scan# 1982
Delta R.T. 0.006 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35




Instrument :
BNA_P
ClientSampleId :
SLCS234



Tgt Ion:109 Resp: 603670
Ion Ratio Lower Upper
109 100
139 154.2 119.2 178.8
65 131.5 110.7 166.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022



#56
Dibenzofuran
Concen: 39.038 ng/ul
RT: 15.045 min Scan# 2016
Delta R.T. 0.000 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35


Tgt Ion:168 Resp: 5585489
Ion Ratio Lower Upper
168 100
139 34.2 27.7 41.5





#57
2,4-Dinitrotoluene
Concen: 49.733 ng/ul
RT: 15.004 min Scan# 2009
Delta R.T. 0.000 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35




Instrument :
BNA_P
ClientSampleId :
SLCS234




Tgt Ion:165 Resp: 1391322
Ion Ratio Lower Upper
165 100
63 37.7 31.0 46.4
182 2.7 2.0 3.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022




#58
2,3,4,6-Tetrachlorophenol
Concen: 41.067 ng/ul
RT: 15.269 min Scan# 2054
Delta R.T. 0.000 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35




Tgt Ion:232 Resp: 1274645
Ion Ratio Lower Upper
232 100
131 37.5 32.0 48.0
130 2.0 1.5 2.3
166 28.4 23.2 34.8





#59
Diethylphthalate
Concen: 39.871 ng/ul
RT: 15.469 min Scan# 2087
Delta R.T. 0.006 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35




Instrument :
BNA_P
ClientSampleId :
SLCS234



Tgt Ion:149 Resp: 4675214
Ion Ratio Lower Upper
149 100
177 24.1 18.1 27.1
150 12.4 9.8 14.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022



#61
Fluorene
Concen: 38.695 ng/ul
RT: 15.698 min Scan# 2127
Delta R.T. 0.000 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35


Tgt Ion:166 Resp: 4439912
Ion Ratio Lower Upper
166 100
165 97.0 78.2 117.2
167 13.4 10.9 16.3





#62
4-Chlorophenyl-phenylether
Concen: 39.627 ng/ul
RT: 15.687 min Scan# 2125
Delta R.T. 0.000 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35




Instrument :
BNA_P
ClientSampleId :
SLCS234




Tgt Ion:204 Resp: 2359110
Ion Ratio Lower Upper
204 100
206 33.0 26.0 39.0
141 49.2 40.7 61.1

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022




#63
4-Nitroaniline
Concen: 52.416 ng/ul
RT: 15.716 min Scan# 2130
Delta R.T. 0.006 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35




Tgt Ion:138 Resp: 980863
Ion Ratio Lower Upper
138 100
92 47.7 38.6 57.8
108 68.0 56.4 84.6





#66
4,6-Dinitro-2-methylphenol
Concen: 46.264 ng/ul
RT: 15.769 min Scan# 2138
Delta R.T. 0.006 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35




Instrument :
BNA_P
ClientSampleId :
SLCS234



Tgt Ion:198 Resp: 851485
Ion Ratio Lower Upper
198 100
182 3.8 4.1 6.1
77 20.4 18.9 28.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022



#67
N-Nitrosodiphenylamine
Concen: 38.936 ng/ul
RT: 15.904 min Scan# 2162
Delta R.T. 0.000 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35

Tgt Ion:169 Resp: 3969761
Ion Ratio Lower Upper
169 100
168 65.5 53.2 79.8
167 36.2 29.6 44.4









#68
4-Bromophenyl-phenylether
Concen: 42.925 ng/ul
RT: 16.586 min Scan# 2278
Delta R.T. 0.000 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35

Instrument :
BNA_P
ClientSampleId :
SLCS234

Tgt Ion:248 Resp: 1515626
Ion Ratio Lower Upper
248 100
250 97.3 78.6 118.0
141 60.3 48.7 73.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022



#69
Hexachlorobenzene
Concen: 39.870 ng/ul
RT: 16.704 min Scan# 2298
Delta R.T. 0.000 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35

Tgt Ion:284 Resp: 1794263
Ion Ratio Lower Upper
284 100
142 28.3 25.0 37.6
249 30.4 24.6 36.8





#70

Atrazine

Concen: 17.473 ng/ul

RT: 16.851 min Scan# 2323

Delta R.T. 0.000 min

Lab File: BP012872.D

Acq: 30 Nov 2022 02:35

Instrument :

BNA_P

ClientSampleId :

SLCS234

Tgt Ion:200 Resp: 611214

Ion Ratio Lower Upper

200 100





173 26.3 20.3 30.5

215 50.1 40.1 60.1

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022



#71

Pentachlorophenol

Concen: 38.594 ng/ul

RT: 17.045 min Scan# 2356

Delta R.T. 0.000 min

Lab File: BP012872.D

Acq: 30 Nov 2022 02:35




Tgt Ion:266 Resp: 1066580


Ion Ratio Lower Upper

266 100

264 62.1 50.7 76.1


268 64.2 51.8 77.8





#72
Phenanthrene
Concen: 38.409 ng/ul
RT: 17.445 min Scan# 2424
Delta R.T. 0.000 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35




Instrument :
BNA_P
Client Sample Id :
SLCS234




Tgt Ion:178 Resp: 7343742
Ion Ratio Lower Upper
178 100
179 15.8 12.3 18.5
176 19.5 15.4 23.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022




#74
Anthracene
Concen: 38.627 ng/ul
RT: 17.539 min Scan# 2440
Delta R.T. 0.000 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35




Tgt Ion:178 Resp: 7306407
Ion Ratio Lower Upper
178 100
179 15.9 12.4 18.6
176 19.0 14.8 22.2





#75
1,2,3,4-Tetrachlorobenzene
Concen: 37.063 ng/uL
RT: 13.445 min Scan# 1745
Delta R.T. -0.006 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35




Instrument :
BNA_P
ClientSampleId :
SLCS234



Tgt Ion:216 Resp: 2055548
Ion Ratio Lower Upper
216 100
214 78.4 62.3 93.5
218 48.1 39.1 58.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022



#76
Pentachlorobenzene
Concen: 34.722 ng/uL
RT: 14.963 min Scan# 2002
Delta R.T. 0.000 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35


Tgt Ion:250 Resp: 2001799
Ion Ratio Lower Upper
250 100
248 64.0 50.6 76.0
252 64.3 50.5 75.7





#77
 Carbazole
 Concen: 40.918 ng/ul
 RT: 17.804 min Scan# 2485
 Delta R.T. 0.000 min
 Lab File: BP012872.D
 Acq: 30 Nov 2022 02:35




Instrument :
 BNA_P
 ClientSampleId :
 SLCS234




Tgt Ion:167 Resp: 6547611
 Ion Ratio Lower Upper
 167 100
 166 21.5 17.1 25.7
 139 12.1 9.8 14.8

Manual Integrations
 APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2022
 Supervised By :mohammad ahmed 11/30/2022




#78
 Di-n-butylphthalate
 Concen: 41.276 ng/ul
 RT: 18.345 min Scan# 2577
 Delta R.T. -0.006 min
 Lab File: BP012872.D
 Acq: 30 Nov 2022 02:35



Tgt Ion:149 Resp: 7818999
 Ion Ratio Lower Upper
 149 100
 150 9.6 7.2 10.8
 104 4.6 3.7 5.5









#80
Fluoranthene
Concen: 38.018 ng/ul
RT: 19.404 min Scan# 2757
Delta R.T. 0.000 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35

Instrument :
BNA_P
ClientSampleId :
SLCS234

Tgt Ion:202 Resp: 842273
Ion Ratio Lower Upper
202 100
101 10.3 8.7 13.1
100 7.8 6.9 10.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022



#82
Pyrene
Concen: 37.201 ng/ul
RT: 19.757 min Scan# 2817
Delta R.T. 0.000 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35


Tgt Ion:202 Resp: 8524301
Ion Ratio Lower Upper
202 100
101 12.1 10.3 15.5
100 9.9 8.6 13.0





#83
Butylbenzylphthalate
Concen: 59.198 ng/ul
RT: 20.622 min Scan# 2964
Delta R.T. 0.000 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35




Instrument :
BNA_P
ClientSampleId :
SLCS234




Tgt Ion:149 Resp: 3174223
Ion Ratio Lower Upper
149 100
91 65.6 58.2 87.2
206 22.8 16.5 24.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022




#84
3,3'-Dichlorobenzidine
Concen: 37.099 ng/ul
RT: 21.398 min Scan# 3096
Delta R.T. 0.000 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35




Tgt Ion:252 Resp: 2088382
Ion Ratio Lower Upper
252 100
254 65.1 52.3 78.5
126 9.4 7.6 11.4





#85
Benzo(a)anthracene
Concen: 36.224 ng/ul
RT: 21.469 min Scan# 3108
Delta R.T. 0.000 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35




Instrument :
BNA_P
ClientSampleId :
SLCS234



Tgt Ion:228 Resp: 7679972
Ion Ratio Lower Upper
228 100
229 20.3 16.2 24.2
226 27.8 22.2 33.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022



#86
Bis(2-ethylhexyl)phthalate
Concen: 53.885 ng/ul
RT: 21.380 min Scan# 3093
Delta R.T. -0.006 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35

Tgt Ion:149 Resp: 4418663
Ion Ratio Lower Upper
149 100
167 28.3 22.7 34.1
279 4.6 3.8 5.6









#87
Chrysene
Concen: 35.000 ng/ul
RT: 21.527 min Scan# 3118
Delta R.T. 0.000 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35

Instrument :
BNA_P
ClientSampleId :
SLCS234

Tgt Ion:228 Resp: 717484
Ion Ratio Lower Upper
228 100
226 30.9 24.3 36.5
229 20.3 15.9 23.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022



#89
Di-n-octyl phthalate
Concen: 49.270 ng/ul
RT: 22.345 min Scan# 3257
Delta R.T. 0.000 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35


Tgt Ion:149 Resp: 6776705





#90
Benzo(b)fluoranthene
Concen: 37.272 ng/ul
RT: 23.227 min Scan# 3407
Delta R.T. 0.000 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35




Instrument :
BNA_P
ClientSampleId :
SLCS234




Tgt Ion:252 Resp: 6647662
Ion Ratio Lower Upper
252 100
253 22.2 17.6 26.4
125 9.0 7.4 11.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022




#91
Benzo(k)fluoranthene
Concen: 37.401 ng/ul
RT: 23.280 min Scan# 3416
Delta R.T. 0.000 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35




Tgt Ion:252 Resp: 6658620
Ion Ratio Lower Upper
252 100
253 22.2 17.8 26.8
125 8.8 7.8 11.8





#93
Benzo(a)pyrene
Concen: 38.054 ng/ul
RT: 23.886 min Scan# 3519
Delta R.T. 0.000 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35




Instrument :
BNA_P
ClientSampleId :
SLCS234



Tgt Ion:252 Resp: 5581148
Ion Ratio Lower Upper
252 100
253 22.3 17.6 26.4
125 10.0 8.5 12.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022



#94
Indeno(1,2,3-cd)pyrene
Concen: 42.732 ng/ul
RT: 26.627 min Scan# 3985
Delta R.T. 0.000 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35

Tgt Ion:276 Resp: 6046420
Ion Ratio Lower Upper
276 100
138 22.2 18.3 27.5
277 25.6 19.8 29.8









#95
Dibenzo(a,h)anthracene
Concen: 41.463 ng/ul
RT: 26.645 min Scan# 3989
Delta R.T. -0.006 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35

Instrument :
BNA_P
ClientSampleId :
SLCS234

Tgt Ion:278 Resp: 5323583
Ion Ratio Lower Upper
278 100
139 14.7 12.6 18.8
279 23.7 19.4 29.2

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 11/30/2022
Supervised By :mohammad ahmed 11/30/2022



#96
Benzo(g,h,i)perylene
Concen: 42.272 ng/ul
RT: 27.433 min Scan# 4122
Delta R.T. -0.006 min
Lab File: BP012872.D
Acq: 30 Nov 2022 02:35

Tgt Ion:276 Resp: 5144745
Ion Ratio Lower Upper
276 100
138 20.0 16.7 25.1
277 23.9 19.1 28.7

