

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120123\
 Data File : BP018480.D
 Acq On : 01 Dec 2023 17:21
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020701

Quant Time: Dec 01 22:36:05 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Dec 01 21:56:43 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.869	152	356200	20.000	ng/ul	0.00
20) Naphthalene-d8	10.663	136	1530831	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.493	164	973525	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.239	188	1925021	20.000	ng/ul	0.00
79) Chrysene-d12	21.327	240	1243105	20.000	ng/ul	0.00
88) Perylene-d12	23.704	264	1430452	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.275	96	83830	8.020	ng/uL	0.00
4) Pyridine-d5	3.693	84	593308	21.191	ng/ul	0.00
7) Phenol-d5	7.034	99	734278	20.542	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.199	67	443563	20.936	ng/ul	0.00
11) 2-Chlorophenol-d4	7.393	132	565053	20.258	ng/ul	0.00
15) 4-Methylphenol-d8	8.581	113	598577	20.692	ng/ul	0.00
21) Nitrobenzene-d5	9.028	128	286562	21.032	ng/ul	0.00
24) 2-Nitrophenol-d4	9.746	143	306623	21.707	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.287	165	598014	20.704	ng/ul	0.00
31) 4-Chloroaniline-d4	10.804	131	817925	21.668	ng/ul	0.00
46) Dimethylphthalate-d6	13.910	166	1731500	19.928	ng/ul	0.00
49) Acenaphthylene-d8	14.187	160	1984893	20.746	ng/ul	0.00
54) 4-Nitrophenol-d4	14.693	143	276600	19.394	ng/ul	0.00
60) Fluorene-d10	15.487	176	1447133	19.842	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.604	200	228970	19.837	ng/ul	0.00
73) Anthracene-d10	17.339	188	2049962	20.256	ng/ul	0.00
81) Pyrene-d10	19.575	212	1986750	20.532	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.545	264	1670172	20.025	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.311	88	89566	7.814	ng/uL	98
5) Pyridine	3.717	79	598234	21.158	ng/ul	96
6) Benzaldehyde	7.005	77	442556	23.979	ng/ul	98
8) Phenol	7.058	94	727730	20.760	ng/ul	99
10) Bis(2-Chloroethyl)ether	7.293	93	609083	21.157	ng/ul	99
12) 2-Chlorophenol	7.428	128	569019	20.157	ng/ul	99
13) 2-Methylphenol	8.310	108	555034	20.681	ng/ul	98
14) 2,2'-oxybis(1-Chloropr...	8.399	45	756879	20.720	ng/ul	98
16) Acetophenone	8.693	105	921752	21.380	ng/ul	99
17) N-Nitroso-di-n-propyla...	8.681	70	466656	19.676	ng/ul	99
18) 4-Methylphenol	8.640	108	608846	20.776	ng/ul	99
19) Hexachloroethane	8.946	117	235976	20.390	ng/ul	96
22) Nitrobenzene	9.069	77	683376	20.629	ng/ul	98
23) Isophorone	9.599	82	1360226	20.139	ng/ul	99
25) 2-Nitrophenol	9.781	139	326116	21.562	ng/ul	99
26) 2,4-Dimethylphenol	9.846	107	605114	20.457	ng/ul	100
27) Bis(2-Chloroethoxy)met...	10.081	93	851134	19.960	ng/ul	99
29) 2,4-Dichlorophenol	10.310	162	576331	20.715	ng/ul	99
30) Naphthalene	10.716	128	1876880	20.211	ng/ul	100
32) 4-Chloroaniline	10.828	127	780852	21.704	ng/ul	100
33) Hexachlorobutadiene	10.999	225	388366	20.843	ng/ul	98
34) Caprolactam	11.604	113	163320	19.144	ng/ul	97
35) 4-Chloro-3-methylphenol	11.951	107	623347	20.000	ng/ul	99
36) 2-Methylnaphthalene	12.322	142	1311917	19.941	ng/ul	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120123\
 Data File : BP018480.D
 Acq On : 01 Dec 2023 17:21
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020701


Quant Time: Dec 01 22:36:05 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Dec 01 21:56:43 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.540	142	1318672	19.847	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.687	216	741728	21.000	ng/ul	99
40) Hexachlorocyclopentadiene	12.669	237	462111	22.073	ng/ul	97
41) 2,4,6-Trichlorophenol	12.928	196	482413	21.231	ng/ul	98
42) 2,4,5-Trichlorophenol	12.999	196	522810	21.224	ng/ul	98
43) 1,1'-Biphenyl	13.334	154	1728403	20.528	ng/ul	99
44) 2-Chloronaphthalene	13.375	162	1380972	20.767	ng/ul	98
45) 2-Nitroaniline	13.575	65	369760	20.511	ng/ul	96
47) Dimethylphthalate	13.957	163	1702023	19.900	ng/ul	100
48) 2,6-Dinitrotoluene	14.075	165	334760	20.717	ng/ul	96
50) Acenaphthylene	14.216	152	2213333	20.576	ng/ul	100
51) 3-Nitroaniline	14.404	138	343450	21.292	ng/ul	98
52) Acenaphthene	14.557	153	1452239	20.419	ng/ul	99
53) 2,4-Dinitrophenol	14.604	184	154269	18.014	ng/ul	98
55) 4-Nitrophenol	14.704	109	208396	19.340	ng/ul	97
56) Dibenzofuran	14.893	168	1993027	20.162	ng/ul	99
57) 2,4-Dinitrotoluene	14.857	165	458838	20.186	ng/ul	98
58) 2,3,4,6-Tetrachlorophenol	15.116	232	418554	20.257	ng/ul	99
59) Diethylphthalate	15.322	149	1636917	19.558	ng/ul	100
61) Fluorene	15.540	166	1568593	20.054	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.540	204	830020	20.233	ng/ul	97
63) 4-Nitroaniline	15.563	138	315095	20.505	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.616	198	251170	20.353	ng/ul	96
67) N-Nitrosodiphenylamine	15.751	169	1338754	20.987	ng/ul	100
68) 4-Bromophenyl-phenylether	16.434	248	513360	21.124	ng/ul	96
69) Hexachlorobenzene	16.539	284	563506	20.768	ng/ul	98
70) Atrazine	16.710	200	413696	21.814	ng/ul	98
71) Pentachlorophenol	16.887	266	326546	20.235	ng/ul	100
72) Phenanthrene	17.281	178	2334524	20.080	ng/ul	99
74) Anthracene	17.375	178	2364546	20.290	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.293	216	736479	22.126	ng/ul	99
76) Pentachlorobenzene	14.810	250	711764	21.769	ng/ul	100
77) Carbazole	17.645	167	1956335	21.063	ng/ul	99
78) Di-n-butylphthalate	18.204	149	2406931	19.462	ng/ul	100
80) Fluoranthene	19.245	202	2360268	20.635	ng/ul	99
82) Pyrene	19.598	202	2369004	20.530	ng/ul	99
83) Butylbenzylphthalate	20.480	149	846326	20.035	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.245	252	669989	22.179	ng/ul	99
85) Benzo(a)anthracene	21.310	228	2029907	20.118	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.245	149	1176650	20.266	ng/ul	99
87) Chrysene	21.363	228	1895771	20.402	ng/ul	100
89) Di-n-octyl phthalate	22.163	149	1896952	18.163	ng/ul	100
90) Benzo(b)fluoranthene	22.974	252	1975656	18.732	ng/ul	99
91) Benzo(k)fluoranthene	23.022	252	1990518	19.506	ng/ul	100
93) Benzo(a)pyrene	23.592	252	1932941	20.058	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.174	276	2581980	22.424	ng/ul	98
95) Dibenzo(a,h)anthracene	26.198	278	2145076	22.395	ng/ul	99
96) Benzo(g,h,i)perylene	26.933	276	2115655	22.832	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Dec 01 22:36:05 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120123.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Dec 01 21:56:43 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 7.814 ng/uL
RT: 3.311 min Scan# 38
Delta R.T. 0.000 min
Lab File: BP018480.D
Acq: 01 Dec 2023 17:21




Instrument :
BNA_P
ClientSampleId :
SSTD020701


Tgt Ion:	88	Resp:	89566
Ion	Ratio	Lower	Upper
88	100		
43	24.6	21.8	32.6
58	71.4	56.5	84.7



#5
Pyridine
Concen: 21.158 ng/uL
RT: 3.717 min Scan# 107
Delta R.T. 0.006 min
Lab File: BP018480.D
Acq: 01 Dec 2023 17:21

Tgt Ion:	79	Resp:	598234
Ion	Ratio	Lower	Upper
79	100		
52	67.3	57.4	86.0
51	32.0	26.2	39.4









#6
Benzaldehyde
Concen: 23.979 ng/ul
RT: 7.005 min Scan# 666
Delta R.T. 0.000 min
Lab File: BP018480.D
Acq: 01 Dec 2023 17:21




Instrument :
BNA_P
ClientSampleId :
SSTD020701


Tgt Ion: 77 Resp: 442556
Ion Ratio Lower Upper
77 100
105 99.4 77.9 116.9
106 94.2 74.6 111.8



#8
Phenol
Concen: 20.760 ng/ul
RT: 7.058 min Scan# 675
Delta R.T. 0.000 min
Lab File: BP018480.D
Acq: 01 Dec 2023 17:21

Tgt Ion: 94 Resp: 727730
Ion Ratio Lower Upper
94 100
65 26.9 21.0 31.6
66 35.5 28.6 43.0









#10
Bis(2-Chloroethyl)ether
Concen: 21.157 ng/ul
RT: 7.293 min Scan# 716
Delta R.T. -0.006 min
Lab File: BP018480.D
Acq: 01 Dec 2023 17:21




Instrument :
BNA_P
ClientSampleId :
SSTD020701


Tgt Ion: 93 Resp: 609083
Ion Ratio Lower Upper
93 100
63 74.5 58.3 87.5
95 32.7 25.9 38.9



#12
2-Chlorophenol
Concen: 20.157 ng/ul
RT: 7.428 min Scan# 738
Delta R.T. 0.000 min
Lab File: BP018480.D
Acq: 01 Dec 2023 17:21

Tgt Ion:128 Resp: 569019
Ion Ratio Lower Upper
128 100
64 46.1 37.0 55.4
130 32.6 26.8 40.2





#13

2-Methylphenol

Concen: 20.681 ng/ul

RT: 8.310 min Scan# 888

Delta R.T. 0.000 min

Lab File: BP018480.D

Acq: 01 Dec 2023 17:21

Instrument :

BNA_P

ClientSampleId :





SSTD020701

Tgt Ion:108 Resp: 555034

Ion Ratio Lower Upper

108 100

107 90.1 70.6 105.8



#14

2,2'-oxybis(1-Chloropropane)

Concen: 20.720 ng/ul

RT: 8.399 min Scan# 903

Delta R.T. -0.006 min

Lab File: BP018480.D

Acq: 01 Dec 2023 17:21




Tgt Ion: 45 Resp: 756879


Ion Ratio Lower Upper

45 100

77 16.4 12.7 19.1

79 12.2 10.5 15.7









#16
Acetophenone
Concen: 21.380 ng/ul
RT: 8.693 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP018480.D
Acq: 01 Dec 2023 17:21




Instrument :
BNA_P
ClientSampleId :
SSTD020701


Tgt Ion:105 Resp: 921752
Ion Ratio Lower Upper
105 100
77 75.4 59.4 89.0
51 23.5 19.4 29.0



#17
N-Nitroso-di-n-propylamine
Concen: 19.676 ng/ul
RT: 8.681 min Scan# 951
Delta R.T. 0.000 min
Lab File: BP018480.D
Acq: 01 Dec 2023 17:21

Tgt Ion: 70 Resp: 466656
Ion Ratio Lower Upper
70 100
42 50.4 41.0 61.4
101 9.7 7.8 11.8
130 21.4 16.9 25.3









#18
4-Methylphenol
Concen: 20.776 ng/ul
RT: 8.640 min Scan# 944
Delta R.T. 0.000 min
Lab File: BP018480.D
Acq: 01 Dec 2023 17:21




Instrument :
BNA_P
ClientSampleId :
SSTD020701


Tgt Ion:108 Resp: 608846
Ion Ratio Lower Upper
108 100
107 111.8 90.6 136.0



#19
Hexachloroethane
Concen: 20.390 ng/ul
RT: 8.946 min Scan# 996
Delta R.T. 0.000 min
Lab File: BP018480.D
Acq: 01 Dec 2023 17:21

Tgt Ion:117 Resp: 235976
Ion Ratio Lower Upper
117 100
201 108.1 83.7 125.5
199 69.0 52.3 78.5









#22
Nitrobenzene
Concen: 20.629 ng/ul
RT: 9.069 min Scan# 1017
Delta R.T. 0.000 min
Lab File: BP018480.D
Acq: 01 Dec 2023 17:21




Instrument :
BNA_P
ClientSampleId :
SSTD020701


Tgt Ion: 77 Resp: 683376
Ion Ratio Lower Upper
77 100
123 46.6 35.8 53.8
65 13.1 10.5 15.7



#23
Isophorone
Concen: 20.139 ng/ul
RT: 9.599 min Scan# 1107
Delta R.T. 0.000 min
Lab File: BP018480.D
Acq: 01 Dec 2023 17:21

Tgt Ion: 82 Resp: 1360226
Ion Ratio Lower Upper
82 100
95 7.0 5.5 8.3
138 17.4 13.5 20.3









#25
2-Nitrophenol
Concen: 21.562 ng/ul
RT: 9.781 min Scan# 1138
Delta R.T. 0.000 min
Lab File: BP018480.D
Acq: 01 Dec 2023 17:21




Instrument :
BNA_P
ClientSampleId :
SSTD020701


Tgt Ion:139 Resp: 326116
Ion Ratio Lower Upper
139 100
65 49.3 39.1 58.7
109 30.2 23.8 35.6



#26
2,4-Dimethylphenol
Concen: 20.457 ng/ul
RT: 9.846 min Scan# 1149
Delta R.T. 0.000 min
Lab File: BP018480.D
Acq: 01 Dec 2023 17:21

Tgt Ion:107 Resp: 605114
Ion Ratio Lower Upper
107 100
121 50.8 40.0 60.0
122 86.5 69.1 103.7









#27
 Bis(2-Chloroethoxy)methane
 Concen: 19.960 ng/ul
 RT: 10.081 min Scan# 1190
 Delta R.T. -0.006 min
 Lab File: BP018480.D
 Acq: 01 Dec 2023 17:21




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020701


Tgt Ion: 93 Resp: 851134
 Ion Ratio Lower Upper
 93 100
 95 32.1 26.1 39.1
 123 10.6 8.9 13.3



#29
 2,4-Dichlorophenol
 Concen: 20.715 ng/ul
 RT: 10.310 min Scan# 1228
 Delta R.T. 0.000 min
 Lab File: BP018480.D
 Acq: 01 Dec 2023 17:21

Tgt Ion: 162 Resp: 576331
 Ion Ratio Lower Upper
 162 100
 164 65.0 51.7 77.5
 98 35.9 29.2 43.8









#30
Naphthalene
Concen: 20.211 ng/ul
RT: 10.716 min Scan# 1297
Delta R.T. 0.000 min
Lab File: BP018480.D
Acq: 01 Dec 2023 17:21




Instrument :
BNA_P
ClientSampleId :
SSTD020701


Tgt Ion:128 Resp: 1876880
Ion Ratio Lower Upper
128 100
129 10.9 8.9 13.3
127 13.3 10.5 15.7



#32
4-Chloroaniline
Concen: 21.704 ng/ul
RT: 10.828 min Scan# 1316
Delta R.T. 0.000 min
Lab File: BP018480.D
Acq: 01 Dec 2023 17:21

Tgt Ion:127 Resp: 780852
Ion Ratio Lower Upper
127 100
129 31.8 25.6 38.4





#33

Hexachlorobutadiene

Concen: 20.843 ng/ul

RT: 10.999 min Scan# 11

Delta R.T. -0.006 min

Lab File: BP018480.D

Acq: 01 Dec 2023 17:21

Instrument :

BNA_P

ClientSampleId :

SSTD020701





Tgt Ion:225 Resp: 388366

Ion Ratio Lower Upper

225 100

223 64.1 49.5 74.3

227 64.8 50.9 76.3



#34

Caprolactam

Concen: 19.144 ng/ul

RT: 11.604 min Scan# 1448

Delta R.T. 0.006 min

Lab File: BP018480.D

Acq: 01 Dec 2023 17:21




Tgt Ion:113 Resp: 163320


Ion Ratio Lower Upper

113 100

55 160.0 133.7 200.5

56 121.5 97.4 146.2









#35
4-Chloro-3-methylphenol
Concen: 20.000 ng/ul
RT: 11.951 min Scan# 1507
Delta R.T. 0.000 min
Lab File: BP018480.D
Acq: 01 Dec 2023 17:21




Instrument :
BNA_P
ClientSampleId :
SSTD020701


Tgt Ion:107 Resp: 623347
Ion Ratio Lower Upper
107 100
144 25.9 21.0 31.6
142 79.0 63.6 95.4



#36
2-Methylnaphthalene
Concen: 19.941 ng/ul
RT: 12.322 min Scan# 1570
Delta R.T. 0.000 min
Lab File: BP018480.D
Acq: 01 Dec 2023 17:21

Tgt Ion:142 Resp: 1311917
Ion Ratio Lower Upper
142 100
141 88.9 71.4 107.0









#37
1-Methylnaphthalene
Concen: 19.847 ng/ul
RT: 12.540 min Scan# 1608
Delta R.T. -0.006 min
Lab File: BP018480.D
Acq: 01 Dec 2023 17:21




Instrument :
BNA_P
ClientSampleId :
SSTD020701


Tgt Ion:142 Resp: 1318672
Ion Ratio Lower Upper
142 100
141 91.6 73.5 110.3
116 3.9 2.8 4.2



#39
1,2,4,5-Tetrachlorobenzene
Concen: 21.000 ng/ul
RT: 12.687 min Scan# 1632
Delta R.T. -0.006 min
Lab File: BP018480.D
Acq: 01 Dec 2023 17:21

Tgt Ion:216 Resp: 741728
Ion Ratio Lower Upper
216 100
214 78.7 63.9 95.9
179 20.4 16.3 24.5
108 15.9 12.6 19.0





#40

Hexachlorocyclopentadiene

Concen: 22.073 ng/ul

RT: 12.669 min Scan# 1629

Delta R.T. 0.000 min

Lab File: BP018480.D

Acq: 01 Dec 2023 17:21

Instrument :

BNA_P

ClientSampleId :

SSTD020701





Tgt Ion:237 Resp: 462111

Ion Ratio Lower Upper

237 100

235 64.3 49.4 74.0

272 12.8 9.8 14.8



#41

2,4,6-Trichlorophenol

Concen: 21.231 ng/ul

RT: 12.928 min Scan# 1673

Delta R.T. 0.000 min

Lab File: BP018480.D

Acq: 01 Dec 2023 17:21




Tgt Ion:196 Resp: 482413


Ion Ratio Lower Upper

196 100

198 94.4 76.9 115.3

200 30.8 25.0 37.6





#42

2,4,5-Trichlorophenol

Concen: 21.224 ng/ul

RT: 12.999 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP018480.D

Acq: 01 Dec 2023 17:21

Instrument :

BNA_P

ClientSampleId :

SSTD020701





Tgt Ion:196 Resp: 522810

Ion Ratio Lower Upper

196 100

198 95.4 77.9 116.9

200 31.2 25.2 37.8



#43

1,1'-Biphenyl

Concen: 20.528 ng/ul

RT: 13.334 min Scan# 1742

Delta R.T. 0.000 min

Lab File: BP018480.D

Acq: 01 Dec 2023 17:21




Tgt Ion:154 Resp: 1728403


Ion Ratio Lower Upper

154 100

153 40.7 32.8 49.2

76 12.9 10.7 16.1









#44
2-Chloronaphthalene
Concen: 20.767 ng/ul
RT: 13.375 min Scan# 1749
Delta R.T. 0.000 min
Lab File: BP018480.D
Acq: 01 Dec 2023 17:21




Instrument :
BNA_P
ClientSampleId :
SSTD020701


Tgt Ion:162 Resp: 1380972
Ion Ratio Lower Upper
162 100
164 33.0 25.7 38.5
127 37.0 30.3 45.5



#45
2-Nitroaniline
Concen: 20.511 ng/ul
RT: 13.575 min Scan# 1783
Delta R.T. -0.006 min
Lab File: BP018480.D
Acq: 01 Dec 2023 17:21

Tgt Ion: 65 Resp: 369760
Ion Ratio Lower Upper
65 100
92 73.1 56.9 85.3
138 104.9 88.2 132.2









#47
Dimethylphthalate
Concen: 19.900 ng/ul
RT: 13.957 min Scan# 1848
Delta R.T. -0.006 min
Lab File: BP018480.D
Acq: 01 Dec 2023 17:21




Instrument :
BNA_P
ClientSampleId :
SSTD020701


Tgt Ion:163 Resp: 1702023
Ion Ratio Lower Upper
163 100
194 4.0 3.2 4.8
164 10.1 8.0 12.0



#48
2,6-Dinitrotoluene
Concen: 20.717 ng/ul
RT: 14.075 min Scan# 1868
Delta R.T. 0.000 min
Lab File: BP018480.D
Acq: 01 Dec 2023 17:21

Tgt Ion:165 Resp: 334760
Ion Ratio Lower Upper
165 100
89 54.5 46.6 70.0
121 20.5 17.3 25.9









#50
Acenaphthylene
Concen: 20.576 ng/ul
RT: 14.216 min Scan# 1892
Delta R.T. 0.000 min
Lab File: BP018480.D
Acq: 01 Dec 2023 17:21




Instrument :
BNA_P
ClientSampleId :
SSTD020701


Tgt Ion:152 Resp: 2213333
Ion Ratio Lower Upper
152 100
151 19.8 15.9 23.9
153 13.1 10.6 16.0



#51
3-Nitroaniline
Concen: 21.292 ng/ul
RT: 14.404 min Scan# 1924
Delta R.T. 0.000 min
Lab File: BP018480.D
Acq: 01 Dec 2023 17:21

Tgt Ion:138 Resp: 343450
Ion Ratio Lower Upper
138 100
108 11.4 9.5 14.3
92 113.8 93.1 139.7









#52
Acenaphthene
Concen: 20.419 ng/ul
RT: 14.557 min Scan# 1950
Delta R.T. 0.000 min
Lab File: BP018480.D
Acq: 01 Dec 2023 17:21




Instrument :
BNA_P
ClientSampleId :
SSTD020701


Tgt Ion:153 Resp: 1452239
Ion Ratio Lower Upper
153 100
152 47.4 38.1 57.1
154 87.9 71.7 107.5



#53
2,4-Dinitrophenol
Concen: 18.014 ng/ul
RT: 14.604 min Scan# 1958
Delta R.T. 0.000 min
Lab File: BP018480.D
Acq: 01 Dec 2023 17:21

Tgt Ion:184 Resp: 154269
Ion Ratio Lower Upper
184 100
63 61.4 48.2 72.4
154 62.7 49.1 73.7









#55
4-Nitrophenol
Concen: 19.340 ng/ul
RT: 14.704 min Scan# 1975
Delta R.T. 0.000 min
Lab File: BP018480.D
Acq: 01 Dec 2023 17:21




Instrument :
BNA_P
ClientSampleId :
SSTD020701


Tgt Ion:109 Resp: 208396
Ion Ratio Lower Upper
109 100
139 133.3 109.0 163.4
65 127.9 106.2 159.4



#56
Dibenzofuran
Concen: 20.162 ng/ul
RT: 14.893 min Scan# 2007
Delta R.T. 0.000 min
Lab File: BP018480.D
Acq: 01 Dec 2023 17:21

Tgt Ion:168 Resp: 1993027
Ion Ratio Lower Upper
168 100
139 36.1 29.1 43.7









#57
2,4-Dinitrotoluene
Concen: 20.186 ng/ul
RT: 14.857 min Scan# 2001
Delta R.T. 0.000 min
Lab File: BP018480.D
Acq: 01 Dec 2023 17:21




Instrument :
BNA_P
ClientSampleId :
SSTD020701


Tgt Ion:165 Resp: 458838
Ion Ratio Lower Upper
165 100
63 39.6 32.7 49.1
182 2.8 2.1 3.1



#58
2,3,4,6-Tetrachlorophenol
Concen: 20.257 ng/ul
RT: 15.116 min Scan# 2045
Delta R.T. 0.000 min
Lab File: BP018480.D
Acq: 01 Dec 2023 17:21

Tgt Ion:232 Resp: 418554
Ion Ratio Lower Upper
232 100
131 41.9 33.8 50.8
130 2.2 1.8 2.8
166 29.7 24.7 37.1









#59
Diethylphthalate
Concen: 19.558 ng/ul
RT: 15.322 min Scan# 2081
Delta R.T. -0.006 min
Lab File: BP018480.D
Acq: 01 Dec 2023 17:21




Instrument :
BNA_P
ClientSampleId :
SSTD020701


Tgt Ion:149 Resp: 1636917
Ion Ratio Lower Upper
149 100
177 22.5 18.2 27.4
150 12.3 9.9 14.9



#61
Fluorene
Concen: 20.054 ng/ul
RT: 15.540 min Scan# 2117
Delta R.T. -0.006 min
Lab File: BP018480.D
Acq: 01 Dec 2023 17:21

Tgt Ion:166 Resp: 1568593
Ion Ratio Lower Upper
166 100
165 96.5 78.0 117.0
167 13.6 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 20.233 ng/ul
RT: 15.540 min Scan# 2117
Delta R.T. 0.000 min
Lab File: BP018480.D
Acq: 01 Dec 2023 17:21




Instrument :
BNA_P
ClientSampleId :
SSTD020701


Tgt Ion:204 Resp: 830020
Ion Ratio Lower Upper
204 100
206 33.1 27.1 40.7
141 54.7 45.8 68.8



#63
4-Nitroaniline
Concen: 20.505 ng/ul
RT: 15.563 min Scan# 2121
Delta R.T. 0.000 min
Lab File: BP018480.D
Acq: 01 Dec 2023 17:21

Tgt Ion:138 Resp: 315095
Ion Ratio Lower Upper
138 100
92 52.5 42.3 63.5
108 80.2 62.0 93.0









#66
4,6-Dinitro-2-methylphenol
Concen: 20.353 ng/ul
RT: 15.616 min Scan# 2130
Delta R.T. 0.000 min
Lab File: BP018480.D
Acq: 01 Dec 2023 17:21




Instrument :
BNA_P
ClientSampleId :
SSTD020701


Tgt Ion:198 Resp: 251170
Ion Ratio Lower Upper
198 100
182 4.6 3.4 5.2
77 23.5 20.6 30.8



#67
N-Nitrosodiphenylamine
Concen: 20.987 ng/ul
RT: 15.751 min Scan# 2153
Delta R.T. -0.006 min
Lab File: BP018480.D
Acq: 01 Dec 2023 17:21

Tgt Ion:169 Resp: 1338754
Ion Ratio Lower Upper
169 100
168 66.6 53.1 79.7
167 36.5 29.1 43.7









#68
4-Bromophenyl-phenylether
Concen: 21.124 ng/ul
RT: 16.434 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP018480.D
Acq: 01 Dec 2023 17:21




Instrument :
BNA_P
ClientSampleId :
SSTD020701


Tgt Ion:248 Resp: 513360
Ion Ratio Lower Upper
248 100
250 96.8 79.3 118.9
141 66.8 57.2 85.8



#69
Hexachlorobenzene
Concen: 20.768 ng/ul
RT: 16.539 min Scan# 2287
Delta R.T. -0.006 min
Lab File: BP018480.D
Acq: 01 Dec 2023 17:21

Tgt Ion:284 Resp: 563506
Ion Ratio Lower Upper
284 100
142 32.9 25.8 38.6
249 32.3 24.9 37.3





#70

Atrazine

Concen: 21.814 ng/ul

RT: 16.710 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP018480.D

Acq: 01 Dec 2023 17:21

Instrument :

BNA_P

ClientSampleId :

SSTD020701





Tgt Ion:200 Resp: 413696

Ion Ratio Lower Upper

200 100

173 25.3 21.0 31.4

215 48.7 38.2 57.2



#71

Pentachlorophenol

Concen: 20.235 ng/ul

RT: 16.887 min Scan# 2346

Delta R.T. 0.000 min

Lab File: BP018480.D

Acq: 01 Dec 2023 17:21




Tgt Ion:266 Resp: 326546


Ion Ratio Lower Upper

266 100

264 62.2 50.2 75.2

268 63.0 50.5 75.7





#72

Phenanthrene

Concen: 20.080 ng/ul

RT: 17.281 min Scan# 2414

Delta R.T. -0.006 min

Lab File: BP018480.D

Acq: 01 Dec 2023 17:21

Instrument :

BNA_P

ClientSampleId :

SSTD020701





Tgt Ion:178 Resp: 2334524

Ion Ratio Lower Upper

178 100

179 15.2 12.3 18.5

176 19.0 15.4 23.2



#74

Anthracene

Concen: 20.290 ng/ul

RT: 17.375 min Scan# 2429

Delta R.T. 0.000 min

Lab File: BP018480.D

Acq: 01 Dec 2023 17:21




Tgt Ion:178 Resp: 2364546


Ion Ratio Lower Upper

178 100

179 15.4 12.5 18.7

176 18.7 15.0 22.6





#75

1,2,3,4-Tetrachlorobenzene

Concen: 22.126 ng/uL

RT: 13.293 min Scan# 1735

Delta R.T. -0.006 min

Lab File: BP018480.D

Acq: 01 Dec 2023 17:21

Instrument :

BNA_P

ClientSampleId :

SSTD020701





Tgt Ion:216 Resp: 736479

Ion Ratio Lower Upper

216 100

214 77.8 61.4 92.2

218 48.4 39.2 58.8



#76

Pentachlorobenzene

Concen: 21.769 ng/uL

RT: 14.810 min Scan# 1993

Delta R.T. 0.000 min

Lab File: BP018480.D

Acq: 01 Dec 2023 17:21




Tgt Ion:250 Resp: 711764


Ion Ratio Lower Upper

250 100

248 62.6 50.6 75.8

252 64.2 51.3 76.9









#77
Carbazole
Concen: 21.063 ng/ul
RT: 17.645 min Scan# 2476
Delta R.T. -0.006 min
Lab File: BP018480.D
Acq: 01 Dec 2023 17:21




Instrument :
BNA_P
ClientSampleId :
SSTD020701


Tgt Ion:167 Resp: 1956335
Ion Ratio Lower Upper
167 100
166 21.3 17.1 25.7
139 12.3 10.2 15.4



#78
Di-n-butylphthalate
Concen: 19.462 ng/ul
RT: 18.204 min Scan# 2570
Delta R.T. -0.006 min
Lab File: BP018480.D
Acq: 01 Dec 2023 17:21

Tgt Ion:149 Resp: 2406931
Ion Ratio Lower Upper
149 100
150 9.3 7.4 11.2
104 5.0 3.9 5.9





#80

Fluoranthene

Concen: 20.635 ng/uI

RT: 19.245 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP018480.D

Acq: 01 Dec 2023 17:21

Instrument :

BNA_P

ClientSampleId :

SSTD020701





Tgt Ion:202 Resp: 2360268

Ion Ratio Lower Upper

202 100

101 11.2 9.1 13.7

100 8.6 7.1 10.7



#82

Pyrene

Concen: 20.530 ng/uI

RT: 19.598 min Scan# 2807

Delta R.T. -0.006 min

Lab File: BP018480.D

Acq: 01 Dec 2023 17:21




Tgt Ion:202 Resp: 2369004


Ion Ratio Lower Upper

202 100

101 13.4 11.0 16.4

100 11.0 9.1 13.7





#83

Butylbenzylphthalate

Concen: 20.035 ng/ul

RT: 20.480 min Scan# 2958

Delta R.T. -0.006 min

Lab File: BP018480.D

Acq: 01 Dec 2023 17:21

Instrument :

BNA_P

ClientSampleId :

SSTD020701





Tgt Ion:149 Resp: 846326

Ion Ratio Lower Upper

149 100

91 72.5 56.6 85.0

206 21.5 17.4 26.0



#84

3,3'-Dichlorobenzidine

Concen: 22.179 ng/ul

RT: 21.245 min Scan# 3087

Delta R.T. -0.006 min

Lab File: BP018480.D

Acq: 01 Dec 2023 17:21




Tgt Ion:252 Resp: 669989


Ion Ratio Lower Upper

252 100

254 64.7 51.0 76.4

126 10.4 7.8 11.6





#85

Benzo(a)anthracene

Concen: 20.118 ng/ul

RT: 21.310 min Scan# 3099

Delta R.T. -0.006 min

Lab File: BP018480.D

Acq: 01 Dec 2023 17:21

Instrument :

BNA_P

ClientSampleId :

SSTD020701





Tgt Ion:228 Resp: 2029907

Ion Ratio Lower Upper

228 100

229 19.6 15.7 23.5

226 26.9 21.9 32.9



#86

Bis(2-ethylhexyl)phthalate

Concen: 20.266 ng/ul

RT: 21.245 min Scan# 3087

Delta R.T. 0.000 min

Lab File: BP018480.D

Acq: 01 Dec 2023 17:21




Tgt Ion:149 Resp: 1176650


Ion Ratio Lower Upper

149 100

167 28.3 22.2 33.2

279 4.5 3.3 4.9









#87
Chrysene
Concen: 20.402 ng/ul
RT: 21.363 min Scan# 3107
Delta R.T. -0.006 min
Lab File: BP018480.D
Acq: 01 Dec 2023 17:21




Instrument :
BNA_P
ClientSampleId :
SSTD020701


Tgt Ion:228 Resp: 1895771
Ion Ratio Lower Upper
228 100
226 29.7 23.9 35.9
229 19.4 15.6 23.4



#89
Di-n-octyl phthalate
Concen: 18.163 ng/ul
RT: 22.163 min Scan# 3243
Delta R.T. 0.000 min
Lab File: BP018480.D
Acq: 01 Dec 2023 17:21

Tgt Ion:149 Resp: 1896952





#90

Benzo(b)fluoranthene

Concen: 18.732 ng/ul

RT: 22.974 min Scan# 3381

Delta R.T. -0.006 min

Lab File: BP018480.D

Acq: 01 Dec 2023 17:21

Instrument :

BNA_P

ClientSampleId :

SSTD020701





Tgt Ion:252 Resp: 1975656

Ion Ratio Lower Upper

252 100

253 21.9 17.3 25.9

125 10.1 7.8 11.6



#91

Benzo(k)fluoranthene

Concen: 19.506 ng/ul

RT: 23.022 min Scan# 3389

Delta R.T. -0.006 min

Lab File: BP018480.D

Acq: 01 Dec 2023 17:21




Tgt Ion:252 Resp: 1990518


Ion Ratio Lower Upper

252 100

253 21.9 17.7 26.5

125 10.2 8.0 12.0









#93
 Benzo(a)pyrene
 Concen: 20.058 ng/ul
 RT: 23.592 min Scan# 3486
 Delta R.T. -0.006 min
 Lab File: BP018480.D
 Acq: 01 Dec 2023 17:21

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020701

Tgt Ion:252 Resp: 1932941




Ion	Ratio	Lower	Upper
252	100		
253	22.0	17.8	26.8
125	10.8	9.2	13.8




#94
 Indeno(1,2,3-cd)pyrene
 Concen: 22.424 ng/ul
 RT: 26.174 min Scan# 3925
 Delta R.T. -0.006 min
 Lab File: BP018480.D
 Acq: 01 Dec 2023 17:21

Tgt Ion:276 Resp: 2581980

Ion	Ratio	Lower	Upper
276	100		
138	22.8	19.4	29.2
277	25.2	19.8	29.6





#95

Dibenzo(a,h)anthracene

Concen: 22.395 ng/ul

RT: 26.198 min Scan# 3929

Delta R.T. 0.000 min

Lab File: BP018480.D

Acq: 01 Dec 2023 17:21

Instrument :

BNA_P

ClientSampleId :

SSTD020701





Tgt Ion:278 Resp: 2145076

Ion Ratio Lower Upper

278 100

139 15.9 13.0 19.6

279 23.6 18.7 28.1



#96

Benzo(g,h,i)perylene

Concen: 22.832 ng/ul

RT: 26.933 min Scan# 4054

Delta R.T. -0.006 min

Lab File: BP018480.D

Acq: 01 Dec 2023 17:21

Tgt Ion:276 Resp: 2115655

Ion Ratio Lower Upper

276 100

138 21.6 18.1 27.1

277 23.6 19.4 29.0

