

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120123\
 Data File : BP018534.D
 Acq On : 03 Dec 2023 01:40
 Operator : MA/JU
 Sample : 05411-15
 Misc :
 ALS Vial : 64 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 C0QK0

Quant Time: Dec 03 02:14:55 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Dec 01 21:56:43 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.858	152	373915	20.000	ng/u1	-0.01
20) Naphthalene-d8	10.657	136	1510553	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.487	164	899018	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.234	188	1888068	20.000	ng/u1	0.00
79) Chrysene-d12	21.316	240	1838877	20.000	ng/u1	-0.01
88) Perylene-d12	23.692	264	2179743	20.000	ng/u1	-0.01

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.270	96	36408	3.318	ng/uL	0.00
4) Pyridine-d5	3.693	84	178373	6.069	ng/u1	0.00
7) Phenol-d5	7.022	99	176155	4.694	ng/u1	-0.01
9) Bis-(2-Chloroethyl)eth...	7.193	67	573872	25.804	ng/u1	-0.01
11) 2-Chlorophenol-d4	7.387	132	527595	18.019	ng/u1	-0.01
15) 4-Methylphenol-d8	8.569	113	322820	10.631	ng/u1	-0.01
21) Nitrobenzene-d5	9.022	128	331537	24.660	ng/u1	0.00
24) 2-Nitrophenol-d4	9.740	143	305404	21.910	ng/u1	-0.01
28) 2,4-Dichlorophenol-d3	10.275	165	580459	20.366	ng/u1	-0.01
31) 4-Chloroaniline-d4	10.793	131	722688	19.402	ng/u1	-0.01
46) Dimethylphthalate-d6	13.904	166	2041527	25.443	ng/u1	0.00
49) Acenaphthylene-d8	14.181	160	2272211	25.717	ng/u1	-0.01
54) 4-Nitrophenol-d4	14.681	143	46561	3.535	ng/u1	-0.01
60) Fluorene-d10	15.481	176	1822577	27.061	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.592	200	221346	19.552	ng/u1	-0.01
73) Anthracene-d10	17.334	188	2960284	29.824	ng/u1	0.00
81) Pyrene-d10	19.569	212	3527770	24.645	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.539	264	3776534	29.714	ng/u1	-0.01

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120123\
Data File : BP018534.D
Acq On : 03 Dec 2023 01:40
Operator : MA/JU
Sample : 05411-15
Misc :
ALS Vial : 64 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
C0QK0

Quant Time: Dec 03 02:14:55 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120123.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Dec 01 21:56:43 2023
Response via : Initial Calibration

