

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120123\
 Data File : BP018483.D
 Acq On : 01 Dec 2023 19:03
 Operator : MA/JU
 Sample : 05332-07
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 YCGA5

Quant Time: Dec 01 22:37:05 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Dec 01 21:56:43 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.863	152	360671	20.000	ng/u1	0.00
20) Naphthalene-d8	10.663	136	1558637	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.492	164	1061933	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.239	188	2308626	20.000	ng/u1	0.00
79) Chrysene-d12	21.321	240	1529439	20.000	ng/u1	0.00
88) Perylene-d12	23.698	264	1414627	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.275	96	38210	3.610	ng/uL	0.00
4) Pyridine-d5	3.699	84	183550	6.474	ng/u1	0.00
7) Phenol-d5	7.028	99	187791	5.188	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.199	67	517141	24.107	ng/u1	0.00
11) 2-Chlorophenol-d4	7.393	132	491643	17.408	ng/u1	0.00
15) 4-Methylphenol-d8	8.575	113	365207	12.468	ng/u1	0.00
21) Nitrobenzene-d5	9.022	128	324809	23.414	ng/u1	0.00
24) 2-Nitrophenol-d4	9.745	143	327549	22.774	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.281	165	614948	20.910	ng/u1	0.00
31) 4-Chloroaniline-d4	10.798	131	671352	17.468	ng/u1	0.00
46) Dimethylphthalate-d6	13.910	166	2480900	26.176	ng/u1	0.00
49) Acenaphthylene-d8	14.186	160	2527891	24.222	ng/u1	0.00
54) 4-Nitrophenol-d4	14.686	143	71970	4.626	ng/u1	0.00
60) Fluorene-d10	15.486	176	2036244	25.595	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.598	200	327028	23.625	ng/u1	0.00
73) Anthracene-d10	17.339	188	3534696	29.124	ng/u1	0.00
81) Pyrene-d10	19.574	212	3776182	31.718	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.545	264	2479860	30.065	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120123\
Data File : BP018483.D
Acq On : 01 Dec 2023 19:03
Operator : MA/JU
Sample : 05332-07
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
YCGA5

Quant Time: Dec 01 22:37:05 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120123.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Dec 01 21:56:43 2023
Response via : Initial Calibration

