

Data Path : Z:\SVOASRV\HPCHEM1\BNA P\DATA\BP120319\  
 Data File : BP001167.D  
 Acq On : 03 Dec 2019 21:08  
 Operator : JU  
 Sample : PB125109BL  
 Misc :  
 ALS Vial : 12 Sample Multiplier: 1

Instrument :  
 BNA\_P  
 ClientSampleId :  
 PB125109BL

Quant Time: Dec 04 05:01:51 2019  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270-BP112719.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Wed Nov 27 17:38:46 2019  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|-----------------------------|-------|------|----------|-------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4   | 8.31  | 152  | 114890   | 20.00 | ng    | -0.02    |
| 21) Naphthalene-d8          | 10.35 | 136  | 403115   | 20.00 | ng    | -0.02    |
| 39) Acenaphthene-d10        | 13.21 | 164  | 221350   | 20.00 | ng    | -0.02    |
| 64) Phenanthrene-d10        | 15.60 | 188  | 433334   | 20.00 | ng    | -0.02    |
| 76) Chrysene-d12            | 19.25 | 240  | 343331   | 20.00 | ng    | -0.02    |
| 87) Perylene-d12            | 21.02 | 264  | 356820   | 20.00 | ng    | -0.03    |
| System Monitoring Compounds |       |      |          |       |       |          |
| 5) 2-Fluorophenol           | 6.22  | 112  | 638312   | 92.18 | ng    | 0.00     |
| 7) Phenol-d6                | 7.76  | 99   | 871876   | 94.51 | ng    | -0.01    |
| 23) Nitrobenzene-d5         | 9.19  | 82   | 837484   | 90.14 | ng    | -0.02    |
| 42) 2,4,6-Tribromophenol    | 14.50 | 330  | 205773   | 96.99 | ng    | -0.02    |
| 45) 2-Fluorobiphenyl        | 12.13 | 172  | 1371015  | 90.65 | ng    | -0.02    |
| 79) Terphenyl-d14           | 17.90 | 244  | 1617636  | 87.17 | ng    | -0.02    |

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA P\DATA\BP120319\  
Data File : BP001167.D  
Acq On : 03 Dec 2019 21:08  
Operator : JU  
Sample : PB125109BL  
Misc :  
ALS Vial : 12 Sample Multiplier: 1

Instrument :  
BNA\_P  
ClientSampleId :  
PB125109BL

Quant Time: Dec 04 05:01:51 2019  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270-BP112719.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Wed Nov 27 17:38:46 2019  
Response via : Initial Calibration

