

Data Path : Z:\svoasrv\HPCHEM1\BNA P\Data\BP120319\
 Data File : BP001199.D
 Acq On : 04 Dec 2019 14:34
 Operator : JU
 Sample : K6085-01
 Misc :
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 MW-2

Quant Time: Dec 12 13:58:53 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270-BP112719.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 27 17:38:46 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.32	152	93116	20.00	ng	-0.01
21) Naphthalene-d8	10.35	136	325217	20.00	ng	-0.01
39) Acenaphthene-d10	13.21	164	189890	20.00	ng	-0.02
64) Phenanthrene-d10	15.60	188	363694	20.00	ng	-0.02
76) Chrysene-d12	19.25	240	287342	20.00	ng	-0.02
87) Perylene-d12	21.03	264	312353	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	6.22	112	518027	92.30	ng	0.00
7) Phenol-d6	7.76	99	491168	65.69	ng	0.00
23) Nitrobenzene-d5	9.20	82	729323	97.30	ng	0.00
42) 2,4,6-Tribromophenol	14.50	330	299236	164.41	ng	-0.02
45) 2-Fluorobiphenyl	12.13	172	1208575	93.15	ng	-0.02
79) Terphenyl-d14	17.90	244	1505151	96.91	ng	-0.02
Target Compounds						
10) Phenol	7.78	94	30725	3.613	ng	94
20) 3+4-Methylphenols	8.98	107	47145	7.021	ng	94
27) 2,4-Dimethylphenol	9.83	122	25645m	5.930	ng	
31) Naphthalene	10.39	128	1811719	109.076	ng	99
37) 2-Methylnaphthalene	11.51	142	186770	16.479	ng	94
38) 1-Methylnaphthalene	11.67	142	382454	35.722	ng	# 99
50) Dimethylphthalate	12.80	163	237779	16.743	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA P\Data\BP120319\
Data File : BP001199.D
Acq On : 04 Dec 2019 14:34
Operator : JU
Sample : K6085-01
Misc :
ALS Vial : 42 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
MW-2

Quant Time: Dec 12 13:58:53 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270-BP112719.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 27 17:38:46 2019
Response via : Initial Calibration

