


Data Path : Z:\SVOASRV\HPCHEM1\BNA P\DATA\BP120319\  
 Data File : BP001179.D  
 Acq On : 04 Dec 2019 04:35  
 Operator : JU  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_P  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA\_P\METHODS\8270-BP112719.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Wed Nov 27 17:38:46 2019



AutoFind: Scans 2381, 2382, 2383; Background Corrected with Scan 2373

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 44.5      | 29396   | PASS             |
| 68          | 69           | 0.00         | 2            | 1.6       | 487     | PASS             |
| 69          | 198          | 0.00         | 100          | 46.1      | 30442   | PASS             |
| 70          | 69           | 0.00         | 2            | 0.5       | 159     | PASS             |
| 127         | 198          | 10           | 80           | 49.4      | 32624   | PASS             |
| 197         | 198          | 0.00         | 2            | 0.4       | 278     | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 66013   | PASS             |
| 199         | 198          | 5            | 9            | 6.6       | 4364    | PASS             |
| 275         | 198          | 10           | 60           | 24.4      | 16094   | PASS             |
| 365         | 198          | 1            | 100          | 2.8       | 1875    | PASS             |
| 441         | 198          | 0.01         | 100          | 10.8      | 7140    | PASS             |
| 442         | 198          | 50           | 100          | 69.5      | 45909   | PASS             |
| 443         | 442          | 15           | 24           | 20.7      | 9483    | PASS             |