




Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120325\
Data File : BP026223.D
Acq On : 03 Dec 2025 12:13
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Dec 03 13:04:39 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 19 01:25:31 2025
Response via : Initial Calibration



TIC: BP026223.D\data.ms

(70) Pentachlorophenol (C)




16.725min (-0.002) 23877.06 ng m

response 803180

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.70	62.35
264.00	62.70	62.02
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120325\
Data File : BP026223.D
Acq On : 03 Dec 2025 12:13
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Dec 03 13:08:50 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 19 01:25:31 2025
Response via : Initial Calibration



TIC: BP026223.D\data.ms

(77) Benzidine

19.424min (-0.010) 582608.37 ng m

response 3235664

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.10	14.14
183.00	11.30	11.91
0.00	0.00	0.00