

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120325\
 Data File : BP026224.D
 Acq On : 03 Dec 2025 12:53
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 03 13:21:51 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 19 01:25:31 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	109	-0.01
2	1,4-Dioxane	0.505	0.489	3.2	109	0.00
3	Pyridine	1.461	1.382	5.4	104	0.00
4	n-Nitrosodimethylamine	0.546	0.528	3.3	106	0.00
5 S	2-Fluorophenol	1.246	1.193	4.3	104	0.00
6	Aniline	2.094	1.951	6.8	100	0.00
7 S	Phenol-d6	1.586	1.518	4.3	102	-0.01
8	2-Chlorophenol	1.418	1.417	0.1	108	-0.01
9	Benzaldehyde	0.831	1.056	-27.1#	124	0.00
10 C	Phenol	1.709	1.631	4.6	101	0.00
11	bis(2-Chloroethyl)ether	1.338	1.298	3.0	104	0.00
12	1,3-Dichlorobenzene	1.528	1.528	0.0	109	-0.02
13 C	1,4-Dichlorobenzene	1.540	1.539	0.1	109	0.00
14	1,2-Dichlorobenzene	1.479	1.480	-0.1	109	0.00
15	Benzyl Alcohol	1.162	1.124	3.3	103	-0.01
16	2,2'-oxybis(1-Chloropropane	1.879	1.856	1.2	105	-0.02
17	2-Methylphenol	1.163	1.133	2.6	103	-0.02
18	Hexachloroethane	0.560	0.572	-2.1	110	-0.02
19 P	n-Nitroso-di-n-propylamine	0.985	0.955	3.0	101	-0.01
20	3+4-Methylphenols	1.612	1.526	5.3	102	-0.02
21 I	Naphthalene-d8	1.000	1.000	0.0	106	-0.03
22	Acetophenone	0.508	0.496	2.4	104	-0.02
23 S	Nitrobenzene-d5	0.352	0.349	0.9	104	-0.02
24	Nitrobenzene	0.356	0.355	0.3	105	-0.02
25	Isophorone	0.698	0.670	4.0	100	-0.02
26 C	2-Nitrophenol	0.180	0.188	-4.4	106	-0.02
27	2,4-Dimethylphenol	0.325	0.281	13.5	91	-0.02
28	bis(2-Chloroethoxy)methane	0.437	0.426	2.5	103	-0.02
29 C	2,4-Dichlorophenol	0.325	0.325	0.0	104	-0.02
30	1,2,4-Trichlorobenzene	0.331	0.340	-2.7	109	-0.02
31	Naphthalene	1.081	1.061	1.9	105	-0.02
32	Benzoic acid	0.274	0.238	13.1	91	-0.01
33	4-Chloroaniline	0.460	0.429	6.7	96	-0.02
34 C	Hexachlorobutadiene	0.215	0.229	-6.5	113	-0.04
35	Caprolactam	0.116	0.099	14.7	90	-0.01
36 C	4-Chloro-3-methylphenol	0.345	0.332	3.8	99	-0.02
37	2-Methylnaphthalene	0.766	0.747	2.5	102	-0.03
38	1-Methylnaphthalene	0.749	0.735	1.9	102	-0.02
39 I	Acenaphthene-d10	1.000	1.000	0.0	99	-0.05
40	1,2,4,5-Tetrachlorobenzene	0.605	0.646	-6.8	107	-0.04
41 P	Hexachlorocyclopentadiene	0.396	0.314	20.7	78	-0.03
42 S	2,4,6-Tribromophenol	0.259	0.254	1.9	95	-0.04
43 C	2,4,6-Trichlorophenol	0.415	0.430	-3.6	101	-0.03
44	2,4,5-Trichlorophenol	0.463	0.475	-2.6	100	-0.03
45 S	2-Fluorobiphenyl	1.365	1.409	-3.2	104	-0.04
46	1,1'-Biphenyl	1.506	1.525	-1.3	101	-0.03
47	2-Chloronaphthalene	1.184	1.205	-1.8	102	-0.04

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120325\
 Data File : BP026224.D
 Acq On : 03 Dec 2025 12:53
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 03 13:21:51 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 19 01:25:31 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.329	0.317	3.6	93	-0.04
49	Acenaphthylene	1.953	1.927	1.3	98	-0.04
50	Dimethylphthalate	1.494	1.449	3.0	96	-0.04
51	2,6-Dinitrotoluene	0.296	0.305	-3.0	95	-0.04
52 C	Acenaphthene	1.145	1.130	1.3	98	-0.04
53	3-Nitroaniline	0.349	0.313	10.3	85	-0.04
54 P	2,4-Dinitrophenol	0.179	0.163	8.9	87	-0.05
55	Dibenzofuran	1.773	1.738	2.0	97	-0.04
56 P	4-Nitrophenol	0.315	0.254	19.4	79	-0.04
57	2,4-Dinitrotoluene	0.442	0.422	4.5	89	-0.05
58	Fluorene	1.401	1.333	4.9	95	-0.04
59	2,3,4,6-Tetrachlorophenol	0.393	0.384	2.3	95	-0.04
60	Diethylphthalate	1.505	1.424	5.4	92	-0.04
61	4-Chlorophenyl-phenylether	0.697	0.692	0.7	99	-0.04
62	4-Nitroaniline	0.362	0.297	18.0	80	-0.04
63	Azobenzene	1.269	1.192	6.1	91	-0.04
64 I	Phenanthrene-d10	1.000	1.000	0.0	88	-0.05
65	4,6-Dinitro-2-methylphenol	0.127	0.130	-2.4	85	-0.04
66 c	n-Nitrosodiphenylamine	0.619	0.657	-6.1	93	-0.04
67	4-Bromophenyl-phenylether	0.221	0.243	-10.0	94	-0.04
68	Hexachlorobenzene	0.257	0.281	-9.3	94	-0.05
69	Atrazine	0.233	0.223	4.3	83	-0.04
70 C	Pentachlorophenol	0.180	0.179	0.6	86	-0.05
71	Phenanthrene	1.127	1.123	0.4	88	-0.05
72	Anthracene	1.149	1.164	-1.3	88	-0.04
73	Carbazole	1.091	0.998	8.5	80	-0.04
74	Di-n-butylphthalate	1.294	1.224	5.4	80	-0.05
75 C	Fluoranthene	1.369	1.215	11.2	78	-0.04
76 I	Chrysene-d12	1.000	1.000	0.0	72	-0.03
77	Benzydine	0.635	0.623	1.9	73	-0.04
78	Pyrene	1.311	1.430	-9.1	77	-0.04
79 S	Terphenyl-d14	0.981	1.089	-11.0	79	-0.04
80	Butylbenzylphthalate	0.578	0.574	0.7	68	-0.04
81	Benzo(a)anthracene	1.374	1.350	1.7	71	-0.04
82	3,3'-Dichlorobenzidine	0.500	0.484	3.2	70	-0.05
83	Chrysene	1.295	1.297	-0.2	73	-0.04
84	Bis(2-ethylhexyl)phthalate	0.874	0.844	3.4	65	-0.05
85 c	Di-n-octyl phthalate	1.529	1.460	4.5	66	-0.06
86 I	Perylene-d12	1.000	1.000	0.0	78	-0.07
87	Indeno(1,2,3-cd)pyrene	1.500	1.615	-7.7	83	-0.11
88	Benzo(b)fluoranthene	1.218	1.185	2.7	74	-0.07
89	Benzo(k)fluoranthene	1.217	1.230	-1.1	79	-0.08
90 C	Benzo(a)pyrene	1.154	1.166	-1.0	78	-0.08
91	Dibenzo(a,h)anthracene	1.228	1.316	-7.2	83	-0.14
92	Benzo(g,h,i)perylene	1.220	1.312	-7.5	83	-0.14

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120325\
Data File : BP026224.D
Acq On : 03 Dec 2025 12:53
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Dec 03 13:21:51 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 19 01:25:31 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0