

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120423\
 Data File : BP018647.D
 Acq On : 06 Dec 2023 20:53
 Operator : MA/JU
 Sample : 05439-14
 Misc :
 ALS Vial : 84 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 C0AD6

Quant Time: Dec 06 21:44:54 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Dec 04 21:39:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.852	152	337345	20.000	ng/u1	0.00
20) Naphthalene-d8	10.646	136	1265367	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.481	164	744817	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.228	188	1619733	20.000	ng/u1	0.00
79) Chrysene-d12	21.316	240	1714130	20.000	ng/u1	0.00
88) Perylene-d12	23.686	264	2025938	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.270	96	34505	3.486	ng/uL	0.00
4) Pyridine-d5	3.687	84	194487	7.335	ng/u1	0.00
7) Phenol-d5	7.016	99	150589	4.448	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.187	67	386758	19.275	ng/u1	0.00
11) 2-Chlorophenol-d4	7.381	132	483057	18.286	ng/u1	0.00
15) 4-Methylphenol-d8	8.563	113	294985	10.767	ng/u1	0.00
21) Nitrobenzene-d5	9.010	128	303174	26.920	ng/u1	0.00
24) 2-Nitrophenol-d4	9.734	143	304333	26.064	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.269	165	531064	22.243	ng/u1	0.00
31) 4-Chloroaniline-d4	10.787	131	592193	18.979	ng/u1	0.00
46) Dimethylphthalate-d6	13.898	166	1738524	26.153	ng/u1	0.00
49) Acenaphthylene-d8	14.175	160	1945524	26.579	ng/u1	0.00
54) 4-Nitrophenol-d4	14.681	143	50706	4.647	ng/u1	0.00
60) Fluorene-d10	15.475	176	1533511	27.483	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.592	200	221186	22.775	ng/u1	0.00
73) Anthracene-d10	17.328	188	2561326	30.079	ng/u1	0.00
81) Pyrene-d10	19.563	212	3253575	24.384	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.533	264	3529510	29.879	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120423\
Data File : BP018647.D
Acq On : 06 Dec 2023 20:53
Operator : MA/JU
Sample : 05439-14
Misc :
ALS Vial : 84 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
C0AD6

Quant Time: Dec 06 21:44:54 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120123.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Dec 04 21:39:40 2023
Response via : Initial Calibration

