

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120423\
 Data File : BP018623.D
 Acq On : 06 Dec 2023 02:27
 Operator : MA/JU
 Sample : 05424-08MSD
 Misc :
 ALS Vial : 59 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 YCG91MSD

Quant Time: Dec 06 03:08:09 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Dec 04 21:39:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.852	152	390754	20.000	ng/u1	0.00
20) Naphthalene-d8	10.646	136	1722300	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.481	164	1145141	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.228	188	2533571	20.000	ng/u1	0.00
79) Chrysene-d12	21.316	240	2012629	20.000	ng/u1	0.00
88) Perylene-d12	23.686	264	2015824	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.264	96	45590	3.976	ng/uL	0.00
4) Pyridine-d5	3.687	84	346746	11.289	ng/u1	0.00
7) Phenol-d5	7.017	99	282097	7.194	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.187	67	776814	33.424	ng/u1	0.00
11) 2-Chlorophenol-d4	7.381	132	761928	24.901	ng/u1	0.00
15) 4-Methylphenol-d8	8.569	113	550867	17.358	ng/u1	0.00
21) Nitrobenzene-d5	9.016	128	525397	34.275	ng/u1	0.00
24) 2-Nitrophenol-d4	9.734	143	582166	36.631	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.269	165	1046448	32.201	ng/u1	0.00
31) 4-Chloroaniline-d4	10.787	131	1161861	27.357	ng/u1	0.00
46) Dimethylphthalate-d6	13.899	166	3648517	35.698	ng/u1	0.00
49) Acenaphthylene-d8	14.175	160	3845404	34.169	ng/u1	0.00
54) 4-Nitrophenol-d4	14.681	143	142656	8.503	ng/u1	0.00
60) Fluorene-d10	15.475	176	3014084	35.133	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.598	200	616372	40.574	ng/u1	0.00
73) Anthracene-d10	17.328	188	4837211	36.317	ng/u1	0.00
81) Pyrene-d10	19.563	212	5478361	34.968	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.533	264	4475800	38.080	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.305	88	56734	4.512	ng/uL	98
5) Pyridine	3.705	79	420597	13.560	ng/u1	98
6) Benzaldehyde	6.993	77	736382	36.372	ng/u1	97
8) Phenol	7.046	94	322855	8.396	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.281	93	1126963	35.684	ng/u1	100
12) 2-Chlorophenol	7.417	128	812814	26.247	ng/u1	99
13) 2-Methylphenol	8.299	108	657531	22.334	ng/u1	97
14) 2,2'-oxybis(1-Chloropr...	8.381	45	1354249	33.795	ng/u1	98
16) Acetophenone	8.681	105	1751468	37.033	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.669	70	873335	33.566	ng/u1	98
18) 4-Methylphenol	8.628	108	619693	19.276	ng/u1	100
19) Hexachloroethane	8.928	117	317901	25.040	ng/u1	96
22) Nitrobenzene	9.058	77	1309649	35.140	ng/u1	97
23) Isophorone	9.587	82	2639328	34.733	ng/u1	99
25) 2-Nitrophenol	9.769	139	620836	36.485	ng/u1	97
26) 2,4-Dimethylphenol	9.828	107	738497	22.190	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.069	93	1651732	34.429	ng/u1	99
29) 2,4-Dichlorophenol	10.299	162	1046477	33.432	ng/u1	99
30) Naphthalene	10.699	128	3428508	32.815	ng/u1	100
32) 4-Chloroaniline	10.810	127	1128653	27.883	ng/u1	100
33) Hexachlorobutadiene	10.981	225	646508	30.840	ng/u1	99
34) Caprolactam	11.610	113	55654	5.798	ng/u1	99
35) 4-Chloro-3-methylphenol	11.934	107	1067783	30.451	ng/u1	99
36) 2-Methylnaphthalene	12.310	142	2543781	34.367	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120423\
 Data File : BP018623.D
 Acq On : 06 Dec 2023 02:27
 Operator : MA/JU
 Sample : 05424-08MSD
 Misc :
 ALS Vial : 59 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 YCG91MSD

Quant Time: Dec 06 03:08:09 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Dec 04 21:39:40 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.528	142	2524300	33.770	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.675	216	1510592	36.358	ng/ul	98
40) Hexachlorocyclopentadiene	12.652	237	664348	26.977	ng/ul	98
41) 2,4,6-Trichlorophenol	12.916	196	1004784	37.594	ng/ul	99
42) 2,4,5-Trichlorophenol	12.987	196	1097703	37.884	ng/ul	98
43) 1,1'-Biphenyl	13.322	154	3507274	35.413	ng/ul	98
44) 2-Chloronaphthalene	13.357	162	2797953	35.770	ng/ul	99
45) 2-Nitroaniline	13.569	65	853914	40.269	ng/ul	98
47) Dimethylphthalate	13.951	163	3851524	38.283	ng/ul	99
48) 2,6-Dinitrotoluene	14.069	165	820325	43.158	ng/ul	93
50) Acenaphthylene	14.204	152	4681744	37.000	ng/ul	100
51) 3-Nitroaniline	14.393	138	732175	38.588	ng/ul	99
52) Acenaphthene	14.546	153	3118572	37.276	ng/ul	99
53) 2,4-Dinitrophenol	14.598	184	389016	38.617	ng/ul	96
55) 4-Nitrophenol	14.698	109	113145	8.927	ng/ul	98
56) Dibenzofuran	14.881	168	4325654	37.201	ng/ul	99
57) 2,4-Dinitrotoluene	14.851	165	1177175	44.028	ng/ul	96
58) 2,3,4,6-Tetrachlorophenol	15.104	232	977678	40.225	ng/ul	97
59) Diethylphthalate	15.310	149	3881774	39.430	ng/ul	100
61) Fluorene	15.534	166	3458641	37.591	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.528	204	1862457	38.597	ng/ul	94
63) 4-Nitroaniline	15.557	138	790558	43.736	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.610	198	699684	43.080	ng/ul	95
67) N-Nitrosodiphenylamine	15.740	169	3117366	37.132	ng/ul	99
68) 4-Bromophenyl-phenylether	16.422	248	1267476	39.627	ng/ul	94
69) Hexachlorobenzene	16.528	284	1438141	40.271	ng/ul	99
70) Atrazine	16.698	200	1034036	41.428	ng/ul	98
71) Pentachlorophenol	16.875	266	865051	40.730	ng/ul	99
72) Phenanthrene	17.275	178	5888913	38.486	ng/ul	100
74) Anthracene	17.363	178	5890811	38.407	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.281	216	1550354	35.390	ng/ul	98
76) Pentachlorobenzene	14.798	250	1632404	37.935	ng/ul	100
77) Carbazole	17.640	167	5322892	43.544	ng/ul	99
78) Di-n-butylphthalate	18.192	149	6642288	40.807	ng/ul	100
80) Fluoranthene	19.239	202	6875957	37.130	ng/ul	98
82) Pyrene	19.592	202	6881735	36.836	ng/ul	97
83) Butylbenzylphthalate	20.469	149	2862352	41.851	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.239	252	1531130	31.306	ng/ul	98
85) Benzo(a)anthracene	21.298	228	6447381	39.467	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.228	149	3751371	39.907	ng/ul#	98
87) Chrysene	21.351	228	5929058	39.412	ng/ul	99
89) Di-n-octyl phthalate	22.145	149	6520794	44.305	ng/ul	100
90) Benzo(b)fluoranthene	22.963	252	5954056	40.060	ng/ul	99
91) Benzo(k)fluoranthene	23.010	252	5796624	40.308	ng/ul	100
93) Benzo(a)pyrene	23.586	252	5458174	40.191	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.151	276	6678446	41.159	ng/ul	97
95) Dibenzo(a,h)anthracene	26.174	278	5551903	41.132	ng/ul	98
96) Benzo(g,h,i)perylene	26.915	276	5394151	41.308	ng/ul	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120423\
 Data File : BP018623.D
 Acq On : 06 Dec 2023 02:27
 Operator : MA/JU
 Sample : 05424-08MSD
 Misc :
 ALS Vial : 59 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 YCG91MSD

Quant Time: Dec 06 03:08:09 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Dec 04 21:39:40 2023
 Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 4.512 ng/uL
 RT: 3.305 min Scan# 31
 Delta R.T. 0.000 min
 Lab File: BP018623.D
 Acq: 06 Dec 2023 02:27




Instrument :
 BNA_P
 ClientSampleId :
 YCG91MSD


Tgt Ion: 88 Resp: 56734
 Ion Ratio Lower Upper
 88 100
 43 25.5 21.8 32.6
 58 71.8 56.5 84.7



#5
 Pyridine
 Concen: 13.560 ng/uL
 RT: 3.705 min Scan# 105
 Delta R.T. 0.000 min
 Lab File: BP018623.D
 Acq: 06 Dec 2023 02:27

Tgt Ion: 79 Resp: 420597
 Ion Ratio Lower Upper
 79 100
 52 69.1 57.4 86.0
 51 32.5 26.2 39.4









#6
Benzaldehyde
Concen: 36.372 ng/ul
RT: 6.993 min Scan# 61
Delta R.T. -0.006 min
Lab File: BP018623.D
Acq: 06 Dec 2023 02:27




Instrument :
BNA_P
ClientSampleId :
YCG91MSD


Tgt Ion: 77 Resp: 736382
Ion Ratio Lower Upper
77 100
105 99.4 77.9 116.9
106 96.7 74.6 111.8



#8
Phenol
Concen: 8.396 ng/ul
RT: 7.046 min Scan# 673
Delta R.T. -0.006 min
Lab File: BP018623.D
Acq: 06 Dec 2023 02:27

Tgt Ion: 94 Resp: 322855
Ion Ratio Lower Upper
94 100
65 25.0 21.0 31.6
66 35.1 28.6 43.0









#10
Bis(2-Chloroethyl)ether
Concen: 35.684 ng/ul
RT: 7.281 min Scan# 713
Delta R.T. -0.006 min
Lab File: BP018623.D
Acq: 06 Dec 2023 02:27

Instrument :
BNA_P
ClientSampleId :
YCG91MSD

Tgt Ion: 93 Resp: 1126963




Ion	Ratio	Lower	Upper
93	100		
63	73.1	58.3	87.5
95	32.8	25.9	38.9




#12
2-Chlorophenol
Concen: 26.247 ng/ul
RT: 7.417 min Scan# 736
Delta R.T. 0.000 min
Lab File: BP018623.D
Acq: 06 Dec 2023 02:27

Tgt Ion: 128 Resp: 812814

Ion	Ratio	Lower	Upper
128	100		
64	45.4	37.0	55.4
130	33.2	26.8	40.2









#13
2-Methylphenol
Concen: 22.334 ng/ul
RT: 8.299 min Scan# 886
Delta R.T. -0.006 min
Lab File: BP018623.D
Acq: 06 Dec 2023 02:27




Instrument :
BNA_P
ClientSampleId :
YCG91MSD


Tgt Ion:108 Resp: 657531
Ion Ratio Lower Upper
108 100
107 90.6 70.6 105.8



#14
2,2'-oxybis(1-Chloropropane)
Concen: 33.795 ng/ul
RT: 8.381 min Scan# 900
Delta R.T. -0.006 min
Lab File: BP018623.D
Acq: 06 Dec 2023 02:27

Tgt Ion: 45 Resp: 1354249
Ion Ratio Lower Upper
45 100
77 16.4 12.7 19.1
79 12.2 10.5 15.7









#16
Acetophenone
Concen: 37.033 ng/ul
RT: 8.681 min Scan# 951
Delta R.T. 0.000 min
Lab File: BP018623.D
Acq: 06 Dec 2023 02:27




Instrument :
BNA_P
ClientSampleId :
YCG91MSD


Tgt Ion:105 Resp: 1751468
Ion Ratio Lower Upper
105 100
77 74.0 59.4 89.0
51 23.3 19.4 29.0



#17
N-Nitroso-di-n-propylamine
Concen: 33.566 ng/ul
RT: 8.669 min Scan# 949
Delta R.T. -0.006 min
Lab File: BP018623.D
Acq: 06 Dec 2023 02:27

Tgt Ion: 70 Resp: 873335
Ion Ratio Lower Upper
70 100
42 49.1 41.0 61.4
101 10.1 7.8 11.8
130 21.4 16.9 25.3









#18
4-Methylphenol
Concen: 19.276 ng/ul
RT: 8.628 min Scan# 942
Delta R.T. -0.006 min
Lab File: BP018623.D
Acq: 06 Dec 2023 02:27




Instrument :
BNA_P
ClientSampleId :
YCG91MSD


Tgt Ion:108 Resp: 619693
Ion Ratio Lower Upper
108 100
107 113.1 90.6 136.0



#19
Hexachloroethane
Concen: 25.040 ng/ul
RT: 8.928 min Scan# 993
Delta R.T. -0.006 min
Lab File: BP018623.D
Acq: 06 Dec 2023 02:27

Tgt Ion:117 Resp: 317901
Ion Ratio Lower Upper
117 100
201 108.8 83.7 125.5
199 68.7 52.3 78.5









#22
Nitrobenzene
Concen: 35.140 ng/ul
RT: 9.058 min Scan# 1015
Delta R.T. 0.000 min
Lab File: BP018623.D
Acq: 06 Dec 2023 02:27




Instrument :
BNA_P
ClientSampleId :
YCG91MSD


Tgt Ion: 77 Resp: 1309649
Ion Ratio Lower Upper
77 100
123 47.2 35.8 53.8
65 13.0 10.5 15.7



#23
Isophorone
Concen: 34.733 ng/ul
RT: 9.587 min Scan# 1105
Delta R.T. 0.000 min
Lab File: BP018623.D
Acq: 06 Dec 2023 02:27

Tgt Ion: 82 Resp: 2639328
Ion Ratio Lower Upper
82 100
95 7.0 5.5 8.3
138 17.3 13.5 20.3









#25
2-Nitrophenol
Concen: 36.485 ng/ul
RT: 9.769 min Scan# 1135
Delta R.T. 0.000 min
Lab File: BP018623.D
Acq: 06 Dec 2023 02:27




Instrument :
BNA_P
ClientSampleId :
YCG91MSD


Tgt Ion:139 Resp: 620836
Ion Ratio Lower Upper
139 100
65 46.0 39.1 58.7
109 29.1 23.8 35.6



#26
2,4-Dimethylphenol
Concen: 22.190 ng/ul
RT: 9.828 min Scan# 1146
Delta R.T. -0.006 min
Lab File: BP018623.D
Acq: 06 Dec 2023 02:27

Tgt Ion:107 Resp: 738497
Ion Ratio Lower Upper
107 100
121 51.0 40.0 60.0
122 86.3 69.1 103.7









#27
 Bis(2-Chloroethoxy)methane
 Concen: 34.429 ng/ul
 RT: 10.069 min Scan# 1187
 Delta R.T. -0.006 min
 Lab File: BP018623.D
 Acq: 06 Dec 2023 02:27




Instrument :
 BNA_P
 ClientSampleId :
 YCG91MSD


Tgt Ion: 93 Resp: 1651732
 Ion Ratio Lower Upper
 93 100
 95 32.3 26.1 39.1
 123 10.6 8.9 13.3



#29
 2,4-Dichlorophenol
 Concen: 33.432 ng/ul
 RT: 10.299 min Scan# 1226
 Delta R.T. 0.000 min
 Lab File: BP018623.D
 Acq: 06 Dec 2023 02:27

Tgt Ion:162 Resp: 1046477
 Ion Ratio Lower Upper
 162 100
 164 64.9 51.7 77.5
 98 35.2 29.2 43.8









#30
Naphthalene
Concen: 32.815 ng/ul
RT: 10.699 min Scan# 1294
Delta R.T. -0.006 min
Lab File: BP018623.D
Acq: 06 Dec 2023 02:27




Instrument :
BNA_P
ClientSampleId :
YCG91MSD


Tgt Ion:128 Resp: 3428508
Ion Ratio Lower Upper
128 100
129 11.0 8.9 13.3
127 13.1 10.5 15.7



#32
4-Chloroaniline
Concen: 27.883 ng/ul
RT: 10.810 min Scan# 1313
Delta R.T. -0.006 min
Lab File: BP018623.D
Acq: 06 Dec 2023 02:27

Tgt Ion:127 Resp: 1128653
Ion Ratio Lower Upper
127 100
129 32.2 25.6 38.4









#33
Hexachlorobutadiene
Concen: 30.840 ng/ul
RT: 10.981 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP018623.D
Acq: 06 Dec 2023 02:27




Instrument :
BNA_P
ClientSampleId :
YCG91MSD


Tgt Ion:225 Resp: 646508
Ion Ratio Lower Upper
225 100
223 63.0 49.5 74.3
227 63.1 50.9 76.3



#34
Caprolactam
Concen: 5.798 ng/ul
RT: 11.610 min Scan# 1449
Delta R.T. 0.018 min
Lab File: BP018623.D
Acq: 06 Dec 2023 02:27

Tgt Ion:113 Resp: 55654
Ion Ratio Lower Upper
113 100
55 167.2 133.7 200.5
56 118.5 97.4 146.2









#35
 4-Chloro-3-methylphenol
 Concen: 30.451 ng/ul
 RT: 11.934 min Scan# 1504
 Delta R.T. -0.006 min
 Lab File: BP018623.D
 Acq: 06 Dec 2023 02:27

Instrument :
 BNA_P
 ClientSampleId :
 YCG91MSD

Tgt Ion:107 Resp: 1067783




Ion	Ratio	Lower	Upper
107	100		
144	25.7	21.0	31.6
142	79.8	63.6	95.4




#36
 2-Methylnaphthalene
 Concen: 34.367 ng/ul
 RT: 12.310 min Scan# 1568
 Delta R.T. 0.000 min
 Lab File: BP018623.D
 Acq: 06 Dec 2023 02:27

Tgt Ion:142 Resp: 2543781

Ion	Ratio	Lower	Upper
142	100		
141	88.8	71.4	107.0









#37
1-Methylnaphthalene
Concen: 33.770 ng/ul
RT: 12.528 min Scan# 1605
Delta R.T. 0.000 min
Lab File: BP018623.D
Acq: 06 Dec 2023 02:27




Instrument :
BNA_P
ClientSampleId :
YCG91MSD


Tgt Ion:142 Resp: 2524300
Ion Ratio Lower Upper
142 100
141 92.2 73.5 110.3
116 3.6 2.8 4.2



#39
1,2,4,5-Tetrachlorobenzene
Concen: 36.358 ng/ul
RT: 12.675 min Scan# 1630
Delta R.T. -0.006 min
Lab File: BP018623.D
Acq: 06 Dec 2023 02:27

Tgt Ion:216 Resp: 1510592
Ion Ratio Lower Upper
216 100
214 78.0 63.9 95.9
179 19.8 16.3 24.5
108 15.1 12.6 19.0





#40

Hexachlorocyclopentadiene

Concen: 26.977 ng/ul

RT: 12.652 min Scan# 1627

Delta R.T. -0.006 min

Lab File: BP018623.D

Acq: 06 Dec 2023 02:27

Instrument :

BNA_P

ClientSampleId :

YCG91MSD





Tgt Ion:237 Resp: 664348

Ion Ratio Lower Upper

237 100

235 63.8 49.4 74.0

272 12.7 9.8 14.8



#41

2,4,6-Trichlorophenol

Concen: 37.594 ng/ul

RT: 12.916 min Scan# 1671

Delta R.T. -0.006 min

Lab File: BP018623.D

Acq: 06 Dec 2023 02:27




Tgt Ion:196 Resp: 1004784


Ion Ratio Lower Upper

196 100

198 95.6 76.9 115.3

200 30.6 25.0 37.6





#42

2,4,5-Trichlorophenol

Concen: 37.884 ng/ul

RT: 12.987 min Scan# 1683

Delta R.T. 0.000 min

Lab File: BP018623.D

Acq: 06 Dec 2023 02:27

Instrument :

BNA_P

ClientSampleId :

YCG91MSD





Tgt Ion:196 Resp: 1097703

Ion Ratio Lower Upper

196 100

198 95.6 77.9 116.9

200 31.0 25.2 37.8



#43

1,1'-Biphenyl

Concen: 35.413 ng/ul

RT: 13.322 min Scan# 1740

Delta R.T. 0.000 min

Lab File: BP018623.D

Acq: 06 Dec 2023 02:27




Tgt Ion:154 Resp: 3507274


Ion Ratio Lower Upper

154 100

153 39.5 32.8 49.2

76 12.7 10.7 16.1









#44
2-Chloronaphthalene
Concen: 35.770 ng/ul
RT: 13.357 min Scan# 1746
Delta R.T. -0.006 min
Lab File: BP018623.D
Acq: 06 Dec 2023 02:27




Instrument :
BNA_P
ClientSampleId :
YCG91MSD


Tgt Ion:162 Resp: 2797953
Ion Ratio Lower Upper
162 100
164 32.8 25.7 38.5
127 37.4 30.3 45.5



#45
2-Nitroaniline
Concen: 40.269 ng/ul
RT: 13.569 min Scan# 1782
Delta R.T. 0.000 min
Lab File: BP018623.D
Acq: 06 Dec 2023 02:27

Tgt Ion: 65 Resp: 853914
Ion Ratio Lower Upper
65 100
92 72.6 56.9 85.3
138 111.7 88.2 132.2









#47
Dimethylphthalate
Concen: 38.283 ng/ul
RT: 13.951 min Scan# 1847
Delta R.T. 0.000 min
Lab File: BP018623.D
Acq: 06 Dec 2023 02:27




Instrument :
BNA_P
ClientSampleId :
YCG91MSD


Tgt Ion:163 Resp: 3851524
Ion Ratio Lower Upper
163 100
194 4.2 3.2 4.8
164 10.2 8.0 12.0



#48
2,6-Dinitrotoluene
Concen: 43.158 ng/ul
RT: 14.069 min Scan# 1867
Delta R.T. 0.000 min
Lab File: BP018623.D
Acq: 06 Dec 2023 02:27

Tgt Ion:165 Resp: 820325
Ion Ratio Lower Upper
165 100
89 52.1 46.6 70.0
121 19.3 17.3 25.9









#50
Acenaphthylene
Concen: 37.000 ng/ul
RT: 14.204 min Scan# 1890
Delta R.T. 0.000 min
Lab File: BP018623.D
Acq: 06 Dec 2023 02:27




Instrument :
BNA_P
ClientSampleId :
YCG91MSD


Tgt Ion:152 Resp: 4681744
Ion Ratio Lower Upper
152 100
151 20.1 15.9 23.9
153 13.5 10.6 16.0



#51
3-Nitroaniline
Concen: 38.588 ng/ul
RT: 14.393 min Scan# 1922
Delta R.T. 0.000 min
Lab File: BP018623.D
Acq: 06 Dec 2023 02:27

Tgt Ion:138 Resp: 732175
Ion Ratio Lower Upper
138 100
108 11.1 9.5 14.3
92 115.9 93.1 139.7









#52
Acenaphthene
Concen: 37.276 ng/ul
RT: 14.546 min Scan# 1948
Delta R.T. -0.006 min
Lab File: BP018623.D
Acq: 06 Dec 2023 02:27




Instrument :
BNA_P
ClientSampleId :
YCG91MSD


Tgt Ion:153 Resp: 3118572
Ion Ratio Lower Upper
153 100
152 47.1 38.1 57.1
154 88.3 71.7 107.5



#53
2,4-Dinitrophenol
Concen: 38.617 ng/ul
RT: 14.598 min Scan# 1957
Delta R.T. 0.000 min
Lab File: BP018623.D
Acq: 06 Dec 2023 02:27

Tgt Ion:184 Resp: 389016
Ion Ratio Lower Upper
184 100
63 54.5 48.2 72.4
154 61.2 49.1 73.7









#55
4-Nitrophenol
Concen: 8.927 ng/ul
RT: 14.698 min Scan# 1974
Delta R.T. 0.000 min
Lab File: BP018623.D
Acq: 06 Dec 2023 02:27




Instrument :
BNA_P
ClientSampleId :
YCG91MSD


Tgt Ion:109 Resp: 113145
Ion Ratio Lower Upper
109 100
139 137.2 109.0 163.4
65 128.5 106.2 159.4



#56
Dibenzofuran
Concen: 37.201 ng/ul
RT: 14.881 min Scan# 2005
Delta R.T. 0.000 min
Lab File: BP018623.D
Acq: 06 Dec 2023 02:27

Tgt Ion:168 Resp: 4325654
Ion Ratio Lower Upper
168 100
139 35.6 29.1 43.7









#57
2,4-Dinitrotoluene
Concen: 44.028 ng/ul
RT: 14.851 min Scan# 2000
Delta R.T. 0.000 min
Lab File: BP018623.D
Acq: 06 Dec 2023 02:27




Instrument :
BNA_P
ClientSampleId :
YCG91MSD


Tgt Ion:165 Resp: 1177175
Ion Ratio Lower Upper
165 100
63 38.1 32.7 49.1
182 2.8 2.1 3.1



#58
2,3,4,6-Tetrachlorophenol
Concen: 40.225 ng/ul
RT: 15.104 min Scan# 2043
Delta R.T. -0.006 min
Lab File: BP018623.D
Acq: 06 Dec 2023 02:27

Tgt Ion:232 Resp: 977678
Ion Ratio Lower Upper
232 100
131 40.4 33.8 50.8
130 2.2 1.8 2.8
166 29.8 24.7 37.1





#59

Diethylphthalate

Concen: 39.430 ng/ul

RT: 15.310 min Scan# 2078

Delta R.T. -0.006 min

Lab File: BP018623.D

Acq: 06 Dec 2023 02:27

Instrument :

BNA_P

ClientSampleId :

YCG91MSD





Tgt Ion:149 Resp: 3881774

Ion Ratio Lower Upper

149 100

177 22.8 18.2 27.4

150 12.6 9.9 14.9



#61

Fluorene

Concen: 37.591 ng/ul

RT: 15.534 min Scan# 2116

Delta R.T. 0.000 min

Lab File: BP018623.D

Acq: 06 Dec 2023 02:27




Tgt Ion:166 Resp: 3458641


Ion Ratio Lower Upper

166 100

165 98.9 78.0 117.0

167 13.7 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 38.597 ng/ul
RT: 15.528 min Scan# 2115
Delta R.T. 0.000 min
Lab File: BP018623.D
Acq: 06 Dec 2023 02:27




Instrument :
BNA_P
ClientSampleId :
YCG91MSD


Tgt Ion:204 Resp: 1862457
Ion Ratio Lower Upper
204 100
206 32.7 27.1 40.7
141 51.6 45.8 68.8



#63
4-Nitroaniline
Concen: 43.736 ng/ul
RT: 15.557 min Scan# 2120
Delta R.T. 0.000 min
Lab File: BP018623.D
Acq: 06 Dec 2023 02:27

Tgt Ion:138 Resp: 790558
Ion Ratio Lower Upper
138 100
92 53.0 42.3 63.5
108 78.6 62.0 93.0









#66
4,6-Dinitro-2-methylphenol
Concen: 43.080 ng/ul
RT: 15.610 min Scan# 2128
Delta R.T. 0.000 min
Lab File: BP018623.D
Acq: 06 Dec 2023 02:27




Instrument :
BNA_P
ClientSampleId :
YCG91MSD


Tgt Ion:198 Resp: 699684
Ion Ratio Lower Upper
198 100
182 3.6 3.4 5.2
77 23.1 20.6 30.8



#67
N-Nitrosodiphenylamine
Concen: 37.132 ng/ul
RT: 15.740 min Scan# 2151
Delta R.T. -0.006 min
Lab File: BP018623.D
Acq: 06 Dec 2023 02:27

Tgt Ion:169 Resp: 3117366
Ion Ratio Lower Upper
169 100
168 67.6 53.1 79.7
167 36.3 29.1 43.7









#68
4-Bromophenyl-phenylether
Concen: 39.627 ng/ul
RT: 16.422 min Scan# 2267
Delta R.T. 0.000 min
Lab File: BP018623.D
Acq: 06 Dec 2023 02:27




Instrument :
BNA_P
ClientSampleId :
YCG91MSD


Tgt Ion:248 Resp: 1267476
Ion Ratio Lower Upper
248 100
250 97.6 79.3 118.9
141 61.8 57.2 85.8



#69
Hexachlorobenzene
Concen: 40.271 ng/ul
RT: 16.528 min Scan# 2285
Delta R.T. -0.006 min
Lab File: BP018623.D
Acq: 06 Dec 2023 02:27

Tgt Ion:284 Resp: 1438141
Ion Ratio Lower Upper
284 100
142 31.6 25.8 38.6
249 31.7 24.9 37.3









#70
Atrazine
Concen: 41.428 ng/ul
RT: 16.698 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP018623.D
Acq: 06 Dec 2023 02:27




Instrument :
BNA_P
ClientSampleId :
YCG91MSD


Tgt Ion:200 Resp: 1034036
Ion Ratio Lower Upper
200 100
173 25.9 21.0 31.4
215 49.1 38.2 57.2



#71
Pentachlorophenol
Concen: 40.730 ng/ul
RT: 16.875 min Scan# 2344
Delta R.T. -0.006 min
Lab File: BP018623.D
Acq: 06 Dec 2023 02:27

Tgt Ion:266 Resp: 865051
Ion Ratio Lower Upper
266 100
264 62.9 50.2 75.2
268 63.8 50.5 75.7





#72

Phenanthrene

Concen: 38.486 ng/uI

RT: 17.275 min Scan# 2411

Delta R.T. 0.000 min

Lab File: BP018623.D

Acq: 06 Dec 2023 02:27

Instrument :

BNA_P

ClientSampleId :

YCG91MSD





Tgt Ion:178 Resp: 5888913

Ion Ratio Lower Upper

178 100

179 15.6 12.3 18.5

176 19.4 15.4 23.2



#74

Anthracene

Concen: 38.407 ng/uI

RT: 17.363 min Scan# 2427

Delta R.T. 0.000 min

Lab File: BP018623.D

Acq: 06 Dec 2023 02:27




Tgt Ion:178 Resp: 5890811


Ion Ratio Lower Upper

178 100

179 15.8 12.5 18.7

176 19.1 15.0 22.6





#75

1,2,3,4-Tetrachlorobenzene

Concen: 35.390 ng/uL

RT: 13.281 min Scan# 1733

Delta R.T. -0.006 min

Lab File: BP018623.D

Acq: 06 Dec 2023 02:27

Instrument :

BNA_P

ClientSampleId :

YCG91MSD





Tgt Ion:216 Resp: 1550354

Ion Ratio Lower Upper

216 100

214 78.8 61.4 92.2

218 48.0 39.2 58.8



#76

Pentachlorobenzene

Concen: 37.935 ng/uL

RT: 14.798 min Scan# 1991

Delta R.T. 0.000 min

Lab File: BP018623.D

Acq: 06 Dec 2023 02:27




Tgt Ion:250 Resp: 1632404


Ion Ratio Lower Upper

250 100

248 62.7 50.6 75.8

252 63.9 51.3 76.9









#77
Carbazole
Concen: 43.544 ng/ul
RT: 17.640 min Scan# 2474
Delta R.T. 0.000 min
Lab File: BP018623.D
Acq: 06 Dec 2023 02:27




Instrument :
BNA_P
ClientSampleId :
YCG91MSD


Tgt Ion:167 Resp: 5322892
Ion Ratio Lower Upper
167 100
166 22.2 17.1 25.7
139 12.5 10.2 15.4



#78
Di-n-butylphthalate
Concen: 40.807 ng/ul
RT: 18.192 min Scan# 2568
Delta R.T. 0.000 min
Lab File: BP018623.D
Acq: 06 Dec 2023 02:27

Tgt Ion:149 Resp: 6642288
Ion Ratio Lower Upper
149 100
150 9.5 7.4 11.2
104 4.9 3.9 5.9









#80
Fluoranthene
Concen: 37.130 ng/u1
RT: 19.239 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP018623.D
Acq: 06 Dec 2023 02:27




Instrument :
BNA_P
ClientSampleId :
YCG91MSD


Tgt Ion:202 Resp: 6875957
Ion Ratio Lower Upper
202 100
101 10.7 9.1 13.7
100 8.2 7.1 10.7



#82
Pyrene
Concen: 36.836 ng/u1
RT: 19.592 min Scan# 2806
Delta R.T. 0.000 min
Lab File: BP018623.D
Acq: 06 Dec 2023 02:27

Tgt Ion:202 Resp: 6881735
Ion Ratio Lower Upper
202 100
101 12.3 11.0 16.4
100 10.1 9.1 13.7





#83

Butylbenzylphthalate

Concen: 41.851 ng/ul

RT: 20.469 min Scan# 2955

Delta R.T. -0.006 min

Lab File: BP018623.D

Acq: 06 Dec 2023 02:27

Instrument :

BNA_P

ClientSampleId :

YCG91MSD





Tgt Ion:149 Resp: 2862352

Ion Ratio Lower Upper

149 100

91 69.3 56.6 85.0

206 22.6 17.4 26.0



#84

3,3'-Dichlorobenzidine

Concen: 31.306 ng/ul

RT: 21.239 min Scan# 3086

Delta R.T. 0.000 min

Lab File: BP018623.D

Acq: 06 Dec 2023 02:27




Tgt Ion:252 Resp: 1531130


Ion Ratio Lower Upper

252 100

254 65.1 51.0 76.4

126 9.0 7.8 11.6





#85

Benzo(a)anthracene

Concen: 39.467 ng/ul

RT: 21.298 min Scan# 3096

Delta R.T. -0.006 min

Lab File: BP018623.D

Acq: 06 Dec 2023 02:27

Instrument :

BNA_P

ClientSampleId :

YCG91MSD





Tgt Ion:228 Resp: 6447381

Ion Ratio Lower Upper

228 100

229 20.5 15.7 23.5

226 28.2 21.9 32.9



#86

Bis(2-ethylhexyl)phthalate

Concen: 39.907 ng/ul

RT: 21.228 min Scan# 3084

Delta R.T. -0.006 min

Lab File: BP018623.D

Acq: 06 Dec 2023 02:27




Tgt Ion:149 Resp: 3751371


Ion Ratio Lower Upper

149 100

167 28.6 22.2 33.2

279 5.0 3.3 4.9#





#87

Chrysene

Concen: 39.412 ng/ul

RT: 21.351 min Scan# 3105

Delta R.T. 0.000 min

Lab File: BP018623.D

Acq: 06 Dec 2023 02:27

Instrument :

BNA_P

ClientSampleId :

YCG91MSD





Tgt Ion:228 Resp: 5929058

Ion Ratio Lower Upper

228 100

226 30.6 23.9 35.9

229 20.2 15.6 23.4



#89

Di-n-octyl phthalate

Concen: 44.305 ng/ul




RT: 22.145 min Scan# 3240


Delta R.T. -0.006 min

Lab File: BP018623.D

Acq: 06 Dec 2023 02:27

Tgt Ion:149 Resp: 6520794





#90

Benzo(b)fluoranthene

Concen: 40.060 ng/u1

RT: 22.963 min Scan# 3379

Delta R.T. 0.000 min

Lab File: BP018623.D

Acq: 06 Dec 2023 02:27

Instrument :

BNA_P

ClientSampleId :

YCG91MSD





Tgt Ion:252 Resp: 5954056

Ion Ratio Lower Upper

252 100

253 22.3 17.3 25.9

125 9.3 7.8 11.6



#91

Benzo(k)fluoranthene

Concen: 40.308 ng/u1

RT: 23.010 min Scan# 3387

Delta R.T. -0.006 min

Lab File: BP018623.D

Acq: 06 Dec 2023 02:27




Tgt Ion:252 Resp: 5796624


Ion Ratio Lower Upper

252 100

253 22.1 17.7 26.5

125 9.4 8.0 12.0





#93

Benzo(a)pyrene

Concen: 40.191 ng/ul

RT: 23.586 min Scan# 3484

Delta R.T. 0.000 min

Lab File: BP018623.D

Acq: 06 Dec 2023 02:27

Instrument :

BNA_P

ClientSampleId :

YCG91MSD





Tgt Ion:252 Resp: 5458174

Ion Ratio Lower Upper

252 100

253 22.2 17.8 26.8

125 9.9 9.2 13.8



#94

Indeno(1,2,3-cd)pyrene

Concen: 41.159 ng/ul

RT: 26.151 min Scan# 3921

Delta R.T. -0.006 min

Lab File: BP018623.D

Acq: 06 Dec 2023 02:27




Tgt Ion:276 Resp: 6678446


Ion Ratio Lower Upper

276 100

138 21.1 19.4 29.2

277 24.9 19.8 29.6









#95
Dibenzo(a,h)anthracene
Concen: 41.132 ng/ul
RT: 26.174 min Scan# 3924
Delta R.T. -0.006 min
Lab File: BP018623.D
Acq: 06 Dec 2023 02:27

Instrument :
BNA_P
ClientSampleId :
YCG91MSD

Tgt Ion:278 Resp: 5551903
Ion Ratio Lower Upper
278 100
139 15.1 13.0 19.6
279 24.1 18.7 28.1



#96
Benzo(g,h,i)perylene
Concen: 41.308 ng/ul
RT: 26.915 min Scan# 4051
Delta R.T. 0.000 min
Lab File: BP018623.D
Acq: 06 Dec 2023 02:27

Tgt Ion:276 Resp: 5394151
Ion Ratio Lower Upper
276 100
138 20.2 18.1 27.1
277 24.0 19.4 29.0

