

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120525\
 Data File : BP026239.D
 Acq On : 05 Dec 2025 15:13
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 05 15:39:36 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 05 14:52:44 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	-0.02
2	1,4-Dioxane	0.506	0.499	1.4	103	-0.01
3	Pyridine	1.453	1.400	3.6	101	-0.02
4	n-Nitrosodimethylamine	0.542	0.524	3.3	102	-0.01
5 S	2-Fluorophenol	1.244	1.259	-1.2	103	-0.02
6	Aniline	2.061	1.856	9.9	99	-0.02
7 S	Phenol-d6	1.570	1.482	5.6	99	-0.03
8	2-Chlorophenol	1.414	1.386	2.0	100	-0.03
9	Benzaldehyde	0.847	1.006	-18.8	100	-0.03
10 C	Phenol	1.689	1.585	6.2	99	-0.04
11	bis(2-Chloroethyl)ether	1.324	1.262	4.7	101	-0.02
12	1,3-Dichlorobenzene	1.530	1.540	-0.7	101	-0.03
13 C	1,4-Dichlorobenzene	1.541	1.560	-1.2	101	-0.02
14	1,2-Dichlorobenzene	1.480	1.475	0.3	100	-0.02
15	Benzyl Alcohol	1.142	1.056	7.5	100	-0.03
16	2,2'-oxybis(1-Chloropropane	1.848	1.701	8.0	101	-0.03
17	2-Methylphenol	1.146	1.056	7.9	99	-0.02
18	Hexachloroethane	0.560	0.567	-1.2	102	-0.02
19 P	n-Nitroso-di-n-propylamine	0.968	0.855	11.7	98	-0.02
20	3+4-Methylphenols	1.583	1.408	11.1	98	-0.02
21 I	Naphthalene-d8	1.000	1.000	0.0	98	-0.02
22	Acetophenone	0.507	0.500	1.4	98	-0.02
23 S	Nitrobenzene-d5	0.351	0.351	0.0	98	-0.02
24	Nitrobenzene	0.355	0.356	-0.3	99	-0.02
25	Isophorone	0.689	0.655	4.9	99	-0.02
26 C	2-Nitrophenol	0.180	0.186	-3.3	99	-0.02
27	2,4-Dimethylphenol	0.318	0.275	13.5	97	-0.02
28	bis(2-Chloroethoxy)methane	0.433	0.411	5.1	98	-0.01
29 C	2,4-Dichlorophenol	0.324	0.325	-0.3	98	-0.02
30	1,2,4-Trichlorobenzene	0.333	0.348	-4.5	98	-0.02
31	Naphthalene	1.081	1.079	0.2	98	-0.01
32	Benzoic acid	0.255	0.307	-20.4	185#	0.02
33	4-Chloroaniline	0.453	0.419	7.5	96	-0.02
34 C	Hexachlorobutadiene	0.217	0.236	-8.8	100	-0.01
35	Caprolactam	0.114	0.103	9.6	95	-0.02
36 C	4-Chloro-3-methylphenol	0.341	0.317	7.0	96	-0.02
37	2-Methylnaphthalene	0.762	0.745	2.2	98	-0.01
38	1-Methylnaphthalene	0.744	0.716	3.8	96	-0.01
39 I	Acenaphthene-d10	1.000	1.000	0.0	96	0.00
40	1,2,4,5-Tetrachlorobenzene	0.611	0.654	-7.0	99	0.00
41 P	Hexachlorocyclopentadiene	0.381	0.437	-14.7	137	0.00
42 S	2,4,6-Tribromophenol	0.260	0.275	-5.8	98	0.00
43 C	2,4,6-Trichlorophenol	0.415	0.423	-1.9	96	-0.01
44	2,4,5-Trichlorophenol	0.462	0.476	-3.0	99	0.00
45 S	2-Fluorobiphenyl	1.376	1.418	-3.1	96	0.00
46	1,1'-Biphenyl	1.510	1.543	-2.2	97	0.00
47	2-Chloronaphthalene	1.190	1.214	-2.0	96	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120525\
 Data File : BP026239.D
 Acq On : 05 Dec 2025 15:13
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 05 15:39:36 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 05 14:52:44 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.326	0.316	3.1	97	0.00
49	Acenaphthylene	1.953	1.943	0.5	96	0.00
50	Dimethylphthalate	1.490	1.491	-0.1	97	0.00
51	2,6-Dinitrotoluene	0.295	0.308	-4.4	97	0.00
52 C	Acenaphthene	1.142	1.121	1.8	96	0.00
53	3-Nitroaniline	0.344	0.333	3.2	96	0.00
54 P	2,4-Dinitrophenol	0.157	0.204	-29.9#	351#	0.00
55	Dibenzofuran	1.776	1.798	-1.2	97	0.00
56 P	4-Nitrophenol	0.307	0.263	14.3	94	0.00
57	2,4-Dinitrotoluene	0.438	0.448	-2.3	98	0.00
58	Fluorene	1.405	1.442	-2.6	99	0.00
59	2,3,4,6-Tetrachlorophenol	0.391	0.400	-2.3	98	0.00
60	Diethylphthalate	1.495	1.500	-0.3	99	0.01
61	4-Chlorophenyl-phenylether	0.700	0.725	-3.6	97	0.00
62	4-Nitroaniline	0.357	0.336	5.9	97	0.00
63	Azobenzene	1.257	1.228	2.3	98	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	97	0.00
65	4,6-Dinitro-2-methylphenol	0.118	0.137	-16.1	165#	0.00
66 c	n-Nitrosodiphenylamine	0.618	0.618	0.0	96	0.00
67	4-Bromophenyl-phenylether	0.222	0.232	-4.5	97	0.00
68	Hexachlorobenzene	0.258	0.272	-5.4	97	0.00
69	Atrazine	0.233	0.237	-1.7	99	0.00
70 C	Pentachlorophenol	0.179	0.183	-2.2	100	0.00
71	Phenanthrene	1.127	1.164	-3.3	100	0.00
72	Anthracene	1.148	1.186	-3.3	98	0.01
73	Carbazole	1.087	1.073	1.3	96	0.00
74	Di-n-butylphthalate	1.284	1.322	-3.0	99	0.00
75 C	Fluoranthene	1.372	1.445	-5.3	100	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	99	0.00
77	Benzydine	0.628	0.562	10.5	96	0.00
78	Pyrene	1.303	1.261	3.2	97	0.00
79 S	Terphenyl-d14	0.979	0.984	-0.5	100	0.00
80	Butylbenzylphthalate	0.572	0.565	1.2	99	0.00
81	Benzo(a)anthracene	1.372	1.350	1.6	97	0.00
82	3,3'-Dichlorobenzidine	0.499	0.490	1.8	97	0.00
83	Chrysene	1.299	1.295	0.3	97	0.00
84	Bis(2-ethylhexyl)phthalate	0.860	0.872	-1.4	103	0.00
85 c	Di-n-octyl phthalate	1.517	1.516	0.1	99	0.00
86 I	Perylene-d12	1.000	1.000	0.0	96	0.02
87	Indeno(1,2,3-cd)pyrene	1.512	1.576	-4.2	95	0.00
88	Benzo(b)fluoranthene	1.215	1.231	-1.3	96	0.01
89	Benzo(k)fluoranthene	1.216	1.217	-0.1	96	0.00
90 C	Benzo(a)pyrene	1.153	1.165	-1.0	96	0.00
91	Dibenzo(a,h)anthracene	1.236	1.301	-5.3	96	0.04
92	Benzo(g,h,i)perylene	1.232	1.299	-5.4	95	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120525\
Data File : BP026239.D
Acq On : 05 Dec 2025 15:13
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Dec 05 15:39:36 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 05 14:52:44 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0