

DDT Breakdown

Date	Instrument Name	DFTPP Data File
12/2/2021	BNA_N	<u>BN017681.D</u>
Compound Name	Response	Retention Time
DDT	2022749	20.59
DDD	25604	20.15
DDE	894	19.63
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
26498	2049247	1.29