

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120723\
 Data File : BP018708.D
 Acq On : 08 Dec 2023 12:40
 Operator : MA/JU
 Sample : PB157521BL
 Misc :
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK521

Quant Time: Dec 08 13:27:13 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Dec 04 21:39:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.846	152	140063	20.000	ng/ul	-0.01
20) Naphthalene-d8	10.640	136	498452	20.000	ng/ul	-0.01
38) Acenaphthene-d10	14.469	164	310991	20.000	ng/ul	-0.02
64) Phenanthrene-d10	17.222	188	762772	20.000	ng/ul	0.00
79) Chrysene-d12	21.310	240	893610	20.000	ng/ul	# 0.00
88) Perylene-d12	23.674	264	1116877	20.000	ng/ul	-0.01

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.264	96	28456	6.923	ng/uL	0.00
4) Pyridine-d5	3.681	84	383621	34.845	ng/ul	0.00
7) Phenol-d5	7.010	99	445383	31.687	ng/ul	-0.01
9) Bis-(2-Chloroethyl)eth...	7.181	67	253361	30.413	ng/ul	-0.01
11) 2-Chlorophenol-d4	7.375	132	348386	31.765	ng/ul	-0.01
15) 4-Methylphenol-d8	8.552	113	343555	30.202	ng/ul	-0.02
21) Nitrobenzene-d5	9.005	128	158845	35.805	ng/ul	-0.01
24) 2-Nitrophenol-d4	9.722	143	171924	37.379	ng/ul	-0.02
28) 2,4-Dichlorophenol-d3	10.257	165	338177	35.957	ng/ul	-0.02
31) 4-Chloroaniline-d4	10.781	131	483075	39.302	ng/ul	-0.01
46) Dimethylphthalate-d6	13.887	166	1115495	40.189	ng/ul	-0.02
49) Acenaphthylene-d8	14.169	160	1190914	38.965	ng/ul	-0.01
54) 4-Nitrophenol-d4	14.675	143	217596	47.759	ng/ul	-0.01
60) Fluorene-d10	15.463	176	968954	41.589	ng/ul	-0.01
65) 4,6-Dinitro-2-methylph...	15.587	200	170325	37.241	ng/ul	0.00
73) Anthracene-d10	17.322	188	1754756	43.759	ng/ul	0.00
81) Pyrene-d10	19.557	212	2383821	34.270	ng/ul	-0.01
92) Benzo(a)pyrene-d12	23.527	264	2896791	44.483	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120723\
Data File : BP018708.D
Acq On : 08 Dec 2023 12:40
Operator : MA/JU
Sample : PB157521BL
Misc :
ALS Vial : 43 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SBLK521

Quant Time: Dec 08 13:27:13 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120123.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Dec 04 21:39:40 2023
Response via : Initial Calibration

