

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120723\
 Data File : BP018671.D
 Acq On : 07 Dec 2023 13:13
 Operator : MA/JU
 Sample : 04929-02DL 5X
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 C0AB5DL

Quant Time: Dec 07 13:45:42 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Dec 04 21:39:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.840	152	205775	20.000	ng/u1	-0.02
20) Naphthalene-d8	10.640	136	802404	20.000	ng/u1	-0.01
38) Acenaphthene-d10	14.475	164	509858	20.000	ng/u1	-0.01
64) Phenanthrene-d10	17.228	188	1189986	20.000	ng/u1	0.00
79) Chrysene-d12	21.316	240	1293967	20.000	ng/u1	0.00
88) Perylene-d12	23.692	264	1539189	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.258	96	5074	0.840	ng/uL	-0.01
4) Pyridine-d5	0.000	84	0d	0.000	ng/u1	
7) Phenol-d5	7.005	99	102620	4.969	ng/u1	-0.02
9) Bis-(2-Chloroethyl)eth...	7.175	67	59937	4.897	ng/u1	-0.02
11) 2-Chlorophenol-d4	7.369	132	77189	4.790	ng/u1	-0.02
15) 4-Methylphenol-d8	8.552	113	80508	4.817	ng/u1	-0.02
21) Nitrobenzene-d5	8.999	128	36017	5.043	ng/u1	-0.02
24) 2-Nitrophenol-d4	9.722	143	35609	4.809	ng/u1	-0.02
28) 2,4-Dichlorophenol-d3	10.257	165	107628	7.109	ng/u1	-0.02
31) 4-Chloroaniline-d4	10.781	131	83602	4.225	ng/u1	-0.01
46) Dimethylphthalate-d6	13.893	166	264704	5.817	ng/u1	-0.01
49) Acenaphthylene-d8	14.169	160	279158	5.571	ng/u1	-0.01
54) 4-Nitrophenol-d4	14.681	143	40928	5.479	ng/u1	0.00
60) Fluorene-d10	15.469	176	242030	6.336	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.587	200	25281	3.543	ng/u1	0.00
73) Anthracene-d10	17.322	188	387556	6.195	ng/u1	0.00
81) Pyrene-d10	19.563	212	510721	5.070	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.533	264	545462	6.078	ng/u1	0.00
Target Compounds						
					Qvalue	
8) Phenol	7.034	94	712463	35.181	ng/u1	98
12) 2-Chlorophenol	7.399	128	174936	10.727	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.646	70	243577	17.777	ng/u1	97
19) Hexachloroethane	8.916	117	101742	15.218	ng/u1	92
23) Isophorone	9.569	82	1398857	39.512	ng/u1	99
25) 2-Nitrophenol	9.752	139	240941	30.392	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.057	93	529355	23.684	ng/u1	99
29) 2,4-Dichlorophenol	10.287	162	650614	44.614	ng/u1	98
33) Hexachlorobutadiene	10.975	225	288836	29.574	ng/u1	98
35) 4-Chloro-3-methylphenol	11.928	107	543438	33.265	ng/u1	100
36) 2-Methylnaphthalene	12.299	142	384786	11.158	ng/u1	98
41) 2,4,6-Trichlorophenol	12.910	196	207279	17.419	ng/u1	98
42) 2,4,5-Trichlorophenol	12.975	196	150369	11.656	ng/u1	98
44) 2-Chloronaphthalene	13.351	162	674607	19.370	ng/u1	99
47) Dimethylphthalate	13.940	163	1949626	43.525	ng/u1	99
50) Acenaphthylene	14.198	152	1122083	19.917	ng/u1	100
52) Acenaphthene	14.540	153	1268346	34.050	ng/u1	99
55) 4-Nitrophenol	14.692	109	190328	33.726	ng/u1	98
59) Diethylphthalate	15.304	149	881715	20.116	ng/u1	100
61) Fluorene	15.528	166	859317	20.977	ng/u1	99
62) 4-Chlorophenyl-phenyle...	15.522	204	1089776	50.724	ng/u1	96
68) 4-Bromophenyl-phenylether	16.422	248	683422	45.491	ng/u1	93
69) Hexachlorobenzene	16.528	284	152320	9.081	ng/u1	97
71) Pentachlorophenol	16.875	266	128863	12.918	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120723\
 Data File : BP018671.D
 Acq On : 07 Dec 2023 13:13
 Operator : MA/JU
 Sample : 04929-02DL 5X
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 C0AB5DL

Quant Time: Dec 07 13:45:42 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Dec 04 21:39:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


72) Phenanthrene	17.269	178	1876044	26.103	ng/ul	99
78) Di-n-butylphthalate	18.192	149	667550	8.732	ng/ul	99
83) Butylbenzylphthalate	20.475	149	359142	8.168	ng/ul	96
86) Bis(2-ethylhexyl)phtha...	21.233	149	992297	16.419	ng/ul	100
87) Chrysene	21.351	228	771026	7.972	ng/ul	99
89) Di-n-octyl phthalate	22.145	149	2783320	24.767	ng/ul	100
91) Benzo(k)fluoranthene	23.016	252	2747555	25.022	ng/ul	99
93) Benzo(a)pyrene	23.586	252	1147128	11.062	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	26.180	276	371299	2.997	ng/ul#	76
95) Dibenzo(a,h)anthracene	26.180	278	1575391	15.286	ng/ul	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120723\
 Data File : BP018671.D
 Acq On : 07 Dec 2023 13:13
 Operator : MA/JU
 Sample : 04929-02DL 5X
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 C0AB5DL

Quant Time: Dec 07 13:45:42 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Dec 04 21:39:40 2023
 Response via : Initial Calibration









#8
Phenol
Concen: 35.181 ng/ul
RT: 7.034 min Scan# 61
Delta R.T. -0.018 min
Lab File: BP018671.D
Acq: 07 Dec 2023 13:13




Instrument :
BNA_P
ClientSampleId :
C0AB5DL


Tgt Ion: 94 Resp: 712463
Ion Ratio Lower Upper
94 100
65 25.9 21.0 31.6
66 34.4 28.6 43.0



#12
2-Chlorophenol
Concen: 10.727 ng/ul
RT: 7.399 min Scan# 733
Delta R.T. -0.018 min
Lab File: BP018671.D
Acq: 07 Dec 2023 13:13

Tgt Ion:128 Resp: 174936
Ion Ratio Lower Upper
128 100
64 45.5 37.0 55.4
130 33.5 26.8 40.2









#17
N-Nitroso-di-n-propylamine
Concen: 17.777 ng/ul
RT: 8.646 min Scan# 947
Delta R.T. -0.029 min
Lab File: BP018671.D
Acq: 07 Dec 2023 13:13

Instrument :
BNA_P
ClientSampleId :
C0AB5DL

Tgt Ion: 70 Resp: 243577




Ion	Ratio	Lower	Upper
70	100		
42	48.4	41.0	61.4
101	9.6	7.8	11.8
130	21.4	16.9	25.3




#19
Hexachloroethane
Concen: 15.218 ng/ul
RT: 8.916 min Scan# 991
Delta R.T. -0.018 min
Lab File: BP018671.D
Acq: 07 Dec 2023 13:13

Tgt Ion: 117 Resp: 101742

Ion	Ratio	Lower	Upper
117	100		
201	113.6	83.7	125.5
199	71.0	52.3	78.5









#23
Isophorone
Concen: 39.512 ng/ul
RT: 9.569 min Scan# 1102
Delta R.T. -0.018 min
Lab File: BP018671.D
Acq: 07 Dec 2023 13:13




Instrument :
BNA_P
ClientSampleId :
C0AB5DL


Tgt Ion: 82 Resp: 1398857
Ion Ratio Lower Upper
82 100
95 6.9 5.5 8.3
138 17.8 13.5 20.3



#25
2-Nitrophenol
Concen: 30.392 ng/ul
RT: 9.752 min Scan# 1133
Delta R.T. -0.018 min
Lab File: BP018671.D
Acq: 07 Dec 2023 13:13

Tgt Ion:139 Resp: 240941
Ion Ratio Lower Upper
139 100
65 48.9 39.1 58.7
109 31.6 23.8 35.6









#27
 Bis(2-Chloroethoxy)methane
 Concen: 23.684 ng/ul
 RT: 10.057 min Scan# 1185
 Delta R.T. -0.018 min
 Lab File: BP018671.D
 Acq: 07 Dec 2023 13:13




Instrument :
 BNA_P
 ClientSampleId :
 C0AB5DL


Tgt Ion: 93 Resp: 529355
 Ion Ratio Lower Upper
 93 100
 95 32.2 26.1 39.1
 123 10.8 8.9 13.3



#29
 2,4-Dichlorophenol
 Concen: 44.614 ng/ul
 RT: 10.287 min Scan# 1224
 Delta R.T. -0.012 min
 Lab File: BP018671.D
 Acq: 07 Dec 2023 13:13

Tgt Ion: 162 Resp: 650614
 Ion Ratio Lower Upper
 162 100
 164 65.9 51.7 77.5
 98 35.3 29.2 43.8









#33
Hexachlorobutadiene
Concen: 29.574 ng/ul
RT: 10.975 min Scan# 1341
Delta R.T. -0.012 min
Lab File: BP018671.D
Acq: 07 Dec 2023 13:13




Instrument :
BNA_P
ClientSampleId :
C0AB5DL


Tgt Ion:225 Resp: 288836
Ion Ratio Lower Upper
225 100
223 63.3 49.5 74.3
227 65.3 50.9 76.3



#35
4-Chloro-3-methylphenol
Concen: 33.265 ng/ul
RT: 11.928 min Scan# 1503
Delta R.T. -0.012 min
Lab File: BP018671.D
Acq: 07 Dec 2023 13:13

Tgt Ion:107 Resp: 543438
Ion Ratio Lower Upper
107 100
144 26.3 21.0 31.6
142 79.5 63.6 95.4









#36
2-Methylnaphthalene
Concen: 11.158 ng/ul
RT: 12.299 min Scan# 1566
Delta R.T. -0.012 min
Lab File: BP018671.D
Acq: 07 Dec 2023 13:13




Instrument :
BNA_P
ClientSampleId :
C0AB5DL


Tgt Ion:142 Resp: 384786
Ion Ratio Lower Upper
142 100
141 87.5 71.4 107.0



#41
2,4,6-Trichlorophenol
Concen: 17.419 ng/ul
RT: 12.910 min Scan# 1670
Delta R.T. -0.012 min
Lab File: BP018671.D
Acq: 07 Dec 2023 13:13

Tgt Ion:196 Resp: 207279
Ion Ratio Lower Upper
196 100
198 94.0 76.9 115.3
200 32.0 25.0 37.6









#42
2,4,5-Trichlorophenol
Concen: 11.656 ng/ul
RT: 12.975 min Scan# 1681
Delta R.T. -0.012 min
Lab File: BP018671.D
Acq: 07 Dec 2023 13:13




Instrument :
BNA_P
ClientSampleId :
C0AB5DL


Tgt Ion:196 Resp: 150369
Ion Ratio Lower Upper
196 100
198 100.1 77.9 116.9
200 30.8 25.2 37.8



#44
2-Chloronaphthalene
Concen: 19.370 ng/ul
RT: 13.351 min Scan# 1745
Delta R.T. -0.012 min
Lab File: BP018671.D
Acq: 07 Dec 2023 13:13

Tgt Ion:162 Resp: 674607
Ion Ratio Lower Upper
162 100
164 32.4 25.7 38.5
127 36.7 30.3 45.5









#47
Dimethylphthalate
Concen: 43.525 ng/ul
RT: 13.940 min Scan# 1845
Delta R.T. -0.012 min
Lab File: BP018671.D
Acq: 07 Dec 2023 13:13




Instrument :
BNA_P
ClientSampleId :
C0AB5DL


Tgt Ion:163 Resp: 1949626
Ion Ratio Lower Upper
163 100
194 4.2 3.2 4.8
164 10.4 8.0 12.0



#50
Acenaphthylene
Concen: 19.917 ng/ul
RT: 14.198 min Scan# 1889
Delta R.T. -0.006 min
Lab File: BP018671.D
Acq: 07 Dec 2023 13:13

Tgt Ion:152 Resp: 1122083
Ion Ratio Lower Upper
152 100
151 19.9 15.9 23.9
153 13.4 10.6 16.0









#52
Acenaphthene
Concen: 34.050 ng/ul
RT: 14.540 min Scan# 1947
Delta R.T. -0.012 min
Lab File: BP018671.D
Acq: 07 Dec 2023 13:13




Instrument :
BNA_P
ClientSampleId :
C0AB5DL


Tgt Ion:153 Resp: 1268346
Ion Ratio Lower Upper
153 100
152 47.1 38.1 57.1
154 88.3 71.7 107.5



#55
4-Nitrophenol
Concen: 33.726 ng/ul
RT: 14.692 min Scan# 1973
Delta R.T. -0.006 min
Lab File: BP018671.D
Acq: 07 Dec 2023 13:13

Tgt Ion:109 Resp: 190328
Ion Ratio Lower Upper
109 100
139 136.7 109.0 163.4
65 128.1 106.2 159.4









#59
Diethylphthalate
Concen: 20.116 ng/ul
RT: 15.304 min Scan# 2077
Delta R.T. -0.012 min
Lab File: BP018671.D
Acq: 07 Dec 2023 13:13




Instrument :
BNA_P
ClientSampleId :
C0AB5DL


Tgt Ion:149 Resp: 881715
Ion Ratio Lower Upper
149 100
177 22.8 18.2 27.4
150 12.4 9.9 14.9



#61
Fluorene
Concen: 20.977 ng/ul
RT: 15.528 min Scan# 2115
Delta R.T. -0.006 min
Lab File: BP018671.D
Acq: 07 Dec 2023 13:13

Tgt Ion:166 Resp: 859317
Ion Ratio Lower Upper
166 100
165 98.2 78.0 117.0
167 13.6 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 50.724 ng/ul
RT: 15.522 min Scan# 2114
Delta R.T. -0.006 min
Lab File: BP018671.D
Acq: 07 Dec 2023 13:13




Instrument :
BNA_P
ClientSampleId :
C0AB5DL


Tgt Ion:204 Resp: 1089776
Ion Ratio Lower Upper
204 100
206 32.7 27.1 40.7
141 53.3 45.8 68.8



#68
4-Bromophenyl-phenylether
Concen: 45.491 ng/ul
RT: 16.422 min Scan# 2267
Delta R.T. 0.000 min
Lab File: BP018671.D
Acq: 07 Dec 2023 13:13

Tgt Ion:248 Resp: 683422
Ion Ratio Lower Upper
248 100
250 95.7 79.3 118.9
141 61.1 57.2 85.8





#69

Hexachlorobenzene

Concen: 9.081 ng/ul

RT: 16.528 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP018671.D

Acq: 07 Dec 2023 13:13

Instrument :

BNA_P

ClientSampleId :

C0AB5DL





Tgt Ion:284 Resp: 152320

Ion Ratio Lower Upper

284 100

142 30.2 25.8 38.6

249 30.0 24.9 37.3



#71

Pentachlorophenol

Concen: 12.918 ng/ul

RT: 16.875 min Scan# 2344

Delta R.T. -0.006 min

Lab File: BP018671.D

Acq: 07 Dec 2023 13:13




Tgt Ion:266 Resp: 128863


Ion Ratio Lower Upper

266 100

264 64.3 50.2 75.2

268 63.7 50.5 75.7





#72

Phenanthrene

Concen: 26.103 ng/ul

RT: 17.269 min Scan# 2411

Delta R.T. -0.006 min

Lab File: BP018671.D

Acq: 07 Dec 2023 13:13

Instrument :

BNA_P

ClientSampleId :

C0AB5DL





Tgt Ion:178 Resp: 1876044

Ion Ratio Lower Upper

178 100

179 15.3 12.3 18.5

176 18.8 15.4 23.2



#78

Di-n-butylphthalate

Concen: 8.732 ng/ul

RT: 18.192 min Scan# 2568

Delta R.T. 0.000 min

Lab File: BP018671.D

Acq: 07 Dec 2023 13:13




Tgt Ion:149 Resp: 667550


Ion Ratio Lower Upper

149 100

150 8.8 7.4 11.2

104 4.7 3.9 5.9





#83

Butylbenzylphthalate

Concen: 8.168 ng/ul

RT: 20.475 min Scan# 2955

Delta R.T. 0.000 min

Lab File: BP018671.D

Acq: 07 Dec 2023 13:13

Instrument :

BNA_P

ClientSampleId :

C0AB5DL





Tgt Ion:149 Resp: 359142

Ion Ratio Lower Upper

149 100

91 67.0 56.6 85.0

206 21.5 17.4 26.0



#86

Bis(2-ethylhexyl)phthalate

Concen: 16.419 ng/ul

RT: 21.233 min Scan# 3085

Delta R.T. 0.000 min

Lab File: BP018671.D

Acq: 07 Dec 2023 13:13




Tgt Ion:149 Resp: 992297


Ion Ratio Lower Upper

149 100

167 27.5 22.2 33.2

279 4.2 3.3 4.9





#87

Chrysene

Concen: 7.972 ng/ul

RT: 21.351 min Scan# 3105

Delta R.T. 0.000 min

Lab File: BP018671.D

Acq: 07 Dec 2023 13:13

Instrument :

BNA_P

ClientSampleId :

C0AB5DL





Tgt Ion:228 Resp: 771026

Ion Ratio Lower Upper

228 100

226 30.4 23.9 35.9

229 19.8 15.6 23.4



#89

Di-n-octyl phthalate

Concen: 24.767 ng/ul




RT: 22.145 min Scan# 3240


Delta R.T. -0.006 min

Lab File: BP018671.D

Acq: 07 Dec 2023 13:13

Tgt Ion:149 Resp: 2783320





#91

Benzo(k)fluoranthene

Concen: 25.022 ng/ul

RT: 23.016 min Scan# 31

Delta R.T. 0.000 min

Lab File: BP018671.D

Acq: 07 Dec 2023 13:13

Instrument :

BNA_P

ClientSampleId :

C0AB5DL





Tgt Ion:252 Resp: 2747555

Ion Ratio Lower Upper

252 100

253 21.9 17.7 26.5

125 9.0 8.0 12.0



#93

Benzo(a)pyrene

Concen: 11.062 ng/ul

RT: 23.586 min Scan# 3485

Delta R.T. 0.000 min

Lab File: BP018671.D

Acq: 07 Dec 2023 13:13




Tgt Ion:252 Resp: 1147128


Ion Ratio Lower Upper

252 100

253 21.5 17.8 26.8

125 10.7 9.2 13.8









#94
Indeno(1,2,3-cd)pyrene
Concen: 2.997 ng/ul
RT: 26.180 min Scan# 3921
Delta R.T. 0.024 min
Lab File: BP018671.D
Acq: 07 Dec 2023 13:13

Instrument :
BNA_P
ClientSampleId :
C0AB5DL

Tgt Ion:276 Resp: 371299
Ion Ratio Lower Upper
276 100
138 43.4 19.4 29.2#
277 29.4 19.8 29.6



#95
Dibenzo(a,h)anthracene
Concen: 15.286 ng/ul
RT: 26.180 min Scan# 3926
Delta R.T. 0.000 min
Lab File: BP018671.D
Acq: 07 Dec 2023 13:13

Tgt Ion:278 Resp: 1575391
Ion Ratio Lower Upper
278 100
139 15.2 13.0 19.6
279 24.3 18.7 28.1

